

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

This checklist incorporates references to both TNI 2016 Standards, where applicable, and specific method, state and / or federal regulatory requirements.

**Directions:** Place a mark (e.g., /, √ or X) in the appropriate column (Yes (Y), No (N), or Not Applicable (NA)). If it is an observation on areas for possible improvement, place a mark under the Suggestion (S) column. In database, use code "SGST."

Lab ID: \_\_\_\_\_

Assessment ID: \_\_\_\_\_

Lab Name: \_\_\_\_\_

Personnel Interviewed:

Reports Reviewed:

_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____

At the time of the assessment, a question marked 'yes' indicates that no evidence of a deficiency was observed.

Assessment Date(s): \_\_\_\_\_ Assessor (Signature): \_\_\_\_\_

If this was a team assessment, indicate the Lead Assessor's name. \_\_\_\_\_

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Microbiological Testing Detailed Method Review	Data Records observed	Comments
<b>Method Number:</b> <b>SOP Number:</b> <b>Rev.:</b> <b>SOP date:</b> <b>Personnel records observed:</b>		
<b>Method Number:</b> <b>SOP Number:</b> <b>Rev.:</b> <b>SOP date:</b> <b>Personnel records observed:</b>		
<b>Method Number:</b> <b>SOP Number:</b> <b>Rev.:</b> <b>SOP date:</b> <b>Personnel records observed:</b>		
<b>Method Number:</b> <b>SOP Number:</b> <b>Rev.:</b> <b>SOP date:</b> <b>Personnel records observed:</b>		

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Microbiological Testing Detailed Method Review	Data Records observed	Comments
<p><b>Method Number:</b>  <b>SOP Number:</b>  <b>Rev.:</b>  <b>SOP date:</b>  <b>Personnel records observed:</b></p>		
<p><b>Method Number:</b>  <b>SOP Number:</b>  <b>Rev.:</b>  <b>SOP date:</b>  <b>Personnel records observed:</b></p>		
<p><b>Method Number:</b>  <b>SOP Number:</b>  <b>Rev.:</b>  <b>SOP date:</b>  <b>Personnel records observed:</b></p>		
<p><b>Method Number:</b>  <b>SOP Number:</b>  <b>Rev.:</b>  <b>SOP date:</b>  <b>Personnel records observed:</b></p>		

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
The laboratory is in adherence to the Quality Control procedures and program requirements specified in the NELAC standard, method, regulation, and project.	M5,1.2					000d30	
<p>SM9215A, 5 &amp; 7-8: <b>Heterotrophic Plate Count</b> SM9215B: <b>Pour Plate Method</b></p> <p>1__ All dilution plates analyzed in duplicate.</p> <p>2__ Incubated at 35.0 ± 0.5 °C for 48 ± 3 hours.</p> <p>3__ Colonies counted with a dark-field colony counter, or one with equivalent magnification &amp; illumination. (SM9215A, 8.a. &amp; b.&amp; ANSI/AAMI RD52:2004, 7.2.3)</p> <p>4__ Incubated at 35.0 +/- 0.5 degrees Celsius for 72 ± 4 hours for finished bottled water. (EPA 600/8-78-017, Part III, Sec. 5.5.2)</p> <p>5__ Incubated at 35-37 °C for 48 hours (for dialysis product water – ANSI/AAMI RD52:2004, 7.2.3 and RD62:2006, 5.1)</p> <p>6__ Sample volume chosen yields between 30 and 300 colonies. (SM9215A, 8.a. &amp; b. &amp; ANSI/AAMI RD52:2004, 7.2.3)</p> <p><b>SimPlate</b></p> <p>7__ Inverted and incubated at 35.0 ± 0.5 °C for 48 hours.</p> <p>8__ When doing unit dose, 10 ± 0.2 mL sample is added to media tube.</p> <p>9__ When doing multi dose, 1 mL of sample and 9 mL of rehydrated media is pipetted onto center of the plate.</p> <p>10__ 6 W, 365 nm, light held 6-12 inches above plate used to count fluorescent wells.</p> <p>SM9215C &amp; ANSI/AAMI RD52:2004 &amp; RD62:2006: <b>Spread Plate</b></p> <p>11__ An inoculum of at least 0.5 mL of sample spread equally over the surface of the agar. (ANSI/AAMI RD52:2004, 7.2.3)</p> <p>12__ Inoculated agar plate with glass rod or pipette. Calibrated loop is not allowed. (ANSI/AAMI RD52:2004, 7.2. &amp; RD62:2006, 5.1.1)</p> <p>13__ Incubated at 35-37 °C for 48 hours. (ANSI/AAMI RD52:2004, 7.2.3 and RD62:2006, 5.1)</p> <p>14__ Colonies counted with a dark-field colony counter, or one with equivalent magnification &amp; illumination. (SM9215A, 8.a. &amp; b.&amp;</p>						<p>0d31a1</p> <p>0d31a2</p> <p>0d31a3</p> <p>0d31a4</p> <p>0d31a10</p> <p>0d31a101</p> <p>0d31a5</p> <p>0d31a6</p> <p>0d31a7</p> <p>0d31a</p> <p>0d31a8</p> <p>0d31a9</p> <p>0d31a10</p> <p>0d31a33</p>	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
<p>ANSI/AAMI RD52:2004, 7.2.3)</p> <p>15__ Sample volume chosen yields between 30 and 300 colonies. (SM9215A, 8.a. &amp; b. &amp; ANSI/AAMI RD52:2004, 7.2.3)</p> <p><b>Note:</b> If colony yield is not met, lab can use smaller or larger volumes. Smaller volumes can be reached by doing 1:10 serial dilutions using sterile phosphate buffer. If larger volumes are required, the MF method should be used</p>						0d31a101	
<p>SM9215D &amp; ANSI/AAMI RD52:2004 &amp; RD62:2006: <b>Membrane Filter Method</b></p> <p>16__ Dispensed 5-mL portion of sterile agar into 50- x 9- mm petri dishes <b>Note:</b> m-HPC agar may not be sterile.</p> <p>17__ Incubated at 35-37 °C for 48 hours. (ANSI/AAMI RD52:2004, 7.2.3 and RD62:2006, 5.1)</p> <p>18__ Colonies counted with a stereoscopic microscope at 10 to 15 x magnification. (SM9215A, 8. b.&amp; ANSI/AAMI RD52:2004, 7.2.3)</p> <p>19__ Sample volume chosen yields between 20 and 200 cfu. (SM9215A, 8b. &amp; ANSI/AAMI RD52:2004, 7.2.3)</p> <p><b>Note:</b> If colony yield is not met, lab can use smaller or larger volumes. Smaller volumes can be reached by doing 1:10 serial dilutions using sterile phosphate buffer. If larger volumes are required, the MF method should be used.</p>						0d31a12 0d31a10 0d31a13 0d31a11	
<p><b>b. SM9221A&amp;B, 1.b.: Total Coliform Multiple Tube Fermentation with Lauryl Tryptose Medium</b></p> <p>1__ SDWA: 100 mL sample analyzed. (five 20 mL tubes, ten 10 mL tubes, or one 100 mL bottle)</p> <p>2__ CWA: 5-tube per dilution for each sample.</p> <p>3__ Incubated at 35.0 ± 0.5 °C for 24 +/- 2 hours.</p> <p>4__ SDWA: If no gas detected after 24 hours, incubate for another 24 hours.</p> <p><b>Note:</b> For other waters (NW), pull positives after 24 +/- 2 hours, transfer them, and still check the ones that are negative after 24 hours at 48 +/- 3 hours.</p> <p>5__ All samples producing turbid cultures with no gas production invalidated, with another sample requested.</p>	40 CFR 141.21(f)(1)					0d31b1 0d31b2 0d31b3 0d31b4  0d31b5	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
<p>6_ Commercially available Lactose Broth (LB) may be used in lieu of LTB if system conducts at least 25 parallel tests using water normally tested and false positive and false negative results using LB is &lt;10%. Note: No requirement to run completed phase</p>	<p>40 CFR 141.21(f)(3) &amp; 40 CFR 136.3(a) Table IA 40 CFR 141.21(f)(3)</p>					0d31c	
<p>c. SM9221D, 1.a. &amp; b.: <b>Total Coliform with Presence/Absence Medium</b>            1__ 100 mL sample analyzed            2__ Incubated at 35.0 ± 0.5 °C for 24 hours            3__ If purple color indicator does not turn yellow, incubate for another 24 hours            4__ All samples producing turbid cultures with no color change invalidated, with another sample requested  <b>Note:</b> Media 6x formulation strength may be used if media is filter sterilized instead of autoclaved.</p>	<p>40 CFR 141.21(f)(3)</p>					0d31c1 0d31c2 0d31c3  0d31c4	
<p>d. SM9221E, 1.a. &amp; b: Thermotolerant (<b>Fecal</b>) <b>Coliform Most Probable Number with EC Medium</b>            1__ 3-dilution (sample volumes), 5-tube (per sample volume) technique for each sample            2__ Each tube inoculated from positive culture grown on m-Endo or LTB medium            3__ Incubated at 44.5 ± 0.2 °C for 24 ± 2 hours            4__ Gas formation indicates Fecal Coliform, no further verification needed            5__ Must be incubated in a water bath</p>	<p>SM and 40 CFR 141 Subpart C(f)(5)</p>					0d31d1  0d31d2  0d31d3 0d31d4  0d31d5	
<p>e. SM9221E, 2.a. &amp; b: Thermotolerant (<b>Fecal</b>) <b>Coliform Most Probable Number with A-1 Medium</b>            1__ 3-dilution (sample volumes), 5-tube (per sample volume) technique for each sample            2__ Direct inoculation with sample possible            3__ Incubated at 35.0 ± 0.5 °C for 3 hours, then at 44.5 ± 0.2 °C for 21 ± 2 hours            4__ Gas formation indicates Thermotolerant (Fecal) Coliform; no further verification needed</p>						0d31e1  0d31e2  0d31e3  0d31e4	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
<p><b>f. SM9221F PW/NW E. coli enumeration &amp; NW Thermotolerant coliform with EC-MUG</b></p> <p>1__ Tube contains inverted Durham tube</p> <p>2__ Enrich in presumptive medium for Total Coliform using 9221 B.2-2014 prior to EC-MUG</p> <p>3__ Presumptive tubes or bottles showing any amount of gas, growth, or acidity within 48 ± 3 h transferred to EC-MUG using sterile 3- or 3.5 mm diameter sterile loop or sterile wooden applicator inserted at least 2.5 cm to transfer growth from fermentation tube to culture tube.</p> <p><b>Note:</b> Wooden applicator must be plunged to bottom of EC-MUG tube.</p> <p>4__ Incubate 44.5 ± 0.2° C for 24 ± 2 hours.</p> <p>5__ Growth and gas indicates thermotolerant coliform</p> <p>6__ Blue fluorescence under 6 W 365-366 nm UV light indicates E. coli</p> <p><b>g. SM9222B, 5.a.-d.: Total Coliform by Membrane Filtration</b></p> <p>1__ SDWA: 100 mL sample filtered</p> <p>2__ CWA: Filter 3 different sample volumes so that at least one dilution will give 20-80 colonies, but not more than 200 colonies.</p> <p>3__ Enhancement recovery required for stressed organisms in chlorinated samples (e.g., spas and swimming pools).</p> <p>4__ Incubated at 35.0 ± 0.5 °C for 22-24 hours</p> <p><b>h. SM9222D, 2.a.-d.: Thermotolerant (Fecal) Coliform by Membrane Filtration</b></p> <p>1__ Filter volumes or dilutions that will give 20-60 fecal coliform colonies per membrane filter</p> <p>2__ Incubated at 44.5 ± 0.2 °C for 24 ± 2 hours</p>	<p>40 CFR 136</p> <p>40 CFR 136</p>					<p>0d32a</p> <p>0d32f</p> <p>0d32b</p>  <p>0d32c</p> <p>0d32d</p> <p>0d32e</p>  <p>0d31f1</p> <p>0d31f2</p> <p>0d31f3</p> <p>0d31f4</p>  <p>0d31g1</p> <p>0d31g2</p>	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
<p><b>i. SM9223B, 2: Total Coliform by MMO-MUG</b></p> <p>1__ SDWA: 100 mL sample analyzed (for drinking waters)</p> <p>2__ Colilert: Incubated at 35.0 ± 0.5 °C for 24 hours.</p> <p>3__ Colilert: When indeterminate after 24 hours, incubate for another 4 hours.</p> <p>4__ Colisure: Incubated at 35.0 ± 0.5 °C for ≥ hours, but ≤ 48 hours.</p> <p>5__ Colilert-18: Incubated at 35.0 ± 0.5°C for 18 hours (up to 22 hours if indeterminate after 18 hours); first 20 minutes MUST be in 35 °C water bath or 7-10 minutes in 44.5 °C water bath.</p> <p>6__ ReadyCult: Incubated at 35.0 ± 0.5 °C for 24 hours ±1 hour.</p> <p>7__ Fluorocult LMX: Incubated at 35.0 ± 0.5 °C for 24 hours ± 1 hour.</p> <p>8__ Colitag: Incubated at 35.0 ± 0.5 °C for 16-48 hours. If results are Read before 22 hours, sample is prewarmed for 7-10 min in 44.5 ±0.2°C water bath.</p> <p>9__ E*Colite: Incubated at 35.0 ± 0.5 °C for 28 hours.</p> <p>10__ Color change indicates Total Coliform; 365-366 nm, 6 W UV light used to determine fluorescence to indicate E. coli; and no further verification needed.</p> <p>11__ When enumerating coliforms using Colilert, the lab uses a Quanti-Tray for each sample dilution tested.</p> <p>12__ The lab checks the Quanti-Tray sealer monthly by adding a dye to the water.</p> <p>The lab reports quantitative (aka estimate of bacterial Density or enumeration) data for E. coli for source water under the SDWA Surface Treatment Rule.</p>	40 CFR 141.21(f)(1)					<p>0d31h1</p> <p>0d31h2</p> <p>0d31h3</p> <p>0d31h4</p> <p>0d31h5</p> <p>0d31h6</p> <p>0d31h7</p> <p>0d31h8</p> <p>0d31h9</p> <p>0d31h10</p> <p>0d31h11</p> <p>0d31h12</p>	



**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
<p><b>j. Fecal Coliform by Colilert-18</b></p> <p>1_ Incubated 44.5 °C ± 0.2 °C for 18 hours (up to 22 hours if indeterminant after 18 hours)</p> <p>2_ Water bath incubator must be used</p>	<p>40 CFR 136.3 Table 1A</p>					<p>0d31z1</p> <p>0d31z2</p>	
<p><b>k. Enterococci by Enterolert</b></p> <p>1__ 100 mL sample analyzed (for drinking waters)</p> <p>2__ Incubated at 41.0 ± 0.5 °C for 24 hours (up to 28 hours if indeterminate after 24 hours)</p> <p><b>I. EPA 1600, 9.5.2, 11.5 &amp; 11.8: Enterococci by Membrane Filtration with mEI Medium</b></p> <p>1__ Filter volumes or dilutions that will give 20-60 enterococci colonies per membrane filter</p> <p>2__ Incubated at 41.0 ± 0.5 °C for 24 hours +/- 2 hours</p> <p><b>m. SM 9230C: Enterococci by Membrane Filtration with mE → EIA Medium</b></p> <p>1__ If mE agar is used, incubated inverted plate for 48 hours at 41°±0.5°C, ± 3 hours and then transfer filter to EIA medium. Incubated at 41°± 0.5°C for 20 minutes.</p>						<p>0d31i1</p> <p>0d31i2</p> <p>0d31j1</p> <p>0d31j2</p> <p>0d31j3</p>	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
<p><b>Method Validation:</b></p> <p>1 For methods other than reference methods, validation must comply with V1M2 and include the following minimum requirements from M5:</p> <p>    <b>a</b>__ Determine <b>accuracy</b> by comparison of at least one positive reference culture result to that of a reference method;</p> <p>    <b>b</b>__ Determine <b>precision</b> by analyzing a minimum of ten replicate analyses spiked with the target microorganism with both the proposed and reference method and determine that the proposed method is statistically equivalent or better than the reference method;</p> <p>    <b>c</b>__ Determine <b>selectivity</b> by analyzing a minimum of ten spiked samples using mixed cultures that include the target microorganisms and at various concentrations. Calculate the number of false positive and false negative results.</p> <p>2 For both reference and non-standard methods, the laboratory participates in proficiency testing programs, where available.</p> <p>3 The laboratory maintains documentation of the validation procedure for as long as the method is in use and for at least five years past the last date of use.</p>	<p>M5,1.5 (a)</p> <p>M5,1.5.1</p> <p>M5,1.5.2</p> <p>M5,1.5.3</p> <p>M5,1.5 (b)</p> <p>M5,1.5 (c)</p>					<p>0d364ar</p> <p>0d364br</p> <p>0d364bb</p> <p>00d320a</p> <p>00d319r</p>	
<p>The quality control protocols specified by the laboratory's method manual are followed by all analysts.</p>	<p>M2,5.9.3 (c)</p>					<p>000d12</p>	
<p>All essential quality control measures are incorporated in the lab's method manual.</p>	<p>M2,5.9.3(c)</p>					<p>000d13</p>	
<p>All quality control measures are assessed and evaluated on an on-going basis and quality control acceptance criteria are used to determine the validity of the data.</p>	<p>M2,5.9.3(b)</p>					<p>000d14</p>	
<p>The laboratory has procedures for developing acceptance/rejection criteria for each test where no method or regulatory criteria exist.</p>	<p>M2,5.9.3(c)</p>					<p>000d15</p>	
<p>Samples are stored, handled, and prepared in accordance with written procedures and in a manner to avoid deterioration, loss, or damage. See also Quality Systems Checklist deficiency 51126.</p> <p>Examples of deterioration, loss, or damage include, but are not limited to, "pouring off" sample collection container to reduce volume to allow analysis in collection container and dipping a test strip into the sample.</p>	<p>M2,5.8.4</p>					<p>000d20</p>	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

<b>Relevant Aspect of Standards</b>	<b>TNI 2016/ELAP/ Regulation/Method Reference</b>	<b>Y</b>	<b>N</b>	<b>N/A</b>	<b>S</b>	<b>ELAP citation codes</b>	<b>Comments</b>
<p>Microbiological samples from <b>known chlorinated water sources, unknown sources where disinfectant usage is suspected, and all potable water supplies</b> are checked in the laboratory for absence of residual chlorine, unless all of the following conditions are met:</p> <p><b>a</b>__ sufficient sodium thiosulfate is added to each container to neutralize at minimum 5 mg/L of chlorine for drinking water and 15 mg/L chlorine for wastewater.</p> <p><b>b</b>__ one container from each batch of laboratory prepared containers or lot of purchased ready-to-use containers is checked to ensure efficacy of the sodium thiosulfate to 5 mg/L chlorine or 15 mg/L chlorine as appropriate and the check is documented,</p> <p><b>c</b>__ chlorine residual is checked in the field and actual concentration is documented with sample submission.</p> <p><b>d</b>__ the laboratory can show that the received sample containers are from its laboratory or have been appropriately tested and documented.</p>	M5,1.7.5.2(a-d)					55818  55818ar  55818br  55818cr  55818dr	
<p>Samples requiring thermal preservation are acceptable if the arrival temperature of a representative sample container meets the method or mandated temperature requirement. Samples delivered to the laboratory on the same day as collection that do not meet the requirement <b>may</b> be considered acceptable if the samples are received on ice with evidence that the cooling process has begun.</p>	M5,1.7.5.1					00d370	
<p>The <b>maximum hold time</b> has not been exceeded for the bacteriological samples analyzed by the laboratory. <b>Note:</b> Refer to ELAP Certification Manual Item 245.</p>	SWTR, BWR, TCR, GWR, NPDES, AAMI/ANSI					00d335z	
<p>The laboratory has documented procedures, which refer to applicable reference methods, for the calibration, verification, and QC of support equipment including conductivity meters, oxygen meters, pH meters, hygrometers, and other similar measurements instruments.</p>	M5,1.7.1.1					00d3677	
<p>The following support equipment associated with microbiological testing is checked with NIST traceable materials (where possible):</p> <p><b>a</b>__ pH meter</p> <p><b>b</b>__ Balance(s)</p> <p><b>c</b>__ Conductivity meter</p> <p><b>d</b>__ Refrigerator(s) for sample storage and/or media storage</p>	M2,5.5.13.1; ELAP Certification Manual Item 231					5916 or 00d34ar 00d34br 00d34cr 00d34dr	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

<b>Relevant Aspect of Standards</b>	<b>TNI 2016/ELAP/ Regulation/Method Reference</b>	<b>Y</b>	<b>N</b>	<b>N/A</b>	<b>S</b>	<b>ELAP citation codes</b>	<b>Comments</b>
e__ Incubators f__ Water baths g__ Freezers						00d34er 00d34fr 00d34gr	
<b>Quality Control</b>							
The lab demonstrates and documents the quality of reagents and media used is appropriate for the test concerned including, but not limited to, test conditions and incubation times.	M5,1.7.3.1					000d37d	
<b>Sterility Checks:</b> All materials and supplies that are needed to process the sample and are required to be sterile prior to use (whether sterilized in the laboratory or purchased as sterilized) are checked by the laboratory once per purchased or prepared lot using non-selective growth media.	M5,1.7.3.1(a)					000d37ar	
Certificates of Analysis (COA) provided by vendors documenting sterility are <b>verified by the laboratory</b> and retained in accordance with V1M2 5.6.4.2.a.	M5,1.7.3.1(a)					000d37br	
<b>Sterility Checks:</b> 1__ The laboratory performs a sterility check for each lot of pre-prepared, ready-to-use media (including chromofluorogenic reagent) and for each batch of media prepared in the laboratory, prior to use: <b>a</b> __ For chromo/fluorogenic media, the media is added to sterile deionized water and incubated at the appropriate temperature and time for the method used and documented. <b>b</b> __ For all other media, the media is incubated uninoculated at working strength (single strength). 2__ The laboratory performs a sterility check on one (1) funnel per lot of pre-sterilized single use funnels and 1 funnel per batch of laboratory-sterilized funnels with non-selective growth media. 3__ <b>a</b> __ The laboratory performs a sterility check on at least one (1) container for each lot of purchased, pre-sterilized sample containers with non-selective growth media. <b>b</b> __ The laboratory performs a sterility check on one (1) container/object per sterilization batch sterilized in the laboratory with non-selective growth media. 4__ The laboratory performs a sterility check on each batch of dilution water prepared in the laboratory and on each lot of pre-prepared, ready-to-use dilution water with non-selective growth media.	M5,1.7.3.1(a)(i)  M5,1.7.3.1(a)(i)(a)  M5,1.7.3.1(a)(i)(b)  M5,1.7.3.1.(a)(ii)  M5,1.7.3.1(a)(iii)  M5,1.7.3.1(a)(iv)					00d381r  00d382r  0d382a1r  00d388r  00d390  00d391  00d385r	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
<p>The concentration of the non-selective growth media shall be single strength after addition of the dilution water.</p> <p><b>5</b>__ The laboratory performs a sterility check on at least one filter from each new lot of membrane filters with non-selective growth media.</p> <p>(When using a non-selective broth, incubate at 35 +/- 0.5 ° C for 24 and 48 hours and check for growth)</p>	M5,1.7.3.1.(a)(v)					00d387r  00d387a	
<p><b>Media:</b></p> <p><b>1</b>__ Media is tested for performance (e.g., selectivity, sensitivity, sterility, growth promotion, and growth inhibition) at a minimum with first use.</p> <p><b>2</b>__ Media is used within the expiration date or shelf-life provided by the manufacturer.</p> <p><b>3</b>__ Laboratory prepared media is used within the holding time specified in the accredited method.</p> <p>For SM Methods:</p> <p><b>a</b>_ Broth in screw-cap flask is used within 96 h and stored at 2-8°C.</p> <p><b>b</b>_ Agar plates with tight-fitting covers used within 2 weeks and stored at 2-8°C</p> <p><b>c</b>_ Agar or broth loose-capped tubes is used within 2 weeks and stored at 2-8°C</p> <p><b>d</b>_ Agar or broth tightly closed screw-cap tubes is used within 3 months. Note: Hold at &lt;30° C.</p> <p><b>e</b>_ Poured agar plates with loose-fitting covers, sealed in plastic bags are used within 2 weeks and stored at 2-8°C.</p> <p><b>f</b>_ Large volume of agar in tightly closed screw cap flask or bottle is used within 3 months and stored at 2-8°C.</p> <p><b>g</b>_ Tubes or plates with growth and/or bubbles are discarded.</p> <p><b>h</b>_ Media in tubes and plates stored &gt;2 weeks with evaporation exceeding 10% original volume or weight is discarded.</p> <p><b>i</b>_ Refrigerated medium is warmed to room temperature before use.</p> <p><b>j</b>_ Prepared, ready to use media with expiration dates later than noted above is verified weekly by testing recoveries with known densities of control cultures that also meet QC check requirements.</p> <p><b>4</b>__ The laboratory has detailed testing criteria information defined in the laboratory's methods, SOPs, or similar documentation.</p>	<p>M5,1.7.3.1(b)(i)</p> <p>M5,1.7.3.1(b)(ii)</p> <p>M5,1.7.3.1(b)(iii)</p> <p>SM 9020B, 4.i.4 and, Table 9020:V</p> <p>M5,1.7.3.1(b)(iv)</p>					00d377a  00d377b  00d377c  00d336a 00d336b  00d336c  00d336d  00d336e  00d336f  00d336g 00d336h  00d336i 00d336j  00d337a	



**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

<b>Relevant Aspect of Standards</b>	<b>TNI 2016/ELAP/ Regulation/Method Reference</b>	<b>Y</b>	<b>N</b>	<b>N/A</b>	<b>S</b>	<b>ELAP citation codes</b>	<b>Comments</b>
<p>e_ Heterotrophic plate count &lt; 500 CFU/mL or MPN &lt;500/mL  f_ Use Test (new water source) Student's t ≤ 2.78  g_ Water Quality Test 0.8 – 3.0 ratio</p> <p>6 Records of reagent water purchased from an outside source and used in the preparation of media, solutions, and buffers meet the criteria specified in 3 and 4 above are documented.</p> <p>7 Reagent water that has been opened for longer than the testing intervals specified above or in the accredited method is re-tested or discarded.</p>	<p>M5,1.7.3.1(d)(v)</p> <p>M5,1.7.3.1(d)(vi)</p>					<p>0d334cr 0d334ir 0d334dr</p> <p>000d334</p> <p>000d335</p>	
<p>The quality of dilution water, including buffer water and/or peptone water, is monitored for sterility, pH, and volume once per batch whether lab prepared or purchased.</p>	<p>M5,1.7.3.1(e)</p>					<p>0d334jr</p>	
<p>Documentation for <b>media and reagents prepared in the laboratory</b> includes the following:</p> <p>a___ Date of preparation,  b___ Preparer's initials,  c___ Type and amount of media prepared,  d___ Manufacturer and Lot #,  e___ Final pH of the media, and  f___ Expiration date</p>	<p>M5,1.7.3.1(f)</p>					<p>0d337ar 0d337br 0d337cr 0d337dr 0d337er 0d337fr</p>	
<p>Documentation for media purchased pre-prepared, ready-to-use (including reagent water purchased from outside sources) includes the following:</p> <p>a___ Manufacturer,  b___ Lot #,  c___ Type and amount of media received,  d___ Date of receipt  e___ Expiration date of the media, and  f___ pH of the media</p>	<p>M5,1.7.3.1(f)</p>					<p>0d338ar 0d338br 0d338cr 0d338dr 0d338er 0d338fr</p>	
<p><b>Method Blanks:</b></p> <p>1 For filtration techniques, method blanks are conducted per the analytical method and at a minimum of beginning and ending blank for each filtration series.</p> <p>2 Filtration series is ended when more than 30 minutes elapses between successive filtrations.</p> <p>3 Filter funnels are rinsed with three 20-30 mL portions of sterile rinse water</p>	<p>M5,1.7.3.2(a)</p> <p>M5,1.7.3.2(b)</p> <p>M5,1.7.3.2(b)</p>					<p>0d382e</p> <p>0d382f</p> <p>0d382cr</p>	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

<b>Relevant Aspect of Standards</b>	<b>TNI 2016/ELAP/ Regulation/Method Reference</b>	<b>Y</b>	<b>N</b>	<b>N/A</b>	<b>S</b>	<b>ELAP citation codes</b>	<b>Comments</b>
<p>after each filtration.</p> <p><b>4</b> After every ten samples, a method blank is inserted, or the filtration unit is sanitized with UV light (254 nm) after every sample filtration.</p> <p><b>5</b> For pour plate methods, at least one uninoculated plate is poured for each lot of pre-prepared, ready to-to-use medium and for each batch of laboratory prepared medium.</p>	<p>M5,1.7.3.2.(b)</p> <p>M5,1.7.3.2.(c)</p>					<p>0d382d</p> <p>00d383r</p>	
<p><b>Test Variability/Reproducibility:</b> For test methods that specify counts (i.e. cfu/100ml or MPN/100 ml), such as membrane filter, plated media, or other methods which specify a quantitative result, <b>duplicate counts</b> are performed <b>monthly</b> on one positive sample for each month that the test is performed:</p> <p><b>a</b> If the lab has two or more analysts, each analyst counts typical results on the same sample for each month the test is performed.</p> <p><b>i)</b> __ Counts shall be within <b>10% difference</b> to be acceptable.</p> <p><b>b</b> In a lab with one microbiology analyst, the same sample is counted twice by the analyst for <b>each month</b> the test is performed.</p> <p><b>ii)</b> __ Counts with no more than <b>5% difference</b> are between the counts are acceptable.</p>	<p>M5,1.7.3.3</p>					<p>0d3161r</p> <p>0d3162r</p> <p>0d3162ar</p> <p>0d3163r</p> <p>0d1363ar</p>	
<p><b>Sample Specific Controls:</b></p> <p><b>1</b> Matrix spikes are performed per the method requirements.</p> <p><b>2</b> Matrix duplicates are performed per the method requirements.</p>	<p>M5,1.7.3.4(a)</p> <p>M5,1.7.3.4(b)</p>					<p>0d390a</p> <p>0d390b</p>	
<p><b>Data Reduction:</b> The calculations, data reduction and statistical interpretations specified by each method are followed.</p> <p>a. 9221D - Reported result as presence-absence test positive or negative for total coliforms in 100 mL of sample.</p> <p>b. 9222B - Compute the count, using membrane filters with 20 to 80 coliform colonies and not more than 200 colonies of all types per membrane, by the following equation: (Total) coliforms/100 mL = (coliform colonies counted x 100)/mL sample filtered</p> <p>c. 9215B - To compute the heterotrophic plate count, CFU/mL, divide total number of colonies or average number (if duplicate plates of the same dilution) per plate by the sample volume.</p> <p>d. 9223B - If performing an MPN procedure, calculate the MPN value for</p>	<p>M5,1.7.3.5</p>					<p>00d326r</p>	



**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

<b>Relevant Aspect of Standards</b>	<b>TNI 2016/ELAP/ Regulation/Method Reference</b>	<b>Y</b>	<b>N</b>	<b>N/A</b>	<b>S</b>	<b>ELAP citation codes</b>	<b>Comments</b>
<p>total coliforms and E. coli from the number of positive tubes as described in Section 9221C. If using the presence-absence procedure, report results as total coliform and E. coli present or absent in 100-mL sample.</p> <p>e. EPA 1600 – Compute the count per 100 mL of sample by dividing the # of enterococci colonies by the volume of sample filtered and then multiplying by 100. Refer to rules in Appendix B of method. For example, if there is &gt; 1 dilution, calculate the arithmetic mean for those results in the acceptable counting range.</p> <p>f. ISO 11731:2017 (E) – For enumeration, select the plate or set of plates from the same culture showing the maximum number of confirmed colonies per water volume and taking any dilutions into account. Do not average the counts from different methods, treatments, or culture media as these are not replicates. Calculate the # of colonies in original water per liter using the equations in section 9 for direct plating, MF, indirect filtration, and plating after dilution.</p>							
<p><b>Selectivity:</b></p> <p>1. All growth and recovery media are checked to assure that the target organisms respond in an acceptable and predictable manner once per lot or batch.</p> <p>2. In order to ensure identity and traceability, reference cultures used for positive and negative controls are obtained from a recognized national collection, organization, or manufacturer recognized by the accreditation body.</p> <p>3. Reference cultures are single use preparations or cultures maintained for their use by documented procedures that demonstrate continued purity and viability of the organisms.</p> <p>    a. Reference cultures are revived (if freeze dried) or transferred from slants and sub-cultured once to provide reference stocks.</p> <p>    b. The reference stocks are preserved by a technique that maintains the desired characteristics of the strains.</p> <p>    c. Reference stocks are used to prepare working stocks for routine work.</p> <p>    d. When reference stocks are thawed, they are not re-frozen and re-used.</p> <p>    e. Working stocks are not sequentially cultured more than 5 times.</p> <p>    f. Working stocks are not sub-cultured to replace reference stocks.</p>	<p>M5,1.7.3.6(a)</p> <p>M5,1.7.3.6(c)</p> <p>M5,1.7.3.6(c)</p> <p>M5,1.7.3.6(c)(i)</p> <p>M5,1.7.3.6(c)(i)</p> <p>M5,1.7.3.6(c)(i)</p> <p>M5,1.7.3.6(c)(i)</p> <p>M5,1.7.3.6(c)(ii)</p> <p>M5,1.7.3.6(c)(ii)</p>					<p>00d325r</p> <p>00d341r</p> <p>00d343r</p> <p>00d342r</p> <p>00d344r</p> <p>00d345r</p> <p>00d346r</p> <p>00d348r</p> <p>00d349r</p>	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
<p>4. Culture Controls:</p> <p>a. Each batch of pre-prepared, ready-to-use lot of selective medium (including chromofluorogenic reagent) and each batch of selective medium prepared in the laboratory is analyzed with at least one known <b>negative culture control</b> as appropriate to the method prior to first use of the medium.</p> <p>b. Each batch of pre-prepared, ready-to-use lot of selective medium (including chromofluorogenic reagent) and each batch of selective medium prepared in the laboratory is analyzed with at least one known <b>positive culture control</b> as appropriate to the method prior to first use of the medium.</p> <p>5. To ensure that analysis results are accurate, a target organisms' identity is verified as specified in the method (e.g. by use of the completed test, or by use of secondary verification tests such as a catalase test or by the use of a selective medium such as Brilliant Green Bile Broth (BGLG) or EC or EC+MUG broth).</p> <p><b>Note:</b> See below for selected methods. Additional method verifications can be found in the promulgated method.</p> <p><b>a. SM9221B, 2b; SM9221D, 2b: Total Coliform by Fermentation Broth method</b></p> <p>i__ Each positive culture from LTB (gas formation or color change) is inoculated onto BGLB (Note: If all 5 tubes produced gas in 2 or more sample dilutions, only the 5 tubes with gas from the highest dilution need be confirmed)</p> <p>ii__ Incubated at 35.0 ± 0.5 °C for 24 ± 2 hours</p> <p>iii__ If no gas formation, re-incubate for additional 24 hours (total of 48 ± 3 hours)</p> <p>iv__ Gas formation in BGLB confirms Total Coliform for purposes of MPN calculation or Presence-Absence reporting</p> <p>v_ SDWA samples also tested according to SM9221E or EPA 1104 Note: No requirement to run completed phase</p>	<p>M5,1.7.3.6(d)(i)(b)</p> <p>M5,1.7.3.6(d)(ii)(b)</p> <p>M5,1.7.3.6(b)</p>					<p>00d311r</p> <p>00d312r</p> <p>0d325ar</p> <p>d340a11</p> <p>d340a2r d340a3r</p> <p>d340a4r</p> <p>0d340a5r</p>	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
<p><b>b. SM9222B, 5f: Total Coliform by Membrane Filter method</b></p> <p>i__ Inoculate at least 10 colonies from filter into LTB &amp; BGLB</p> <p>ii__ SDWA: Inoculate all colonies (can swab entire filter) into one LTB tube &amp; one BGLB tube</p> <p>iii__ Incubate at 35.0 ± 0.5 °C for 48 hours</p> <p>iv__ Gas production in LTB &amp; BGLB confirms Total Coliform</p> <p>v __ SM9222B: May use rapid-test or commercial multi-test verification systems that utilize test reactions for cytochrome oxidase &amp; b-galactosidase; negative reaction for cytochrome oxidase &amp; positive reaction for b-galactosidase confirms Total Coliform</p> <p>vi__ SDWA: Positive cultures from LTB or membrane filter colonies also tested according to SM9221E, EPA 1104, or EPA 1105. (Note: May inoculate m-Endo colonies directly into BGLB medium. However, if gas is observed in LTB, but not in the corresponding BGLB tube, another BGLB tube must be inoculated &amp; tested with the positive culture from the LTB tube</p> <p>vii__ SM 9020B, 9.b.1 Membrane Filter Method Confirmation:</p> <p>a)__ For drinking water, all colonies from positive samples on m-Endo medium are verified.</p> <p>b)__ If there are no positives, at least one known positive source water is tested quarterly.</p> <p>c)__ For other waters, at least 10 sheen colonies are verified monthly using LTB and BGLB, followed by count adjustment based on these results.</p> <p>d)__ For other waters, non-sheen colonies are verified monthly using LTB.</p> <p><b>c. SM9221E, 1b: Fecal Coliform with EC Medium (A-1 is not allowed for SDWA)</b></p> <p>i__ Incubated at 44.5 ± 0.2 °C for 24 ± 2 hours</p> <p>ii__ Gas formation confirms that the Total Coliform is a Fecal Coliform</p>	<p>40 CFR 136 Table IA</p> <p>40 CFR 136 Table IA</p>					<p>d340b1 d340b2</p> <p>d340b3 d340b4 d340b5</p> <p>d340b6</p> <p>d340b7a d340b7b d340b7c d340b7d</p> <p>d340c1 d340c2</p>	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
<p><b>d. SM 9222G/EPA 1104, 11: E. coli by EC + MUG Tube Procedure</b></p> <p>i__ Incubated at 44.5 ± 0.2 °C for 24 ± 2 hours</p> <p>ii__ 366-nm blue-light fluorescence confirms that the Total Coliform is E. coli</p> <p>iii__ Membrane filter is transferred in its entirety to EC + MUG medium.</p> <p><b>e. SM 9222G/EPA 1105, 11: E. coli by Nutrient Agar + MUG Membrane Filter Procedure</b></p> <p>i__ Membrane filter transferred in its entirety to NA + MUG medium (Note: Some colonies removed for LTB &amp; BGLB tests.)</p> <p>ii__ Incubated at 35.0 ± 0.5 °C for 4 hours</p> <p>iii__ 366-nm blue-light fluorescent halos around MF colonies confirm that Total Coliform is E. coli</p> <p><b>f. SM 9222B/SM 9222I: NW E. coli by Membrane Filtration</b></p> <p>i__ Positive samples determined by SM 9222B subject to SM 9222I using NA-MUG</p> <p><b>g. SM 9222D/SM9020B, 9b: Fecal Coliform by Membrane Filtration</b></p> <p>i__ Inoculate at least 10 colonies from filter into LTB</p> <p>ii__ Incubated at 35.0 ± 0.5 °C for 24 hours (48 hours if no gas production after 24 hr.)</p> <p>iii__ Positive cultures from LTB (gas formation) inoculated into EC medium</p> <p>iv__ EC tubes incubated at 44.5 ± 0.2 °C for 24 hours</p> <p>v__ Positives verified monthly (by picking at least 10 blue colonies from one positive sample); and false negatives picked and verified (SM9020B, 9.b.2)</p> <p>vi__ Count is adjusted based on these results.</p> <p><b>Note:</b> May inoculate m-FC colonies directly into EC medium. However, if gas is observed in LTB, but not in the corresponding EC tube, another EC tube must be inoculated &amp; tested with the positive culture from the LTB tube.</p>	<p align="center">40 CFR 136.3 Table IA</p>					<p>d340d1 d340d2  d340d3   d340e1   d340e2 d340e3    d340   d340f1 d340f2  d340f3  d340f4 d340f5  d340f6</p>	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

<b>Relevant Aspect of Standards</b>	<b>TNI 2016/ELAP/ Regulation/Method Reference</b>	<b>Y</b>	<b>N</b>	<b>N/A</b>	<b>S</b>	<b>ELAP citation codes</b>	<b>Comments</b>
<p><b><i>Constant and Consistent Test Conditions:</i></b>  <u>Laboratory Facility:</u>                      1. Floors and work surfaces are non-absorbent and easy to clean and disinfect.                      2. Work surfaces are adequately sealed.                      3. Sufficient storage space is provided                      4. Laboratory is clean and free from dust accumulation.</p>	M5,1.7.3.7(a)					00d354r  00d353r 00d355ar 00d355cr	
<p><u>Laboratory Equipment:</u>                      1. Temperature Measuring Devices:                          a. The temperature measurement devices have the appropriate quality needed to achieve the specification in the test method.                          b. The temperature measurement devices are verified to national or international standards for temperature at least annually.                          Note: See ELAP Certification Manual Item 231 for required frequency.                          c. Temperature measuring device verification may be accomplished using a single point that represents the method mandated temperature and use conditions.                          d. The graduation and range of the temperature measuring devices are appropriate for the required accuracy of measurement.                      2. Sterilization Equipment:                          a. The performance of each autoclave is initially evaluated by establishing its functional properties and performance, i.e. heat distribution characteristics established with respect to typical uses.                          b. Autoclaves shall meet specified temperature tolerances.                          Note: Pressure cookers cannot be used for sterilization of growth media.                          c. Proper sterilization temperature is demonstrated by use of continuous temperature recording device or use of a maximum registering thermometer each autoclave cycle.                          d. Effective sterilization is demonstrated using appropriate biological indicators at least once each month the autoclave is used.                          e. The biological indicator used in the autoclave shall be effective at the sterilization temperature and time needed to sterilize lactose-based media.                          f. Temperature sensitive tape is used each autoclave run to indicate the contents have been processed.</p>	M5,1.7.3.7(b)(i)  M5,1.7.3.7(b)(i)  M5,1.7.3.7(b)(i)  M5,1.7.3.7(b)(i)  M5, 1.7.3.7(b)(ii)(a)(1)  M5, 1.7.3.7(b)(ii)(a)(1) M5, 1.7.3.7(b)(ii)(a)(2)  M5, 1.7.3.7(b)(ii)(a)(2) M5, 1.7.3.7(b)(ii)(a)(2)  M5, 1.7.3.7(b)(ii)(a)(2)					00d356r  00d357r  0d357a  00d358  00d360r  00d361r  00d361a  00d361b  00d361dr  00d361c	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
<p>g. Records of autoclave operation maintained for every cycle include the following:</p> <ul style="list-style-type: none"> <li>i. ___ Date,</li> <li>ii. ___ Contents,</li> <li>iii. ___ Maximum temperature reached,</li> <li>iv. ___ Time in sterilization mode,</li> <li>v. ___ Total run time (may be recorded as time in and time out),</li> <li>vi. ___ Analyst's initials, and</li> <li>vii. ___ Pressure</li> </ul> <p>h. Annual autoclave maintenance, internally or by service contract, is performed and records are maintained.</p> <p>i. Annual autoclave maintenance includes a pressure check and verification of temperature device. Note: If it has been determined that the autoclave has no leaks, pressure checks can be documented using the formula PV=nRT.</p> <p>j. The autoclave mechanical timing device is checked quarterly against a stopwatch and the actual time elapsed is documented.</p> <p>k. Ovens used for sterilization are checked for sterilization effectiveness monthly with appropriate biological indicator.</p> <p>l. Records of oven operation maintained for every cycle include the following:</p> <ul style="list-style-type: none"> <li>i. ___ Date,</li> <li>ii. ___ Cycle time,</li> <li>iii. ___ Temperature,</li> <li>iv. ___ contents, and</li> <li>v. ___ Analyst's initials.</li> </ul> <p>m. Temperature sensitive tape is used each oven run to indicate the contents have been processed.</p> <p>3. Volumetric Equipment:</p> <ul style="list-style-type: none"> <li>a. Volumetric equipment with movable parts such as automatic dispensers, dispensers/diluters, and mechanical hand pipettes are verified by the laboratory for accuracy quarterly and documented.</li> <li>b. Volumetric equipment such as filter funnels, bottles, non-Class A glassware, and other marked containers are verified by the laboratory once per lot prior to first use.</li> </ul>	<p>M5, 1.7.3.7(b)(ii)(a)(3)</p> <p>M5, 1.7.3.7(b)(ii)(a)(4)</p> <p>M5, 1.7.3.7(b)(ii)(a)(4)</p> <p>M5, 1.7.3.7(b)(ii)(a)(5) M5,1.7.3.7(b)(ii)(b)</p> <p>M5,1.7.3.7(b)(ii)(b)</p> <p>M5,1.7.3.7(b)(ii)(b)</p> <p>M5,1.7.3.7(b)(iii)(a)</p> <p>M5,1.7.3.7(b)(iii)(b)</p>					<p>00d361er or 0d362ar 0d362br 0d362cr 0d362dr 0d362er 0d362fr 0d362gr</p> <p>00d363zr</p> <p>0d363ar</p> <p>0d363br</p> <p>0d367ar</p> <p>d367bar d367bbr d367bcr d367bdr d367ber d367bfr</p> <p>00d364r</p> <p>00d365r</p>	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
c. The volume of disposable volumetric equipment such as sample bottles, disposable pipettes, and micropipette tips are checked by the laboratory once per lot.	M5,1.7.3.7(b)(iii)(c)					0d365ar	
d. Verification of volume is within 2.5% of expected volume. Note: This verification can be volumetric as compared to Class A or gravimetric.	M5,1.7.3.7(b)(iii)(d)					0d365br	
4. UV Instruments:							
a. UV instruments used for sanitization are tested quarterly for effectiveness with an appropriate UV light meter or by agar spread plate irradiation test.	M5,1.7.3.7(b)(iv); ELAP Certification Manual Item 231					0d366ar	
b. Bulbs are replaced if output is less than 70% of original for light tests or if count reduction is less than 99% for a plate containing 200 to 300 organisms.	M5,1.7.3.7(b)(iv); ELAP Certification Manual Item 231					00d366r	
5. Incubators:							
a. The uniformity of temperature distribution and time required to achieve equilibrium conditions in incubators and water baths is established prior to first use after installation and service. The equilibrium check includes the time required after test sample addition to re-establish equilibrium conditions under full capacity load for intended use. Note: Position, space between and height of stacks of Petri dishes established. Dishes are not to be stacked more than 4 high. SimPlate plates can be stacked higher than 4.	M5,1.7.3.7(b)(v)(a)					000d32a	
b. When samples are under test, temperatures of incubators and water baths are recorded twice daily separated by <b>at least 4 hours</b> . Data loggers, continuous temperature monitoring devices, or other temperature monitoring equipment can be used if it is calibrated in accordance with the TNI V1M2 Section 5.5.13.1 and ELAP Certification Manual Item No. 231 requirements. <b>Note:</b> "Under test" is defined as the time period that the sample is in the incubation phase of the method. <b>Note:</b> There is no intent to take incubator temperatures when there are no samples under test.	M5,1.7.3.7(b)(v)(b)					000d32r	
6. Labware:							
a. The laboratory has a documented procedure for washing labware, if applicable. Note: Labware is glassware and plasticware.	M5,1.7.3.7(b)(vi)(a)					0d367dr	
b. Only detergents designed for labware are used.	M5,1.7.3.7(b)(vi)(a)					0d367er	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
<p>c. Glassware used is made of borosilicate or other non-corrosive material, free of chips and cracks, and has readable measurement marks.</p> <p>d. The laboratory tests labware that is washed and reused for the possible presence of residues which may inhibit or promote growth of microorganisms by performing the Inhibitory Residue Test initially, and each time the lab changes detergent formulation or washing procedures.</p> <p>e. Washed labware is tested at least once daily, each day of washing, for possible acid or alkaline residue by testing one piece of glassware with a suitable pH indicator such as bromothymol blue. Records of tests are maintained.</p>	M5,1.7.3.7(b)(vi)(a)					00d368r	
	M5,1.7.3.7(b)(vi)(b)						
	M5,1.7.3.7(b)(vi)(c)					00d369r	
	M5,1.7.3.7(b)(vi)(c)						
Original containers of reagents and media are labeled with an expiration date.	M2,5.6.4.2(b)					51026r	
<p><b>Media, solution and reagent requirements:</b></p> <p><b>a. ___ Heterotrophic Plate Count Medium</b> (SM9215A, 6, SM9215B, 3a, SM9215C, 2-3, and SM9215D, 2-3):</p> <p>___ Plate count agar autoclaved at 121 °C for 15 minutes. R2A agar heated and sterilized at 121 °C for 15 minutes.</p> <p>___ Final pH 6.8-7.2 for Plate Count Agar, 7.2 for R2A Agar.</p> <p>___ Sterile agar medium melted not more than once.</p> <p>___ Melted agar used within 3 hours; agar tempered at 44-46 °C before pouring.</p> <p>___ A separate “temperature” container exposed to same heating and cooling as medium. (Do not depend on the sense of touch.)</p> <p>___ Blood agar and chocolate agar are not used with dialysis related product water (ANSI/AAMI RD52:2004, 7.2.3 and RD62:2006, 5.1.1).</p> <p><b>b. ___ Phosphate buffer</b> (SM9050C):</p> <p>___ Stock buffer autoclaved at 121 C for 15 minutes.</p> <p>___ Stock buffer final pH 7.2 ± 0.5</p> <p>___ Dilution rinse water prepared from stock buffer &amp; MgCl<sub>2</sub>.</p> <p>___ A commercially available preparation can also be used.</p>						00d335a	
<p><b>c. ___ Peptone water</b> (SM9050B, 1b):</p> <p>___ 10% peptone stock solution autoclaved or filter sterilized.</p> <p>___ 0.1% peptone water prepared as dilution rinse water.</p> <p>___ Final pH 6.6-7.0.</p>						00d335c	



**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
<p><b>d. ___ m-Endo Medium (SM9222B, 2):</b>            ___ Medium brought to a boil, but not boiled, removed immediately; and not autoclaved.            ___ Ethanol used is not denatured.            ___ Prepared in sterilized flask.            ___ Final pH 7.0-7.4 for m-Endo Agar LES; 7.0-7.4 for m-Endo medium.            ___ Uninoculated media discarded if growth or surface sheen observed.</p> <p><b>e. ___ Lauryl Tryptose (Lauryl Sulfate) or Lactose Broth (SM9221B, SM9222B, 1a):</b>            ___ Formulated so that concentration is single strength after sample addition.            ___ Autoclaved at 121 °C for 12-15 minutes.            ___ Final pH 6.6-7.0.            ___ Inverted vials in sterilized media, one-third to one-half covered by media, &amp; free of air bubbles.</p> <p><b>f. ___ Brilliant Green Lactose Bile Broth (SM9221B, SM9222B, 2a):</b>            ___ Autoclaved at 121 °C for 12-15 minutes.            ___ Final pH 7.0-7.4.</p> <p><b>g. ___ Presence-Absence Test Medium (SM9221D, 1a):</b>            ___ Autoclaved at 121 °C for 12 minutes, with space allowed between bottles.            ___ Final pH 6.6-7.0.</p> <p>Note: Media 6x formulation strength may be used if media is filter sterilized instead of autoclaved.</p> <p><b>h. ___ EC Medium (SM9221E, 1a):</b>            ___ Autoclaved at 121 °C for 12-15 minutes.            ___ Final pH 6.7-7.1.            ___ Inverted tubes one-third to one-half covered by media &amp; free of air bubbles.</p> <p><b>i. ___ MMO-MUG Medium (Colilert, Colisure, E*Colite, ReadyCult, Fluorocult, or Colitag) (SM9223B, 1):</b>            ___ Commercial preparation used.            ___ Protected from light.            ___ Not autoclaved.</p>	<p>40 CFR 141.21(f)(3)</p>					<p>00d335d</p> <p>00d335e</p> <p>00d335f</p> <p>00d335g</p> <p>00d335h</p> <p>00d335i</p>	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
<p><b>j. ___ EC Medium + MUG (EPA 1104, 7):</b>            ___ Autoclaved at 121 °C for 12-15 minutes.            ___ Final MUG concentration 50 ug/ml.            ___ Final pH 6.7-7.1.            ___ Inverted vial in test tube not used.            ___ Checked for absence of fluorescence prior to use (with 366-nm UV light).</p> <p><b>k. ___ Nutrient Agar Medium + MUG (EPA 1105, 7):</b>            ___ Autoclaved in 100-mL volumes at 121 °C for 15 minutes.            ___ Final MUG concentration 100 ug/ml.            ___ Final pH 6.6-7.0.</p> <p><b>l. ___ A-1 Medium (SM9221E, 2a):</b>            ___ Autoclaved at 121 °C for 10 minutes.            ___ Final pH 6.8-7.0.            ___ Inverted tubes one-third to one-half covered by media &amp; free of air bubbles.            (Note: May be stored in the dark at room temperature, but must be used within 1 week.)</p> <p><b>m. ___ Chromocult (used with Membrane Filtration):</b>            ___ Not autoclaved or overheated.            ___ Final pH 6.6-7.0.</p> <p><b>n. ___ Coliscan (used with Membrane Filtration):</b>            ___ Not autoclaved or overheated.            ___ Final pH 6.6-7.0.</p> <p><b>o. ___ m-ColiBlue-24 (used with Membrane Filtration):</b>            ___ Inverted 2-3 times to mix contents before opening broth ampule.            ___ Final pH 6.8-7.2.</p> <p><b>p. ___ Enterolert:</b>            ___ Commercial preparation used.            ___ Protected from light.            ___ Not autoclaved.</p> <p><b>q. ___ mEI (EPA 1600, 7.5):</b>            ___ Filter-sterilized solution.            ___ Final pH 6.9-7.3.</p>						<p>00d335j</p> <p>00d335k</p> <p>00d335l</p> <p>00d335m</p> <p>00d335n</p> <p>00d335o</p> <p>00d335p</p> <p>00d335q</p>	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
<p><b>r. ___ m-FC Broth or Agar (SM9222D, 1a &amp; EPA-600/8-78-017, Part II-B, 5.2.1):</b>            ___ Medium brought to boiling point, removed immediately, not autoclaved, and final pH of 7.2-7.6.            (Note: m-FC Broth or Agar medium may be used without 1% rosolic acid provided no background growth.)</p> <p><b>s ___ SimPlate:</b>            ___ Commercial preparation used.            ___ Stored at 2-30 °C and protected from light.            ___ Commercial preparation used.            ___ pH 6.7 - 7.3</p>						00d335rr  00d335s	
<p><b>ISO 11731:2017(E), 8.2 – 8.5: Legionella</b></p> <p align="center"><u>Concentration of Water Samples</u></p> <p>1__ Filtered sample using vacuum filtration or positive pressure filtration, or centrifuged sample where concentration by filtration is not possible.</p> <p>2__ Filtered an appropriate volume of sample based on particulate content or desired detection level.</p> <p>3__ MF and direct plating: Filtered water sample (without treatment, after acid treatment, and if required, after heat treatment) through <b>cellulose nitrate or mixed cellulose esters membrane</b> filter, and placed filter (right-side up) directly onto culture media, ensuring no air bubble is trapped.</p> <p>4__ MF followed by washing: Filtered water through <b>polycarbonate or polyethersulfone</b> membrane filter.            Note: Placed filter right side down in a screw cap, sterile container with or without sterile beads.</p> <p>5__ Washed filter using 5 to 10 ml of sterile diluent, or sample, and vortexed for at least 2 minutes, or alternatively, placed the container in an ultrasonic bath, ensuring the level of diluent is below the level of the water in the bath, for an optimum time interval for maximum recovery.            Note: Filters may be cut into pieces using sterile scissors to aid elution.            Also, refer to NOTES 1-3 in method.</p> <p>6__ Divided concentrate into one portion untreated, one portion with heat, and one portion for treatment with acid solution.</p>						0d31k1  0d31k2  0d31k3  0d31k4  0d31k5  0d31k6	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
<p align="center"><u>Sample Pre-Treatment</u></p> <p>7__ Heat: Added sample (concentrated or unconcentrated) into a sterile container and placed in a water bath at 50 ± 1 °C for 30 ± 2 min. Note: Small volumes (&lt;= 5 ml) should be used.</p> <p>8__ Acid: Diluted one volume of the sample (concentrated or unconcentrated) with nine volumes of an acid solution, mixed well and left to stand for 5.0 ± 0.5 min.</p> <p align="center">OR</p> <p>9__ Acid: Transferred around 30 ml acid solution onto membrane filter, left for 5.0 ± 0.5 min, and rinsed the filter with at least 20 ml of the diluent.</p> <p align="center"><u>Plating and Inoculation</u></p> <p>10__ For samples expected with high concentration of Legionella (&gt;10<sup>4</sup> cfu/l) and low concentration of interfering microorganisms: plated sample directly and inoculated and spread 0.1 ml to 0.5 ml sample onto one plate of BCYE and one plate of BCYE+AB agar.</p> <p>11__ For samples expected with low concentration of Legionella and low concentration of interfering microorganism: placed untreated, filtered sample onto one plate of BCYE agar. For acid treated filters, the samples are placed on one or more plates of BCYE+AB agar or GVPC or MWY agar.</p> <p align="center">OR</p> <p>12__ Samples expected with low concentration of Legionella and low concentration of interfering microorganism: inoculated and spread 0.1 ml to 0.5 ml of each concentrated portion of untreated, heat treated and acid treated sample from membrane filtration with washing onto one plate of BCYE agar and on one or more plates of BCYE+AB agar or GVPC or MWY agar.</p> <p>13__ Samples expected with a high concentration of interfering microorganisms: spread 0.1 ml to 0.5 ml of each portion of untreated, heat treated, and acid treated subsample onto one plate of GVPC or MWY agar.</p> <p>14__ Samples expected with an extremely high concentration of interfering microorganisms: spread 0.1 ml to 0.5 ml of each portion of subsample, that has been heat treated then acid treated and mixed well by a vortex mixer or in an ultrasonic water bath, onto one plate of GVPC or MWY agar.</p>						<p>0d31k7</p> <p>0d31k8</p> <p>0d31k9</p> <p>0d31k10</p> <p>0d31k11</p> <p>0d31k12</p> <p>0d31k13</p> <p>0d31k14</p>	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
<p align="center"><u>Incubation</u></p> <p>15__ Plates inverted and incubated plates at 36 ± 2 °C for 10 d in a humid atmosphere to prevent desiccation of plates. Note: Inoculated media left to stand until inoculated volume is absorbed. <b>Note:</b> <i>Environmental samples (including but not limited to surface, potable, non-potable, industrial, waste, and cooling tower samples) showing new growth at 7 days must be incubated up to 10 days to ensure colony is large enough for isolation. The laboratory must have language in their SOP regarding their incubation duration procedures.</i></p>						0d31k15	
<p align="center"><u>Examination of Plates</u></p> <p>16__ Plates inspected for the first time on day 2, 3, 4, or 5 followed by a final inspection at the end of the incubation period (i.e. day 7 or day 10, dependent on the nature of the sample), and the # of each colony type recorded. Check the plates on day 2 to determine if dilutions are needed. Note: With outbreak investigations, it is advisable for samples with expected high concentration of interfering microorganisms to check the plates on day 2.</p>						0d31k16	
<p align="center"><u>Subculturing/Confirmation</u></p> <p>17__ Subculture 3 presumptive colonies from plate(s) showing the highest counts when there is only one colony type. First inoculate BCYE-cys (or alternate media in note 6.1.2) and then BCYE.</p>						0d31k17	
<p>18__ Subculture at least 1 colony of each type if more than 1 presumptive type of colony is present. First inoculate BCYE-cys (or alternate media in note 6.1.2) and then BCYE.</p>						0d31k18	
<p>19__ Subcultured plates incubated at 36 ± 2 °C for 2 d to 5 d in a humid atmosphere to prevent desiccation of plates. Note: It is acceptable to stop the incubation at day 2 for those samples that are easily confirmed.</p>						0d31k19	
<p>20__ With outbreak investigations, subculture and incubated at least 5 presumptive colonies if only one morphology is present, or 2 presumptive colonies for each type of morphology if multiple morphologies present.</p>						0d31k20	

**NYSDOH ELAP MICROBIOLOGY CHECKLIST**

Relevant Aspect of Standards	TNI 2016/ELAP/ Regulation/Method Reference	Y	N	N/A	S	ELAP citation codes	Comments
<p align="center"><u>Recording Results</u></p> <p>21__ Record the results of all plates. Regard as Legionella those colonies that grow on BCYE agar but fail to grow on BCYE-cys agar.</p> <p>22__ Record volume filtered.</p> <p>23__ Record volume concentrated and final volume.</p> <p>24__ Record the inoculated volume.</p> <p>Note: Record issues can also be cited using a code in the Quality System checklist, Section 13 – Control of Records.</p> <p align="center"><u>Legionella Reagents and Media:</u></p> <p><b>a__ Phosphate Buffered Saline (ISO 11731, 6.2 Annex C):</b></p> <p>___ Commercially available preparation used</p> <p>___ Reconstituted according to manufacturer’s instructions</p> <p>___ Final pH 7.5</p> <p>___ Sterility check performed once per lot using double-strength non-selective medium, 35 °C, 24 hour</p> <p><b>b__ BCYE (ISO 11731:2017(E), Annex B.1):</b></p> <p>___ L-cysteine and iron solutions prepared fresh, sterilized through filtration, and stored at -20 ± 3 °C for not more than 3 months.</p> <p>___ ACES buffer is prepared by mixing 2 solutions – 1) ACES granules dissolved in 500 ml distilled water using a water bath (45-50 °C) and 2) KOH pellets dissolved in 480 ml distilled water using gentle shaking.</p> <p>___ Charcoal, yeast extract and α-ketoglutarate added sequentially to ACES buffer.</p> <p>___ H2SO4or KOH used to adjust pH to 6.8 ± 0.2.</p> <p>___ Agar added and mixed to ACES solution, autoclaved at 121 ± 3 °C for 15 ± 1 min, and cooled in a water bath to 48 ± 3 °C.</p> <p>___ L-cysteine and iron solutions added aseptically, mixing well between additions.</p> <p>___ Final pH is 6.8 ± 0.2 at 25 °C.</p> <p>___ Stored at 5 ± 3 °C in airtight containers and protected from light for 3 months.</p> <p><b>c__ BCYE-cys (ISO 11731:2017(E), Annex B.2):</b></p> <p>___ Prepared as noted above for BYCE, except that L-cysteine is omitted.</p> <p>___ Stored at 5 ± 3 °C in airtight containers in the dark for 3 months.</p>						<p>0d31k21</p> <p>0d31k22</p> <p>0d31k23</p> <p>0d31k24</p> <p>00d335y</p> <p>00d335t</p> <p>00d335u</p>	



