

**NEW YORK**  
state department of  
**HEALTH**

Nirav R. Shah, M.D., M.P.H.  
Commissioner

Sue Kelly  
Executive Deputy Commissioner

### Clinical Chemistry Proficiency Testing – September 10, 2012

Enclosed are results from the clinical chemistry proficiency survey shipped September 10, 2012. Test samples were prepared in-house by the quantitative transfer of constituents to pooled human serum. The material was subsequently sterile filtered, dispensed into aliquots and stored frozen at -80 °C. Five specimens (**C76, C77, C78, C79, C80**) were distributed to each participant for analysis.

Outlined below is a description of the process utilized in the evaluation of your laboratory's proficiency test results. A summary of your laboratory's performance for the three most recent surveys is also included with your report.

**Target Value:** In general, targets utilized are derived from all-participant mean values calculated by robust statistical technique. In some cases, however, it is recognized that method, reagent, and/or instrument specific targets may be required and "peer group" specific targets are used where appropriate. Should an alternate target be required to evaluate your laboratory's data, an asterisk will be placed adjacent to the method, reagent, and/or instrument listing(s) corresponding to the peer group utilized.

**Acceptable Range:** Represents limits established using criteria specified by CLIA '88 regulations, allowing for rounding to appropriate significant digits. Results falling within this range are scored as 100%. Any result exceeding these limits is considered unsatisfactory and receives a score of 0%. Laboratories must achieve an overall analyte score  $\geq 80\%$  in order to meet performance criteria for that analyte.

**Range Plots:** Plots relative distance of your laboratory's result (represented by an "x") from the target for each sample analyzed. Any result exceeding the high or low limit by  $>20\%$  of the acceptable range is indicated by an asterisk (\*).

**Not Gradable:** Results for graded analytes for a few laboratories using unique instrument, reagent, or instrument/reagent combinations were considered "not gradable". For these laboratories, pass credit (100%) has been issued. **If a laboratory is unable to participate in the NYS clinical chemistry proficiency test event as a graded participant for any analyte, it is the responsibility of that laboratory to establish alternate means to verify the accuracy and precision of the test system for any non-graded analyte(s).**

**Troponin I, Troponin T, and Estimated Glomerular Filtration Rate:** These analytes were included in the September 2012 test event for evaluation only. Although results were not graded, targets and acceptable ranges appropriate for your laboratory's methodology are provided.

The attached statistical report provides a summary of participant data for the five survey specimens. Results for individual instrument and reagent systems where the number of laboratories using those systems is three or greater are provided. Mean and Standard Deviation (1 SD) values shown are calculated by a robust statistical technique that does not assume a Gaussian distribution. These statistical reports are also available on the internet at: <http://www.wadsworth.org/chemheme>

Should you have any questions regarding these reports or wish to obtain an additional copy, please contact the Clinical Chemistry Section at (518) 474-5582.

Note: The use of brand and/or trade names in this report does not constitute an endorsement of the products on the part of the Wadsworth Center or the New York State Department of Health

## Summary of Participant Performance (Mean and Standard Deviation)

## Glucose (mg/dL)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
144.2 ± 3.67	92.5 ± 3.19	44.9 ± 2.20	113.5 ± 3.43	291.4 ± 6.86	n = 385	[---] All Methods & Instruments
						<Instruments>
146.7 ± 0.51	95.7 ± 0.51	48.7 ± 0.51	116.0 ± 0.90	291.0 ± 0.90	n = 3	[AXA] Abaxis Piccolo
142.3 ± 1.89	91.3 ± 1.31	43.3 ± 0.64	112.6 ± 1.44	294.7 ± 3.18	n = 22	[ABJ] Abbott Architect c System
142.8 ± 3.52	91.0 ± 2.31	44.5 ± 1.33	112.3 ± 2.87	289.1 ± 6.86	n = 57	[OLC] Beckman Coulter AU Chemistry System
144.4 ± 6.47	93.2 ± 5.71	44.3 ± 2.47	113.9 ± 8.55	291.0 ± 20.49	n = 4	[BCS] Beckman Coulter CX
145.8 ± 3.42	93.1 ± 2.35	45.7 ± 1.51	115.9 ± 4.25	293.1 ± 6.20	n = 4	[BCX] Beckman Coulter LX-20
144.3 ± 3.64	91.8 ± 2.65	44.8 ± 1.87	113.4 ± 2.36	290.8 ± 7.72	n = 17	[BCG] Beckman Coulter UniCel DxC 600
143.2 ± 2.01	90.6 ± 1.96	44.0 ± 1.47	113.4 ± 2.01	291.5 ± 5.69	n = 14	[BCH] Beckman Coulter UniCel DxC 800
141.5 ± 6.32	107.0 ± 4.51	63.2 ± 2.36	113.3 ± 4.96	311.4 ± 6.45	n = 3	[HEB] HemoCue B-Glucose
157.5 ± 9.45	125.2 ± 4.16	75.4 ± 3.79	132.9 ± 7.40	319.7 ± 5.24	n = 4	[HEC] HemoCue Glucose 201
144.0 ± 0.00	90.2 ± 0.73	43.5 ± 0.57	109.0 ± 0.55	290.6 ± 3.48	n = 6	[IAA] i-STAT
145.5 ± 3.47	91.9 ± 1.64	43.5 ± 1.81	112.9 ± 2.76	290.1 ± 5.21	n = 11	[JJE] Ortho Vitros 250/350/950
143.3 ± 2.73	90.8 ± 2.02	42.4 ± 1.03	110.8 ± 2.28	289.1 ± 4.36	n = 20	[JJF] Ortho Vitros 5,1FS
142.3 ± 3.00	89.9 ± 2.19	41.9 ± 1.16	110.1 ± 2.18	287.3 ± 6.66	n = 18	[JJG] Ortho Vitros 5600
144.3 ± 3.00	92.0 ± 1.83	44.9 ± 1.17	113.7 ± 2.65	293.5 ± 4.80	n = 17	[ROC] Roche cobas c501
143.3 ± 2.26	91.0 ± 0.90	44.3 ± 0.51	113.0 ± 0.90	292.1 ± 5.72	n = 3	[ROH] Roche cobas c701
142.8 ± 1.14	91.2 ± 1.20	44.7 ± 0.69	112.7 ± 0.94	290.1 ± 3.91	n = 8	[ROS] Roche Cobas INTEGRA 400
141.2 ± 3.10	89.1 ± 1.88	43.3 ± 0.82	110.5 ± 2.30	283.2 ± 6.56	n = 4	[ROT] Roche Cobas INTEGRA 800
144.6 ± 3.93	92.4 ± 2.45	45.0 ± 1.31	114.0 ± 2.98	293.9 ± 6.77	n = 31	[ROD] Roche MODULAR D/P
143.2 ± 2.63	92.4 ± 1.77	44.6 ± 0.97	112.9 ± 2.05	289.7 ± 5.55	n = 22	[BYE] Siemens ADVIA 1800
140.5 ± 1.86	90.7 ± 1.37	43.7 ± 0.51	110.8 ± 1.54	284.1 ± 2.86	n = 3	[BYB] Siemens ADVIA 2400
147.3 ± 1.78	96.4 ± 1.82	47.5 ± 1.13	117.3 ± 1.89	294.9 ± 4.56	n = 14	[DUE] Siemens Dimension EXL
149.1 ± 3.38	98.1 ± 2.46	48.5 ± 1.47	118.4 ± 3.23	296.9 ± 6.86	n = 24	[DUR] Siemens Dimension RxL
144.6 ± 2.77	94.9 ± 2.23	46.3 ± 1.20	114.9 ± 2.85	289.4 ± 5.58	n = 40	[DUT] Siemens Dimension Vista
146.8 ± 3.08	96.4 ± 1.99	47.3 ± 1.22	116.6 ± 2.21	292.3 ± 6.24	n = 21	[DUX] Siemens Dimension Xpand
						<Reagents>
146.7 ± 0.51	95.7 ± 0.51	48.7 ± 0.51	116.0 ± 0.90	291.0 ± 0.90	n = 3	[AX1] Abaxis
142.3 ± 1.89	91.3 ± 1.31	43.3 ± 0.64	112.6 ± 1.44	294.7 ± 3.18	n = 22	[AB1] Abbott
143.8 ± 3.14	91.3 ± 2.39	44.4 ± 1.81	113.3 ± 2.54	290.8 ± 7.16	n = 38	[BC1] Beckman Coulter
142.6 ± 3.31	90.9 ± 2.26	44.4 ± 1.21	112.2 ± 2.75	288.9 ± 6.83	n = 53	[OL1] Beckman Coulter AU Series
144.8 ± 6.95	95.0 ± 6.42	46.3 ± 1.37	116.6 ± 10.30	298.9 ± 18.79	n = 3	[CR1] Carolina
150.2 ± 11.59	117.6 ± 10.71	70.0 ± 7.44	124.3 ± 12.28	315.9 ± 7.17	n = 7	[HE1] HemoCue
144.0 ± 0.00	90.2 ± 0.73	43.5 ± 0.57	109.0 ± 0.55	290.6 ± 3.48	n = 6	[IA1] i-STAT
143.4 ± 3.27	90.8 ± 2.17	42.4 ± 1.25	111.1 ± 2.64	288.9 ± 5.60	n = 50	[JJ1] Ortho Clinical Diagnostics
144.0 ± 2.76	91.7 ± 1.69	44.7 ± 1.09	113.5 ± 2.31	292.9 ± 4.92	n = 22	[RO4] Roche cobas c311/c501/c502/c701
144.6 ± 3.93	92.4 ± 2.45	45.0 ± 1.31	114.0 ± 2.98	293.9 ± 6.77	n = 31	[RO2] Roche Hitachi and Modular D/P
142.2 ± 2.08	90.6 ± 1.71	44.4 ± 1.01	112.1 ± 1.91	288.1 ± 6.09	n = 12	[RO1] Roche Integra and MIRA
143.1 ± 3.10	92.4 ± 2.03	44.6 ± 1.05	112.7 ± 2.31	289.2 ± 6.17	n = 27	[BY1] Siemens ADVIA/ADVISIA Centaur
146.5 ± 3.45	96.2 ± 2.50	47.1 ± 1.57	116.4 ± 3.03	292.7 ± 6.90	n = 99	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Urea Nitrogen (mg/dL)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
19.2 ± 1.31	10.9 ± 0.72	26.9 ± 1.45	52.1 ± 2.69	22.3 ± 1.49	n = 368	[---] All Methods & Instruments
						<Instruments>
19.0 ± 0.00	11.3 ± 0.51	25.7 ± 0.51	52.7 ± 1.37	21.7 ± 0.51	n = 3	[AXA] Abaxis Piccolo
19.3 ± 0.51	11.0 ± 0.00	27.1 ± 0.54	52.8 ± 1.10	22.5 ± 0.57	n = 21	[ABJ] Abbott Architect c System
19.5 ± 0.84	11.1 ± 0.43	27.5 ± 0.94	52.4 ± 1.79	22.7 ± 0.80	n = 55	[OLC] Beckman Coulter AU Chemistry System
20.0 ± 0.75	11.2 ± 1.46	27.9 ± 1.13	53.4 ± 3.54	23.1 ± 1.13	n = 4	[BCS] Beckman Coulter CX
18.2 ± 1.27	9.1 ± 1.13	25.0 ± 0.75	48.7 ± 2.43	19.9 ± 1.88	n = 4	[BCX] Beckman Coulter LX-20
19.8 ± 0.88	11.5 ± 0.57	27.6 ± 0.96	52.8 ± 1.58	22.9 ± 1.08	n = 16	[BCG] Beckman Coulter UniCel DxC 600
18.5 ± 1.08	9.9 ± 0.58	25.6 ± 1.02	50.0 ± 0.00	20.7 ± 0.82	n = 15	[BCH] Beckman Coulter UniCel DxC 800
22.0 ± 0.00	11.0 ± 0.00	28.4 ± 0.79	59.0 ± 0.90	26.5 ± 0.57	n = 6	[IAA] i-STAT
16.0 ± 0.60	10.0 ± 0.00	24.0 ± 0.79	47.0 ± 0.95	19.7 ± 0.76	n = 11	[JJE] Ortho Vitros 250/350/950
15.6 ± 0.55	9.8 ± 0.48	23.3 ± 0.52	46.4 ± 0.86	18.8 ± 0.54	n = 20	[JJF] Ortho Vitros 5,1FS
15.6 ± 0.63	9.8 ± 0.46	23.7 ± 0.61	46.2 ± 0.79	18.8 ± 0.65	n = 18	[JJG] Ortho Vitros 5600
19.1 ± 0.61	11.0 ± 0.00	27.1 ± 0.92	52.2 ± 1.49	22.4 ± 0.69	n = 17	[ROC] Roche cobas c501
18.7 ± 0.51	11.0 ± 0.00	27.0 ± 0.90	52.7 ± 0.51	22.7 ± 0.51	n = 3	[ROH] Roche cobas c701
19.3 ± 0.72	11.0 ± 0.00	26.9 ± 0.59	52.9 ± 1.82	22.3 ± 0.72	n = 6	[ROS] Roche Cobas INTEGRA 400
19.0 ± 0.00	10.8 ± 0.41	26.8 ± 0.41	53.0 ± 0.00	22.0 ± 0.00	n = 4	[ROT] Roche Cobas INTEGRA 800
19.6 ± 0.82	11.4 ± 0.63	27.5 ± 0.82	52.8 ± 1.14	23.1 ± 0.86	n = 30	[ROD] Roche MODULAR D/P
19.9 ± 0.59	11.0 ± 0.00	27.5 ± 0.61	53.0 ± 1.13	22.5 ± 0.73	n = 22	[BYE] Siemens ADVIA 1800
20.0 ± 0.00	11.0 ± 0.00	27.3 ± 0.51	53.7 ± 0.51	23.0 ± 0.00	n = 3	[BYB] Siemens ADVIA 2400
19.4 ± 0.73	11.0 ± 0.00	27.0 ± 0.00	52.5 ± 1.13	22.5 ± 0.75	n = 14	[DUE] Siemens Dimension EXL
19.6 ± 0.70	11.0 ± 0.71	27.6 ± 1.14	53.4 ± 1.99	22.7 ± 1.09	n = 24	[DUR] Siemens Dimension RxL
19.0 ± 0.67	11.0 ± 0.00	26.9 ± 0.75	52.4 ± 1.52	22.4 ± 0.79	n = 40	[DUT] Siemens Dimension Vista
19.7 ± 0.94	11.1 ± 0.47	27.7 ± 1.18	53.8 ± 1.62	22.9 ± 1.01	n = 20	[DUX] Siemens Dimension Xpand
						<Reagents>
19.0 ± 0.00	11.3 ± 0.51	25.7 ± 0.51	52.7 ± 1.37	21.7 ± 0.51	n = 3	[AX1] Abaxis
19.3 ± 0.51	11.0 ± 0.00	27.1 ± 0.54	52.8 ± 1.10	22.5 ± 0.57	n = 21	[AB1] Abbott
19.1 ± 1.13	10.5 ± 1.28	26.5 ± 1.52	50.9 ± 2.47	21.6 ± 1.66	n = 37	[BC1] Beckman Coulter
19.5 ± 0.84	11.1 ± 0.44	27.4 ± 0.95	52.4 ± 1.81	22.7 ± 0.81	n = 54	[OL1] Beckman Coulter AU Series
22.0 ± 0.00	11.0 ± 0.00	28.4 ± 0.79	59.0 ± 0.90	26.5 ± 0.57	n = 6	[IA1] i-STAT
15.7 ± 0.63	10.0 ± 0.00	23.6 ± 0.71	46.5 ± 0.99	19.0 ± 0.75	n = 51	[JJ1] Ortho Clinical Diagnostics
19.1 ± 0.62	11.0 ± 0.00	27.1 ± 0.87	52.2 ± 1.36	22.4 ± 0.67	n = 22	[RO4] Roche cobas c311/c501/c502/c701
19.6 ± 0.82	11.4 ± 0.63	27.5 ± 0.82	52.8 ± 1.14	23.1 ± 0.86	n = 30	[RO2] Roche Hitachi and Modular D/P
19.0 ± 0.00	11.0 ± 0.00	27.0 ± 0.00	52.9 ± 1.21	22.0 ± 0.00	n = 10	[RO1] Roche Integra and MIRA
19.9 ± 0.49	11.0 ± 0.00	27.5 ± 0.61	53.1 ± 1.06	22.5 ± 0.71	n = 27	[BY1] Siemens ADVIA/ADVIS Centaur
19.3 ± 0.80	11.0 ± 0.51	27.2 ± 0.95	52.9 ± 1.70	22.6 ± 0.93	n = 98	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Creatinine (mg/dL)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
2.12 ± 0.13	1.17 ± 0.14	2.41 ± 0.12	5.19 ± 0.24	1.11 ± 0.15	n = 373	[---] All Methods & Instruments
2.12 ± 0.14	1.15 ± 0.14	2.39 ± 0.13	5.19 ± 0.27	1.08 ± 0.14	n = 206	[---] All IDMS Traceable Methods
2.13 ± 0.11	1.20 ± 0.13	2.43 ± 0.11	5.20 ± 0.21	1.16 ± 0.15	n = 164	[---] All Non-IDMS Traceable Methods
2.12 ± 0.09	1.22 ± 0.12	2.43 ± 0.10	5.19 ± 0.16	1.19 ± 0.12	n = 136	[‐G‐] Alkaline picrate/Jaffe
2.08 ± 0.09	1.18 ± 0.15	2.35 ± 0.09	5.12 ± 0.18	1.12 ± 0.13	n = 139	[‐H‐] Alkaline picrate/Jaffe-IDMS calibration
2.22 ± 0.24	1.09 ± 0.12	2.48 ± 0.18	5.42 ± 0.41	0.98 ± 0.13	n = 28	[‐I‐] Enzymatic
2.24 ± 0.21	1.09 ± 0.11	2.49 ± 0.16	5.45 ± 0.36	0.98 ± 0.12	n = 67	[‐J‐] Enzymatic-IDMS-traceable calibration
<Instruments>						
1.97 ± 0.14	1.08 ± 0.16	2.30 ± 0.09	5.01 ± 0.07	0.88 ± 0.03	n = 3	[AXA] Abaxis Piccolo
2.21 ± 0.05	1.36 ± 0.05	2.36 ± 0.05	5.40 ± 0.09	1.35 ± 0.05	n = 21	[ABJ] Abbott Architect c System
2.08 ± 0.05	1.18 ± 0.05	2.37 ± 0.06	5.08 ± 0.14	1.09 ± 0.03	n = 57	[OLC] Beckman Coulter AU Chemistry System
2.09 ± 0.05	1.15 ± 0.06	2.45 ± 0.06	5.20 ± 0.15	1.20 ± 0.08	n = 4	[BCS] Beckman Coulter CX
2.05 ± 0.06	1.11 ± 0.02	2.40 ± 0.07	5.23 ± 0.09	1.05 ± 0.05	n = 4	[BCX] Beckman Coulter LX-20
2.08 ± 0.07	0.98 ± 0.07	2.34 ± 0.10	5.14 ± 0.18	0.98 ± 0.07	n = 17	[BCG] Beckman Coulter UniCel DxC 600
2.06 ± 0.05	1.10 ± 0.00	2.39 ± 0.05	5.21 ± 0.06	1.06 ± 0.08	n = 15	[BCH] Beckman Coulter UniCel DxC 800
2.40 ± 0.00	1.14 ± 0.06	2.82 ± 0.08	6.00 ± 0.32	1.00 ± 0.00	n = 5	[IAA] i-STAT
2.39 ± 0.06	1.16 ± 0.05	2.58 ± 0.05	5.72 ± 0.10	1.06 ± 0.05	n = 11	[JJE] Ortho Vitros 250/350/950
2.36 ± 0.11	1.12 ± 0.08	2.58 ± 0.12	5.64 ± 0.24	1.02 ± 0.09	n = 20	[JJF] Ortho Vitros 5,1FS
2.34 ± 0.10	1.11 ± 0.10	2.55 ± 0.11	5.58 ± 0.16	1.02 ± 0.11	n = 18	[JJG] Ortho Vitros 5600
2.06 ± 0.08	0.99 ± 0.04	2.40 ± 0.08	5.13 ± 0.15	0.99 ± 0.11	n = 18	[ROC] Roche cobas c501
2.06 ± 0.07	1.15 ± 0.11	2.38 ± 0.08	5.10 ± 0.23	1.10 ± 0.07	n = 3	[ROH] Roche cobas c701
2.02 ± 0.09	1.08 ± 0.07	2.36 ± 0.09	4.93 ± 0.13	1.00 ± 0.00	n = 7	[ROS] Roche Cobas INTEGRA 400
2.01 ± 0.05	1.00 ± 0.08	2.38 ± 0.04	5.15 ± 0.24	0.94 ± 0.12	n = 4	[ROT] Roche Cobas INTEGRA 800
2.13 ± 0.12	1.24 ± 0.20	2.42 ± 0.10	5.19 ± 0.17	1.14 ± 0.19	n = 30	[ROD] Roche MODULAR D/P
2.10 ± 0.08	1.33 ± 0.11	2.27 ± 0.08	5.05 ± 0.11	1.27 ± 0.08	n = 22	[BYE] Siemens ADVIA 1800
2.03 ± 0.09	1.28 ± 0.06	2.15 ± 0.11	4.82 ± 0.23	1.16 ± 0.12	n = 3	[BYB] Siemens ADVIA 2400
2.12 ± 0.06	1.25 ± 0.07	2.46 ± 0.06	5.21 ± 0.13	1.21 ± 0.05	n = 14	[DUE] Siemens Dimension EXL
2.13 ± 0.12	1.21 ± 0.14	2.43 ± 0.13	5.15 ± 0.17	1.19 ± 0.14	n = 24	[DUR] Siemens Dimension RxL
2.09 ± 0.11	1.19 ± 0.11	2.42 ± 0.12	5.19 ± 0.15	1.17 ± 0.11	n = 40	[DUT] Siemens Dimension Vista
2.13 ± 0.11	1.24 ± 0.12	2.46 ± 0.12	5.15 ± 0.14	1.24 ± 0.13	n = 20	[DUX] Siemens Dimension Xpand
<Reagents>						
1.97 ± 0.14	1.08 ± 0.16	2.30 ± 0.09	5.01 ± 0.07	0.88 ± 0.03	n = 3	[AX1] Abaxis
2.21 ± 0.06	1.36 ± 0.06	2.36 ± 0.05	5.40 ± 0.09	1.35 ± 0.05	n = 22	[AB1] Abbott
2.07 ± 0.06	1.06 ± 0.09	2.38 ± 0.08	5.19 ± 0.12	1.03 ± 0.09	n = 38	[BC1] Beckman Coulter
2.08 ± 0.06	1.18 ± 0.05	2.37 ± 0.06	5.07 ± 0.14	1.09 ± 0.03	n = 54	[OL1] Beckman Coulter AU Series
2.11 ± 0.02	1.17 ± 0.06	2.43 ± 0.05	5.26 ± 0.10	1.20 ± 0.09	n = 3	[CR1] Carolina
2.42 ± 0.04	1.12 ± 0.04	2.85 ± 0.06	6.10 ± 0.18	0.98 ± 0.04	n = 4	[IA1] i-STAT
2.37 ± 0.10	1.13 ± 0.08	2.57 ± 0.11	5.66 ± 0.21	1.03 ± 0.09	n = 51	[JJ1] Ortho Clinical Diagnostics
2.05 ± 0.08	1.03 ± 0.12	2.39 ± 0.08	5.11 ± 0.16	1.01 ± 0.11	n = 24	[RO4] Roche cobas c311/c501/c701
2.13 ± 0.12	1.24 ± 0.20	2.42 ± 0.10	5.19 ± 0.17	1.14 ± 0.19	n = 30	[RO2] Roche Hitachi and Modular D/P
2.01 ± 0.07	1.06 ± 0.08	2.37 ± 0.08	4.99 ± 0.19	1.01 ± 0.10	n = 11	[RO1] Roche Integra and MIRA
2.09 ± 0.07	1.32 ± 0.10	2.26 ± 0.09	5.04 ± 0.11	1.24 ± 0.12	n = 27	[BY1] Siemens ADVIA/ADVISIA Centaur
2.11 ± 0.11	1.21 ± 0.12	2.44 ± 0.12	5.18 ± 0.15	1.20 ± 0.12	n = 98	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

Estimated Glomerular Filtration Rate (mL/min/1.73 m<sup>2</sup>)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
27.7 ± 2.60	53.6 ± 7.55	24.0 ± 1.88	9.8 ± 0.69	56.3 ± 8.92	n = 287	[---] All Methods & Instruments
27.1 ± 2.60	53.5 ± 6.91	23.5 ± 1.87	9.6 ± 0.74	56.9 ± 8.22	n = 165	[-A-] IDMS-traceable MDRD Study Equation
27.9 ± 1.90	51.4 ± 5.85	24.4 ± 1.31	10.0 ± 0.00	51.8 ± 4.80	n = 98	[-B-] Original MDRD Study Equation (4-variable)
30.8 ± 2.27	63.1 ± 9.55	26.2 ± 1.54	10.5 ± 0.75	69.2 ± 12.43	n = 20	[-F-] CKD-EPI Equation
27.9 ± 3.72	60.6 ± 6.14	23.3 ± 2.26	9.7 ± 1.37	64.5 ± 6.32	n = 3	[-Z-] Other

## Target values and allowable ranges for Estimated Glomerular Filtration Rate (eGFR):

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Method
27 (21-33)	55 (43-66)	24 (18-29)	10 (7-12)	59 (47-71)	IDMS-traceable MDRD Study Equation
29 (22-35)	55 (44-67)	25 (19-30)	10 (8-13)	58 (46-69)	Original MDRD Study Equation
30 (24-36)	63 (50-76)	26 (20-32)	10 (8-13)	68 (54-82)	CKD-EPI Equation
42 (33-51)	76 (61-92)	37 (29-45)	17 (13-21)	80 (64-97)	Cockcroft-Gault Equation

Laboratories were asked to report Estimated Glomerular Filtration Rate (eGFR) for samples C76-C80 for a 32-year-old non-African American woman weighing 70 kg.

Target values for eGFR calculated by the MDRD Study Equations were derived from participant mean values for serum creatinine for both conventional and isotope dilution mass spectroscopy (IDMS) calibration methods and application of the appropriate 4-variable MDRD Study equation.

Target values for eGFR calculated by the CKD-EPI equation were derived from participant mean values for IDMS-traceable serum creatinine methods and application of the CKD-EPI equation.

Target values for eGFR calculated by the Cockcroft-Gault equation were derived from all-method mean values for serum creatinine and application of the Cockcroft-Gault formula.

Allowable ranges are ± 20% of the target eGFR for all samples/equations.

Note: the NKDEP recommends reporting estimated GFR values greater than or equal to 60 mL/min/1.73 m<sup>2</sup> as "> 60 mL/min/1.73 m<sup>2</sup>" and not as an exact number. However, ranges exceeding 60 mL/min are provided as a relative indicator of acceptability for laboratories that report numeric results above that threshold. Note that some laboratories reported results > 60 mL/min/1.73 m<sup>2</sup> for samples C77 and C80. These data were removed from the calculations of mean and standard deviation since their inclusion would have skewed results. Participant results for specimens C77 and C80 reported as > 60 mL/min/1.73 m<sup>2</sup> were considered acceptable performance.

## Summary of Participant Performance (Mean and Standard Deviation)

## Uric Acid (mg/dL)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
10.88 ± 0.50	3.03 ± 0.21	4.45 ± 0.21	5.98 ± 0.26	7.36 ± 0.35	n = 333	[---] All Methods & Instruments
						<Instruments>
11.46 ± 0.39	2.98 ± 0.06	4.51 ± 0.11	6.15 ± 0.17	7.61 ± 0.22	n = 21	[ABJ] Abbott Architect c System
11.43 ± 0.25	3.36 ± 0.08	4.94 ± 0.11	6.45 ± 0.14	7.96 ± 0.18	n = 54	[OLC] Beckman Coulter AU Chemistry System
10.77 ± 0.25	2.98 ± 0.20	4.48 ± 0.13	5.96 ± 0.43	7.22 ± 0.31	n = 4	[BCS] Beckman Coulter CX
10.58 ± 0.28	2.97 ± 0.08	4.43 ± 0.09	5.83 ± 0.16	7.10 ± 0.18	n = 4	[BCX] Beckman Coulter LX-20
10.67 ± 0.23	2.89 ± 0.04	4.35 ± 0.07	5.82 ± 0.12	7.01 ± 0.14	n = 15	[BCG] Beckman Coulter UniCel DxC 600
10.58 ± 0.13	2.86 ± 0.06	4.33 ± 0.05	5.79 ± 0.06	7.03 ± 0.07	n = 15	[BCH] Beckman Coulter UniCel DxC 800
10.75 ± 0.22	2.83 ± 0.09	4.35 ± 0.12	5.87 ± 0.12	7.21 ± 0.15	n = 8	[JJE] Ortho Vitros 250/350/950
10.72 ± 0.18	2.82 ± 0.07	4.29 ± 0.10	5.80 ± 0.13	7.18 ± 0.12	n = 19	[JJF] Ortho Vitros 5,1FS
10.69 ± 0.29	2.84 ± 0.11	4.32 ± 0.11	5.82 ± 0.14	7.15 ± 0.18	n = 18	[JJG] Ortho Vitros 5600
11.06 ± 0.36	2.97 ± 0.09	4.50 ± 0.14	6.08 ± 0.16	7.48 ± 0.25	n = 17	[ROC] Roche cobas c501
10.89 ± 0.11	2.88 ± 0.04	4.38 ± 0.04	5.90 ± 0.08	7.25 ± 0.12	n = 4	[ROT] Roche Cobas INTEGRA 800
11.01 ± 0.21	2.92 ± 0.07	4.46 ± 0.08	5.99 ± 0.12	7.39 ± 0.13	n = 29	[ROD] Roche MODULAR D/P
11.01 ± 0.11	3.00 ± 0.00	4.50 ± 0.00	6.01 ± 0.08	7.39 ± 0.06	n = 22	[BYE] Siemens ADVIA 1800
11.03 ± 0.05	2.83 ± 0.05	4.37 ± 0.05	6.00 ± 0.09	7.33 ± 0.05	n = 3	[BYB] Siemens ADVIA 2400
10.78 ± 0.23	3.18 ± 0.09	4.50 ± 0.10	6.03 ± 0.14	7.39 ± 0.15	n = 13	[DUE] Siemens Dimension EXL
10.79 ± 0.23	3.20 ± 0.11	4.49 ± 0.11	6.00 ± 0.15	7.45 ± 0.20	n = 22	[DUR] Siemens Dimension RxL
9.94 ± 0.14	3.09 ± 0.08	4.31 ± 0.11	5.73 ± 0.10	7.01 ± 0.12	n = 39	[DUT] Siemens Dimension Vista
10.72 ± 0.18	3.20 ± 0.07	4.48 ± 0.11	6.00 ± 0.11	7.41 ± 0.12	n = 14	[DUX] Siemens Dimension Xpand
						<Reagents>
11.46 ± 0.39	2.98 ± 0.06	4.51 ± 0.11	6.15 ± 0.17	7.61 ± 0.22	n = 21	[AB1] Abbott
10.61 ± 0.18	2.88 ± 0.07	4.35 ± 0.07	5.80 ± 0.10	7.02 ± 0.12	n = 37	[BC1] Beckman Coulter
11.43 ± 0.25	3.36 ± 0.08	4.94 ± 0.11	6.45 ± 0.14	7.97 ± 0.17	n = 53	[OL1] Beckman Coulter AU Series
10.72 ± 0.23	2.83 ± 0.09	4.31 ± 0.11	5.82 ± 0.14	7.17 ± 0.15	n = 45	[JJ1] Ortho Clinical Diagnostics
11.10 ± 0.34	2.96 ± 0.09	4.50 ± 0.13	6.07 ± 0.15	7.47 ± 0.23	n = 21	[RO4] Roche cobas c311/c501/c502/c701
11.01 ± 0.21	2.92 ± 0.07	4.46 ± 0.08	5.99 ± 0.12	7.39 ± 0.13	n = 29	[RO2] Roche Hitachi and Modular D/P
10.90 ± 0.14	2.90 ± 0.00	4.40 ± 0.06	5.94 ± 0.08	7.31 ± 0.12	n = 6	[RO1] Roche Integra and MIRA
11.02 ± 0.10	2.97 ± 0.07	4.46 ± 0.08	6.01 ± 0.09	7.39 ± 0.08	n = 27	[BY1] Siemens ADVIA/ADVIS Centaur
10.39 ± 0.49	3.15 ± 0.10	4.41 ± 0.14	5.88 ± 0.19	7.25 ± 0.26	n = 88	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Bilirubin (mg/dL)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
0.78 ± 0.13	4.75 ± 0.28	1.55 ± 0.15	2.26 ± 0.17	0.91 ± 0.14	n = 358	[---] All Methods & Instruments
0.83 ± 0.05	5.02 ± 0.15	1.63 ± 0.05	2.34 ± 0.10	0.97 ± 0.05	n = 3	<Instruments>
0.88 ± 0.08	5.16 ± 0.26	1.75 ± 0.14	2.47 ± 0.12	0.98 ± 0.08	n = 21	[AXA] Abaxis Piccolo
0.85 ± 0.07	4.43 ± 0.15	1.59 ± 0.05	2.22 ± 0.06	1.00 ± 0.00	n = 55	[ABJ] Abbott Architect c System
1.08 ± 0.04	5.00 ± 0.43	1.77 ± 0.16	2.59 ± 0.23	1.16 ± 0.18	n = 4	[OLC] Beckman Coulter AU Chemistry System
1.07 ± 0.20	4.99 ± 0.19	1.75 ± 0.23	2.45 ± 0.23	1.07 ± 0.08	n = 4	[BCS] Beckman Coulter CX
1.02 ± 0.17	4.90 ± 0.16	1.70 ± 0.14	2.45 ± 0.11	1.06 ± 0.14	n = 16	[BCX] Beckman Coulter LX-20
1.06 ± 0.14	4.85 ± 0.17	1.75 ± 0.14	2.46 ± 0.14	1.12 ± 0.14	n = 15	[BCG] Beckman Coulter UniCel DxC 600
0.79 ± 0.08	4.99 ± 0.16	1.65 ± 0.08	2.42 ± 0.09	1.07 ± 0.08	n = 10	[BCH] Beckman Coulter UniCel DxC 800
0.74 ± 0.10	4.84 ± 0.24	1.59 ± 0.09	2.35 ± 0.13	0.99 ± 0.10	n = 20	[JJE] Ortho Vitros 250/350/950
0.71 ± 0.06	4.76 ± 0.15	1.55 ± 0.11	2.30 ± 0.10	0.95 ± 0.10	n = 18	[JJF] Ortho Vitros 5,1FS
0.62 ± 0.08	4.51 ± 0.10	1.30 ± 0.04	2.03 ± 0.08	0.71 ± 0.05	n = 15	[JVG] Ortho Vitros 5600
0.58 ± 0.04	4.40 ± 0.18	1.25 ± 0.06	1.93 ± 0.09	0.62 ± 0.04	n = 4	[ROC] Roche cobas c501
0.60 ± 0.08	4.35 ± 0.15	1.30 ± 0.08	2.03 ± 0.07	0.68 ± 0.05	n = 7	[ROH] Roche cobas c701
0.60 ± 0.08	4.48 ± 0.15	1.33 ± 0.09	2.02 ± 0.13	0.70 ± 0.08	n = 4	[ROS] Roche Cobas INTEGRA 400
0.68 ± 0.05	4.67 ± 0.14	1.42 ± 0.07	2.13 ± 0.08	0.78 ± 0.06	n = 30	[ROT] Roche Cobas INTEGRA 800
0.82 ± 0.08	5.09 ± 0.14	1.60 ± 0.07	2.40 ± 0.09	0.91 ± 0.08	n = 22	[ROD] Roche MODULAR D/P
0.83 ± 0.05	5.17 ± 0.05	1.60 ± 0.00	2.40 ± 0.00	0.93 ± 0.05	n = 3	[BYE] Siemens ADVIA 1800
0.70 ± 0.00	4.73 ± 0.10	1.50 ± 0.00	2.18 ± 0.07	0.83 ± 0.06	n = 14	[BYB] Siemens ADVIA 2400
0.75 ± 0.06	4.85 ± 0.14	1.52 ± 0.07	2.24 ± 0.09	0.87 ± 0.09	n = 24	[DUE] Siemens Dimension EXL
0.80 ± 0.00	4.76 ± 0.09	1.53 ± 0.07	2.22 ± 0.07	0.90 ± 0.00	n = 40	[DUR] Siemens Dimension RxL
0.76 ± 0.08	4.82 ± 0.15	1.52 ± 0.09	2.24 ± 0.11	0.87 ± 0.06	n = 20	[DUT] Siemens Dimension Vista
0.83 ± 0.05	5.02 ± 0.15	1.63 ± 0.05	2.34 ± 0.10	0.97 ± 0.05	n = 3	[DUX] Siemens Dimension Xpand
0.88 ± 0.08	5.16 ± 0.26	1.75 ± 0.14	2.47 ± 0.12	0.98 ± 0.08	n = 21	<Reagents>
1.05 ± 0.16	4.89 ± 0.17	1.73 ± 0.15	2.46 ± 0.13	1.09 ± 0.14	n = 37	[AX1] Abaxis
0.85 ± 0.07	4.43 ± 0.15	1.59 ± 0.04	2.22 ± 0.06	1.00 ± 0.00	n = 54	[AB1] Abbott
1.00 ± 0.09	5.18 ± 0.50	1.78 ± 0.15	2.61 ± 0.29	1.01 ± 0.20	n = 3	[BC1] Beckman Coulter
0.74 ± 0.09	4.85 ± 0.22	1.59 ± 0.10	2.35 ± 0.13	0.99 ± 0.11	n = 50	[CR1] Carolina
0.61 ± 0.07	4.50 ± 0.12	1.30 ± 0.06	2.02 ± 0.09	0.69 ± 0.06	n = 21	[JJ1] Ortho Clinical Diagnostics
0.68 ± 0.05	4.67 ± 0.14	1.42 ± 0.07	2.13 ± 0.08	0.78 ± 0.06	n = 30	[RO4] Roche cobas c311/c501/c502/c701
0.60 ± 0.08	4.40 ± 0.16	1.31 ± 0.09	2.03 ± 0.09	0.68 ± 0.06	n = 11	[RO2] Roche Hitachi and Modular D/P
0.83 ± 0.07	5.10 ± 0.14	1.60 ± 0.07	2.41 ± 0.08	0.92 ± 0.08	n = 27	[RO1] Roche Integra and MIRA
0.76 ± 0.06	4.78 ± 0.12	1.52 ± 0.07	2.22 ± 0.08	0.87 ± 0.06	n = 98	[BY1] Siemens ADVIA/ADVISIA Centaur
0.83 ± 0.05	5.02 ± 0.15	1.52 ± 0.09	2.24 ± 0.11	0.87 ± 0.06	n = 98	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Phosphorus (mg/dL)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
4.33 ± 0.23	2.68 ± 0.21	3.42 ± 0.17	5.25 ± 0.26	3.16 ± 0.20	n = 337	[---] All Methods & Instruments
						<Instruments>
4.27 ± 0.11	2.63 ± 0.10	3.41 ± 0.13	5.19 ± 0.11	3.12 ± 0.12	n = 20	[ABJ] Abbott Architect c System
4.16 ± 0.11	2.59 ± 0.11	3.34 ± 0.09	5.06 ± 0.14	3.06 ± 0.09	n = 52	[OLC] Beckman Coulter AU Chemistry System
4.85 ± 0.42	3.15 ± 0.30	3.78 ± 0.28	5.68 ± 0.54	3.60 ± 0.30	n = 4	[BCS] Beckman Coulter CX
4.50 ± 0.08	2.77 ± 0.09	3.48 ± 0.04	5.57 ± 0.16	3.28 ± 0.13	n = 4	[BCX] Beckman Coulter LX-20
4.36 ± 0.11	2.85 ± 0.17	3.54 ± 0.13	5.32 ± 0.20	3.28 ± 0.12	n = 16	[BCG] Beckman Coulter UniCel DxC 600
4.58 ± 0.11	2.80 ± 0.11	3.51 ± 0.09	5.62 ± 0.12	3.28 ± 0.11	n = 15	[BCH] Beckman Coulter UniCel DxC 800
4.80 ± 0.10	3.13 ± 0.14	3.79 ± 0.13	5.75 ± 0.15	3.66 ± 0.13	n = 9	[JJE] Ortho Vitros 250/350/950
4.86 ± 0.15	3.13 ± 0.14	3.76 ± 0.13	5.70 ± 0.16	3.62 ± 0.12	n = 19	[JJF] Ortho Vitros 5,1FS
4.89 ± 0.13	3.20 ± 0.14	3.83 ± 0.12	5.80 ± 0.15	3.68 ± 0.12	n = 18	[JJG] Ortho Vitros 5600
4.36 ± 0.09	2.69 ± 0.09	3.45 ± 0.07	5.28 ± 0.12	3.20 ± 0.10	n = 18	[ROC] Roche cobas c501
4.23 ± 0.14	2.60 ± 0.09	3.33 ± 0.05	5.16 ± 0.10	3.06 ± 0.10	n = 3	[ROH] Roche cobas c701
4.30 ± 0.06	2.64 ± 0.06	3.40 ± 0.00	5.22 ± 0.08	3.10 ± 0.00	n = 5	[ROS] Roche Cobas INTEGRA 400
4.38 ± 0.04	2.75 ± 0.06	3.45 ± 0.06	5.30 ± 0.08	3.18 ± 0.04	n = 4	[ROT] Roche Cobas INTEGRA 800
4.31 ± 0.14	2.70 ± 0.12	3.44 ± 0.12	5.24 ± 0.14	3.19 ± 0.12	n = 29	[ROD] Roche MODULAR D/P
4.28 ± 0.06	2.69 ± 0.09	3.41 ± 0.06	5.20 ± 0.09	3.16 ± 0.09	n = 22	[BYE] Siemens ADVIA 1800
4.17 ± 0.05	2.60 ± 0.00	3.27 ± 0.05	5.03 ± 0.05	3.03 ± 0.05	n = 3	[BYB] Siemens ADVIA 2400
4.25 ± 0.09	2.55 ± 0.07	3.32 ± 0.07	5.06 ± 0.08	3.04 ± 0.06	n = 13	[DUE] Siemens Dimension EXL
4.27 ± 0.08	2.57 ± 0.08	3.36 ± 0.07	5.14 ± 0.08	3.08 ± 0.10	n = 22	[DUR] Siemens Dimension RxL
4.27 ± 0.17	2.52 ± 0.10	3.30 ± 0.11	5.14 ± 0.16	3.05 ± 0.11	n = 39	[DUT] Siemens Dimension Vista
4.28 ± 0.10	2.57 ± 0.09	3.35 ± 0.09	5.13 ± 0.12	3.06 ± 0.07	n = 17	[DUX] Siemens Dimension Xpand
						<Reagents>
4.27 ± 0.11	2.63 ± 0.10	3.41 ± 0.13	5.19 ± 0.11	3.12 ± 0.12	n = 20	[AB1] Abbott
4.49 ± 0.18	2.83 ± 0.15	3.53 ± 0.12	5.50 ± 0.21	3.29 ± 0.14	n = 37	[BC1] Beckman Coulter
4.17 ± 0.10	2.60 ± 0.10	3.34 ± 0.09	5.07 ± 0.14	3.06 ± 0.09	n = 51	[OL1] Beckman Coulter AU Series
4.86 ± 0.14	3.15 ± 0.14	3.79 ± 0.13	5.75 ± 0.16	3.65 ± 0.13	n = 46	[JJ1] Ortho Clinical Diagnostics
4.35 ± 0.10	2.68 ± 0.09	3.44 ± 0.08	5.26 ± 0.12	3.18 ± 0.11	n = 22	[RO4] Roche cobas c311/c501/c502/c701
4.31 ± 0.14	2.70 ± 0.12	3.44 ± 0.12	5.24 ± 0.14	3.19 ± 0.12	n = 29	[RO2] Roche Hitachi and Modular D/P
4.34 ± 0.07	2.69 ± 0.08	3.40 ± 0.00	5.26 ± 0.09	3.14 ± 0.06	n = 9	[RO1] Roche Integra and MIRA
4.27 ± 0.08	2.67 ± 0.08	3.40 ± 0.07	5.18 ± 0.10	3.14 ± 0.09	n = 27	[BY1] Siemens ADVIA/ADVISIA Centaur
4.27 ± 0.12	2.55 ± 0.09	3.33 ± 0.09	5.13 ± 0.12	3.06 ± 0.09	n = 91	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Calcium (mg/dL)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
8.90 ± 0.25	14.61 ± 0.41	6.99 ± 0.18	8.39 ± 0.23	11.39 ± 0.30	n = 365	[---] All Methods & Instruments
8.57 ± 0.05	14.10 ± 0.36	6.97 ± 0.14	8.52 ± 0.24	11.23 ± 0.05	n = 3	<Instruments>
8.99 ± 0.14	14.78 ± 0.33	6.99 ± 0.10	8.41 ± 0.09	11.43 ± 0.18	n = 21	[AXA] Abaxis Piccolo
9.06 ± 0.16	14.52 ± 0.33	7.01 ± 0.13	8.41 ± 0.16	11.34 ± 0.19	n = 56	[ABJ] Abbott Architect c System
8.95 ± 0.37	14.52 ± 0.79	7.12 ± 0.28	8.51 ± 0.58	11.25 ± 0.56	n = 4	[OLC] Beckman Coulter AU Chemistry System
8.90 ± 0.08	14.30 ± 0.08	6.99 ± 0.11	8.25 ± 0.06	11.15 ± 0.06	n = 4	[BCS] Beckman Coulter CX
8.85 ± 0.16	14.20 ± 0.28	6.95 ± 0.12	8.20 ± 0.16	11.11 ± 0.23	n = 17	[BCX] Beckman Coulter LX-20
8.89 ± 0.11	14.33 ± 0.17	7.03 ± 0.14	8.34 ± 0.16	11.21 ± 0.18	n = 15	[BCG] Beckman Coulter UniCel DxC 600
8.90 ± 0.12	14.81 ± 0.41	6.92 ± 0.15	8.57 ± 0.19	11.73 ± 0.19	n = 11	[BCH] Beckman Coulter UniCel DxC 800
8.84 ± 0.18	14.51 ± 0.44	6.86 ± 0.13	8.53 ± 0.19	11.57 ± 0.26	n = 20	[JJE] Ortho Vitros 250/350/950
8.82 ± 0.18	14.48 ± 0.41	6.95 ± 0.17	8.58 ± 0.22	11.63 ± 0.26	n = 18	[JJF] Ortho Vitros 5,1FS
9.01 ± 0.29	15.33 ± 0.46	6.99 ± 0.23	8.43 ± 0.25	11.68 ± 0.34	n = 18	[JGJ] Ortho Vitros 5600
9.00 ± 0.18	15.00 ± 0.18	6.97 ± 0.14	8.47 ± 0.14	11.77 ± 0.23	n = 3	[ROC] Roche cobas c501
8.84 ± 0.15	15.22 ± 0.31	6.76 ± 0.16	8.16 ± 0.13	11.67 ± 0.21	n = 7	[ROH] Roche cobas c701
8.90 ± 0.23	15.12 ± 0.34	6.86 ± 0.23	8.20 ± 0.23	11.54 ± 0.30	n = 4	[ROS] Roche Cobas INTEGRA 400
9.09 ± 0.22	14.79 ± 0.39	7.13 ± 0.19	8.53 ± 0.20	11.52 ± 0.25	n = 30	[ROT] Roche MODULAR D/P
9.15 ± 0.19	14.58 ± 0.34	7.07 ± 0.16	8.52 ± 0.19	11.42 ± 0.20	n = 22	[BYE] Siemens ADVIA 1800
8.88 ± 0.32	14.21 ± 0.20	6.90 ± 0.27	8.39 ± 0.29	11.03 ± 0.34	n = 3	[BYB] Siemens ADVIA 2400
8.67 ± 0.25	14.61 ± 0.29	6.96 ± 0.24	8.23 ± 0.23	11.33 ± 0.22	n = 14	[DUE] Siemens Dimension EXL
8.75 ± 0.17	14.72 ± 0.25	7.04 ± 0.15	8.33 ± 0.14	11.32 ± 0.21	n = 24	[DUR] Siemens Dimension RxL
8.67 ± 0.19	14.56 ± 0.27	6.93 ± 0.20	8.25 ± 0.21	11.19 ± 0.30	n = 40	[DUT] Siemens Dimension Vista
8.68 ± 0.20	14.65 ± 0.28	6.98 ± 0.18	8.21 ± 0.20	11.32 ± 0.23	n = 20	[DUX] Siemens Dimension Xpand
8.57 ± 0.05	14.10 ± 0.36	6.97 ± 0.14	8.52 ± 0.24	11.23 ± 0.05	n = 3	<Reagents>
8.99 ± 0.14	14.78 ± 0.33	6.99 ± 0.10	8.41 ± 0.09	11.43 ± 0.18	n = 21	[AX1] Abaxis
8.87 ± 0.13	14.25 ± 0.24	6.98 ± 0.14	8.25 ± 0.18	11.14 ± 0.21	n = 38	[AB1] Abbott
9.06 ± 0.16	14.53 ± 0.33	7.01 ± 0.13	8.40 ± 0.16	11.34 ± 0.19	n = 55	[BC1] Beckman Coulter
8.86 ± 0.17	14.56 ± 0.44	6.91 ± 0.17	8.56 ± 0.20	11.64 ± 0.26	n = 50	[OL1] Beckman Coulter AU Series
9.02 ± 0.23	15.28 ± 0.40	6.98 ± 0.17	8.46 ± 0.22	11.70 ± 0.26	n = 21	[JJ1] Ortho Clinical Diagnostics
9.09 ± 0.23	14.81 ± 0.37	7.13 ± 0.19	8.53 ± 0.21	11.53 ± 0.25	n = 29	[RO4] Roche cobas c311/c501/c502/c701
8.86 ± 0.17	15.19 ± 0.32	6.79 ± 0.18	8.17 ± 0.17	11.63 ± 0.25	n = 11	[RO2] Roche Hitachi and Modular D/P
8.93 ± 0.32	13.97 ± 0.34	6.97 ± 0.14	8.47 ± 0.32	11.03 ± 0.31	n = 3	[R01] Roche Integra and MIRA
9.15 ± 0.23	14.54 ± 0.36	7.08 ± 0.19	8.52 ± 0.21	11.42 ± 0.25	n = 27	[GZ1] Sekisui Diagnostics (Genzyme)
8.69 ± 0.20	14.63 ± 0.28	6.97 ± 0.19	8.26 ± 0.21	11.27 ± 0.27	n = 98	[BY1] Siemens ADVIA/ADVIS Centaur
						[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Magnesium (mg/dL)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
1.64 ± 0.10	3.42 ± 0.14	2.39 ± 0.10	1.00 ± 0.09	4.68 ± 0.19	n = 332	[---] All Methods & Instruments
1.70 ± 0.07	3.23 ± 0.10	2.28 ± 0.05	0.93 ± 0.09	4.38 ± 0.17	n = 20	<Instruments>
1.65 ± 0.06	3.33 ± 0.09	2.35 ± 0.08	1.00 ± 0.00	4.56 ± 0.13	n = 53	[ABJ] Abbott Architect c System
1.82 ± 0.04	3.40 ± 0.18	2.40 ± 0.18	1.10 ± 0.08	4.65 ± 0.34	n = 4	[OLC] Beckman Coulter AU Chemistry System
1.80 ± 0.08	3.55 ± 0.12	2.50 ± 0.08	1.12 ± 0.04	4.87 ± 0.09	n = 4	[BCS] Beckman Coulter CX
1.76 ± 0.06	3.50 ± 0.08	2.47 ± 0.07	1.07 ± 0.05	4.77 ± 0.11	n = 17	[BCX] Beckman Coulter LX-20
1.73 ± 0.06	3.50 ± 0.08	2.45 ± 0.07	1.09 ± 0.04	4.75 ± 0.08	n = 15	[BCG] Beckman Coulter UniCel DxC 600
1.60 ± 0.00	3.53 ± 0.05	2.47 ± 0.05	1.00 ± 0.00	4.79 ± 0.14	n = 6	[BCH] Beckman Coulter UniCel DxC 800
1.56 ± 0.06	3.48 ± 0.07	2.42 ± 0.06	0.96 ± 0.07	4.70 ± 0.11	n = 20	[JJE] Ortho Vitros 250/350/950
1.58 ± 0.06	3.45 ± 0.08	2.43 ± 0.08	0.94 ± 0.06	4.72 ± 0.09	n = 17	[JJF] Ortho Vitros 5,1FS
1.67 ± 0.06	3.31 ± 0.07	2.39 ± 0.06	1.08 ± 0.04	4.53 ± 0.09	n = 16	[JJG] Ortho Vitros 5600
1.63 ± 0.05	3.40 ± 0.09	2.33 ± 0.05	0.97 ± 0.05	4.43 ± 0.05	n = 3	[ROC] Roche cobas c501
1.68 ± 0.04	3.38 ± 0.04	2.38 ± 0.04	1.05 ± 0.06	4.52 ± 0.04	n = 4	[ROH] Roche cobas c701
1.64 ± 0.06	3.41 ± 0.05	2.39 ± 0.06	1.00 ± 0.00	4.59 ± 0.11	n = 28	[ROT] Roche Cobas INTEGRA 800
1.74 ± 0.07	3.65 ± 0.13	2.46 ± 0.10	1.15 ± 0.08	4.74 ± 0.15	n = 21	[ROD] Roche MODULAR D/P
1.73 ± 0.14	3.74 ± 0.10	2.43 ± 0.14	1.10 ± 0.09	4.76 ± 0.10	n = 3	[BYE] Siemens ADVIA 1800
1.54 ± 0.08	3.35 ± 0.08	2.33 ± 0.06	0.93 ± 0.06	4.72 ± 0.08	n = 12	[BYB] Siemens ADVIA 2400
1.57 ± 0.05	3.39 ± 0.12	2.35 ± 0.07	0.91 ± 0.06	4.75 ± 0.09	n = 23	[DUE] Siemens Dimension EXL
1.60 ± 0.09	3.54 ± 0.11	2.48 ± 0.11	0.99 ± 0.09	4.92 ± 0.12	n = 40	[DUR] Siemens Dimension RxL
1.57 ± 0.08	3.36 ± 0.12	2.34 ± 0.09	0.90 ± 0.00	4.68 ± 0.09	n = 19	[DUT] Siemens Dimension Vista
1.70 ± 0.07	3.23 ± 0.10	2.28 ± 0.05	0.93 ± 0.09	4.38 ± 0.17	n = 20	[DUX] Siemens Dimension Xpand
1.76 ± 0.06	3.50 ± 0.09	2.47 ± 0.07	1.09 ± 0.05	4.78 ± 0.11	n = 37	<Reagents>
1.65 ± 0.07	3.34 ± 0.09	2.35 ± 0.08	1.00 ± 0.00	4.57 ± 0.13	n = 53	[BC1] Beckman Coulter
1.57 ± 0.06	3.48 ± 0.08	2.43 ± 0.07	0.96 ± 0.06	4.72 ± 0.11	n = 43	[OL1] Beckman Coulter AU Series
1.66 ± 0.06	3.32 ± 0.08	2.38 ± 0.06	1.07 ± 0.06	4.51 ± 0.09	n = 19	[JJ1] Ortho Clinical Diagnostics
1.64 ± 0.06	3.41 ± 0.05	2.39 ± 0.06	1.00 ± 0.00	4.60 ± 0.11	n = 27	[RO4] Roche cobas c311/c501/c502/c701
1.66 ± 0.06	3.38 ± 0.05	2.38 ± 0.05	1.06 ± 0.06	4.52 ± 0.07	n = 7	[RO2] Roche Hitachi and Modular D/P
1.74 ± 0.09	3.65 ± 0.13	2.47 ± 0.11	1.15 ± 0.08	4.75 ± 0.14	n = 27	[RO1] Roche Integra and MIRA
1.58 ± 0.08	3.44 ± 0.14	2.39 ± 0.11	0.94 ± 0.08	4.81 ± 0.16	n = 93	[BY1] Siemens ADVIA/ADVISIA Centaur
						[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

Iron ( $\mu\text{g/dL}$ )

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
116.1 $\pm$ 5.60	101.6 $\pm$ 4.16	69.2 $\pm$ 2.89	84.9 $\pm$ 4.56	100.9 $\pm$ 4.99	n = 269	[---] All Methods & Instruments
						<Instruments>
110.5 $\pm$ 13.24	97.9 $\pm$ 4.54	67.9 $\pm$ 2.86	81.3 $\pm$ 7.98	98.8 $\pm$ 9.67	n = 15	[ABJ] Abbott Architect c System
119.7 $\pm$ 3.13	103.5 $\pm$ 3.10	70.9 $\pm$ 2.29	86.6 $\pm$ 2.54	101.9 $\pm$ 2.75	n = 46	[OLC] Beckman Coulter AU Chemistry System
101.2 $\pm$ 10.49	103.7 $\pm$ 8.57	70.3 $\pm$ 5.97	80.2 $\pm$ 6.79	96.4 $\pm$ 10.90	n = 3	[BCS] Beckman Coulter CX
110.1 $\pm$ 5.61	104.4 $\pm$ 3.91	69.8 $\pm$ 3.10	81.8 $\pm$ 4.08	99.1 $\pm$ 5.61	n = 4	[BCX] Beckman Coulter LX-20
110.6 $\pm$ 3.36	103.7 $\pm$ 1.79	70.1 $\pm$ 2.56	82.6 $\pm$ 2.20	99.1 $\pm$ 3.71	n = 11	[BCG] Beckman Coulter UniCel DxC 600
109.4 $\pm$ 3.08	103.1 $\pm$ 2.15	69.9 $\pm$ 1.19	81.6 $\pm$ 2.34	98.4 $\pm$ 2.07	n = 13	[BCH] Beckman Coulter UniCel DxC 800
141.5 $\pm$ 1.86	102.3 $\pm$ 5.86	62.8 $\pm$ 5.00	96.4 $\pm$ 2.56	109.0 $\pm$ 1.80	n = 3	[JJE] Ortho Vitros 250/350/950
140.2 $\pm$ 6.25	107.0 $\pm$ 7.00	65.9 $\pm$ 4.75	97.8 $\pm$ 5.81	112.5 $\pm$ 5.80	n = 18	[JJF] Ortho Vitros 5,1FS
140.0 $\pm$ 5.08	104.3 $\pm$ 4.71	63.6 $\pm$ 4.12	96.0 $\pm$ 4.59	112.1 $\pm$ 4.71	n = 18	[JJG] Ortho Vitros 5600
119.5 $\pm$ 4.76	103.5 $\pm$ 3.53	71.0 $\pm$ 2.49	86.6 $\pm$ 2.29	102.1 $\pm$ 2.54	n = 10	[ROC] Roche cobas c501
117.0 $\pm$ 0.00	102.7 $\pm$ 0.51	74.3 $\pm$ 6.85	85.7 $\pm$ 1.37	104.7 $\pm$ 4.06	n = 3	[ROS] Roche Cobas INTEGRA 400
117.8 $\pm$ 3.23	103.7 $\pm$ 3.16	72.3 $\pm$ 1.37	87.6 $\pm$ 1.02	103.3 $\pm$ 2.26	n = 3	[ROT] Roche Cobas INTEGRA 800
116.6 $\pm$ 2.36	101.6 $\pm$ 2.06	70.0 $\pm$ 1.43	84.9 $\pm$ 1.97	100.7 $\pm$ 2.20	n = 27	[ROD] Roche MODULAR D/P
117.7 $\pm$ 2.44	100.2 $\pm$ 2.43	70.0 $\pm$ 1.87	84.5 $\pm$ 2.16	100.9 $\pm$ 2.15	n = 21	[BYE] Siemens ADVIA 1800
117.3 $\pm$ 3.37	98.8 $\pm$ 2.36	68.8 $\pm$ 1.54	83.8 $\pm$ 1.54	99.1 $\pm$ 2.05	n = 3	[BYB] Siemens ADVIA 2400
113.8 $\pm$ 2.28	99.0 $\pm$ 2.16	69.2 $\pm$ 1.49	83.0 $\pm$ 1.66	97.9 $\pm$ 2.34	n = 8	[DUE] Siemens Dimension EXL
113.1 $\pm$ 2.22	98.5 $\pm$ 1.70	68.3 $\pm$ 1.23	82.0 $\pm$ 1.90	97.1 $\pm$ 1.80	n = 14	[DUR] Siemens Dimension RxL
113.9 $\pm$ 2.50	99.2 $\pm$ 2.49	68.7 $\pm$ 1.88	82.4 $\pm$ 2.05	97.7 $\pm$ 2.35	n = 36	[DUT] Siemens Dimension Vista
114.0 $\pm$ 3.16	98.8 $\pm$ 2.80	68.2 $\pm$ 1.27	82.0 $\pm$ 2.45	97.2 $\pm$ 2.58	n = 4	[DUX] Siemens Dimension Xpand
						<Reagents>
116.2 $\pm$ 2.74	100.1 $\pm$ 3.11	69.0 $\pm$ 1.05	84.8 $\pm$ 2.03	103.2 $\pm$ 3.04	n = 10	[AB3] Abbott-Iron/6K95
87.0 $\pm$ 0.76	93.1 $\pm$ 2.60	64.7 $\pm$ 1.38	67.7 $\pm$ 0.83	82.5 $\pm$ 0.57	n = 5	[AB2] Abbott-Iron/7D68
109.8 $\pm$ 3.72	103.6 $\pm$ 2.46	69.9 $\pm$ 2.02	82.1 $\pm$ 2.34	98.7 $\pm$ 3.11	n = 29	[BC1] Beckman Coulter
120.3 $\pm$ 2.49	104.2 $\pm$ 2.77	71.3 $\pm$ 2.29	87.1 $\pm$ 2.18	102.5 $\pm$ 2.40	n = 37	[OL1] Beckman Coulter AU Series
140.2 $\pm$ 5.42	105.2 $\pm$ 6.12	64.5 $\pm$ 4.65	96.8 $\pm$ 5.08	112.0 $\pm$ 5.16	n = 39	[JJ1] Ortho Clinical Diagnostics
118.6 $\pm$ 4.25	103.2 $\pm$ 3.37	70.6 $\pm$ 2.16	86.4 $\pm$ 2.12	102.1 $\pm$ 2.32	n = 14	[RO4] Roche cobas c311/c501/c502/c701
116.6 $\pm$ 2.36	101.6 $\pm$ 2.06	70.0 $\pm$ 1.43	84.9 $\pm$ 1.97	100.7 $\pm$ 2.20	n = 27	[RO2] Roche Hitachi and Modular D/P
117.0 $\pm$ 0.00	103.0 $\pm$ 2.11	72.4 $\pm$ 3.73	86.7 $\pm$ 1.48	103.9 $\pm$ 3.28	n = 6	[RO1] Roche Integra and MIRA
115.5 $\pm$ 3.10	100.3 $\pm$ 2.75	69.6 $\pm$ 1.68	83.8 $\pm$ 2.19	99.0 $\pm$ 2.14	n = 8	[GZ1] Sekisui Diagnostics (Genzyme)
117.5 $\pm$ 2.37	100.0 $\pm$ 2.49	69.9 $\pm$ 1.66	84.4 $\pm$ 1.92	100.7 $\pm$ 2.06	n = 25	[BY1] Siemens ADVIA/ADVIS Centaur
113.7 $\pm$ 2.48	99.0 $\pm$ 2.31	68.6 $\pm$ 1.62	82.4 $\pm$ 2.01	97.6 $\pm$ 2.22	n = 62	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Sodium (mmol/L)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
146.8 ± 2.50	148.5 ± 2.31	128.8 ± 1.75	155.6 ± 2.15	153.9 ± 2.54	n = 371	[---] All Methods & Instruments
145.2 ± 2.36	145.5 ± 1.86	124.7 ± 3.37	155.6 ± 1.02	151.5 ± 3.63	n = 3	<Instruments>
147.2 ± 0.75	148.4 ± 0.86	128.0 ± 0.72	156.1 ± 1.50	154.3 ± 0.99	n = 21	[AXA] Abaxis Piccolo
145.8 ± 1.38	147.4 ± 1.15	128.3 ± 1.20	155.1 ± 1.06	152.4 ± 1.26	n = 56	[ABJ] Abbott Architect c System
145.9 ± 3.97	148.4 ± 6.10	127.6 ± 3.91	156.9 ± 8.25	154.1 ± 6.05	n = 4	[OLC] Beckman Coulter AU Chemistry System
146.7 ± 0.90	148.7 ± 0.90	129.9 ± 1.13	155.7 ± 1.58	153.4 ± 1.80	n = 4	[BCS] Beckman Coulter CX
146.9 ± 1.91	147.9 ± 1.81	128.2 ± 1.03	155.8 ± 2.26	153.5 ± 2.30	n = 17	[BCX] Beckman Coulter LX-20
146.0 ± 1.07	147.4 ± 1.09	128.3 ± 1.09	155.1 ± 1.22	153.3 ± 1.15	n = 15	[BCG] Beckman Coulter UniCel DxC 600
143.6 ± 0.96	145.0 ± 0.47	126.0 ± 0.82	150.0 ± 0.00	152.2 ± 0.66	n = 7	[BCH] Beckman Coulter UniCel DxC 800
152.2 ± 1.57	152.7 ± 1.74	128.9 ± 1.62	158.5 ± 1.93	161.0 ± 2.43	n = 11	[IAA] i-STAT
152.0 ± 1.57	152.7 ± 1.29	128.8 ± 1.14	158.2 ± 1.57	160.4 ± 1.66	n = 20	[JJE] Ortho Vitros 250/350/950
152.4 ± 1.88	153.1 ± 2.16	129.2 ± 1.94	158.9 ± 1.94	161.2 ± 2.44	n = 18	[JJF] Ortho Vitros 5,1FS
146.1 ± 1.60	148.3 ± 1.09	127.8 ± 1.28	156.1 ± 1.53	153.4 ± 1.81	n = 17	[JJG] Ortho Vitros 5600
147.2 ± 2.36	148.9 ± 2.05	128.5 ± 1.86	156.6 ± 2.56	153.2 ± 2.36	n = 3	[ROC] Roche cobas c501
146.5 ± 1.02	146.7 ± 0.97	127.5 ± 1.02	154.8 ± 0.91	152.6 ± 0.94	n = 6	[ROH] Roche cobas c701
146.0 ± 1.50	145.3 ± 0.82	127.0 ± 0.75	153.5 ± 1.71	152.3 ± 0.90	n = 4	[ROS] Roche Cobas INTEGRA 400
147.8 ± 1.39	149.3 ± 1.44	128.9 ± 1.51	156.6 ± 1.16	154.2 ± 1.65	n = 29	[ROT] Roche Cobas INTEGRA 800
148.3 ± 0.74	149.7 ± 1.04	130.4 ± 0.67	157.2 ± 1.01	154.6 ± 0.81	n = 22	[ROD] Roche MODULAR D/P
148.0 ± 0.00	149.0 ± 0.00	129.7 ± 0.51	156.7 ± 0.51	154.7 ± 0.51	n = 3	[BYE] Siemens ADVIA 1800
145.6 ± 1.01	148.0 ± 0.44	129.4 ± 1.60	154.8 ± 1.31	153.5 ± 1.88	n = 15	[BYB] Siemens ADVIA 2400
145.0 ± 1.65	146.7 ± 1.81	128.6 ± 1.54	153.7 ± 2.07	152.4 ± 1.82	n = 23	[DUR] Siemens Dimension RxL
145.6 ± 1.85	149.1 ± 1.62	130.7 ± 1.21	154.5 ± 1.62	154.8 ± 1.52	n = 40	[DUT] Siemens Dimension Vista
146.1 ± 1.17	148.2 ± 1.41	129.6 ± 1.32	155.1 ± 1.07	153.9 ± 1.36	n = 20	[DUX] Siemens Dimension Xpand
145.2 ± 2.36	145.5 ± 1.86	124.7 ± 3.37	155.6 ± 1.02	151.5 ± 3.63	n = 3	<Reagents>
147.2 ± 0.76	148.4 ± 0.97	127.9 ± 0.79	156.0 ± 1.74	154.0 ± 1.32	n = 22	[AX1] Abaxis
146.3 ± 1.56	147.6 ± 1.49	128.4 ± 1.49	155.2 ± 1.89	153.3 ± 1.76	n = 38	[AB1] Abbott
145.7 ± 1.32	147.4 ± 1.16	128.2 ± 1.17	155.0 ± 1.06	152.4 ± 1.27	n = 55	[BC1] Beckman Coulter
143.8 ± 1.00	145.0 ± 0.55	126.2 ± 0.73	150.0 ± 0.00	152.2 ± 0.73	n = 6	[OL1] Beckman Coulter AU Series
146.0 ± 0.90	149.8 ± 1.54	127.0 ± 0.90	153.6 ± 1.02	156.8 ± 1.54	n = 3	[IA1] i-STAT
152.2 ± 1.69	152.9 ± 1.77	128.9 ± 1.57	158.5 ± 1.79	160.9 ± 2.19	n = 50	[IL1] Instrumentation Lab
146.3 ± 1.70	148.4 ± 1.23	128.0 ± 1.35	156.1 ± 1.63	153.5 ± 1.82	n = 22	[JJ1] Ortho Clinical Diagnostics
147.8 ± 1.39	149.3 ± 1.44	128.9 ± 1.51	156.6 ± 1.16	154.2 ± 1.65	n = 29	[RO4] Roche cobas c311/c501/c502/c701
146.3 ± 1.22	146.2 ± 1.19	127.3 ± 0.94	154.5 ± 1.42	152.5 ± 0.89	n = 10	[RO2] Roche Hitachi and Modular D/P
148.2 ± 0.72	149.5 ± 1.02	130.3 ± 0.69	157.1 ± 0.98	154.6 ± 0.76	n = 26	[RO1] Roche Integra and MIRA
145.6 ± 1.63	148.2 ± 1.84	129.9 ± 1.69	154.5 ± 1.67	153.9 ± 1.95	n = 97	[BY1] Siemens ADVIA/ADVISIA Centaur
						[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Potassium (mmol/L)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
4.15 ± 0.11	3.64 ± 0.09	2.81 ± 0.08	5.69 ± 0.11	4.94 ± 0.10	n = 371	[---] All Methods & Instruments
4.33 ± 0.05	4.00 ± 0.09	2.97 ± 0.05	5.99 ± 0.20	5.35 ± 0.27	n = 3	<Instruments>
4.13 ± 0.08	3.63 ± 0.07	2.80 ± 0.00	5.64 ± 0.09	4.90 ± 0.09	n = 21	[AXA] Abaxis Piccolo
4.16 ± 0.06	3.68 ± 0.05	2.83 ± 0.05	5.70 ± 0.04	4.92 ± 0.06	n = 56	[ABJ] Abbott Architect c System
4.10 ± 0.17	3.64 ± 0.19	2.77 ± 0.09	5.76 ± 0.32	4.99 ± 0.23	n = 4	[OLC] Beckman Coulter AU Chemistry System
4.15 ± 0.06	3.65 ± 0.06	2.80 ± 0.00	5.77 ± 0.09	4.97 ± 0.09	n = 4	[BCS] Beckman Coulter CX
4.16 ± 0.09	3.64 ± 0.06	2.76 ± 0.06	5.73 ± 0.10	4.96 ± 0.10	n = 17	[BCX] Beckman Coulter LX-20
4.12 ± 0.04	3.63 ± 0.06	2.75 ± 0.06	5.73 ± 0.06	4.99 ± 0.04	n = 15	[BCG] Beckman Coulter UniCel DxC 600
4.10 ± 0.00	3.52 ± 0.05	2.70 ± 0.00	5.54 ± 0.06	4.88 ± 0.05	n = 7	[BCH] Beckman Coulter UniCel DxC 800
4.37 ± 0.08	3.80 ± 0.06	2.88 ± 0.06	5.87 ± 0.08	5.19 ± 0.10	n = 11	[IAA] i-STAT
4.34 ± 0.07	3.80 ± 0.08	2.85 ± 0.06	5.84 ± 0.09	5.13 ± 0.09	n = 20	[JJE] Ortho Vitros 250/350/950
4.35 ± 0.10	3.79 ± 0.07	2.89 ± 0.07	5.83 ± 0.09	5.14 ± 0.07	n = 18	[JJF] Ortho Vitros 5,1FS
4.06 ± 0.10	3.54 ± 0.09	2.71 ± 0.10	5.62 ± 0.10	4.86 ± 0.10	n = 17	[JJG] Ortho Vitros 5600
4.10 ± 0.09	3.57 ± 0.05	2.74 ± 0.10	5.67 ± 0.05	4.87 ± 0.05	n = 3	[ROC] Roche cobas c501
4.21 ± 0.06	3.65 ± 0.06	2.80 ± 0.00	5.70 ± 0.06	4.93 ± 0.05	n = 6	[ROH] Roche cobas c701
4.20 ± 0.08	3.62 ± 0.04	2.80 ± 0.00	5.70 ± 0.00	4.95 ± 0.06	n = 4	[ROS] Roche Cobas INTEGRA 400
4.10 ± 0.08	3.59 ± 0.06	2.77 ± 0.11	5.66 ± 0.08	4.87 ± 0.09	n = 29	[ROT] Roche Cobas INTEGRA 800
4.20 ± 0.00	3.72 ± 0.04	2.90 ± 0.00	5.78 ± 0.05	5.00 ± 0.00	n = 22	[ROD] Roche MODULAR D/P
4.20 ± 0.00	3.73 ± 0.05	2.90 ± 0.00	5.77 ± 0.05	5.00 ± 0.00	n = 3	[BYE] Siemens ADVIA 1800
4.10 ± 0.00	3.60 ± 0.00	2.77 ± 0.06	5.70 ± 0.00	4.94 ± 0.06	n = 14	[BYB] Siemens ADVIA 2400
4.10 ± 0.05	3.60 ± 0.00	2.76 ± 0.06	5.64 ± 0.07	4.91 ± 0.06	n = 24	[DUE] Siemens Dimension EXL
4.07 ± 0.05	3.60 ± 0.00	2.80 ± 0.00	5.57 ± 0.07	4.90 ± 0.06	n = 40	[DUR] Siemens Dimension RxL
4.10 ± 0.00	3.60 ± 0.00	2.75 ± 0.06	5.70 ± 0.05	4.94 ± 0.06	n = 20	[DUT] Siemens Dimension Vista
						[DUX] Siemens Dimension Xpand
4.33 ± 0.05	4.00 ± 0.09	2.97 ± 0.05	5.99 ± 0.20	5.35 ± 0.27	n = 3	<Reagents>
4.13 ± 0.08	3.62 ± 0.08	2.80 ± 0.04	5.63 ± 0.09	4.90 ± 0.08	n = 22	[AX1] Abaxis
4.14 ± 0.07	3.63 ± 0.06	2.76 ± 0.06	5.73 ± 0.09	4.97 ± 0.08	n = 38	[AB1] Abbott
4.16 ± 0.06	3.68 ± 0.05	2.83 ± 0.05	5.70 ± 0.04	4.92 ± 0.06	n = 55	[BC1] Beckman Coulter
4.10 ± 0.00	3.53 ± 0.05	2.70 ± 0.00	5.53 ± 0.05	4.87 ± 0.05	n = 6	[OL1] Beckman Coulter AU Series
4.10 ± 0.00	3.50 ± 0.00	2.67 ± 0.05	5.60 ± 0.00	4.90 ± 0.00	n = 3	[IA1] i-STAT
4.35 ± 0.08	3.80 ± 0.07	2.87 ± 0.07	5.85 ± 0.09	5.15 ± 0.09	n = 50	[IL1] Instrumentation Lab
4.07 ± 0.09	3.55 ± 0.08	2.71 ± 0.09	5.62 ± 0.08	4.86 ± 0.09	n = 22	[JJ1] Ortho Clinical Diagnostics
4.10 ± 0.08	3.59 ± 0.06	2.77 ± 0.11	5.66 ± 0.08	4.87 ± 0.09	n = 29	[RO4] Roche cobas c311/c501/c502/c701
4.20 ± 0.07	3.64 ± 0.06	2.80 ± 0.00	5.70 ± 0.00	4.94 ± 0.06	n = 10	[RO2] Roche Hitachi and Modular D/P
4.20 ± 0.00	3.72 ± 0.04	2.90 ± 0.00	5.77 ± 0.05	5.00 ± 0.00	n = 26	[RO1] Roche Integra and MIRA
4.09 ± 0.05	3.60 ± 0.00	2.78 ± 0.05	5.64 ± 0.09	4.91 ± 0.06	n = 98	[BY1] Siemens ADVIA/ADVIA Centaur
						[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Chloride (mmol/L)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
105.2 ± 2.67	112.1 ± 1.95	94.0 ± 2.08	112.6 ± 2.32	122.9 ± 2.67	n = 366	[---] All Methods & Instruments
						<Instruments>
107.3 ± 0.51	111.5 ± 1.86	93.3 ± 1.37	109.5 ± 1.86	120.6 ± 1.02	n = 3	[AXA] Abaxis Piccolo
106.2 ± 0.78	112.6 ± 0.82	94.8 ± 0.61	112.2 ± 0.78	124.2 ± 0.96	n = 20	[ABJ] Abbott Architect c System
103.4 ± 0.97	111.1 ± 1.03	93.5 ± 1.04	110.9 ± 1.14	121.9 ± 1.30	n = 55	[OLC] Beckman Coulter AU Chemistry System
109.3 ± 3.89	116.3 ± 5.74	99.2 ± 7.38	115.9 ± 6.49	126.4 ± 6.21	n = 4	[BCS] Beckman Coulter CX
107.6 ± 1.80	114.0 ± 1.65	96.4 ± 2.31	114.0 ± 1.65	125.9 ± 2.35	n = 4	[BCX] Beckman Coulter LX-20
106.9 ± 1.90	113.7 ± 1.16	95.6 ± 1.27	113.5 ± 1.31	124.8 ± 2.09	n = 17	[BCG] Beckman Coulter UniCel DxC 600
106.8 ± 0.85	113.9 ± 0.93	96.2 ± 0.76	113.8 ± 0.99	125.7 ± 0.95	n = 15	[BCH] Beckman Coulter UniCel DxC 800
113.8 ± 0.73	119.2 ± 0.73	96.9 ± 1.13	118.5 ± 0.57	131.0 ± 1.37	n = 6	[IAA] i-STAT
107.9 ± 1.49	113.9 ± 1.63	95.5 ± 1.81	115.3 ± 1.41	126.2 ± 1.85	n = 11	[JJE] Ortho Vitros 250/350/950
107.7 ± 1.57	113.6 ± 1.38	95.4 ± 1.29	114.5 ± 1.61	125.5 ± 1.55	n = 20	[JJF] Ortho Vitros 5,1FS
107.6 ± 1.88	113.6 ± 1.72	95.5 ± 1.77	114.7 ± 2.12	125.4 ± 2.15	n = 18	[JJG] Ortho Vitros 5600
99.5 ± 1.41	108.9 ± 1.31	89.5 ± 1.19	108.8 ± 1.15	120.1 ± 1.57	n = 17	[ROC] Roche cobas c501
100.3 ± 1.37	109.7 ± 0.51	89.7 ± 0.51	109.0 ± 0.90	119.7 ± 0.51	n = 3	[ROH] Roche cobas c701
105.5 ± 1.02	112.6 ± 0.79	94.2 ± 0.73	113.1 ± 1.13	123.5 ± 1.02	n = 6	[ROS] Roche Cobas INTEGRA 400
106.0 ± 1.14	111.7 ± 0.90	94.5 ± 1.22	112.5 ± 1.22	123.0 ± 0.75	n = 4	[ROT] Roche Cobas INTEGRA 800
102.2 ± 1.03	111.2 ± 0.94	91.6 ± 0.81	110.7 ± 0.89	122.2 ± 1.37	n = 29	[ROD] Roche MODULAR D/P
105.4 ± 1.08	112.5 ± 0.95	93.4 ± 1.18	111.2 ± 1.69	122.4 ± 2.13	n = 22	[BYE] Siemens ADVIA 1800
105.3 ± 0.51	112.0 ± 0.90	92.5 ± 1.86	110.5 ± 1.86	121.2 ± 2.36	n = 3	[BYB] Siemens ADVIA 2400
105.1 ± 0.93	111.1 ± 0.52	94.1 ± 0.90	113.4 ± 1.10	120.7 ± 1.22	n = 14	[DUE] Siemens Dimension EXL
104.3 ± 1.73	111.0 ± 1.50	92.9 ± 1.49	113.3 ± 1.73	121.0 ± 1.84	n = 24	[DUR] Siemens Dimension RxL
106.5 ± 1.33	113.4 ± 1.47	94.7 ± 1.21	113.8 ± 1.18	123.9 ± 1.73	n = 40	[DUT] Siemens Dimension Vista
104.7 ± 1.13	110.7 ± 1.31	93.6 ± 1.27	113.0 ± 0.94	120.0 ± 1.32	n = 20	[DUX] Siemens Dimension Xpand
						<Reagents>
107.3 ± 0.51	111.5 ± 1.86	93.3 ± 1.37	109.5 ± 1.86	120.6 ± 1.02	n = 3	[AX1] Abaxis
106.1 ± 0.86	112.5 ± 0.88	94.8 ± 0.65	112.2 ± 0.90	124.1 ± 1.07	n = 21	[AB1] Abbott
106.9 ± 1.53	113.8 ± 1.19	96.0 ± 1.29	113.6 ± 1.22	125.3 ± 1.78	n = 38	[BC1] Beckman Coulter
103.4 ± 0.94	111.1 ± 1.00	93.4 ± 1.00	110.9 ± 1.14	121.8 ± 1.30	n = 54	[OL1] Beckman Coulter AU Series
113.8 ± 0.73	119.2 ± 0.73	96.9 ± 1.13	118.5 ± 0.57	131.0 ± 1.37	n = 6	[IA1] i-STAT
107.7 ± 1.75	113.7 ± 1.64	95.5 ± 1.58	114.8 ± 1.87	125.6 ± 1.97	n = 50	[JJ1] Ortho Clinical Diagnostics
99.6 ± 1.38	109.1 ± 1.21	89.5 ± 1.04	108.8 ± 1.11	120.1 ± 1.53	n = 21	[RO4] Roche cobas c311/c501/c502/c701
102.2 ± 1.03	111.2 ± 0.94	91.6 ± 0.81	110.7 ± 0.89	122.2 ± 1.37	n = 29	[RO2] Roche Hitachi and Modular D/P
105.7 ± 1.15	112.1 ± 1.11	94.2 ± 1.08	112.7 ± 1.41	123.1 ± 1.18	n = 11	[RO1] Roche Integra and MIRA
105.3 ± 1.04	112.3 ± 1.10	93.1 ± 1.42	111.1 ± 1.72	122.2 ± 2.18	n = 26	[BY1] Siemens ADVIA/ADVISIA Centaur
105.4 ± 1.66	111.9 ± 1.84	94.0 ± 1.48	113.5 ± 1.31	121.9 ± 2.46	n = 98	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Albumin (g/dL)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
4.88 ± 0.17	4.03 ± 0.21	2.77 ± 0.21	3.41 ± 0.20	4.00 ± 0.22	n = 356	[---] All Methods & Instruments
4.67 ± 0.05	4.37 ± 0.05	3.13 ± 0.05	3.64 ± 0.10	4.37 ± 0.05	n = 3	<Instruments>
4.66 ± 0.15	3.80 ± 0.15	2.64 ± 0.09	3.25 ± 0.16	3.79 ± 0.15	n = 21	[AXA] Abaxis Piccolo
4.88 ± 0.11	4.07 ± 0.08	2.87 ± 0.08	3.48 ± 0.06	4.07 ± 0.09	n = 56	[ABJ] Abbott Architect c System
4.73 ± 0.16	3.75 ± 0.23	2.60 ± 0.08	3.15 ± 0.17	3.73 ± 0.20	n = 4	[OLC] Beckman Coulter AU Chemistry System
4.73 ± 0.09	3.67 ± 0.08	2.55 ± 0.06	3.08 ± 0.04	3.63 ± 0.09	n = 4	[BCS] Beckman Coulter CX
4.72 ± 0.14	3.62 ± 0.08	2.52 ± 0.07	3.06 ± 0.08	3.61 ± 0.09	n = 16	[BCX] Beckman Coulter LX-20
4.80 ± 0.07	3.70 ± 0.00	2.60 ± 0.00	3.14 ± 0.06	3.71 ± 0.05	n = 15	[BCG] Beckman Coulter UniCel DxC 600
4.80 ± 0.15	3.82 ± 0.09	2.44 ± 0.11	3.43 ± 0.11	3.79 ± 0.13	n = 10	[BCH] Beckman Coulter UniCel DxC 800
4.80 ± 0.17	3.86 ± 0.15	2.41 ± 0.10	3.36 ± 0.13	3.78 ± 0.13	n = 19	[JJE] Ortho Vitros 250/350/950
4.75 ± 0.10	3.84 ± 0.08	2.42 ± 0.04	3.35 ± 0.09	3.76 ± 0.09	n = 18	[JJF] Ortho Vitros 5,1FS
5.03 ± 0.12	4.28 ± 0.14	3.03 ± 0.10	3.73 ± 0.10	4.26 ± 0.17	n = 16	[JVG] Ortho Vitros 5600
4.97 ± 0.14	4.17 ± 0.05	2.93 ± 0.05	3.63 ± 0.05	4.23 ± 0.05	n = 3	[ROC] Roche cobas c501
4.90 ± 0.00	4.20 ± 0.00	2.92 ± 0.07	3.60 ± 0.00	4.20 ± 0.09	n = 6	[ROH] Roche cobas c701
4.83 ± 0.09	4.07 ± 0.09	2.87 ± 0.09	3.55 ± 0.06	4.07 ± 0.09	n = 4	[ROS] Roche Cobas INTEGRA 400
4.98 ± 0.13	4.22 ± 0.14	2.95 ± 0.12	3.69 ± 0.13	4.21 ± 0.11	n = 30	[ROT] Roche Cobas INTEGRA 800
4.80 ± 0.08	4.20 ± 0.05	2.94 ± 0.06	3.59 ± 0.04	4.20 ± 0.05	n = 22	[ROD] Roche MODULAR D/P
4.70 ± 0.09	4.17 ± 0.05	2.90 ± 0.09	3.53 ± 0.05	4.13 ± 0.05	n = 3	[BYE] Siemens ADVIA 1800
4.97 ± 0.06	4.03 ± 0.06	2.74 ± 0.06	3.29 ± 0.06	3.99 ± 0.05	n = 14	[BYB] Siemens ADVIA 2400
5.01 ± 0.14	4.07 ± 0.11	2.77 ± 0.09	3.32 ± 0.10	4.03 ± 0.09	n = 24	[DUE] Siemens Dimension EXL
5.03 ± 0.10	4.07 ± 0.11	2.81 ± 0.07	3.37 ± 0.09	4.05 ± 0.09	n = 40	[DUR] Siemens Dimension RxL
4.97 ± 0.08	4.05 ± 0.08	2.76 ± 0.06	3.32 ± 0.06	4.00 ± 0.08	n = 20	[DUT] Siemens Dimension Vista
						[DUX] Siemens Dimension Xpand
4.67 ± 0.05	4.37 ± 0.05	3.13 ± 0.05	3.64 ± 0.10	4.37 ± 0.05	n = 3	<Reagents>
4.66 ± 0.15	3.80 ± 0.15	2.64 ± 0.09	3.25 ± 0.16	3.79 ± 0.15	n = 21	[AX1] Abaxis
4.74 ± 0.12	3.66 ± 0.08	2.55 ± 0.07	3.09 ± 0.08	3.64 ± 0.09	n = 36	[AB1] Abbott
4.89 ± 0.10	4.07 ± 0.08	2.87 ± 0.07	3.48 ± 0.06	4.07 ± 0.09	n = 55	[BC1] Beckman Coulter
4.77 ± 0.14	3.84 ± 0.12	2.42 ± 0.08	3.36 ± 0.12	3.77 ± 0.12	n = 48	[OL1] Beckman Coulter AU Series
5.02 ± 0.13	4.26 ± 0.13	3.01 ± 0.10	3.70 ± 0.09	4.25 ± 0.15	n = 21	[JJ1] Ortho Clinical Diagnostics
4.97 ± 0.13	4.23 ± 0.13	2.95 ± 0.12	3.70 ± 0.11	4.21 ± 0.10	n = 29	[RO4] Roche cobas c311/c501/c502/c701
4.88 ± 0.05	4.15 ± 0.08	2.90 ± 0.08	3.59 ± 0.05	4.15 ± 0.11	n = 10	[RO2] Roche Hitachi and Modular D/P
4.80 ± 0.08	4.20 ± 0.06	2.94 ± 0.06	3.59 ± 0.06	4.20 ± 0.06	n = 27	[BY1] Roche Integra and MIRA
5.00 ± 0.11	4.06 ± 0.09	2.78 ± 0.08	3.33 ± 0.08	4.02 ± 0.09	n = 98	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Total Protein (g/dL)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
7.43 ± 0.22	7.00 ± 0.20	4.86 ± 0.15	6.10 ± 0.18	7.08 ± 0.22	n = 357	[---] All Methods & Instruments
						<Instruments>
7.42 ± 0.15	7.13 ± 0.05	4.87 ± 0.05	6.27 ± 0.05	7.23 ± 0.05	n = 3	[AXA] Abaxis Piccolo
7.45 ± 0.09	6.99 ± 0.07	4.76 ± 0.08	6.07 ± 0.08	7.09 ± 0.07	n = 21	[ABJ] Abbott Architect c System
7.28 ± 0.15	6.82 ± 0.14	4.76 ± 0.11	5.98 ± 0.12	6.93 ± 0.14	n = 56	[OLC] Beckman Coulter AU Chemistry System
7.32 ± 0.04	7.05 ± 0.12	4.82 ± 0.04	6.08 ± 0.20	7.07 ± 0.09	n = 4	[BCS] Beckman Coulter CX
7.16 ± 0.17	6.70 ± 0.09	4.68 ± 0.08	5.78 ± 0.08	6.79 ± 0.14	n = 5	[BCX] Beckman Coulter LX-20
7.33 ± 0.17	6.99 ± 0.17	4.83 ± 0.10	5.98 ± 0.16	7.02 ± 0.14	n = 15	[BCG] Beckman Coulter UniCel DxC 600
7.09 ± 0.16	6.84 ± 0.13	4.69 ± 0.10	5.85 ± 0.12	6.81 ± 0.15	n = 15	[BCH] Beckman Coulter UniCel DxC 800
7.52 ± 0.12	7.05 ± 0.12	4.85 ± 0.12	6.13 ± 0.12	6.96 ± 0.09	n = 10	[JJE] Ortho Vitros 250/350/950
7.44 ± 0.17	6.93 ± 0.12	4.73 ± 0.11	5.98 ± 0.13	6.87 ± 0.13	n = 20	[JJF] Ortho Vitros 5,1FS
7.52 ± 0.16	7.07 ± 0.11	4.85 ± 0.09	6.09 ± 0.15	6.96 ± 0.17	n = 18	[JJG] Ortho Vitros 5600
7.25 ± 0.16	6.86 ± 0.12	4.81 ± 0.10	6.00 ± 0.13	6.99 ± 0.12	n = 16	[ROC] Roche cobas c501
7.17 ± 0.05	6.74 ± 0.10	4.80 ± 0.09	5.97 ± 0.05	6.87 ± 0.05	n = 3	[ROH] Roche cobas c701
7.29 ± 0.11	6.81 ± 0.11	4.76 ± 0.08	6.00 ± 0.09	6.97 ± 0.12	n = 6	[ROS] Roche Cobas INTEGRA 400
7.13 ± 0.15	6.73 ± 0.15	4.82 ± 0.15	5.92 ± 0.15	6.90 ± 0.08	n = 4	[ROT] Roche Cobas INTEGRA 800
7.35 ± 0.12	6.91 ± 0.13	4.85 ± 0.11	6.07 ± 0.09	7.03 ± 0.12	n = 30	[ROD] Roche MODULAR D/P
7.45 ± 0.10	7.14 ± 0.10	4.94 ± 0.08	6.17 ± 0.07	7.15 ± 0.09	n = 22	[BYE] Siemens ADVIA 1800
7.43 ± 0.14	7.13 ± 0.14	4.94 ± 0.10	6.17 ± 0.14	7.13 ± 0.14	n = 3	[BYB] Siemens ADVIA 2400
7.66 ± 0.09	7.17 ± 0.08	5.02 ± 0.07	6.27 ± 0.08	7.34 ± 0.11	n = 14	[DUE] Siemens Dimension EXL
7.68 ± 0.16	7.21 ± 0.15	5.02 ± 0.11	6.31 ± 0.14	7.35 ± 0.12	n = 24	[DUR] Siemens Dimension RxL
7.64 ± 0.10	7.16 ± 0.08	5.00 ± 0.08	6.26 ± 0.07	7.31 ± 0.08	n = 40	[DUT] Siemens Dimension Vista
7.64 ± 0.13	7.16 ± 0.11	5.00 ± 0.09	6.28 ± 0.11	7.33 ± 0.13	n = 20	[DUX] Siemens Dimension Xpand
						<Reagents>
7.42 ± 0.15	7.13 ± 0.05	4.87 ± 0.05	6.27 ± 0.05	7.23 ± 0.05	n = 3	[AX1] Abaxis
7.45 ± 0.09	6.99 ± 0.07	4.76 ± 0.08	6.07 ± 0.08	7.09 ± 0.07	n = 21	[AB1] Abbott
7.21 ± 0.19	6.88 ± 0.18	4.76 ± 0.13	5.89 ± 0.15	6.91 ± 0.19	n = 37	[BC1] Beckman Coulter
7.28 ± 0.16	6.82 ± 0.14	4.76 ± 0.12	5.98 ± 0.12	6.94 ± 0.14	n = 55	[OL1] Beckman Coulter AU Series
7.49 ± 0.16	7.01 ± 0.14	4.81 ± 0.12	6.05 ± 0.15	6.92 ± 0.15	n = 48	[JJ1] Ortho Clinical Diagnostics
7.23 ± 0.14	6.86 ± 0.12	4.81 ± 0.09	6.00 ± 0.10	6.96 ± 0.12	n = 20	[RO4] Roche cobas c311/c501/c502/c701
7.35 ± 0.12	6.91 ± 0.13	4.85 ± 0.11	6.07 ± 0.09	7.03 ± 0.12	n = 30	[RO2] Roche Hitachi and Modular D/P
7.23 ± 0.16	6.78 ± 0.14	4.78 ± 0.11	5.97 ± 0.12	6.94 ± 0.11	n = 10	[RO1] Roche Integra and MIRA
7.46 ± 0.10	7.14 ± 0.10	4.94 ± 0.08	6.18 ± 0.09	7.15 ± 0.09	n = 27	[BY1] Siemens ADVIA/ADVISIA Centaur
7.65 ± 0.12	7.17 ± 0.10	5.01 ± 0.09	6.27 ± 0.10	7.32 ± 0.11	n = 97	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Cholesterol (mg/dL)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
269.6 ± 8.63	149.1 ± 8.79	110.0 ± 5.32	244.9 ± 7.82	204.3 ± 8.49	n = 327	[---] All Methods & Instruments
273.3 ± 1.37	160.2 ± 1.54	113.6 ± 1.02	252.5 ± 1.86	208.0 ± 1.80	n = 3	<Instruments>
275.6 ± 2.39	151.2 ± 1.68	115.0 ± 1.08	249.6 ± 2.39	210.3 ± 1.83	n = 17	[AXA] Abaxis Piccolo
267.5 ± 7.07	144.5 ± 3.90	108.1 ± 2.66	241.3 ± 5.75	202.0 ± 5.43	n = 60	[ABJ] Abbott Architect c System
264.6 ± 5.76	147.9 ± 1.13	108.2 ± 6.92	241.6 ± 13.35	198.6 ± 7.62	n = 5	[OLC] Beckman Coulter AU Chemistry System
271.5 ± 3.77	150.7 ± 4.21	108.6 ± 3.39	244.6 ± 5.64	198.8 ± 5.30	n = 4	[BCS] Beckman Coulter CX
266.3 ± 7.70	149.1 ± 3.73	107.4 ± 3.50	240.1 ± 6.06	197.8 ± 5.00	n = 14	[BCX] Beckman Coulter LX-20
265.2 ± 4.17	147.1 ± 2.55	107.3 ± 1.62	240.4 ± 5.35	196.2 ± 3.65	n = 14	[BCG] Beckman Coulter UniCel DxC 600
281.8 ± 10.39	164.1 ± 3.52	114.3 ± 4.26	250.3 ± 8.12	218.4 ± 6.45	n = 4	[BCH] Beckman Coulter UniCel DxC 800
285.1 ± 7.78	167.6 ± 4.18	115.8 ± 2.14	252.7 ± 5.53	218.8 ± 5.72	n = 18	[JJE] Ortho Vitros 250/350/950
290.1 ± 4.59	170.6 ± 2.42	116.0 ± 2.26	256.3 ± 4.25	219.9 ± 5.62	n = 18	[JJF] Ortho Vitros 5,1FS
274.9 ± 5.73	153.7 ± 4.58	115.6 ± 4.30	249.3 ± 5.61	210.3 ± 5.34	n = 14	[JGJ] Ortho Vitros 5600
268.5 ± 5.43	147.0 ± 2.70	111.7 ± 2.26	243.7 ± 3.16	203.5 ± 2.74	n = 3	[ROC] Roche cobas c501
274.0 ± 1.37	150.0 ± 1.60	112.4 ± 1.02	247.1 ± 3.13	207.9 ± 2.60	n = 6	[ROH] Roche cobas c701
271.8 ± 3.10	150.5 ± 2.98	112.8 ± 1.96	247.8 ± 2.92	205.4 ± 1.80	n = 4	[ROS] Roche Cobas INTEGRA 400
272.6 ± 6.18	151.8 ± 3.13	113.5 ± 2.50	249.7 ± 5.16	208.6 ± 4.88	n = 31	[ROT] Roche Cobas INTEGRA 800
265.8 ± 5.42	156.6 ± 3.45	112.8 ± 2.34	244.1 ± 4.98	202.7 ± 3.85	n = 22	[ROD] Roche MODULAR D/P
263.9 ± 5.22	156.8 ± 2.36	112.3 ± 3.16	242.1 ± 5.63	203.3 ± 9.73	n = 3	[BYE] Siemens ADVIA 1800
263.5 ± 4.86	137.9 ± 3.48	102.2 ± 2.71	237.6 ± 5.57	196.5 ± 4.91	n = 13	[BYB] Siemens ADVIA 2400
266.0 ± 4.04	140.5 ± 1.92	103.8 ± 2.13	240.5 ± 5.29	199.4 ± 4.88	n = 17	[DUE] Siemens Dimension EXL
264.7 ± 5.19	144.4 ± 3.25	106.0 ± 2.46	239.4 ± 5.16	200.9 ± 5.07	n = 35	[DUR] Siemens Dimension RxL
267.5 ± 4.21	141.0 ± 3.33	104.5 ± 2.17	243.9 ± 3.97	199.9 ± 3.09	n = 14	[DUT] Siemens Dimension Vista
273.3 ± 1.37	160.2 ± 1.54	113.6 ± 1.02	252.5 ± 1.86	208.0 ± 1.80	n = 3	[DUX] Siemens Dimension Xpand
275.6 ± 2.39	151.2 ± 1.68	115.0 ± 1.08	249.6 ± 2.39	210.3 ± 1.83	n = 17	<Reagents>
266.0 ± 6.18	148.2 ± 3.44	107.5 ± 3.01	240.5 ± 6.71	197.2 ± 5.06	n = 36	[AX1] Abaxis
267.8 ± 6.90	144.5 ± 3.79	108.2 ± 2.62	241.5 ± 5.64	202.3 ± 5.21	n = 57	[AB1] Abbott
287.3 ± 7.35	168.6 ± 4.17	115.9 ± 2.34	254.4 ± 5.57	219.3 ± 5.74	n = 40	[BC1] Beckman Coulter
274.2 ± 6.11	152.8 ± 5.08	115.0 ± 4.19	248.6 ± 5.72	209.2 ± 5.63	n = 18	[OL1] Beckman Coulter AU Series
272.5 ± 6.34	151.9 ± 3.20	113.5 ± 2.63	249.8 ± 5.46	208.8 ± 4.94	n = 31	[JJ1] Ortho Clinical Diagnostics
273.6 ± 1.34	150.2 ± 2.23	112.4 ± 1.27	247.4 ± 3.07	206.8 ± 2.50	n = 10	[RO4] Roche cobas c311/c501/c502/c701
265.5 ± 6.04	156.7 ± 3.63	112.7 ± 2.64	244.2 ± 5.45	202.7 ± 4.69	n = 27	[RO2] Roche Hitachi and Modular D/P
265.4 ± 4.90	141.9 ± 3.84	104.6 ± 2.68	240.3 ± 5.53	199.6 ± 4.68	n = 79	[BY1] Siemens Integra and MIRA
						[DA5] Siemens Dimension Centaur

## Summary of Participant Performance (Mean and Standard Deviation)

## HDL-Cholesterol (mg/dL)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
78.1 ± 8.15	35.8 ± 3.52	26.0 ± 3.15	92.8 ± 6.94	45.8 ± 5.22	n = 312	[---] All Methods & Instruments
82.4 ± 8.79	35.9 ± 4.54	25.0 ± 2.69	95.1 ± 7.74	47.7 ± 6.31	n = 19	[---] All Precipitation Methods
77.8 ± 7.98	35.8 ± 3.43	26.1 ± 3.16	92.6 ± 6.90	45.7 ± 5.09	n = 293	[---] All Homogeneous (Direct) Methods
80.6 ± 4.04	35.9 ± 1.41	25.2 ± 1.28	91.5 ± 4.08	44.3 ± 2.50	n = 16	[AB1] Abbott
61.7 ± 0.51	21.7 ± 1.37	15.3 ± 0.51	74.8 ± 3.23	31.7 ± 2.26	n = 3	[AX1] Abaxis
90.6 ± 4.61	37.0 ± 1.84	24.8 ± 1.45	101.4 ± 4.57	49.6 ± 2.35	n = 32	[BC1] Beckman Coulter
81.4 ± 2.85	34.7 ± 1.26	23.9 ± 0.96	91.1 ± 3.05	43.0 ± 1.57	n = 38	[OL1] Beckman Coulter AU Series
84.7 ± 1.51	36.5 ± 1.22	25.8 ± 0.41	95.9 ± 2.72	45.8 ± 0.41	n = 4	[GZ1] Sekisui Diagnostics (Genzyme)
90.0 ± 3.61	40.0 ± 1.44	25.6 ± 0.96	100.1 ± 5.75	53.6 ± 1.98	n = 31	[JJ1] Ortho Clinical Diagnostics
72.5 ± 2.19	34.5 ± 1.44	27.1 ± 1.26	90.4 ± 2.94	44.0 ± 1.47	n = 15	[RO4] Roche cobas c311/c501/c502/c701
75.6 ± 2.98	38.1 ± 2.31	30.1 ± 1.86	94.7 ± 3.38	48.3 ± 2.54	n = 27	[RO2] Roche Hitachi and Modular D/P
75.7 ± 2.17	36.4 ± 2.18	28.9 ± 1.98	94.2 ± 3.01	45.8 ± 2.33	n = 10	[RO1] Roche Integra and MIRA
70.5 ± 2.11	26.0 ± 0.94	17.3 ± 0.62	77.8 ± 2.41	34.0 ± 1.15	n = 27	[BY1] Siemens ADVIA/ADVIa Centaur
73.4 ± 2.50	35.1 ± 1.61	27.5 ± 1.32	91.5 ± 3.55	45.4 ± 2.21	n = 72	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## LDL-Cholesterol (mg/dL)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
155.4 ± 15.70	93.6 ± 11.00	66.4 ± 9.26	123.2 ± 13.45	129.1 ± 14.05	n = 299	[---] All Methods & Instruments
161.9 ± 8.89	97.2 ± 9.67	70.7 ± 6.47	127.0 ± 9.31	136.2 ± 8.30	n = 151	[-A-] All Calculated Results Friedewald formula [LDL=TC-HDL-(Trigs+5)]
145.7 ± 18.23	89.2 ± 11.19	61.1 ± 9.22	117.5 ± 15.95	119.7 ± 14.65	n = 147	[---] All Homogeneous (Direct) Methods
138.2 ± 1.90	91.0 ± 1.90	63.0 ± 0.00	113.8 ± 1.55	120.3 ± 1.62	n = 6	[AB1] Abbott
133.4 ± 3.65	86.1 ± 3.63	58.8 ± 2.33	109.3 ± 4.15	113.5 ± 3.82	n = 14	[BC1] Beckman Coulter
123.2 ± 5.53	76.7 ± 3.03	52.0 ± 2.84	98.0 ± 4.92	102.2 ± 5.09	n = 21	[OL1] Beckman Coulter AU Series
163.3 ± 7.62	96.1 ± 4.67	59.9 ± 2.96	135.5 ± 4.88	131.7 ± 6.04	n = 16	[JJ1] Ortho Clinical Diagnostics
172.6 ± 5.90	107.9 ± 3.58	76.9 ± 1.87	142.0 ± 3.51	139.9 ± 3.49	n = 5	[RO4] Roche cobas c311/c501/c502/c701
170.2 ± 5.11	101.6 ± 3.24	72.6 ± 2.16	137.8 ± 3.58	134.9 ± 3.52	n = 12	[RO2] Roche Hitachi and Modular D/P
148.6 ± 4.66	59.3 ± 6.84	45.4 ± 2.31	105.3 ± 3.17	93.2 ± 8.97	n = 4	[RO1] Roche Integra and MIRA
121.5 ± 8.52	75.6 ± 7.02	51.3 ± 4.90	96.4 ± 7.79	99.5 ± 8.76	n = 12	[GZ1] Sekisui Diagnostics (Genzyme)
144.0 ± 4.85	85.6 ± 2.42	56.5 ± 1.73	118.3 ± 3.66	115.9 ± 3.29	n = 14	[BY1] Siemens ADVIA/ADVIA Centaur
151.8 ± 9.20	94.3 ± 6.43	68.1 ± 4.61	121.4 ± 7.22	126.1 ± 7.09	n = 36	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Triglycerides (mg/dL)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
153.2 ± 6.59	93.0 ± 5.86	72.1 ± 4.37	127.5 ± 5.77	118.9 ± 6.15	n = 320	[---] All Methods & Instruments
						<Instruments>
175.7 ± 1.37	106.0 ± 0.90	81.3 ± 0.51	151.0 ± 1.80	138.7 ± 0.51	n = 3	[AXA] Abaxis Piccolo
150.9 ± 2.03	97.0 ± 2.48	73.1 ± 0.97	125.4 ± 2.80	116.0 ± 1.55	n = 18	[ABJ] Abbott Architect c System
151.2 ± 4.90	87.4 ± 3.48	69.8 ± 2.29	125.2 ± 4.03	116.3 ± 3.68	n = 55	[OLC] Beckman Coulter AU Chemistry System
155.8 ± 7.89	97.4 ± 7.84	72.0 ± 5.11	131.6 ± 12.27	123.6 ± 11.08	n = 4	[BCS] Beckman Coulter CX
159.0 ± 6.09	95.1 ± 4.10	72.8 ± 3.73	133.0 ± 6.63	126.7 ± 5.74	n = 4	[BCX] Beckman Coulter LX-20
156.3 ± 6.34	94.1 ± 2.74	70.0 ± 1.92	130.0 ± 5.07	124.6 ± 4.56	n = 13	[BCG] Beckman Coulter UniCel DxC 600
154.1 ± 3.61	93.5 ± 2.64	71.1 ± 2.94	129.4 ± 3.73	123.2 ± 3.43	n = 13	[BCH] Beckman Coulter UniCel DxC 800
165.5 ± 2.32	100.9 ± 1.13	78.8 ± 1.27	135.3 ± 0.82	132.4 ± 2.31	n = 4	[JJE] Ortho Vitros 250/350/950
159.0 ± 3.09	95.2 ± 2.09	74.5 ± 1.76	128.0 ± 2.80	124.1 ± 3.27	n = 18	[JJF] Ortho Vitros 5,1FS
157.0 ± 3.79	95.0 ± 3.22	74.3 ± 2.44	127.3 ± 4.85	123.6 ± 4.67	n = 18	[JJG] Ortho Vitros 5600
151.7 ± 5.05	95.0 ± 3.96	76.1 ± 3.76	129.6 ± 3.55	119.8 ± 3.76	n = 14	[ROC] Roche cobas c501
151.5 ± 1.86	91.7 ± 1.37	72.7 ± 1.37	127.3 ± 2.26	117.5 ± 1.86	n = 3	[ROH] Roche cobas c701
149.2 ± 2.71	87.4 ± 2.55	72.6 ± 1.99	124.2 ± 0.81	116.7 ± 2.60	n = 6	[ROS] Roche Cobas INTEGRA 400
147.8 ± 3.10	88.0 ± 1.50	72.2 ± 1.96	123.9 ± 3.38	116.5 ± 2.67	n = 4	[ROT] Roche Cobas INTEGRA 800
150.0 ± 3.53	94.5 ± 3.13	73.6 ± 2.52	128.7 ± 3.19	117.8 ± 3.52	n = 31	[ROD] Roche MODULAR D/P
153.8 ± 2.19	95.3 ± 1.90	72.3 ± 1.52	129.2 ± 2.08	118.3 ± 2.33	n = 22	[BYE] Siemens ADVIA 1800
147.3 ± 1.37	89.0 ± 0.90	68.5 ± 1.86	122.7 ± 1.37	114.0 ± 2.70	n = 3	[BYB] Siemens ADVIA 2400
147.3 ± 2.36	86.3 ± 3.22	64.9 ± 1.84	121.0 ± 2.42	111.0 ± 1.73	n = 13	[DUE] Siemens Dimension EXL
146.7 ± 3.22	86.9 ± 3.19	65.3 ± 3.29	121.6 ± 3.85	111.2 ± 3.25	n = 17	[DUR] Siemens Dimension RxL
162.4 ± 4.10	99.7 ± 2.79	76.6 ± 2.29	136.5 ± 3.09	124.8 ± 3.39	n = 37	[DUT] Siemens Dimension Vista
147.3 ± 3.06	88.0 ± 2.57	65.4 ± 1.77	122.6 ± 2.51	112.6 ± 2.49	n = 12	[DUX] Siemens Dimension Xpand
						<Reagents>
175.7 ± 1.37	106.0 ± 0.90	81.3 ± 0.51	151.0 ± 1.80	138.7 ± 0.51	n = 3	[AX1] Abaxis
150.9 ± 2.03	97.0 ± 2.48	73.1 ± 0.97	125.4 ± 2.80	116.0 ± 1.55	n = 18	[AB1] Abbott
156.0 ± 5.53	94.1 ± 2.85	70.9 ± 2.83	129.9 ± 5.06	124.1 ± 4.38	n = 33	[BC1] Beckman Coulter
151.6 ± 4.55	87.7 ± 3.01	69.9 ± 2.10	125.4 ± 4.01	116.3 ± 3.48	n = 51	[OL1] Beckman Coulter AU Series
148.7 ± 11.29	93.6 ± 15.38	74.4 ± 5.58	132.3 ± 14.17	124.6 ± 12.69	n = 3	[CR1] Carolina
158.7 ± 4.13	95.7 ± 3.14	74.8 ± 2.49	128.3 ± 4.51	124.6 ± 4.63	n = 40	[JJ1] Ortho Clinical Diagnostics
151.8 ± 4.44	94.1 ± 3.47	75.3 ± 3.62	129.0 ± 3.29	119.5 ± 3.44	n = 18	[RO4] Roche cobas c311/c501/c502/c701
150.0 ± 3.53	94.5 ± 3.13	73.6 ± 2.52	128.7 ± 3.19	117.8 ± 3.52	n = 31	[RO2] Roche Hitachi and Modular D/P
148.5 ± 3.10	87.7 ± 2.17	72.4 ± 2.00	123.7 ± 2.70	116.6 ± 2.64	n = 10	[RO1] Roche Integra and MIRA
153.6 ± 2.83	95.1 ± 2.44	72.1 ± 1.71	128.8 ± 3.09	118.1 ± 2.76	n = 27	[BY1] Siemens ADVIA/ADVIS Centaur
153.9 ± 9.62	93.0 ± 8.21	70.8 ± 7.01	128.6 ± 9.16	117.4 ± 8.45	n = 78	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

Homocysteine ( $\mu\text{mol/L}$ )

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
24.17 $\pm$ 2.93	10.34 $\pm$ 1.31	6.91 $\pm$ 1.04	8.52 $\pm$ 1.23	16.50 $\pm$ 1.88	n = 118	[---] All Methods & Instruments
23.71 $\pm$ 0.55	10.89 $\pm$ 0.62	7.55 $\pm$ 0.63	9.00 $\pm$ 0.79	17.28 $\pm$ 0.88	n = 11	<Instruments>
25.03 $\pm$ 1.95	9.80 $\pm$ 1.09	7.56 $\pm$ 0.71	8.70 $\pm$ 0.64	17.38 $\pm$ 1.66	n = 3	[ABH] Abbott Architect i System
26.14 $\pm$ 1.16	10.91 $\pm$ 0.59	7.64 $\pm$ 0.22	9.35 $\pm$ 0.29	17.63 $\pm$ 0.66	n = 22	[ABB] Abbott AxSym
24.85 $\pm$ 0.55	11.70 $\pm$ 0.57	7.02 $\pm$ 0.26	9.04 $\pm$ 0.38	17.96 $\pm$ 0.60	n = 4	[OLC] Beckman Coulter AU Chemistry System
26.87 $\pm$ 0.27	10.45 $\pm$ 1.10	7.17 $\pm$ 0.75	9.27 $\pm$ 0.42	17.32 $\pm$ 0.60	n = 6	[JJG] Ortho Vitros 5600
26.37 $\pm$ 0.59	10.86 $\pm$ 0.10	7.75 $\pm$ 0.19	9.25 $\pm$ 0.46	17.23 $\pm$ 0.34	n = 3	[ROD] Roche MODULAR D/P
19.79 $\pm$ 1.26	8.84 $\pm$ 0.60	5.52 $\pm$ 0.38	6.95 $\pm$ 0.54	13.90 $\pm$ 0.85	n = 25	[BYE] Siemens ADVIA 1800
22.02 $\pm$ 2.10	8.86 $\pm$ 0.75	6.46 $\pm$ 0.50	7.05 $\pm$ 0.27	13.55 $\pm$ 1.98	n = 7	[DUT] Siemens Dimension Vista
25.65 $\pm$ 1.56	10.71 $\pm$ 0.93	6.84 $\pm$ 0.71	8.65 $\pm$ 0.81	16.94 $\pm$ 0.57	n = 17	[DPD] Siemens Immulite 2000
23.78 $\pm$ 0.74	10.76 $\pm$ 0.74	7.55 $\pm$ 0.64	8.92 $\pm$ 0.76	17.33 $\pm$ 1.05	n = 14	<Reagents>
25.94 $\pm$ 2.17	12.54 $\pm$ 1.19	8.00 $\pm$ 1.18	10.71 $\pm$ 1.10	18.34 $\pm$ 0.90	n = 3	[AB1] Abbott
25.04 $\pm$ 1.44	11.41 $\pm$ 0.74	7.39 $\pm$ 0.49	9.12 $\pm$ 0.51	18.46 $\pm$ 0.20	n = 7	[AS1] Axis-Shield
26.58 $\pm$ 0.76	10.75 $\pm$ 0.49	7.65 $\pm$ 0.39	9.31 $\pm$ 0.35	17.42 $\pm$ 0.56	n = 27	[CR1] Carolina
24.24 $\pm$ 1.13	11.19 $\pm$ 0.94	6.72 $\pm$ 0.57	8.78 $\pm$ 0.56	17.58 $\pm$ 0.83	n = 6	[DZ1] Diazyme
26.33 $\pm$ 2.12	10.87 $\pm$ 1.13	7.65 $\pm$ 0.63	9.51 $\pm$ 0.85	17.42 $\pm$ 1.66	n = 3	[JJ1] Ortho Clinical Diagnostics
19.79 $\pm$ 1.26	8.84 $\pm$ 0.60	5.52 $\pm$ 0.38	6.95 $\pm$ 0.54	13.90 $\pm$ 0.85	n = 25	[GZ1] Sekisui Diagnostics (Genzyme)
22.04 $\pm$ 2.33	8.85 $\pm$ 0.74	6.42 $\pm$ 0.55	7.13 $\pm$ 0.49	13.90 $\pm$ 1.37	n = 6	[DA5] Siemens Dimension
25.50 $\pm$ 1.64	10.68 $\pm$ 0.90	6.87 $\pm$ 0.69	8.66 $\pm$ 0.78	16.89 $\pm$ 0.59	n = 18	[DP5] Siemens Immulite

## Summary of Participant Performance (Mean and Standard Deviation)

## Troponin I (µg/L)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
2.331 ± 0.272	0.018 ± 0.016	0.020 ± 0.019	0.366 ± 0.083	1.405 ± 0.312	n = 226	[---] All Methods & Instruments
14.952 ± 0.547	0.010 ± 0.004	0.010 ± 0.005	2.029 ± 0.089	7.231 ± 0.255	n = 22	<Instruments>
2.212 ± 0.192	0.010 ± 0.000	0.010 ± 0.000	0.323 ± 0.031	1.140 ± 0.094	n = 24	[ABH] Abbott Architect i System
1.190 ± 0.201	0.050 ± 0.000	0.050 ± 0.000	0.057 ± 0.011	0.390 ± 0.065	n = 8	[SAA] Beckman Coulter ACCESS
7.988 ± 1.101	0.034 ± 0.045	0.034 ± 0.045	1.126 ± 0.103	3.674 ± 0.237	n = 4	[BSA] BioSite Triage
8.778 ± 0.232	0.016 ± 0.009	0.014 ± 0.008	1.245 ± 0.046	4.174 ± 0.125	n = 16	[IAG] i-STAT
8.510 ± 0.260	0.010 ± 0.000	0.010 ± 0.000	1.225 ± 0.054	3.999 ± 0.138	n = 12	[JJG] Ortho Vitros 5600
2.570 ± 0.172	0.009 ± 0.006	0.009 ± 0.006	0.370 ± 0.032	1.565 ± 0.087	n = 42	[JJC] Ortho Vitros ECi/ECiQ
2.901 ± 0.216	0.018 ± 0.020	0.018 ± 0.020	0.386 ± 0.023	1.674 ± 0.101	n = 4	[COB] Siemens ADVIA Centaur
2.278 ± 0.082	0.078 ± 0.035	0.078 ± 0.035	0.443 ± 0.031	1.561 ± 0.087	n = 10	[BYP] Siemens ADVIA Centaur CP
2.129 ± 0.180	0.040 ± 0.000	0.040 ± 0.000	0.268 ± 0.024	1.073 ± 0.077	n = 14	[DUE] Siemens Dimension EXL
2.291 ± 0.085	0.020 ± 0.000	0.031 ± 0.006	0.452 ± 0.024	1.606 ± 0.063	n = 38	[DUR] Siemens Dimension RxL
2.286 ± 0.151	0.033 ± 0.016	0.033 ± 0.022	0.286 ± 0.035	1.173 ± 0.088	n = 11	[DUT] Siemens Dimension Vista
6.466 ± 0.513	0.200 ± 0.000	0.200 ± 0.000	0.962 ± 0.122	3.377 ± 0.185	n = 5	[DUX] Siemens Dimension Xpand
14.496 ± 0.525	0.060 ± 0.000	0.060 ± 0.000	2.129 ± 0.139	7.104 ± 0.151	n = 5	[DPD] Siemens Immulite 2000
						[TOM] Tosoh Bioscience
14.979 ± 0.547	0.009 ± 0.006	0.009 ± 0.007	2.027 ± 0.086	7.241 ± 0.252	n = 27	<Reagents>
2.195 ± 0.192	0.009 ± 0.006	0.010 ± 0.000	0.321 ± 0.029	1.138 ± 0.089	n = 26	[AB1] Abbott
1.190 ± 0.201	0.050 ± 0.000	0.050 ± 0.000	0.057 ± 0.011	0.390 ± 0.065	n = 8	[BC1] Beckman Coulter
8.657 ± 0.294	0.012 ± 0.006	0.012 ± 0.006	1.237 ± 0.050	4.094 ± 0.165	n = 28	[BS1] Biosite Diagnostics
1.307 ± 0.095	0.304 ± 0.006	0.304 ± 0.006	0.304 ± 0.006	0.582 ± 0.024	n = 5	[JJ1] Ortho Clinical Diagnostics
2.591 ± 0.196	0.009 ± 0.005	0.009 ± 0.006	0.371 ± 0.031	1.573 ± 0.091	n = 46	[RO3] Roche Elecsys/Modular E/e601/e411
2.218 ± 0.176	0.035 ± 0.015	0.035 ± 0.016	0.275 ± 0.030	1.119 ± 0.107	n = 27	[BY1] Siemens ADVIA/ADVIS Centaur
2.284 ± 0.085	0.020 ± 0.000	0.031 ± 0.007	0.450 ± 0.026	1.597 ± 0.072	n = 46	[DA5] Siemens Dimension
6.466 ± 0.513	0.200 ± 0.000	0.200 ± 0.000	0.962 ± 0.122	3.377 ± 0.185	n = 5	[DA6] Siemens Dimension LOCI
14.630 ± 0.598	0.060 ± 0.000	0.060 ± 0.000	2.106 ± 0.158	7.052 ± 0.108	n = 4	[DP5] Siemens Immulite
						[TO2] Tosoh ST AIA

## Summary of Participant Performance (Mean and Standard Deviation)

Troponin T ( $\mu\text{g/L}$ )

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
1.339 $\pm$ 0.100	0.010 $\pm$ 0.000	0.010 $\pm$ 0.000	0.278 $\pm$ 0.020	0.730 $\pm$ 0.062	n = 32	[---] All Methods & Instruments
1.342 $\pm$ 0.069	0.010 $\pm$ 0.000	0.010 $\pm$ 0.000	0.280 $\pm$ 0.020	0.738 $\pm$ 0.057	n = 11	<Instruments>
1.406 $\pm$ 0.088	0.010 $\pm$ 0.000	0.010 $\pm$ 0.000	0.286 $\pm$ 0.025	0.762 $\pm$ 0.077	n = 9	[ROA] Roche cobas e601
1.283 $\pm$ 0.043	0.010 $\pm$ 0.000	0.010 $\pm$ 0.000	0.273 $\pm$ 0.014	0.704 $\pm$ 0.017	n = 8	[BME] Roche Elecsys
1.337 $\pm$ 0.105	0.010 $\pm$ 0.000	0.010 $\pm$ 0.000	0.279 $\pm$ 0.021	0.733 $\pm$ 0.066	n = 29	[ROE] Roche MODULAR E
						<Reagents>
						[RO3] Roche Elecsys/Modular E/e601/e411

## Summary of Participant Performance (Mean and Standard Deviation)

## Alanine Aminotransferase (U/L 37°C)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
111.2 ± 8.02	49.8 ± 6.66	246.1 ± 15.20	79.5 ± 7.72	170.4 ± 11.11	n = 358	[---] All Methods & Instruments
99.2 ± 1.54	49.3 ± 1.37	215.5 ± 4.53	73.5 ± 1.86	151.0 ± 4.60	n = 3	<Instruments>
111.2 ± 3.29	47.0 ± 1.68	246.5 ± 7.27	77.8 ± 2.67	170.6 ± 5.17	n = 21	[AXA] Abaxis Piccolo
97.6 ± 2.40	44.7 ± 1.50	220.6 ± 5.58	70.3 ± 1.72	153.2 ± 3.70	n = 56	[ABJ] Abbott Architect c System
109.4 ± 4.66	49.3 ± 4.21	244.3 ± 13.40	77.4 ± 5.87	168.7 ± 10.21	n = 4	[OLC] Beckman Coulter AU Chemistry System
108.1 ± 1.88	49.0 ± 1.14	236.0 ± 4.01	76.2 ± 1.27	162.0 ± 4.01	n = 4	[BCS] Beckman Coulter CX
108.8 ± 3.39	48.7 ± 1.23	237.2 ± 3.86	76.6 ± 2.12	164.6 ± 4.00	n = 16	[BCX] Beckman Coulter LX-20
110.0 ± 2.04	49.2 ± 1.11	240.1 ± 3.29	77.4 ± 1.37	166.6 ± 1.86	n = 14	[BCG] Beckman Coulter UniCel DxC 600
116.3 ± 4.08	66.3 ± 4.43	265.4 ± 5.54	93.6 ± 3.65	186.2 ± 4.42	n = 11	[BCH] Beckman Coulter UniCel DxC 800
117.9 ± 5.39	66.7 ± 4.35	267.4 ± 6.53	93.0 ± 3.89	187.3 ± 5.67	n = 20	[JJE] Ortho Vitros 250/350/950
115.2 ± 3.20	64.5 ± 3.01	265.3 ± 7.07	92.8 ± 2.63	183.4 ± 4.84	n = 18	[JJF] Ortho Vitros 5,1FS
108.9 ± 3.36	48.7 ± 1.93	247.6 ± 8.61	78.4 ± 2.81	170.4 ± 4.94	n = 16	[JGJ] Ortho Vitros 5600
105.5 ± 1.86	44.7 ± 1.37	242.3 ± 2.26	75.3 ± 0.51	165.3 ± 0.51	n = 3	[ROC] Roche cobas c501
107.4 ± 3.33	46.2 ± 1.83	247.3 ± 7.48	76.2 ± 2.31	168.5 ± 4.58	n = 7	[ROH] Roche cobas c701
105.3 ± 0.90	46.2 ± 0.41	240.8 ± 1.46	75.0 ± 0.75	165.0 ± 1.65	n = 4	[ROS] Roche Cobas INTEGRA 400
108.9 ± 3.50	47.7 ± 1.90	242.8 ± 8.12	77.3 ± 2.43	167.0 ± 5.25	n = 30	[ROT] Roche Cobas INTEGRA 800
114.4 ± 3.01	47.8 ± 3.25	255.6 ± 5.15	80.9 ± 2.41	174.6 ± 5.68	n = 22	[ROD] Roche MODULAR D/P
112.4 ± 1.02	49.2 ± 4.10	249.9 ± 8.93	79.3 ± 2.26	171.8 ± 2.36	n = 3	[BYE] Siemens ADVIA 1800
119.8 ± 3.63	57.5 ± 3.82	249.2 ± 3.53	86.4 ± 3.68	175.0 ± 3.71	n = 15	[BYB] Siemens ADVIA 2400
121.4 ± 3.68	58.1 ± 3.07	252.6 ± 5.00	87.6 ± 2.74	177.7 ± 3.70	n = 23	[DUE] Siemens Dimension EXL
113.5 ± 2.02	47.9 ± 1.40	247.2 ± 4.22	78.6 ± 1.82	171.2 ± 2.92	n = 40	[DUR] Siemens Dimension RxL
121.0 ± 4.69	56.6 ± 6.03	253.9 ± 5.95	87.3 ± 5.19	178.4 ± 4.59	n = 19	[DUT] Siemens Dimension Vista
99.2 ± 1.54	49.3 ± 1.37	215.5 ± 4.53	73.5 ± 1.86	151.0 ± 4.60	n = 3	[DUX] Siemens Dimension Xpand
111.2 ± 3.29	47.0 ± 1.68	246.5 ± 7.27	77.8 ± 2.67	170.6 ± 5.17	n = 21	<Reagents>
109.2 ± 2.82	48.9 ± 1.20	238.2 ± 4.24	76.8 ± 1.85	165.1 ± 3.50	n = 38	[AB1] Abbott
97.7 ± 2.33	44.7 ± 1.37	220.7 ± 5.48	70.3 ± 1.61	153.3 ± 3.59	n = 54	[BC1] Beckman Coulter
116.3 ± 4.62	65.6 ± 4.21	266.1 ± 6.52	92.9 ± 3.53	185.6 ± 5.30	n = 50	[OL1] Beckman Coulter AU Series
108.4 ± 3.23	48.4 ± 2.27	246.2 ± 7.81	77.8 ± 2.71	169.6 ± 4.77	n = 21	[JJ1] Ortho Clinical Diagnostics
108.9 ± 3.50	47.7 ± 1.90	242.8 ± 8.12	77.3 ± 2.43	167.0 ± 5.25	n = 30	[RO4] Roche cobas c311/c501/c502/c701
106.3 ± 2.78	46.2 ± 1.42	244.1 ± 6.54	75.6 ± 1.92	166.8 ± 4.07	n = 11	[RO2] Roche Hitachi and Modular D/P
113.9 ± 3.04	47.8 ± 3.30	254.9 ± 6.17	80.7 ± 2.49	174.2 ± 5.50	n = 27	[RO1] Roche Integra and MIRA
118.6 ± 5.21	54.5 ± 6.19	250.6 ± 5.56	84.5 ± 5.65	175.6 ± 4.90	n = 73	[DA5] Siemens Dimension
114.3 ± 3.60	47.8 ± 1.20	247.4 ± 4.31	79.6 ± 3.36	171.7 ± 3.45	n = 22	[DA8] Siemens Dimension IFCC Standardized

## Summary of Participant Performance (Mean and Standard Deviation)

## Aspartate Aminotransferase (U/L 37°C)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
214.9 ± 10.44	76.0 ± 4.44	191.1 ± 9.90	126.4 ± 7.26	97.8 ± 4.87	n = 358	[---] All Methods & Instruments
210.5 ± 6.32	77.7 ± 4.22	188.7 ± 5.97	124.0 ± 4.51	97.5 ± 2.74	n = 3	<Instruments>
216.3 ± 4.04	75.5 ± 1.31	192.4 ± 3.85	124.5 ± 1.95	97.2 ± 1.90	n = 21	[AXA] Abaxis Piccolo
191.5 ± 5.46	68.8 ± 1.60	172.2 ± 4.27	111.7 ± 2.99	88.2 ± 2.20	n = 56	[ABJ] Abbott Architect c System
215.0 ± 12.66	76.1 ± 7.00	190.5 ± 13.09	124.8 ± 11.95	98.2 ± 7.45	n = 4	[OLC] Beckman Coulter AU Chemistry System
210.4 ± 4.14	75.2 ± 2.80	187.4 ± 3.05	122.3 ± 3.17	96.7 ± 1.51	n = 4	[BCS] Beckman Coulter CX
212.4 ± 4.12	75.4 ± 2.07	189.2 ± 3.40	123.3 ± 2.88	97.3 ± 2.43	n = 16	[BCX] Beckman Coulter LX-20
212.8 ± 2.82	76.3 ± 1.37	190.6 ± 3.26	124.2 ± 1.44	97.2 ± 1.46	n = 14	[BCG] Beckman Coulter UniCel DxC 600
214.2 ± 5.57	81.2 ± 2.46	184.8 ± 3.84	130.7 ± 2.33	100.8 ± 2.35	n = 11	[BCH] Beckman Coulter UniCel DxC 800
217.4 ± 5.04	81.9 ± 2.04	188.7 ± 3.31	131.1 ± 3.39	100.7 ± 2.80	n = 20	[JJE] Ortho Vitros 250/350/950
218.4 ± 6.33	82.3 ± 3.31	188.0 ± 5.65	131.5 ± 4.55	101.5 ± 3.19	n = 18	[JJF] Ortho Vitros 5,1FS
217.7 ± 8.09	76.3 ± 2.83	196.0 ± 7.17	126.8 ± 4.11	98.7 ± 3.35	n = 15	[JJG] Ortho Vitros 5600
218.3 ± 3.07	75.3 ± 1.37	192.6 ± 3.87	122.6 ± 2.56	98.7 ± 1.37	n = 3	[ROC] Roche cobas c501
224.2 ± 5.04	77.0 ± 1.64	201.5 ± 4.94	128.7 ± 2.29	100.7 ± 1.89	n = 7	[ROH] Roche cobas c701
220.1 ± 4.04	76.3 ± 2.04	199.2 ± 4.86	126.7 ± 0.82	98.0 ± 0.75	n = 5	[ROS] Roche Cobas INTEGRA 400
215.9 ± 5.33	75.7 ± 2.41	191.8 ± 4.49	125.1 ± 3.21	97.7 ± 3.10	n = 30	[ROT] Roche Cobas INTEGRA 800
231.5 ± 4.91	78.4 ± 4.44	205.7 ± 5.10	134.6 ± 3.79	103.6 ± 3.61	n = 22	[ROD] Roche MODULAR D/P
225.9 ± 3.72	81.1 ± 5.63	203.5 ± 5.43	130.8 ± 3.23	103.1 ± 2.86	n = 3	[BYE] Siemens ADVIA 1800
213.9 ± 5.23	76.3 ± 2.08	192.9 ± 4.32	128.1 ± 3.40	97.9 ± 2.83	n = 14	[BYB] Siemens ADVIA 2400
214.4 ± 5.28	76.4 ± 1.95	195.2 ± 4.29	129.0 ± 3.63	98.5 ± 2.74	n = 24	[DUE] Siemens Dimension EXL
218.1 ± 3.74	75.5 ± 2.17	195.8 ± 3.87	129.6 ± 2.73	97.9 ± 2.18	n = 40	[DUR] Siemens Dimension RxL
217.0 ± 4.06	76.8 ± 2.67	196.3 ± 4.10	130.6 ± 2.83	100.0 ± 2.85	n = 19	[DUT] Siemens Dimension Vista
210.5 ± 6.32	77.7 ± 4.22	188.7 ± 5.97	124.0 ± 4.51	97.5 ± 2.74	n = 3	[DUX] Siemens Dimension Xpand
216.3 ± 4.04	75.5 ± 1.31	192.4 ± 3.85	124.5 ± 1.95	97.2 ± 1.90	n = 21	<Reagents>
212.2 ± 3.77	75.7 ± 2.10	189.3 ± 3.65	123.6 ± 2.61	97.1 ± 1.99	n = 38	[AX1] Abaxis
191.6 ± 5.30	68.8 ± 1.54	172.3 ± 4.07	111.8 ± 2.87	88.3 ± 2.08	n = 54	[AB1] Abbott
217.1 ± 5.83	81.8 ± 2.64	187.5 ± 4.78	131.1 ± 3.69	101.0 ± 2.86	n = 50	[BC1] Beckman Coulter
217.2 ± 7.03	76.0 ± 2.46	194.9 ± 6.61	125.6 ± 4.09	98.5 ± 2.95	n = 20	[OL1] Beckman Coulter AU Series
215.9 ± 5.33	75.7 ± 2.41	191.8 ± 4.49	125.1 ± 3.21	97.7 ± 3.10	n = 30	[JJ1] Ortho Clinical Diagnostics
222.7 ± 5.29	76.7 ± 1.85	200.6 ± 5.08	128.1 ± 2.64	100.0 ± 2.48	n = 12	[RO4] Roche cobas c311/c501/c502/c701
230.1 ± 6.12	78.5 ± 4.37	204.6 ± 5.90	133.9 ± 4.17	103.3 ± 3.68	n = 27	[RO2] Roche Hitachi and Modular D/P
216.3 ± 4.75	76.2 ± 2.17	195.3 ± 4.35	129.4 ± 3.08	98.5 ± 2.51	n = 89	[R01] Roche Integra and MIRA
220.0 ± 3.71	76.1 ± 4.54	195.9 ± 3.28	130.0 ± 3.85	98.9 ± 6.27	n = 7	[BY1] Siemens ADVIA/ADVIa Centaur
						[DA5] Siemens Dimension
						[DA8] Siemens Dimension IFCC Standardized

## Summary of Participant Performance (Mean and Standard Deviation)

 $\alpha$ -Amylase (U/L 37°C)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
46.4 ± 6.32	351.0 ± 58.51	220.8 ± 37.74	131.9 ± 16.12	90.5 ± 7.43	n = 317	[---] All Methods & Instruments
47.8 ± 1.21	398.8 ± 5.77	245.2 ± 2.81	143.9 ± 1.65	93.9 ± 1.27	n = 19	<Instruments>
36.9 ± 2.34	308.2 ± 18.63	192.5 ± 11.50	113.3 ± 6.34	73.5 ± 4.10	n = 48	[ABJ] Abbott Architect c System
55.5 ± 1.86	463.1 ± 142.63	280.4 ± 67.48	166.8 ± 39.94	111.3 ± 17.03	n = 3	[OLC] Beckman Coulter AU Chemistry System
51.7 ± 0.90	374.5 ± 8.28	235.6 ± 4.98	142.8 ± 3.09	100.7 ± 2.54	n = 4	[BCS] Beckman Coulter CX
51.0 ± 1.35	371.0 ± 7.34	234.5 ± 5.65	141.5 ± 2.43	98.4 ± 2.22	n = 13	[BCX] Beckman Coulter LX-20
51.6 ± 0.94	375.4 ± 6.15	236.8 ± 4.27	143.4 ± 3.10	99.2 ± 2.03	n = 14	[BCG] Beckman Coulter UniCel DxC 600
49.8 ± 3.94	232.1 ± 7.04	147.5 ± 3.32	109.2 ± 1.69	88.0 ± 4.13	n = 8	[BCH] Beckman Coulter UniCel DxC 800
54.0 ± 5.18	240.9 ± 6.07	145.1 ± 4.70	107.7 ± 4.69	88.9 ± 4.73	n = 19	[JJE] Ortho Vitros 250/350/950
51.6 ± 2.88	237.3 ± 8.71	144.9 ± 3.89	107.6 ± 2.49	85.5 ± 2.92	n = 18	[JJF] Ortho Vitros 5,1FS
50.6 ± 0.97	346.3 ± 5.44	217.7 ± 2.64	132.9 ± 2.90	94.2 ± 1.55	n = 16	[JJG] Ortho Vitros 5600
49.6 ± 1.02	342.1 ± 12.04	217.4 ± 7.08	132.3 ± 4.22	93.1 ± 2.86	n = 3	[ROC] Roche cobas c501
49.5 ± 0.57	333.8 ± 1.27	212.2 ± 1.27	129.5 ± 1.22	91.5 ± 1.22	n = 4	[ROS] Roche Cobas INTEGRA 400
49.7 ± 1.04	339.4 ± 7.36	213.3 ± 4.68	130.6 ± 2.83	92.4 ± 1.76	n = 28	[ROT] Roche Cobas INTEGRA 800
49.6 ± 1.36	351.5 ± 9.02	220.2 ± 6.30	133.8 ± 3.49	94.2 ± 2.53	n = 21	[ROD] Roche MODULAR D/P
48.5 ± 1.86	345.1 ± 12.75	215.6 ± 8.43	131.3 ± 5.97	92.1 ± 3.72	n = 3	[BYE] Siemens ADVIA 1800
43.6 ± 0.97	416.3 ± 6.19	263.1 ± 3.14	149.1 ± 2.35	92.9 ± 1.49	n = 11	[BYB] Siemens ADVIA 2400
43.6 ± 0.85	414.1 ± 5.79	262.8 ± 4.93	149.2 ± 2.12	93.0 ± 1.49	n = 21	[DUE] Siemens Dimension EXL
39.8 ± 0.91	386.8 ± 6.42	246.2 ± 4.53	138.5 ± 3.09	86.1 ± 1.69	n = 39	[DUR] Siemens Dimension RxL
43.8 ± 0.51	419.2 ± 4.95	265.3 ± 3.50	150.4 ± 2.02	93.8 ± 1.55	n = 16	[DUT] Siemens Dimension Vista
						[DUX] Siemens Dimension Xpand
47.8 ± 1.21	398.8 ± 5.77	245.2 ± 2.81	143.9 ± 1.65	93.9 ± 1.27	n = 19	<Reagents>
51.5 ± 1.22	374.5 ± 4.34	235.1 ± 3.07	142.4 ± 1.99	99.9 ± 1.02	n = 8	[AB1] Abbott
36.7 ± 2.23	307.4 ± 18.14	192.0 ± 11.22	113.1 ± 6.20	73.4 ± 4.16	n = 47	[BC1] Beckman Coulter
51.6 ± 1.19	374.3 ± 8.12	237.0 ± 5.49	143.1 ± 3.11	99.2 ± 2.55	n = 25	[OL1] Beckman Coulter AU Series
52.4 ± 4.38	238.3 ± 8.53	145.5 ± 4.24	108.0 ± 3.45	87.5 ± 4.44	n = 47	[BC2] Beckman Coulter IFCC Standardized
50.6 ± 0.89	345.9 ± 5.47	217.4 ± 2.88	132.9 ± 2.72	94.3 ± 1.43	n = 19	[JJ1] Ortho Clinical Diagnostics
49.7 ± 1.03	339.6 ± 7.23	213.4 ± 4.57	130.7 ± 2.76	92.4 ± 1.71	n = 29	[RO4] Roche cobas c311/c501/c502/c701
49.5 ± 0.74	334.5 ± 1.40	212.5 ± 1.52	129.7 ± 1.22	91.5 ± 1.12	n = 7	[RO2] Roche Hitachi and Modular D/P
49.3 ± 1.53	350.3 ± 10.37	219.3 ± 7.15	133.7 ± 3.91	94.0 ± 2.78	n = 25	[RO1] Roche Integra and MIRA
41.9 ± 2.36	403.9 ± 17.67	255.8 ± 10.61	144.7 ± 6.59	90.2 ± 4.26	n = 87	[BY1] Siemens ADVIA/ADVISIA Centaur
						[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Alkaline Phosphatase (U/L 37°C)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
370.9 ± 33.49	57.5 ± 10.95	165.3 ± 14.77	506.7 ± 43.98	119.4 ± 18.45	n = 357	[---] All Methods & Instruments
331.2 ± 13.15	54.0 ± 0.90	159.2 ± 2.36	464.7 ± 18.58	108.7 ± 1.37	n = 3	<Instruments>
382.4 ± 17.05	56.8 ± 2.57	174.4 ± 7.62	537.2 ± 23.07	118.8 ± 5.13	n = 21	[AXA] Abaxis Piccolo
340.0 ± 19.75	49.6 ± 3.04	154.9 ± 9.73	476.1 ± 26.58	105.6 ± 6.42	n = 55	[ABJ] Abbott Architect c System
341.1 ± 15.74	51.5 ± 2.98	154.5 ± 8.61	485.8 ± 33.36	105.0 ± 5.11	n = 4	[OLC] Beckman Coulter AU Chemistry System
341.7 ± 22.21	50.6 ± 4.19	155.0 ± 9.63	477.0 ± 30.25	107.3 ± 7.76	n = 4	[BCS] Beckman Coulter CX
335.5 ± 11.75	48.8 ± 2.05	153.6 ± 5.04	473.7 ± 16.82	104.4 ± 4.05	n = 16	[BCX] Beckman Coulter LX-20
344.0 ± 14.34	50.3 ± 1.95	155.9 ± 5.69	479.0 ± 16.75	106.6 ± 3.97	n = 15	[BCG] Beckman Coulter UniCel DxC 600
401.1 ± 11.99	78.7 ± 2.22	166.6 ± 6.48	519.4 ± 10.49	162.0 ± 4.13	n = 10	[BCH] Beckman Coulter UniCel DxC 800
394.4 ± 18.57	78.8 ± 3.55	160.3 ± 6.06	490.6 ± 24.10	158.4 ± 6.92	n = 20	[JJE] Ortho Vitros 250/350/950
388.9 ± 9.55	77.9 ± 2.46	158.0 ± 5.44	477.4 ± 15.68	156.7 ± 4.65	n = 18	[JJF] Ortho Vitros 5,1FS
372.6 ± 10.79	54.7 ± 1.27	169.4 ± 3.23	515.3 ± 9.73	116.2 ± 2.46	n = 16	[JJG] Ortho Vitros 5600
360.6 ± 2.56	51.3 ± 0.51	162.0 ± 1.80	501.6 ± 1.02	111.0 ± 0.90	n = 3	[ROC] Roche cobas c501
368.3 ± 8.94	52.9 ± 1.13	166.6 ± 4.04	516.6 ± 11.32	114.6 ± 3.30	n = 6	[ROH] Roche cobas c701
375.0 ± 8.89	53.0 ± 1.14	170.2 ± 3.10	529.2 ± 13.13	115.4 ± 1.90	n = 4	[ROS] Roche Cobas INTEGRA 400
355.8 ± 7.45	52.5 ± 1.50	163.1 ± 2.91	490.0 ± 11.71	113.0 ± 2.80	n = 29	[ROT] Roche MODULAR D/P
398.3 ± 11.88	57.6 ± 1.91	180.4 ± 5.14	558.5 ± 18.03	123.6 ± 3.55	n = 22	[ROD] Roche MODULAR D/P
390.8 ± 10.49	57.7 ± 5.86	176.8 ± 8.71	545.6 ± 13.58	121.0 ± 6.42	n = 3	[BYE] Siemens ADVIA 1800
433.0 ± 31.64	76.0 ± 5.63	206.2 ± 11.63	601.1 ± 27.97	142.3 ± 6.80	n = 14	[BYB] Siemens ADVIA 2400
423.2 ± 18.84	71.9 ± 7.66	198.6 ± 9.10	588.5 ± 24.59	139.6 ± 10.44	n = 24	[DUE] Siemens Dimension EXL
355.3 ± 17.98	54.2 ± 2.76	160.8 ± 7.78	492.5 ± 24.55	110.4 ± 5.86	n = 40	[DUR] Siemens Dimension RxL
401.9 ± 20.36	67.1 ± 5.91	186.6 ± 9.55	555.2 ± 21.54	132.5 ± 7.16	n = 19	[DUT] Siemens Dimension Vista
331.2 ± 13.15	54.0 ± 0.90	159.2 ± 2.36	464.7 ± 18.58	108.7 ± 1.37	n = 3	[DUX] Siemens Dimension Xpand
382.4 ± 17.05	56.8 ± 2.57	174.4 ± 7.62	537.2 ± 23.07	118.8 ± 5.13	n = 21	<Reagents>
338.1 ± 13.41	49.5 ± 2.11	153.9 ± 5.64	474.3 ± 17.88	105.1 ± 3.90	n = 37	[AX1] Abaxis
339.5 ± 18.99	49.5 ± 2.93	154.7 ± 9.47	475.2 ± 25.87	105.5 ± 6.20	n = 54	[AB1] Abbott
392.5 ± 15.41	78.4 ± 2.84	160.3 ± 7.04	488.9 ± 25.45	158.3 ± 6.17	n = 49	[BC1] Beckman Coulter
370.7 ± 10.74	54.4 ± 1.63	168.5 ± 3.89	512.8 ± 10.40	115.5 ± 2.88	n = 21	[OL1] Beckman Coulter AU Series
355.8 ± 7.24	52.5 ± 1.52	163.0 ± 2.97	489.9 ± 11.36	113.0 ± 2.72	n = 30	[RO4] Roche cobas c311/c501/c502/c701
370.7 ± 9.07	52.9 ± 1.07	168.0 ± 4.01	520.6 ± 12.80	114.8 ± 2.55	n = 11	[RO2] Roche Hitachi and Modular D/P
396.4 ± 13.32	57.4 ± 2.54	179.5 ± 6.50	555.4 ± 19.90	123.0 ± 4.51	n = 27	[RO1] Roche Integra and MIRA
392.3 ± 40.92	63.9 ± 11.06	180.9 ± 22.68	544.2 ± 57.24	126.7 ± 17.18	n = 97	[BY1] Siemens ADVIA/ADVIA Centaur
						[DAS5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

 $\gamma$ -Glutamyltransferase (U/L 37°C)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
52.4 ± 12.67	132.5 ± 30.11	71.8 ± 17.30	152.9 ± 35.92	34.9 ± 8.07	n = 302	[---] All Methods & Instruments
51.7 ± 2.72	134.3 ± 6.96	71.9 ± 4.05	154.0 ± 8.10	35.0 ± 2.03	n = 17	<Instruments>
39.7 ± 2.00	101.6 ± 5.05	54.9 ± 2.76	115.8 ± 4.91	27.5 ± 1.54	n = 47	[ABJ] Abbott Architect c System
41.5 ± 4.08	111.0 ± 18.74	59.7 ± 9.53	128.6 ± 21.16	28.2 ± 1.96	n = 4	[OLC] Beckman Coulter AU Chemistry System
46.0 ± 1.50	130.4 ± 3.55	68.3 ± 2.47	150.4 ± 4.54	30.2 ± 1.96	n = 4	[BCS] Beckman Coulter CX
48.1 ± 1.74	133.7 ± 2.92	72.0 ± 1.73	154.2 ± 3.24	30.4 ± 1.40	n = 15	[BCX] Beckman Coulter LX-20
47.3 ± 1.55	132.5 ± 2.59	70.6 ± 1.78	152.7 ± 3.65	30.9 ± 1.18	n = 13	[BCG] Beckman Coulter UniCel DxC 600
74.7 ± 2.14	214.8 ± 5.99	118.6 ± 2.83	251.2 ± 6.41	48.8 ± 2.57	n = 7	[BCH] Beckman Coulter UniCel DxC 800
74.0 ± 2.34	211.6 ± 5.52	116.9 ± 3.59	243.8 ± 7.43	48.8 ± 1.89	n = 18	[JJE] Ortho Vitros 250/350/950
73.8 ± 2.53	211.7 ± 6.35	116.9 ± 3.55	245.2 ± 6.43	49.1 ± 1.96	n = 18	[JJF] Ortho Vitros 5,1FS
43.2 ± 1.04	112.5 ± 2.03	60.3 ± 0.85	129.1 ± 2.39	29.0 ± 0.47	n = 14	[JJG] Ortho Vitros 5600
42.3 ± 0.90	112.6 ± 2.31	60.2 ± 1.27	130.0 ± 3.16	28.3 ± 0.90	n = 4	[ROC] Roche cobas c501
42.8 ± 1.46	111.4 ± 1.64	59.7 ± 1.51	129.1 ± 2.33	28.0 ± 0.75	n = 4	[ROS] Roche Cobas INTEGRA 400
43.4 ± 1.23	115.6 ± 3.48	61.6 ± 1.92	132.5 ± 4.24	29.1 ± 1.25	n = 28	[ROT] Roche Cobas INTEGRA 800
48.8 ± 1.68	127.7 ± 2.86	68.2 ± 2.56	145.3 ± 4.21	33.1 ± 1.51	n = 22	[ROD] Roche MODULAR D/P
47.5 ± 1.86	124.9 ± 4.38	66.0 ± 1.80	142.9 ± 4.38	32.7 ± 1.37	n = 3	[BYE] Siemens ADVIA 1800
62.9 ± 1.11	147.5 ± 2.04	81.6 ± 2.15	171.2 ± 2.88	42.7 ± 1.16	n = 11	[BYB] Siemens ADVIA 2400
62.2 ± 2.87	146.4 ± 3.07	81.5 ± 3.23	169.4 ± 4.04	42.5 ± 2.18	n = 17	[DUE] Siemens Dimension EXL
62.0 ± 2.02	152.4 ± 3.28	82.3 ± 2.36	176.9 ± 3.47	39.4 ± 2.06	n = 37	[DUR] Siemens Dimension RxL
64.2 ± 1.80	148.8 ± 2.59	82.7 ± 1.66	171.7 ± 2.54	43.5 ± 1.23	n = 10	[DUT] Siemens Dimension Vista
51.4 ± 2.50	133.7 ± 6.46	71.6 ± 3.87	153.3 ± 7.64	34.8 ± 1.86	n = 16	[DUX] Siemens Dimension Xpand
47.4 ± 1.85	132.7 ± 3.28	71.0 ± 2.30	153.0 ± 4.05	30.5 ± 1.39	n = 34	<Reagents>
39.7 ± 2.03	101.6 ± 5.12	54.9 ± 2.80	115.8 ± 4.98	27.5 ± 1.56	n = 46	[BC1] Beckman Coulter
74.1 ± 2.41	212.1 ± 6.04	117.2 ± 3.52	245.6 ± 7.33	48.9 ± 2.02	n = 43	[OL1] Beckman Coulter AU Series
43.1 ± 0.93	112.4 ± 2.19	60.1 ± 1.04	128.8 ± 2.54	29.0 ± 0.45	n = 18	[JJ1] Ortho Clinical Diagnostics
43.4 ± 1.23	115.6 ± 3.48	61.6 ± 1.92	132.5 ± 4.24	29.1 ± 1.25	n = 28	[RO4] Roche cobas c311/c501/c502/c701
42.5 ± 1.22	111.9 ± 2.04	60.0 ± 1.38	129.4 ± 2.79	28.1 ± 0.85	n = 8	[RO2] Roche Hitachi and Modular D/P
48.6 ± 1.85	126.9 ± 3.65	67.7 ± 2.58	144.7 ± 4.40	33.2 ± 1.60	n = 27	[RO1] Roche Integra and MIRA
62.5 ± 2.15	149.6 ± 3.98	82.1 ± 2.46	173.4 ± 4.81	41.2 ± 2.59	n = 70	[BY1] Siemens ADVIA/ADVISIA Centaur
63.1 ± 2.80	151.6 ± 5.09	82.5 ± 2.90	175.5 ± 4.65	41.5 ± 2.90	n = 5	[DA5] Siemens Dimension
						[DA8] Siemens Dimension IFCC Standardized

## Summary of Participant Performance (Mean and Standard Deviation)

## Creatine Kinase (U/L 37°C)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
248.8 ± 18.72	56.2 ± 5.56	39.1 ± 4.16	267.7 ± 18.44	303.5 ± 22.51	n = 326	[---] All Methods & Instruments
260.9 ± 8.71	61.5 ± 2.16	42.4 ± 1.51	280.3 ± 7.15	321.4 ± 11.98	n = 21	<Instruments>
221.3 ± 10.79	48.2 ± 2.46	33.2 ± 1.98	240.9 ± 11.44	271.4 ± 14.46	n = 52	[ABJ] Abbott Architect c System
231.4 ± 11.69	57.9 ± 7.54	38.9 ± 6.39	265.5 ± 10.51	286.6 ± 15.90	n = 4	[OLC] Beckman Coulter AU Chemistry System
253.0 ± 7.16	60.1 ± 1.88	41.5 ± 1.22	279.4 ± 5.16	310.5 ± 10.31	n = 4	[BCS] Beckman Coulter CX
250.4 ± 6.39	60.9 ± 2.70	41.6 ± 1.47	277.4 ± 7.03	308.8 ± 11.94	n = 14	[BCX] Beckman Coulter LX-20
256.7 ± 9.66	61.9 ± 3.32	42.5 ± 2.45	285.2 ± 9.50	318.4 ± 14.27	n = 15	[BCG] Beckman Coulter UniCel DxC 600
274.9 ± 20.54	51.8 ± 4.61	40.5 ± 3.18	279.2 ± 22.83	317.7 ± 16.52	n = 7	[BCH] Beckman Coulter UniCel DxC 800
276.7 ± 17.03	53.3 ± 2.62	41.0 ± 2.83	287.4 ± 18.31	319.3 ± 23.61	n = 18	[JJE] Ortho Vitros 250/350/950
272.9 ± 18.16	54.0 ± 1.97	41.4 ± 2.51	279.7 ± 12.45	312.0 ± 17.39	n = 18	[JJF] Ortho Vitros 5,1FS
261.0 ± 12.73	55.3 ± 2.28	37.4 ± 1.73	283.7 ± 11.55	316.5 ± 17.36	n = 16	[JJG] Ortho Vitros 5600
254.0 ± 0.00	60.0 ± 0.90	41.7 ± 0.51	269.7 ± 1.37	307.0 ± 3.58	n = 3	[ROC] Roche cobas c501
264.1 ± 9.27	54.2 ± 3.80	37.5 ± 4.86	284.7 ± 6.10	320.9 ± 7.89	n = 4	[ROH] Roche cobas c701
252.7 ± 6.55	61.2 ± 1.73	41.9 ± 1.65	269.3 ± 6.80	310.9 ± 9.54	n = 29	[ROT] Roche Cobas INTEGRA 800
243.2 ± 9.23	52.0 ± 1.52	34.9 ± 1.00	255.9 ± 10.46	289.1 ± 14.25	n = 21	[ROD] Roche MODULAR D/P
237.6 ± 6.45	51.1 ± 2.05	34.3 ± 1.37	254.3 ± 5.97	277.7 ± 17.68	n = 3	[BYE] Siemens ADVIA 1800
249.8 ± 10.47	57.7 ± 2.07	39.3 ± 1.88	268.3 ± 10.25	309.1 ± 17.97	n = 13	[BYB] Siemens ADVIA 2400
248.2 ± 6.59	58.2 ± 2.35	39.0 ± 2.70	269.5 ± 6.96	307.6 ± 8.28	n = 23	[DUE] Siemens Dimension EXL
247.2 ± 7.42	58.9 ± 1.99	40.2 ± 1.90	264.7 ± 6.89	306.3 ± 8.10	n = 39	[DUR] Siemens Dimension RxL
250.7 ± 4.61	56.9 ± 2.26	38.5 ± 2.40	267.0 ± 5.69	306.4 ± 6.99	n = 15	[DUT] Siemens Dimension Vista
						[DUX] Siemens Dimension Xpand
						<Reagents>
260.9 ± 8.71	61.5 ± 2.16	42.4 ± 1.51	280.3 ± 7.15	321.4 ± 11.98	n = 21	[AB1] Abbott
251.8 ± 10.63	61.1 ± 2.90	41.8 ± 2.24	279.6 ± 9.54	310.7 ± 15.71	n = 37	[BC1] Beckman Coulter
221.6 ± 10.12	48.3 ± 2.34	33.3 ± 1.87	241.2 ± 10.73	271.9 ± 13.47	n = 49	[OL1] Beckman Coulter AU Series
274.9 ± 18.21	53.6 ± 2.63	41.1 ± 2.79	283.3 ± 17.37	315.9 ± 20.76	n = 43	[JJ1] Ortho Clinical Diagnostics
260.1 ± 11.65	56.1 ± 2.71	38.3 ± 2.37	281.5 ± 12.41	315.4 ± 15.96	n = 20	[RO4] Roche cobas c311/c501/c502/c701
252.9 ± 6.62	61.3 ± 1.67	42.0 ± 1.67	269.7 ± 6.64	311.2 ± 9.59	n = 28	[RO2] Roche Hitachi and Modular D/P
263.3 ± 6.57	56.5 ± 4.71	40.8 ± 2.85	284.9 ± 4.69	318.4 ± 6.63	n = 6	[RO1] Roche Integra and MIRA
241.7 ± 9.03	51.7 ± 1.73	34.7 ± 1.12	256.0 ± 10.49	285.6 ± 17.04	n = 26	[BY1] Siemens ADVIA/ADVISIA Centaur
247.2 ± 9.40	58.0 ± 2.41	39.2 ± 2.72	262.7 ± 15.31	302.3 ± 19.03	n = 24	[DA5] Siemens Dimension
248.8 ± 6.53	58.3 ± 2.20	39.7 ± 2.17	266.8 ± 6.49	306.9 ± 8.77	n = 64	[DA8] Siemens Dimension IFCC Standardized

## Summary of Participant Performance (Mean and Standard Deviation)

## Creatine Kinase-MB (ng/mL, U/L 37°C, %)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
37.44 ± 5.02	0.55 ± 0.29	0.48 ± 0.31	0.44 ± 0.22	50.42 ± 6.83	n = 204	[-A-] All Methods - Results reported in ng/mL
37.20 ± 2.58	0.44 ± 0.09	0.29 ± 0.06	0.36 ± 0.06	50.03 ± 3.21	n = 23	[AB1] Abbott
43.89 ± 2.63	0.56 ± 0.06	0.40 ± 0.00	0.50 ± 0.00	59.20 ± 3.17	n = 14	[SAA] Beckman Coulter ACCESS
44.36 ± 2.28	0.58 ± 0.06	0.40 ± 0.00	0.50 ± 0.00	59.51 ± 2.93	n = 14	[BC1] Beckman Coulter UniCel
27.88 ± 4.90	1.00 ± 0.00	1.00 ± 0.00	1.00 ± 0.00	33.34 ± 4.48	n = 5	[BS1] Biosite Diagnostics
30.58 ± 2.16	0.32 ± 0.06	0.20 ± 0.00	0.23 ± 0.05	39.67 ± 2.92	n = 26	[JJ1] Ortho Clinical Diagnostics
39.73 ± 1.96	0.89 ± 0.11	0.88 ± 0.16	0.63 ± 0.09	53.09 ± 2.64	n = 26	[RO3] Roche Elecsys/Modular E/e601/e411
37.23 ± 1.85	0.21 ± 0.07	0.20 ± 0.08	0.19 ± 0.07	48.11 ± 2.47	n = 33	[COB] Siemens ADVIA Centaur
39.34 ± 2.84	0.50 ± 0.25	0.48 ± 0.26	0.44 ± 0.22	54.11 ± 4.52	n = 29	[DA5] Siemens Dimension
33.32 ± 1.09	0.74 ± 0.15	0.75 ± 0.18	0.50 ± 0.00	47.81 ± 1.40	n = 27	[DA6] Siemens Dimension LOCI
49.70 ± 3.31	0.75 ± 0.06	0.65 ± 0.17	0.60 ± 0.11	67.10 ± 0.23	n = 2	[TO1] Tosoh
-----	-----	-----	-----	-----	-----	-----
36.90 ± 3.97	2.24 ± 1.08	2.24 ± 1.08	3.20 ± 2.44	37.88 ± 18.78	n = 4	[-B-] All Methods - Results reported in U/L
-----	-----	-----	-----	-----	-----	-----
17.33 ± 4.21	0.00 ± 0.00	0.00 ± 0.00	0.00 ± 0.00	19.82 ± 6.30	n = 4	[-P-] All Methods - Results reported as %
18.91 ± 4.38	0.00 ± 0.00	0.00 ± 0.00	0.00 ± 0.00	22.12 ± 6.58	n = 3	[HL1] Helena Laboratories

## Summary of Participant Performance (Mean and Standard Deviation)

## Lactate Dehydrogenase (U/L 37°C)

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
269.1 ± 26.80	113.5 ± 10.63	79.7 ± 7.55	262.3 ± 22.85	271.4 ± 27.23	n = 264	[-A-] All Methods - Lactate to Pyruvate
692.2 ± 25.38	323.0 ± 17.31	218.8 ± 14.64	686.2 ± 25.18	720.6 ± 23.31	n = 47	[-B-] All Methods - Pyruvate to Lactate
						<Instruments>
275.9 ± 7.57	120.0 ± 5.14	84.0 ± 4.04	274.3 ± 7.56	282.0 ± 7.23	n = 20	[ABJ] Abbott Architect c System
240.6 ± 9.54	103.3 ± 4.82	72.6 ± 4.24	238.4 ± 9.43	241.8 ± 9.68	n = 49	[OLC] Beckman Coulter AU Chemistry System
218.9 ± 4.93	94.2 ± 2.80	68.6 ± 3.21	220.2 ± 6.59	222.1 ± 5.64	n = 4	[BCS] Beckman Coulter CX
224.0 ± 5.99	96.7 ± 2.26	68.5 ± 2.32	223.4 ± 8.38	225.6 ± 8.87	n = 4	[BCX] Beckman Coulter LX-20
225.2 ± 4.24	96.0 ± 3.10	67.8 ± 3.20	224.4 ± 4.79	227.1 ± 5.16	n = 15	[BCG] Beckman Coulter UniCel DxC 600
228.1 ± 4.53	98.9 ± 3.03	69.8 ± 2.14	225.8 ± 5.93	228.5 ± 5.81	n = 15	[BCH] Beckman Coulter UniCel DxC 800
701.0 ± 17.83	326.2 ± 15.62	216.7 ± 14.95	701.7 ± 25.96	738.7 ± 23.73	n = 8	[JJE] Ortho Vitros 250/350/950
691.5 ± 24.98	321.2 ± 14.78	218.6 ± 16.23	685.8 ± 19.95	720.2 ± 17.48	n = 20	[JJF] Ortho Vitros 5,1FS
689.0 ± 25.60	323.1 ± 18.84	218.6 ± 11.60	681.3 ± 24.95	711.8 ± 22.62	n = 18	[JJG] Ortho Vitros 5600
279.4 ± 3.83	115.5 ± 2.99	81.3 ± 1.44	265.9 ± 6.79	280.8 ± 3.44	n = 16	[ROC] Roche cobas c501
287.2 ± 4.97	117.4 ± 2.31	83.5 ± 2.17	271.1 ± 7.00	284.8 ± 6.12	n = 4	[ROT] Roche Cobas INTEGRA 800
277.7 ± 5.05	115.8 ± 2.52	81.0 ± 1.98	267.5 ± 4.73	280.4 ± 5.17	n = 27	[ROD] Roche MODULAR D/P
280.3 ± 6.48	121.6 ± 4.08	85.2 ± 2.94	274.5 ± 7.48	284.9 ± 7.08	n = 22	[BYE] Siemens ADVIA 1800
271.5 ± 8.12	116.9 ± 2.86	82.5 ± 2.74	264.1 ± 7.17	275.5 ± 5.40	n = 3	[BYB] Siemens ADVIA 2400
285.0 ± 7.78	116.3 ± 3.57	83.5 ± 3.59	276.3 ± 7.48	287.2 ± 6.84	n = 12	[DUE] Siemens Dimension EXL
291.7 ± 7.99	121.4 ± 5.36	84.9 ± 4.73	281.0 ± 9.18	293.3 ± 9.49	n = 15	[DUR] Siemens Dimension RxL
289.9 ± 8.80	120.0 ± 6.31	83.8 ± 5.03	278.1 ± 6.57	292.0 ± 7.38	n = 40	[DUT] Siemens Dimension Vista
289.3 ± 6.43	120.3 ± 3.57	84.6 ± 2.09	275.0 ± 5.40	288.0 ± 2.59	n = 11	[DUX] Siemens Dimension Xpand
						<Reagents>
275.9 ± 7.57	120.0 ± 5.14	84.0 ± 4.04	274.3 ± 7.56	282.0 ± 7.23	n = 20	[AB1] Abbott
225.8 ± 5.48	97.0 ± 3.52	68.8 ± 2.68	224.2 ± 6.20	226.9 ± 6.29	n = 35	[BC1] Beckman Coulter
240.5 ± 9.44	103.2 ± 4.67	72.6 ± 4.08	238.3 ± 9.20	241.7 ± 9.51	n = 48	[OL1] Beckman Coulter AU Series
692.5 ± 24.24	322.8 ± 16.66	218.2 ± 14.07	686.7 ± 24.06	720.1 ± 22.42	n = 46	[JJ1] Ortho Clinical Diagnostics
279.5 ± 4.32	115.6 ± 2.47	81.2 ± 1.60	267.1 ± 6.47	281.0 ± 4.12	n = 20	[RO4] Roche cobas c311/c501/c502/c701
277.7 ± 5.05	115.8 ± 2.52	81.0 ± 1.98	267.5 ± 4.73	280.4 ± 5.17	n = 27	[RO2] Roche Hitachi and Modular D/P
286.8 ± 3.90	118.1 ± 2.03	84.3 ± 2.07	273.4 ± 6.39	286.0 ± 4.96	n = 6	[RO1] Roche Integra and MIRA
278.6 ± 7.70	120.4 ± 4.54	84.5 ± 3.23	273.0 ± 8.69	282.5 ± 8.84	n = 27	[BY1] Siemens ADVIA/ADVISIA Centaur
288.7 ± 6.43	116.1 ± 7.53	81.6 ± 5.69	273.0 ± 13.17	285.2 ± 16.00	n = 21	[DA5] Siemens Dimension
289.8 ± 9.20	120.7 ± 5.01	84.8 ± 3.85	278.6 ± 7.95	291.8 ± 8.41	n = 56	[DA8] Siemens Dimension IFCC Standardized

### **Summary of Participant Performance (Mean and Standard Deviation)**

### **LDH Isoenzyme 1 (%)**

Specimen: C76	Specimen: C77	Specimen: C78	Specimen: C79	Specimen: C80	Number	[Code] Instrument or Reagent System
58.5 ± 0.71	32.6 ± 3.38	30.8 ± 1.16	16.2 ± 3.91	58.0 ± 1.42	n = 8	[-P-] All Methods
58.2 ± 1.27	35.3 ± 2.65	32.5 ± 2.30	19.1 ± 3.44	56.6 ± 1.80	n = 4	<Instruments>
58.5 ± 0.57	30.3 ± 1.58	30.0 ± 0.75	13.6 ± 2.31	58.8 ± 0.41	n = 4	[HLS] Helena SPIFE [SEE] Sebia Electrophoresis
58.2 ± 1.27	35.3 ± 2.65	32.5 ± 2.30	19.1 ± 3.44	56.6 ± 1.80	n = 4	<Reagents>
58.5 ± 0.57	30.3 ± 1.58	30.0 ± 0.75	13.6 ± 2.31	58.8 ± 0.41	n = 4	[HL1] Helena Laboratories [SE1] Sebia