

---

# **Wadsworth Center**

NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*

## **TRACE ELEMENTS IN WHOLE BLOOD**

**Event #3, 2012**

**November 30<sup>th</sup>, 2012**

---

Nirav R. Shah, M.D., M.P.H.  
Commissioner

**NEW YORK**  
*state department of*  
**HEALTH**

Sue Kelly  
Executive Deputy Commissioner

November 30, 2012

**Trace Elements in Whole Blood  
Event #3, 2012**

Dear Laboratory Director:

Results from the third proficiency test (PT) event in 2012 for Trace Elements in Whole Blood have been tabulated and summarized. Target values for Arsenic, Cadmium, Mercury and Lead in whole blood have been established along with acceptable ranges. Results are graded using element-specific criteria as indicated in each narrative section. A laboratory with an unacceptable significant analytical bias relative to the target value will be expected to investigate the source of the error. A confidential three-digit code number assigned by the PT program identifies participant laboratories. The data for blood lead were previously reported in the Blood Lead PT Report issued October 18th, 2012, and are reproduced here for completeness.


**PT Materials**

Test materials for the third event were prepared from caprine (goat) whole blood obtained from animals dosed with lead acetate to create physiologically bound lead (Pb). A total of five blood pools were supplemented with arsenic (as inorganic As<sup>3+</sup>), cadmium (as Cd<sup>2+</sup>) and mercury as both inorganic (Hg<sup>2+</sup>) and organic (ethylmercury (CH<sub>3</sub>CH<sub>2</sub>Hg<sup>+</sup>) and methylmercury (CH<sub>3</sub>Hg<sup>+</sup>)) species. In addition to As, Cd, Pb and Hg, blood pools were supplemented with the trace elements manganese (Mn), thallium (Tl), tin (Sn), cobalt (Co), and chromium (Cr).

**The next PT event for trace elements in whole blood is scheduled to be mailed Wednesday, January 16th, 2013.** Please inform our laboratory staff at (518) 474-4484 if the test materials have not arrived within five days of the scheduled mail out date. **The deadline for reporting results is Wednesday, February 6th, 2013.**

Thank you for your participation in this event.

Yours sincerely,



Patrick J. Parsons, Ph.D.  
Chief, Laboratory of Inorganic and Nuclear Chemistry  
Deputy Director, Division of Environmental Health



Mary Frances Verostek, Ph.D.  
Assistant Section Head  
PT Program for Blood Lead /Trace Elements

**New York State Department of Health**  
**Event #3, 2012**

---

**Whole Blood Arsenic**

Test materials for arsenic were prepared from caprine (goat) whole blood preserved with K<sub>2</sub>EDTA anticoagulant. A total of five pools were supplemented with arsenic as inorganic As<sup>3+</sup>.

**The Target Value** assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E Statistical methods for use in proficiency testing by interlaboratory comparisons**. Values for whole blood arsenic range from 6.9 µg/L (0.09 µmol/L) to 72.4 µg/L (0.97 µmol/L).

**Acceptable range:** The acceptable range for arsenic is set at  $\pm 6$  µg/L or  $\pm 20\%$ , whichever is greater. Thus, it is fixed at  $\pm 6$  µg/L for concentrations below 30 µg/L.

**Discussion:** Based upon the above criteria, 93.3% of test results reported were judged as satisfactory, with two of the 21 laboratories (9.5%) reporting 2 or more of the 5 results outside the acceptable ranges.

**New York State Department of Health**  
**Blood Arsenic Test Results, 2012 Event #3**  
**ROBUST STATISTICAL SUMMARY**

---

**TARGET VALUE ASSIGNMENT AND STATISTICS**

	Results ( $\mu\text{g}/\text{L}$ whole blood)				
	BE12-11	BE12-12	BE12-13	BE12-14	BE12-15
<b>Robust Mean</b>	<b>17.2</b>	<b>6.9</b>	<b>19.9</b>	<b>42.7</b>	<b>72.4</b>
Robust Standard Deviation	2.3	1.8	2.2	3.9	7.0
Standard Uncertainty	0.6	0.5	0.6	1.1	1.9
RSD (%)	13.1	26.3	11.3	9.2	9.6
Acceptable Range:					
Upper Limit	23.2	12.9	25.9	51.2	86.9
Lower Limit	11.2	0.9	13.9	34.2	57.9

---

**notes:** Results reported as less than the method detection limit are excluded from statistical calculations.

**New York State Department of Health**  
**Blood Arsenic Test Results, 2012 Event #3**  
**PERFORMANCE OF PARTICIPATING LABORATORIES**

Lab Code	Method	Results ( $\mu\text{g/L}$ whole blood)					Info Only
		BE12-11	BE12-12	BE12-13	BE12-14	BE12-15	
	Target Values:	17.2	6.9	19.9	42.7	72.4	
103	DRC/CC-ICP-MS	17.2	6.2	20.0	45.5	78.7	Info
110	DRC/CC-ICP-MS	18.9	7.4	21.9	47.8	79.6	
114	ICP-MS	17.0	8.0	20.0	47.0	79.0	
147	ICP-MS	15.9	5.9	19.3	41.3	72.1	Info
156	ICP-MS	16.0	0.0 ↓	17.0	42.0	61.0	
159	ICP-MS	20.0	9.0	22.0	45.0	75.0	
164	ICP-MS	16.0	6.0	21.0	44.0	79.0	
179	ICP-MS	19.0	<12.0	20.0	44.0	72.0	
197	DRC/CC-ICP-MS	15.0	<10.0	17.0	40.0	70.0	
200	ICP-MS	17.6	8.5	22.2	46.2	76.9	Info
206	DRC/CC-ICP-MS	19.6	12.9	19.6	41.3	69.3	
208	ICP-MS	20.9	12.3	23.0	42.4	72.7	
293	DRC/CC-ICP-MS	14.6	5.5	17.6	39.7	68.7	Info
305	DRC/CC-ICP-MS	17.0	6.0	20.0	43.0	75.0	
312	DRC/CC-ICP-MS	16.0	7.3	19.6	45.9	72.1	
324	HR-ICP-MS	14.4	5.4	17.8	37.5	66.0	Info
325	ETAAS-Z	4.7 ↓	1.6	6.2 ↓	2.2 ↓	7.6 ↓	Info
339	HR-ICP-MS	16.4	6.0	18.9	43.3	77.1	Info
359	ICP-MS	17.2	11.4	19.9	38.6	68.8	
391	DRC/CC-ICP-MS	18.5	6.9	21.7	48.2	83.7	Info
469	ICP-MS	26.3 ↑	7.0	27.3 ↑	37.7	60.8	

Percent satisfactory results for all participants: 93.3 %

**notes:** ↑ reported outside upper limit  
↓ reported outside lower limit  
↖: Unacceptable result

**notes:** Results reported as less than the method detection limit are excluded from statistical calculations.  
**Info only:** results included for informational purposes only.

**New York State Department of Health**  
**Blood Arsenic Test Results, 2012 Event #3**  
**STATISTICAL SUMMARY BY METHOD**

	Results ( $\mu\text{g/L}$ whole blood)				
	BE12-11	BE12-12	BE12-13	BE12-14	BE12-15
<b>DRC/CC-ICP-MS</b>					
Number of Sample Measurements:	8	7	8	8	8
Mean:	17.1	7.5	19.7	43.9	74.6
Standard Deviation:	1.8	2.5	1.7	3.4	5.5
RSD (%):	10.7	33.5	8.7	7.7	7.4
<b>ETAAS-Z</b>					
Number of Sample Measurements:	1	1	1	0	1
Mean:	4.7	1.6	6.2		7.6
Standard Deviation:	?	?	?		?
RSD (%):	—	—	—	—	—
<b>HR-ICP-MS</b>					
Number of Sample Measurements:	2	2	2	2	2
Mean:	15.4	5.7	18.4	40.4	71.6
Standard Deviation:	1.4	0.4	0.8	4.1	7.8
RSD (%):	—	—	—	—	—
<b>ICP-MS</b>					
Number of Sample Measurements:	10	8	10	10	10
Mean:	18.6	8.5	21.2	42.8	71.7
Standard Deviation:	3.2	2.3	2.7	3.0	6.6
RSD (%):	17.3	27.6	13.0	7.1	9.1
<b>All Laboratories</b>					
Number of Sample Measurements:	21	18	21	20	21
Mean:	17.1	7.4	19.6	43.0	69.8
Standard Deviation:	3.9	2.7	3.9	3.3	15.4
RSD (%):	22.7	36.7	19.6	7.6	22.1

**notes:** ? Insufficient data for calculation.

A Standard Deviation displayed as 0.0 should be interpreted as <0.1

**New York State Department of Health**  
**Event #3, 2012**

---

**Whole Blood Cadmium**

Test materials for cadmium were prepared from caprine (goat) whole blood preserved with K<sub>2</sub>EDTA anticoagulant. A total of five blood pools were supplemented with different amounts of cadmium (as Cd<sup>2+</sup>).

**The Target Value** assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E Statistical methods for use in proficiency testing by interlaboratory comparisons**. Values for whole blood cadmium range from 2.4 µg/L (21 nmol/L) to 11.3 µg/L (101 nmol/L).

**Acceptable ranges** are based on the OSHA criteria of ±15%, or ±1 µg/L around the target value, whichever is greater. So, the range is fixed at ±1 µg/L for concentrations below 6.6 µg/L, where above 6.6 µg/L, it is ±15%.

**Discussion:** Based upon the above criteria, 98.6% of the results reported by all participants were satisfactory, with one of the 29 laboratories (3.4%) reporting 2 or more of the 5 results outside the acceptable ranges.

**New York State Department of Health**  
**Blood Cadmium Test Results, 2012 Event #3**  
**ROBUST STATISTICAL SUMMARY**

---

**TARGET VALUE ASSIGNMENT AND STATISTICS**

	Results ( $\mu\text{g/L}$ whole blood)				
	BE12-11	BE12-12	BE12-13	BE12-14	BE12-15
<b>Robust Mean</b>	<b>6.2</b>	<b>2.4</b>	<b>8.8</b>	<b>3.2</b>	<b>11.3</b>
Robust Standard Deviation	0.4	0.2	0.5	0.3	0.6
Standard Uncertainty	0.1	<0.1	0.1	0.1	0.1
RSD (%)	6.6	6.8	5.3	8.2	5.7
Acceptable Range:					
Upper Limit	7.2	3.4	10.1	4.2	13.0
Lower Limit	5.2	1.4	7.5	2.2	9.6

---

**notes:** Results reported as less than the method detection limit are excluded from statistical calculations.

**New York State Department of Health**  
**Blood Cadmium Test Results, 2012 Event #3**  
**PERFORMANCE OF PARTICIPATING LABORATORIES**

Lab Code	Method	Results ( $\mu\text{g/L}$ whole blood)					Info Only
		BE12-11	BE12-12	BE12-13	BE12-14	BE12-15	
	Target Values:	6.2	2.4	8.8	3.2	11.3	
103	DRC/CC-ICP-MS	6.6	2.4	9.2	3.3	12.1	Info
106	ICP-MS	6.1	2.2	8.7	3.0	11.2	Info
107	DRC/CC-ICP-MS	6.3	2.3	9.2	3.2	11.7	Info
109	ICP-MS	6.5	2.4	9.6	3.2	11.9	Info
110	ICP-MS	6.0	2.1	8.2	2.9	10.9	
114	ICP-MS	5.7	2.1	8.0	2.8	10.6	
116	ICP-MS	6.8	2.8	9.2	3.6	12.0	Info
147	ICP-MS	6.1	2.4	8.8	3.0	10.9	Info
156	ICP-MS	6.0	2.4	8.8	3.0	11.0	
159	ICP-MS	6.3	2.5	8.9	3.2	11.7	
164	ICP-MS	6.1	2.4	8.5	3.2	10.8	
179	ICP-MS	6.2	2.2	8.8	3.1	11.6	
197	DRC/CC-ICP-MS	5.8	2.1	8.8	2.9	10.3	
199	ICP-MS	5.7	2.3	8.1	3.0	10.6	Info
200	ICP-MS	5.8	2.3	9.0	3.2	11.8	Info
206	ICP-MS	5.9	2.5	9.0	3.2	11.1	
208	ICP-MS	6.6	2.5	9.0	3.7	11.1	
293	ICP-MS	6.5	2.3	9.2	3.2	10.8	Info
305	ICP-MS	6.1	2.3	8.8	3.1	11.5	
312	ICP-MS	6.0	2.9	8.4	3.6	11.1	
324	HR-ICP-MS	5.3	2.0	7.7	2.4	10.1	Info
339	HR-ICP-MS	6.2	2.5	8.6	3.2	10.9	Info
359	ICP-MS	5.9	2.4	8.1	3.1	10.5	
366	ETAAS-Z	6.6	2.6	8.9	3.2	11.2	Info
383	ETAAS-Z	7.5 ↑	3.0	9.1	4.3 ↑	12.1	
391	DRC/CC-ICP-MS	6.2	2.3	8.7	3.5	11.0	Info
401	ETAAS-Z	6.7	2.5	9.3	3.3	11.8	Info
410	ICP-MS	6.8	2.4	9.5	3.4	12.6	Info
469	ICP-MS	6.6	2.4	9.3	2.9	11.7	

Percent satisfactory results for all participants: 98.6 %

**notes:** ↑ reported outside upper limit  
↓ reported outside lower limit  
▼: Unacceptable result

**notes:** Results reported as less than the method detection limit are excluded from statistical calculations.  
**Info only:** results included for informational purposes only.

**New York State Department of Health**  
**Blood Cadmium Test Results, 2012 Event #3**  
**STATISTICAL SUMMARY BY METHOD**

	Results ( $\mu\text{g/L}$ whole blood)				
	BE12-11	BE12-12	BE12-13	BE12-14	BE12-15
<b>DRC/CC-ICP-MS</b>					
Number of Sample Measurements:	4	4	4	4	4
Mean:	6.2	2.3	9.0	3.2	11.3
Standard Deviation:	0.3	0.1	0.3	0.3	0.8
RSD (%):	5.3	5.5	2.9	7.8	7.0
<b>ETAAS-Z</b>					
Number of Sample Measurements:	3	3	3	3	3
Mean:	6.9	2.7	9.1	3.6	11.7
Standard Deviation:	0.5	0.3	0.2	0.6	0.5
RSD (%):	—	—	—	—	—
<b>HR-ICP-MS</b>					
Number of Sample Measurements:	2	2	2	2	2
Mean:	5.8	2.3	8.2	2.8	10.5
Standard Deviation:	0.6	0.4	0.6	0.6	0.6
RSD (%):	—	—	—	—	—
<b>ICP-MS</b>					
Number of Sample Measurements:	20	20	20	20	20
Mean:	6.2	2.4	8.8	3.2	11.3
Standard Deviation:	0.3	0.2	0.5	0.2	0.6
RSD (%):	5.5	8.2	5.3	7.7	4.9
<b>All Laboratories</b>					
Number of Sample Measurements:	29	29	29	29	29
Mean:	6.2	2.4	8.8	3.2	11.3
Standard Deviation:	0.4	0.2	0.5	0.3	0.6
RSD (%):	7.1	9.4	5.3	10.6	5.4

**notes:** ? Insufficient data for calculation.

A Standard Deviation displayed as 0.0 should be interpreted as <0.1

**New York State Department of Health**  
**Event #3, 2012**

---

**Whole Blood Mercury**

Test materials for mercury were prepared from caprine (goat) whole blood preserved with K<sub>2</sub>EDTA anticoagulant. A total of five pools were supplemented with different amounts of mercury as both inorganic (Hg<sup>2+</sup>) and organometallic (as both ethylmercury, CH<sub>3</sub>CH<sub>2</sub>Hg<sup>+</sup>, and methylmercury, CH<sub>3</sub>Hg<sup>+</sup>) species.

<b>Sample</b>	<b>Mercury species added</b>
BE12-11	Hg <sup>2+</sup> , CH <sub>3</sub> Hg <sup>+</sup> , and CH <sub>3</sub> CH <sub>2</sub> Hg <sup>+</sup>
BE12-12	Hg <sup>2+</sup> and CH <sub>3</sub> Hg <sup>+</sup>
BE12-13	Hg <sup>2+</sup> and CH <sub>3</sub> Hg <sup>+</sup>
BE12-14	Hg <sup>2+</sup> , CH <sub>3</sub> Hg <sup>+</sup> , and CH <sub>3</sub> CH <sub>2</sub> Hg <sup>+</sup>
BE12-15	Hg <sup>2+</sup> , CH <sub>3</sub> Hg <sup>+</sup> , and CH <sub>3</sub> CH <sub>2</sub> Hg <sup>+</sup>

**The Target Value** assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E Statistical methods for use in proficiency testing by interlaboratory comparisons**. Values for whole blood mercury range from 3.0 µg/L (15 nmol/L) to 43.6 µg/L (217 nmol/L).

**Acceptable ranges** were fixed at ±30%, or ±3 µg/L around the target value, whichever is greater. That is, the range is fixed at ±3 µg/L for concentrations below 10 µg/L, while above 10 µg/L, it is ±30%.

**Discussion:** Based on the above criteria, 97.2% of results reported by all participants were satisfactory, with one of the 29 laboratories (3.4%) reporting 2 or more of the 5 results outside the acceptable ranges.

**New York State Department of Health**  
**Blood Mercury Test Results, 2012 Event #3**  
**ROBUST STATISTICAL SUMMARY**

---

**TARGET VALUE ASSIGNMENT AND STATISTICS**

	Results ( $\mu\text{g}/\text{L}$ whole blood)				
	BE12-11	BE12-12	BE12-13	BE12-14	BE12-15
<b>Robust Mean</b>	<b>9.0</b>	<b>4.5</b>	<b>43.6</b>	<b>3.0</b>	<b>7.1</b>
Robust Standard Deviation	1.2	0.6	5.1	0.3	0.9
Standard Uncertainty	0.3	0.1	1.2	0.1	0.2
RSD (%)	13.8	13.5	11.6	11.0	12.0
Acceptable Range:					
Upper Limit	12.0	7.5	56.7	6.0	10.1
Lower Limit	6.0	1.5	30.5	0.0	4.1

---

**notes:** Results reported as less than the method detection limit are excluded from statistical calculations.

**New York State Department of Health**  
**Blood Mercury Test Results, 2012 Event #3**  
**PERFORMANCE OF PARTICIPATING LABORATORIES**

Lab Code	Method	Results ( $\mu\text{g/L}$ whole blood)					Info Only
		BE12-11	BE12-12	BE12-13	BE12-14	BE12-15	
	Target Values:	9.0	4.5	43.6	3.0	7.1	
103	DRC/CC-ICP-MS	9.0	4.5	42.3	2.8	7.2	Info
106	ICP-MS	9.4	4.7	45.3	3.0	7.6	Info
107	DRC/CC-ICP-MS	9.4	4.7	45.4	2.9	7.3	Info
109	ICP-MS	9.7	4.8	78.4 ↑	3.0	7.6	Info
110	ICP-MS	8.2	4.2	41.5	2.8	6.7	
114	ICP-MS	7.8	3.7	45.9	2.2	7.0	
116	ICP-MS	9.2	4.7	43.3	3.0	7.4	Info
147	ICP-MS	9.0	4.4	43.3	2.8	7.3	Info
156	ICP-MS	7.7	3.7	36.0	0.0	6.4	
159	ICP-MS	10.0	5.0	45.0	3.0	7.0	
164	ICP-MS	11.0	5.0	51.0	<4.0	8.0	
179	ICP-MS	8.0	4.0	45.0	3.0	7.0	
197	DRC/CC-ICP-MS	7.0	<5.0	39.0	<5.0	6.0	
199	ICP-MS	11.6	5.7	52.8	3.7	9.1	Info
200	ICP-MS	8.3	3.7	36.7	2.4	7.0	Info
206	ICP-MS	9.0	5.0	49.0	3.0	8.0	
208	ICP-MS	7.6	<5.0	41.9	<5.0	5.7	
293	ICP-MS	8.9	5.2	40.9	5.1	7.7	Info
305	ICP-MS	9.0	4.0	41.0	3.0	6.0	
312	ICP-MS	8.9	4.5	45.9	3.2	7.0	
324	AFS	11.0	6.0	49.3	3.8	8.1	Info
339	HR-ICP-MS	8.7	4.5	39.2	2.8	6.8	Info
359	ICP-MS	7.7	4.2	39.2	2.7	6.9	
366	ICP-MS	9.4	4.2	36.0	3.0	6.5	Info
391	CV-AAS	11.5	4.4	40.9	4.8	5.7	Info
401	CV-AAS	4.4 ↓	2.7	35.1	1.6	5.2	Info
410	ICP-MS	9.8	4.9	48.1	3.0	7.8	Info
453	CV-AAS	9.4	4.9	46.3	3.0	7.4	Info
469	ICP-MS	12.2 ↑	6.2	69.6 ↑	3.9	9.6	

Percent satisfactory results for all participants: 97.2 %

**notes:** ↑ reported outside upper limit  
↓ reported outside lower limit  
▼: Unacceptable result

**notes:** Results reported as less than the method detection limit are excluded from statistical calculations.  
**Info only:** results included for informational purposes only.

**New York State Department of Health**  
**Blood Mercury Test Results, 2012 Event #3**  
**STATISTICAL SUMMARY BY METHOD**

	Results ( $\mu\text{g/L}$ whole blood)				
	BE12-11	BE12-12	BE12-13	BE12-14	BE12-15
<b>AFS</b>					
Number of Sample Measurements:	1	1	1	1	1
Mean:	11.0	6.0	49.3	3.8	8.1
Standard Deviation:	?	?	?	?	?
RSD (%):	—	—	—	—	—
<b>CV-AAS</b>					
Number of Sample Measurements:	3	3	3	3	3
Mean:	8.4	4.0	40.8	3.1	6.1
Standard Deviation:	3.6	1.2	5.6	1.6	1.2
RSD (%):	—	—	—	—	—
<b>DRC/CC-ICP-MS</b>					
Number of Sample Measurements:	3	2	3	2	3
Mean:	8.5	4.6	42.2	2.9	6.8
Standard Deviation:	1.3	0.1	3.2	0.1	0.7
RSD (%):	—	—	—	—	—
<b>HR-ICP-MS</b>					
Number of Sample Measurements:	1	1	1	1	1
Mean:	8.7	4.5	39.2	2.8	6.8
Standard Deviation:	?	?	?	?	?
RSD (%):	—	—	—	—	—
<b>ICP-MS</b>					
Number of Sample Measurements:	21	20	21	18	21
Mean:	9.2	4.6	46.5	3.1	7.3
Standard Deviation:	1.3	0.7	10.3	0.6	0.9
RSD (%):	13.8	14.5	22.2	20.3	12.5
<b>All Laboratories</b>					
Number of Sample Measurements:	29	27	29	25	29
Mean:	9.1	4.6	45.3	3.1	7.1
Standard Deviation:	1.6	0.7	9.2	0.7	1.0
RSD (%):	17.2	16.1	20.4	23.4	13.4

**notes:** ? Insufficient data for calculation.

A Standard Deviation displayed as 0.0 should be interpreted as <0.1

**New York State Department of Health**  
**Event #3, 2012**

---

**Whole Blood Lead**

Test materials for lead were prepared from caprine (goat) whole blood obtained from animals dosed with lead acetate to create physiologically-bound Pb. Whole blood was collected into collection bags containing K<sub>2</sub>EDTA anticoagulant.

**Target values** were established as the mean of 23 measurements performed by 21 reference laboratories using ICP-MS, ETAAS and ASV methods. Values range from 4 µg/dL to 25 µg/dL. Among the reference group, imprecision (SD) varied from 0.6 - 1.3 µg/dL, increasing with Pb concentration.

**Acceptable ranges** are based on the CLIA '88 criteria (Federal Register Volume 57, Number 40, §§ 493.2 and 493.937, February 28, 1992). The criteria are set at ±10% or ±4 µg/dL, whichever is greater.

**Discussion** Based on the CLIA '88 criteria, 98.5% of results reported by all participants were judged as satisfactory, with one out of 94 participant laboratories (1.1%) reporting 2 or more of the 5 results outside the acceptable ranges.

**New York State Department of Health**  
**Blood Lead Test Results, 2012 Event #3**  
**PERFORMANCE OF PARTICIPATING LABORATORIES**

Lab Code	Method	Results ( $\mu\text{g}/\text{dL}$ whole blood)					Normalized Mean	Info Only
		BE12-11	BE12-12	BE12-13	BE12-14	BE12-15		
	Target values:	9	18	4	25	17		
103	DRC/CC-ICP-MS	9	18	4	25	17	1.00	
103	ASV-LeadCare	10	19	4	28	19	1.10	Info
104	ETAAS-Z	9	18	4	24	16	0.97	
106	ICP-MS	9	18	4	25	17	1.00	Info
107	ASV-LeadCare	9	18	4	30 ↑	20	1.13	Info
107	DRC/CC-ICP-MS	9	18	4	26	17	1.01	
109	ICP-MS	9	18	4	25	17	1.00	
109	ETAAS-Z	8	18	4	26	17	1.01	
109	ASV-LeadCare	9	17	4	26	16	0.98	Info
109	ASV-LeadCare	9	19	4	27	19	1.08	Info
110	ETAAS-Z	10	20	6	27	19	1.10	
110	ASV-LeadCare	8	17	4	24	16	0.95	Info
110	ASV-LeadCare	9	21	4	28	18	1.12	Info
110	ICP-MS	9	18	4	26	17	1.01	
112	ETAAS-Z	9	18	4	25	17	1.00	
114	ICP-MS	10	18	6	30 ↑	20	1.13	
116	ICP-MS	8	18	4	25	16	0.98	Info
121	ETAAS-Z	7	16	4	26	17	0.98	Info
123	ETAAS-Z	9	19	4	25	17	1.02	
126	ETAAS-Z	8	19	4	25	18	1.04	
131	ETAAS-Z	9	18	5	25	19	1.04	
143	ETAAS-Z	8	17	3	25	16	0.96	
144	ETAAS-Z	8	18	4	25	16	0.98	
146	ETAAS-Z	9	18	5	23	17	0.97	
147	ICP-MS	8	17	4	24	16	0.95	
150	ASV-LeadCare	8	16	4	26	16	0.96	
156	ICP-MS	8	16	4	24	16	0.93	
158	ICP-MS	9	18	5	26	18	1.03	
159	ICP-MS	8	17	4	24	16	0.95	

**notes:** ↑ reported value outside upper limit  
↓ reported value outside lower limit

Normalized mean: The average of each reported result divided by the corresponding target value. It measures bias.  
Info only: results included for informational purposes only.

**New York State Department of Health**  
**Blood Lead Test Results, 2012 Event #3**  
**PERFORMANCE OF PARTICIPATING LABORATORIES**

Lab Code	Method	Results ( $\mu\text{g}/\text{dL}$ whole blood)					Normalized Mean	Info Only
		BE12-11	BE12-12	BE12-13	BE12-14	BE12-15		
	Target values:	9	18	4	25	17		
160	ICP-MS	8	18	4	26	19	1.05	
164	ICP-MS	9	18	4	26	17	1.01	
166	ASV-3010	7	18	4	24	16	0.97	
168	ETAAS-Z	9	19	5	26	17	1.03	
179	ICP-MS	9	18	4	25	17	1.00	
197	ICP-MS	8	17	4	23	15	0.92	
198	ETAAS-Z	10	19	5	26	18	1.05	
199	ICP-MS	8	16	4	22	15	0.88	
200	ICP-MS	8	17	4	24	17	0.97	
204	ASV-3010	6	13 ↓	<2	20 ↓	13	0.76	
206	ICP-MS	8	16	6	24	16	0.93	
208	ETAAS-Z	9	18	4	23	17	0.97	
221	ETAAS-Z	9	20	5	29	19	1.13	
232	ASV-3010	8	17	4	27	16	0.99	
237	ETAAS-Z	8	17	4	24	16	0.95	
243	ASV-3010	7	16	4	24	15	0.91	
254	ETAAS-Z	7	15	3	21	14	0.83	
255	ETAAS-Z	8	18	4	26	17	1.01	
261	ETAAS-Z	8	16	4	22	15	0.88	
269	ETAAS-Z	7	15	3	22	14	0.85	
271	ETAAS-Z	9	18	4	25	16	0.98	
272	ETAAS-Z	8	18	4	25	16	0.98	
279	ETAAS-Z	8	17	4	27	16	0.99	
290	ICP-MS	9	19	5	26	16	1.01	
291	ASV-3010	9	17	6	25	18	1.00	
293	ICP-MS	8	17	5	25	17	0.98	
295	ASV-3010	5	17	2	24	15	0.93	
301	ETAAS Other	8	17	4	23	16	0.94	
305	ETAAS-Z	8	16	4	23	16	0.92	

**notes:** ↑ reported value outside upper limit  
 ↓ reported value outside lower limit

Normalized mean: The average of each reported result divided by the corresponding target value. It measures bias.  
 Info only: results included for informational purposes only.

**New York State Department of Health**  
**Blood Lead Test Results, 2012 Event #3**  
**PERFORMANCE OF PARTICIPATING LABORATORIES**

Lab Code	Method	Results ( $\mu\text{g}/\text{dL}$ whole blood)					Normalized Mean	Info Only
		BE12-11	BE12-12	BE12-13	BE12-14	BE12-15		
	Target values:	9	18	4	25	17		
312	ICP-MS	8	19	4	27	18	1.06	
317	ETAAS-Z	10	19	5	26	18	1.05	
324	HR-ICP-MS	9	19	5	27	18	1.06	
325	ETAAS-Z	9	18	5	25	17	1.00	
333	ETAAS-Z	8	17	5	24	16	0.95	
337	ASV-LeadCare	9	18	5	27	16	1.01	
339	HR-ICP-MS	8.1	16.9	4.1	24.1	15.7	0.95	Info
340	ETAAS-Z	9	20	4	26	18	1.07	
343	ASV-LeadCare	9	21	4	31 ↑	19	1.17	Info
348	ETAAS-Z	9	19	5	27	18	1.06	
349	ETAAS-Z	8	17	4	24	16	0.95	
350	ASV-3010	11	21	6	28	19	1.16	
352	ASV-3010	9	18	4	25	16	0.98	
353	ETAAS-Z	7	16	3	23	15	0.90	
365	ETAAS-Z	7	18	4	25	17	1.00	
366	ETAAS-Z	9	19	4	28	18	1.08	Info
368	ASV-3010	9	19	4	25	16	1.00	
369	ASV-3010	6	14	3	22	15	0.85	
374	ASV-3010	7	17	3	27	16	0.99	
383	ETAAS-Z	10	18	5	25	17	1.00	
384	ASV-3010	7	14	3	19 ↓	14	0.79	
388	ASV-3010	6	14	3	22	15	0.85	
389	ETAAS-Z	9	19	5	27	17	1.05	
391	ETAAS-Z	9	19	4	27	17	1.05	Info
393	ASV-LeadCare	11	21	5	28	19	1.16	
401	ETAAS-Z	9	20	5	29	17	1.09	Info
410	ICP-MS	9	18	4	25	17	1.00	Info
461	ASV-3010	7	15	2	22	14	0.85	
463	ASV-LeadCare	9	20	4	30 ↑	19	1.14	

**notes:** ↑ reported value outside upper limit  
↓ reported value outside lower limit

Normalized mean: The average of each reported result divided by the corresponding target value. It measures bias.  
Info only: results included for informational purposes only.

**New York State Department of Health**  
**Blood Lead Test Results, 2012 Event #3**  
**PERFORMANCE OF PARTICIPATING LABORATORIES**

Lab Code	Method	Results ( $\mu\text{g}/\text{dL}$ whole blood)					Normalized Mean	Info Only
		BE12-11	BE12-12	BE12-13	BE12-14	BE12-15		
	Target values:	9	18	4	25	17		
469	ICP-MS	9	18	5	26	17	1.01	
470	ASV-LeadCare	9	19	4	27	17	1.05	
473	ASV-LeadCare	8	19	4	27	20	1.10	
474	ASV-LeadCare	7	16	3	27	16	0.97	
475	ASV-LeadCare	9	20	5	28	19	1.12	
476	ASV-LeadCare	8	17	4	25	17	0.98	
477	ASV-LeadCare	10	18	4	27	18	1.05	

Percent satisfactory results for all participants: 98.5 %

**notes:** ↑ reported value outside upper limit  
 ↓ reported value outside lower limit

Normalized mean: The average of each reported result divided by the corresponding target value. It measures bias.  
 Info only: results included for informational purposes only.

**New York State Department of Health**  
**Blood Lead Test Results, 2012 Event #3**  
**STATISTICAL SUMMARY**

<b>Lab Code</b>	<b>Method</b>	<b>TARGET VALUE ASSIGNMENT AND STATISTICS</b>				
		<b>Results (<math>\mu\text{g/dL}</math> whole blood)</b>				
		<b>BE12-11</b>	<b>BE12-12</b>	<b>BE12-13</b>	<b>BE12-14</b>	<b>BE12-15</b>
103	DRC/CC-ICP-MS	9	18	4	25	17
104	ETAAS-Z	9	18	4	24	16
107	DRC/CC-ICP-MS	9	18	4	26	17
109	ICP-MS	9	18	4	25	17
109	ETAAS-Z	8	18	4	26	17
110	ETAAS-Z	10	20	6	27	19
110	ICP-MS	9	18	4	26	17
112	ETAAS-Z	9	18	4	25	17
147	ICP-MS	8	17	4	24	16
156	ICP-MS	8	16	4	24	16
159	ICP-MS	8	17	4	24	16
160	ICP-MS	8	18	4	26	19
164	ICP-MS	9	18	4	26	17
166	ASV-3010	7	18	4	24	16
179	ICP-MS	9	18	4	25	17
198	ETAAS-Z	10	19	5	26	18
199	ICP-MS	8	16	4	22	15
200	ICP-MS	8	17	4	24	17
243	ASV-3010	7	16	4	24	15
293	ICP-MS	8	17	5	25	17
324	HR-ICP-MS	9	19	5	27	18
325	ETAAS-Z	9	18	5	25	17
350	ASV-3010	11	21	6	28	19
Number of Sample Measurements:		23	23	23	23	23
<b>Mean (target value):</b>		<b>9</b>	<b>18</b>	<b>4</b>	<b>25</b>	<b>17</b>
Standard Deviation:		0.9	1.2	0.6	1.3	1.1
RSD (%):		10.8	6.6	14.9	5.3	6.5
Acceptable Range:						
Upper Limit:		13	22	8	29	21
Lower Limit:		5	14	0	21	13

**notes:** Results reported as less than the detection limits are treated as zero for statistical and grading purposes.

**New York State Department of Health**  
**Blood Lead Test Results, 2012 Event #3**  
**STATISTICAL SUMMARY BY METHOD**

	Results ( $\mu\text{g/dL}$ whole blood)				
	BE12-11	BE12-12	BE12-13	BE12-14	BE12-15
<b>ASV-3010</b>					
Number of Sample Measurements:	14	14	13	14	14
Mean:	7.4	16.4	3.7	23.9	15.6
Standard Deviation:	1.6	2.2	1.3	2.6	1.6
RSD (%):	21.6	13.7	33.9	11.0	10.0
<b>ASV-LeadCare</b>					
Number of Sample Measurements:	17	17	17	17	17
Mean:	8.9	18.6	4.1	27.4	17.9
Standard Deviation:	0.9	1.7	0.5	1.8	1.5
RSD (%):	10.4	8.9	11.8	6.5	8.4
<b>DRC/CC-ICP-MS</b>					
Number of Sample Measurements:	2	2	2	2	2
Mean:	9.0	18.0	4.0	25.5	17.0
Standard Deviation:	0.0	0.0	0.0	0.7	0.0
RSD (%):	—	—	—	—	—
<b>ETAAS Other</b>					
Number of Sample Measurements:	1	1	1	1	1
Mean:	8.0	17.0	4.0	23.0	16.0
Standard Deviation:	?	?	?	?	?
RSD (%):	—	—	—	—	—
<b>ETAAS-Z</b>					
Number of Sample Measurements:	37	37	37	37	37
Mean:	8.5	17.9	4.3	25.2	16.8
Standard Deviation:	0.9	1.3	0.7	1.8	1.2
RSD (%):	10.2	7.4	16.2	7.3	7.2
<b>HR-ICP-MS</b>					
Number of Sample Measurements:	2	2	2	2	2
Mean:	8.6	18.0	4.6	25.6	16.9
Standard Deviation:	0.6	1.5	0.6	2.1	1.6
RSD (%):	—	—	—	—	—
<b>ICP-MS</b>					
Number of Sample Measurements:	21	21	21	21	21
Mean:	8.5	17.6	4.4	25.1	16.9
Standard Deviation:	0.6	0.9	0.7	1.6	1.2
RSD (%):	7.1	5.0	15.3	6.4	7.1
<b>All Laboratories</b>					
Number of Sample Measurements:	94	94	93	94	94
Mean:	8.4	17.7	4.2	25.4	16.8
Standard Deviation:	1.0	1.6	0.8	2.1	1.4
RSD (%):	12.4	8.9	18.3	8.5	8.6

**notes:** ? Insufficient data for calculation.

A Standard Deviation displayed as 0.0 should be interpreted as <0.1 (see DRC/CC-ICP-MS and HR-ICP-MS participants)

**New York State Department of Health**  
**Blood Lead Test Results, 2012 Event #3**  
**STATISTICAL SUMMARY BY CLASS**

	Results ( $\mu\text{g}/\text{dL}$ whole blood)				
	BE12-11	BE12-12	BE12-13	BE12-14	BE12-15
<b>Evaluated</b>					
Number of Sample Measurements:	56	56	55	56	56
Mean:	8.3	17.5	4.1	25.1	16.6
Standard Deviation:	1.1	1.7	0.9	2.3	1.5
RSD (%):	13.7	9.7	21.5	9.1	9.3
<b>Info</b>					
Number of Sample Measurements:	15	15	15	15	15
Mean:	8.7	18.5	4.1	26.9	17.4
Standard Deviation:	0.7	1.5	0.3	2.1	1.3
RSD (%):	8.0	7.9	6.3	7.9	7.6
<b>Reference</b>					
Number of Sample Measurements:	23	23	23	23	23
Mean:	8.7	17.9	4.3	25.1	17.0
Standard Deviation:	0.9	1.2	0.6	1.3	1.1
RSD (%):	10.8	6.6	14.9	5.3	6.5
<b>All Laboratories</b>					
Number of Sample Measurements:	94	94	93	94	94
Mean:	8.4	17.7	4.2	25.4	16.8
Standard Deviation:	1.0	1.6	0.8	2.1	1.4
RSD (%):	12.4	8.9	18.3	8.5	8.6

---

**notes:** ? Insufficient data for calculation.

**New York State Department of Health**  
**Event #3, 2012**

---

**Additional Trace Elements Reported in Whole Blood**

Participant laboratories reported their analytical results for any additional trace elements (other than As, Cd, Hg and Pb) that are routinely reported so that a more complete characterization can be recorded for these proficiency test materials. Results for additional trace elements are reported here, but no target value is implied nor are any acceptable ranges provided. These data are provided solely for educational and informational purposes.

In addition to As, Cd, Pb and Hg, the whole blood pools were supplemented with additional trace elements as indicated below.

**Additional Elements**

Mn, Sn, Tl, Co, Cr

**New York State Department of Health**  
**Whole Blood Additional Elements, 2012 Event #3**  
**Page 1**

**Blood Aluminum ( $\mu\text{g/L}$ )**

<b>Lab Code</b>	<b>Method</b>	<b>BE12-11</b>	<b>BE12-12</b>	<b>BE12-13</b>	<b>BE12-14</b>	<b>BE12-15</b>
147	ICP-MS	< 10.8	< 10.8	< 10.8	< 10.8	< 10.8

**Blood Antimony ( $\mu\text{g/L}$ )**

<b>Lab Code</b>	<b>Method</b>	<b>BE12-11</b>	<b>BE12-12</b>	<b>BE12-13</b>	<b>BE12-14</b>	<b>BE12-15</b>
110	ICP-MS	< 0.10	< 0.10	< 0.10	0.15	< 0.10
147	ICP-MS	< 0.037	< 0.037	< 0.037	< 0.037	< 0.037
206	ICP-MS	<2.0	<2.0	<2.0	<2.0	<2.0

**Blood Barium ( $\mu\text{g/L}$ )**

<b>Lab Code</b>	<b>Method</b>	<b>BE12-11</b>	<b>BE12-12</b>	<b>BE12-13</b>	<b>BE12-14</b>	<b>BE12-15</b>
110	ICP-MS	9.9	17.9	16.4	29.2*	12.9
147	ICP-MS	8.7	14.6	12.5	20.6	11.3
197	ICP-MS	8.3	13.9	11.1	21.5	10.6
312	ICP-MS	7.9	13.7	11.6	21.2	10.1
<b>Arithmetic Mean</b>		<b>8.7</b>	<b>15.0</b>	<b>12.9</b>	<b>21.1</b>	<b>11.2</b>
<i>*Omitted</i>		SD	2.0	2.4	0.5	1.2
n		4	4	4	3	4

**Blood Beryllium ( $\mu\text{g/L}$ )**

<b>Lab Code</b>	<b>Method</b>	<b>BE12-11</b>	<b>BE12-12</b>	<b>BE12-13</b>	<b>BE12-14</b>	<b>BE12-15</b>
110	ICP-MS	< 0.14	< 0.14	< 0.14	< 0.14	< 0.14
147	ICP-MS	< 0.9	< 0.9	< 0.9	< 0.9	< 0.9
197	ICP-MS	<0.2	<0.2	<0.2	<0.2	<0.2

**Blood Bismuth ( $\mu\text{g/L}$ )**

<b>Lab Code</b>	<b>Method</b>	<b>BE12-11</b>	<b>BE12-12</b>	<b>BE12-13</b>	<b>BE12-14</b>	<b>BE12-15</b>
147	ICP-MS	< 0.02	< 0.02	< 0.02	< 0.02	< 0.02

**Blood Cesium ( $\mu\text{g/L}$ )**

<b>Lab Code</b>	<b>Method</b>	<b>BE12-11</b>	<b>BE12-12</b>	<b>BE12-13</b>	<b>BE12-14</b>	<b>BE12-15</b>
110	ICP-MS	0.51	0.59	0.51	0.62	0.67

**Blood Chromium ( $\mu\text{g/L}$ )**

<b>Lab Code</b>	<b>Method</b>	<b>BE12-11</b>	<b>BE12-12</b>	<b>BE12-13</b>	<b>BE12-14</b>	<b>BE12-15</b>
110	DRC/CC-ICP-MS	2.8	6.6	12.4	8.4	1.5
147	ICP-MS	2.7	6.2	12.5	8.6	2.2
159	DRC/CC-ICP-MS	2.4	5.7	11.3	7.9	1.5
164	ICP-MS	2.6	5.3	10.6	7.3	1.5
197	DRC/CC-ICP-MS	2.7	6.3	12.5	8.9	1.6
305	ICP-MS	2.3	4.8	11.0	7.6	0.9
312	DRC/CC-ICP-MS	2.9	7.1	13.2	9.5	2.0
391	DRC/CC-ICP-MS	14.8*	6.1	8.4	23.0*	4.5*
<b>Arithmetic mean</b>		2.6	6.0	11.5	8.3	1.6
<i>*Omitted</i>		SD	0.2	1.5	0.8	0.4
n		6	7	7	6	6

**New York State Department of Health**  
**Whole Blood Additional Elements, 2012 Event #3**  
**Page 2**

**Blood Cobalt ( $\mu\text{g/L}$ )**

Lab Code	Method	BE12-11	BE12-12	BE12-13	BE12-14	BE12-15
110	ICP-MS	3.7	4.0	11.8	10.7	1.5
147	ICP-MS	3.4	3.7	11.2	9.7	1.3
159	ICP-MS	3.8	4.1	12.0	10.9	1.5
164	ICP-MS	3.3	3.7	11.1	10.2	1.3
197	ICP-MS	3.2	3.7	10.7	9.9	1.2
206	ICP-MS	3.3	3.7	10.3	9.8	1.3
305	ICP-MS	3.6	3.8	11.3	10.1	1.4
312	ICP-MS	3.6	4.1	11.3	10.5	1.4
391	DRC/CC-ICP-MS	3.5	3.5	10.6	9.7	1.3
<b>Arithmetic mean</b>		<b>3.5</b>	<b>3.8</b>	<b>11.1</b>	<b>10.2</b>	<b>1.4</b>
SD		0.2	0.2	0.6	0.4	0.1
n		9	9	9	9	9

**Blood Copper ( $\mu\text{g/L}$ )**

Lab Code	Method	BE12-11	BE12-12	BE12-13	BE12-14	BE12-15
110	ICP-MS	1382	1334	1200	1175	1297
147	ICP-MS	1302	1245	1156	1112	1226
197	ICP-MS	1340.0	1330.0	1210.0	1160.0	1270.0
312	ICP-MS	1333.0	1349.0	1270.0	1257.0	1375.0
<b>Arithmetic mean</b>		<b>1339</b>	<b>1315</b>	<b>1209</b>	<b>1176</b>	<b>1292</b>
SD		33	47	47	60	63
n		4	4	4	4	4

**Blood Iodine ( $\mu\text{g/L}$ )**

Lab Code	Method	BE12-11	BE12-12	BE12-13	BE12-14	BE12-15
147	ICP-MS	42.2	33.5	47	72.9	49.1

**Blood Lithium ( $\mu\text{g/L}$ )**

Lab Code	Method	BE12-11	BE12-12	BE12-13	BE12-14	BE12-15
147	ICP-MS	4.3	2.2	5.2	4.2	5

**Blood Manganese ( $\mu\text{g/L}$ )**

Lab Code	Method	BE12-11	BE12-12	BE12-13	BE12-14	BE12-15
103	DRC/CC-ICP-MS	33.1	33.9	73.9	26.5	51.5
107	DRC/CC-ICP-MS	33.0	33.5	74.1	25.5	51.5
110	ETAAS-Z	33.7	35.0	74.8	26.0	51.0
147	ICP-MS	33.5	33	70.9	25.5	50.9
156	ICP-MS	28.0	30.0	67.0	21.0	46.0
179	ETAAS-Z	32.3	32.4	73.5	25.2	49.4
197	DRC/CC-ICP-MS	29.7	32.3	69.2	23.1	43.3
293	ICP-MS	31.0	31.0	70.6	23.0	49.1
305	ICP-MS	32.0	33.0	67.1	24.9	48.7
312	DRC/CC-ICP-MS	33.6	40.0	79.8	33.2*	58.8
391	DRC/CC-ICP-MS	27.3	26.1	55.3*	21.2	38.2
<b>Arithmetic mean</b>		<b>31.6</b>	<b>32.7</b>	<b>72.1</b>	<b>24.2</b>	<b>48.9</b>
*Omitted		2.3	3.4	3.9	2.0	5.2
n		11	11	10	10	11

**New York State Department of Health**  
**Whole Blood Additional Elements, 2012 Event #3**  
**Page 3**

<b>Blood Molybdenum (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE12-11</b>	<b>BE12-12</b>	<b>BE12-13</b>	<b>BE12-14</b>	<b>BE12-15</b>
147	ICP-MS	12.5	16.5	10.6	15.1	8.3
<b>Blood Nickel (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE12-11</b>	<b>BE12-12</b>	<b>BE12-13</b>	<b>BE12-14</b>	<b>BE12-15</b>
147	ICP-MS	0.65	0.7	0.89	0.59	0.6
<b>Blood Platinum (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE12-11</b>	<b>BE12-12</b>	<b>BE12-13</b>	<b>BE12-14</b>	<b>BE12-15</b>
110	ICP-MS	<0.10	<0.10	<0.10	<0.10	<0.10
<b>Blood Selenium (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE12-11</b>	<b>BE12-12</b>	<b>BE12-13</b>	<b>BE12-14</b>	<b>BE12-15</b>
107	DRC/CC-ICP-MS	240	348	269	273	294
109	ICP-MS	235	342	273	268	295
147	ICP-MS	235.00	326	262	254	292
197	ICP-MS	255.0	375.0	277.0	302.0	340.0
305	ICP-MS	244.0	356.0	269.0	272.0	316.0
312	ICP-MS	241.0	377.0	296.0	303.0	337.0
359	ICP-MS	245.0	323.0	267.0	266.0	284.0
391	DRC/CC-ICP-MS	275.5*	396.7	310.1	310.0	352.7
		<b>Arithmetic mean</b>	<b>242</b>	<b>355</b>	<b>278</b>	<b>281</b>
<i>*Omitted</i>		SD	7	26	17	21
		n	7	8	8	8
<b>Blood Tellurium (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE12-11</b>	<b>BE12-12</b>	<b>BE12-13</b>	<b>BE12-14</b>	<b>BE12-15</b>
147	ICP-MS	<0.06	<0.06	<0.06	<0.06	<0.06
<b>Blood Thorium (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE12-11</b>	<b>BE12-12</b>	<b>BE12-13</b>	<b>BE12-14</b>	<b>BE12-15</b>
147	ICP-MS	< 0.12	< 0.12	< 0.12	< 0.12	< 0.12
<b>Blood Thallium (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE12-11</b>	<b>BE12-12</b>	<b>BE12-13</b>	<b>BE12-14</b>	<b>BE12-15</b>
110	ICP-MS	4.2	2.3	6.9	3.9	15.0
147	ICP-MS	3.9	2.3	6.6	3.7	14.4
156	ICP-MS	4.2	2.2	7.0	4.0	15.0
159	ICP-MS	4.1	2.3	6.6	3.9	15.0
179	ICP-MS	4.0	2.0	7.0	4.0	15.0
197	ICP-MS	3.9	2.2	6.5	3.8	14.6
206	ICP-MS	3.8	2.2	6.7	3.8	14.6
305	ICP-MS	3.7	1.9	6.1	3.7	14.5
312	ICP-MS	4.1	2.4	7.2	4.2	16.3*
		<b>Arithmetic mean</b>	<b>4.0</b>	<b>2.2</b>	<b>6.7</b>	<b>3.9</b>
<i>*Omitted</i>		SD	0.2	0.2	0.3	0.2
		n	9	9	9	8

**New York State Department of Health**  
**Whole Blood Additional Elements, 2012 Event #3**  
**Page 4**

**Blood Tin ( $\mu\text{g}/\text{L}$ )**

<b>Lab Code</b>	<b>Method</b>	<b>BE12-11</b>	<b>BE12-12</b>	<b>BE12-13</b>	<b>BE12-14</b>	<b>BE12-15</b>
110	ICP-MS	2.1	3.1	12.1	0.5	19.8
147	ICP-MS	1.9	2.9	10	<0.12	19.2
156	ICP-MS	0.0	0.0	7.8	0.0	17.0
197	ICP-MS	<5.0	<5.0	10.3	<5.0	20.1

**Blood Uranium ( $\mu\text{g}/\text{L}$ )**

<b>Lab Code</b>	<b>Method</b>	<b>BE12-11</b>	<b>BE12-12</b>	<b>BE12-13</b>	<b>BE12-14</b>	<b>BE12-15</b>
103	DRC/CC-ICP-MS	0.0	0.0	0.0	0.0	0.0
110	ICP-MS	<0.02	<0.02	<0.02	<0.02	<0.02
147	ICP-MS	<0.007	<0.007	<0.007	<0.007	<0.007
312	ICP-MS	<0.1	<0.1	<0.1	<0.1	<0.1

**Blood Vanadium ( $\mu\text{g}/\text{L}$ )**

<b>Lab Code</b>	<b>Method</b>	<b>BE12-11</b>	<b>BE12-12</b>	<b>BE12-13</b>	<b>BE12-14</b>	<b>BE12-15</b>
147	ICP-MS	0.039	0.039	0.041	0.032	0.052

**Blood Zinc ( $\mu\text{g}/\text{L}$ )**

<b>Lab Code</b>	<b>Method</b>	<b>BE12-11</b>	<b>BE12-12</b>	<b>BE12-13</b>	<b>BE12-14</b>	<b>BE12-15</b>
110	ICP-MS	2784	1713	2024	2688	1963
147	ICP-MS	2830	1756	2085	2732	2085
197	ICP-MS	2620.0	1680.0	1910.0	2410.0	1870.0
312	ICP-MS	2986.0	1838.0	2325.0	3129.0	2148.0
<b>Arithmetic mean</b>		<b>2805</b>	<b>1747</b>	<b>2086</b>	<b>2740</b>	<b>2017</b>
SD		151	68	175	296	124
n		4	4	4	4	4

**New York State Department of Health**  
**Trace Elements in Whole Blood**  
**METHOD NOTES**

---

***ATOMIC SPECTROMETRY METHODS***

- A-1    ETAAS-Z (Electrothermal atomic absorption spectrometry with Zeeman background correction)
- A-2    ETAAS other (i.e., D<sub>2</sub>, S-H background correction)
- A-3    FAAS (Flame atomic absorption spectrometry)
- A-4    CV-AAS (Cold vapor atomic absorption spectrometry)
- A-5    HG-AAS (Hydride generation atomic absorption spectrometry)
- A-6    AFS (Atomic fluorescence spectrometry)
- A-7    Other

***INDUCTIVELY COUPLED PLASMA***

- P-1    ICP-MS (Inductively coupled plasma - mass spectrometry)
- P-2    DRC/CC-ICP-MS (ICP-MS used in the Dynamic Reaction Cell or Collision Cell mode)
- P-3    ICP-AES/OES (ICP atomic/optical emission spectrometry)
- P-4    HR-ICP-MS (High resolution ICP-MS)
- P-5    ETV-ICP-MS (Electrothermal vaporization ICP-MS)
- P-6    ID-ICP-MS (Isotope dilution ICP-MS)
- P-7    Other

***ELECTROCHEMICAL METHODS***

- E-1    ASV (Anodic stripping voltammetry without digestion)
- E-2    ASV-LeadCare® (Anodic stripping voltammetry using the ESA LeadCare® system)
- E-3    Fluoride specific electrode
- E-4    Other

***MOLECULAR FLUORIMETRY***

- F-1    EtOAc (Ethyl acetate-acetic acid extraction method for determination of erythrocyte protoporphyrin)
- F-2    Aviv hematofluorometry (for determination of EP at hematocrit 35)
- F-3    Helena ZPP (for determination of zinc protoporphyrin in µmol ZPP/mol heme)
- F-4    Other

***OTHER METHODS***

If your method is not listed in the above list, please describe it briefly.

---