
Wadsworth Center

New York State Department of Health

TRACE ELEMENTS IN SERUM

Event #3, 2010

November 22, 2010



STATE OF NEW YORK DEPARTMENT OF HEALTH

Wadsworth Center The Governor Nelson A. Rockefeller Empire State Plaza P.O. Box 509 Albany, New York 12201-0509

Richard F. Daines, M.D.
Commissioner

James W. Clyne, Jr.,
Executive Deputy Commissioner

November 22, 2010

Trace Elements in Serum Event #3, 2010

Dear Laboratory Director:

Results from the third proficiency test (PT) event for 2010 in the category Trace Elements in Serum have been tabulated and are summarized. Target values for Aluminum, Copper, Selenium and Zinc have been established along with acceptable ranges. Results are graded using element-specific criteria as indicated in each narrative section. A laboratory with an unacceptable significant analytical bias relative to the target value will be expected to investigate the source of the error. A confidential three-digit code number assigned by the PT program identifies participant laboratories.

PT Materials

Test materials were prepared from human serum obtained from Tennessee Blood Services, Inc. Serum units were spiked with a suite of additional trace elements as described in each narrative.

Assignment of Target Values for Trace Elements

Except for blood lead, we will implement robust statistics for assigning target values for all trace element panels. Method specific and additional trace element data will continue to be calculated utilizing traditional statistics. The use of robust statistics for assigning target values for proficiency testing pools is one approach that is acceptable under ISO 13528. In collaboration with other trace element PT scheme organizers, we have conducted an evaluation of robust statistics. As a result of our evaluation, we have elected to introduce this approach in our program.

The next PT event for trace elements in serum is scheduled to be mailed Wednesday, January 12th, 2011. Please inform our laboratory staff at (518) 473-0452 if the test materials have not arrived within five days of the scheduled mail out date. **The deadline for reporting results is Wednesday, February 9th, 2011.**

Thank you for your participation.

Sincerely,

Patrick J. Parsons, Ph.D.
Section Head, Trace Elements Proficiency Testing Program

New York State Department of Health
Event #3, 2010

Serum Aluminum

The test materials for serum Al were prepared from human serum obtained from Tennessee Blood Services, Inc. The units were tested by FDA approved methods and found to be Non-reactive for Anti-HIV-1/2, Anti-HCV 3.0 and HBsAg. The serum has also been found to STS (RPR) Non-reactive and Negative for HIV-1 and HCV by PCR. Serum units were dispensed into acid-washed 500-mL polypropylene containers to make up five (5) serum pools. Each pool was spiked with a suite of additional trace elements including aluminum as Al^{3+} at various concentrations.

The Target Value assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E** Statistical methods for use in proficiency testing by interlaboratory comparisons. Values for serum aluminum range from 26 $\mu\text{g/L}$ (0.96 $\mu\text{mol/L}$) to 138 $\mu\text{g/L}$ (5.11 $\mu\text{mol/L}$).

Acceptable ranges for serum aluminum are based on fixed criteria of $\pm 20\%$, or $\pm 5 \mu\text{g/L}$ below 25 $\mu\text{g/L}$. These criteria are based on consensus recommendations from several EQAS organizers (1).

Discussion. Based on the above criteria, 90.0% of test results reported were judged as satisfactory, with three out of 26 participant laboratories (11.5%) reporting 2 or more of the 5 results outside the acceptable ranges.

1. Taylor, A., Angerer, J., Claeys, F., Kristiansen, J., Mazarrasa, O., Menditto, A., Patriarca, M., Pineau, A., Schoeters, I., Sykes, C., Valkonen, S. and Weykamp, C. Comparison of procedures for evaluating laboratory performance in external quality assessment schemes for lead in blood and aluminum in serum demonstrates the need for common quality specifications. Clinical Chemistry 2002 **48** 2000-2007.

New York State Department of Health
Serum Aluminum Test Results, 2010 Event #3
ROBUST STATISTICAL SUMMARY

TARGET VALUE ASSIGNMENT AND STATISTICS

	Results ($\mu\text{g/L}$)				
	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
Robust Mean	138	46	99	54	26
Robust Standard Deviation	9.1	4.9	8.5	6.6	3.7
Standard Uncertainty	2.2	1.2	2.1	1.6	0.9
RSD (%)	6.6	10.6	8.6	12.2	14.5
Acceptable Range:					
Upper Limit:	166	55	119	65	31
Lower Limit:	110	37	79	43	21

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Serum Aluminum Test Results, 2010 Event #3
PERFORMANCE OF PARTICIPATING LABORATORIES

Lab Code	Method	Results (µg/L)					Info Only
		SE10-11	SE10-12	SE10-13	SE10-14	SE10-15	
Target Values:		138	46	99	54	26	
110	ETAAS-Z	134	45	97	52	27	Info
114	ETAAS-Z	210 ↑	55	120 ↑	64	33 ↑	
147	ETAAS-Z	136	48	103	56	25	
156	ICP-MS	155	47	106	55	24	
159	ETAAS-Z	136	48	96	57	28	
160	ETAAS-Z	131	52	97	62	31	Info
164	ICP-MS	143	45	98	49	22	
179	DRC/CC-ICP-MS	134	43	93	49	21	
197	ICP-MS	145	47	101	53	24	
200	DRC/CC-ICP-MS	133	41	89	45	23	
206	ICP-MS	147	53	103	67 ↑	30	Info
287	ETAAS-Z	131	42	96	51	23	
293	ICP-MS	129	41	91	47	24	
301	ETAAS-Z	136	46	100	57	27	
305	ICP-MS	136	44	100	50	24	
324	ICP-MS	122	42	89	49	24	Info
325	ETAAS-Z	430 ↑	70 ↑	150 ↑	83 ↑	36 ↑	Info
355	ICP-MS	148	47	106	56	26	Info
357	ICP-MS	130	41	90	55	23	
358	ICP-MS	143	48	105	54	26	
362	ICP-MS	155	56 ↑	111	61	31	
363	ICP-MS	133	44	98	51	24	
366	ETAAS-Z	143	52	105	63	29	Info
367	ETAAS-Z	136	40	91	43 ↓	19 ↓	Info
401	ICP-AES/OES	127	38	89	46	16 ↓	Info
458	ETAAS Other	132	43	84	50	29	

Percent satisfactory results for all participants: 90.0 %

notes: ↑ reported outside upper limit
↓ reported outside lower limit

Info only: results included for informational purposes only.

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Serum Aluminum Test Results, 2010 Event #3
STATISTICAL SUMMARY BY METHOD

	Results (µg/L)				
	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
DRC/CC-ICP-MS					
Number of Sample Measurements:	2	2	2	2	2
Mean:	134	42	91	47	22
Standard Deviation:	1	1	3	3	1
RSD (%):	—	—	—	—	—
ETAAS Other					
Number of Sample Measurements:	1	1	1	1	1
Mean:	132	43	84	50	29
Standard Deviation:	?	?	?	?	?
RSD (%):	—	—	—	—	—
ETAAS-Z					
Number of Sample Measurements:	9	10	10	10	10
Mean:	144	50	106	59	28
Standard Deviation:	25	8	18	11	5
RSD (%):	17.5	17.0	16.6	18.0	17.6
ICP-AES/OES					
Number of Sample Measurements:	1	1	1	1	1
Mean:	127	38	89	46	16
Standard Deviation:	?	?	?	?	?
RSD (%):	—	—	—	—	—
ICP-MS					
Number of Sample Measurements:	12	12	12	12	12
Mean:	141	46	100	54	25
Standard Deviation:	11	5	7	6	3
RSD (%):	7.5	9.9	7.0	10.4	10.8
All Laboratories					
Number of Sample Measurements:	25	26	26	26	26
Mean:	140	47	100	55	26
Standard Deviation:	17	7	13	8	4
RSD (%):	11.9	14.2	12.8	15.3	17.0

notes: ? Insufficient data for calculation.

New York State Department of Health
Event #3, 2010

Serum Copper

The test materials for serum Cu were prepared from human serum obtained from Tennessee Blood Services, Inc. The units were tested by FDA approved methods and found to be Non-reactive for Anti-HIV-1/2, Anti-HCV 3.0 and HBsAg. The serum has also been found to STS (RPR) Non-reactive and Negative for HIV-1 and HCV by PCR. Serum units were dispensed into acid-washed 500-mL polypropylene containers to make up five (5) serum pools. Each pool was spiked with a suite of additional trace elements including copper as Cu^{2+} at various concentrations.

The Target Value assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E** Statistical methods for use in proficiency testing by interlaboratory comparisons. Values for serum copper range from 1051 $\mu\text{g/L}$ (16.54 $\mu\text{mol/L}$) to 2836 $\mu\text{g/L}$ (44.63 $\mu\text{mol/L}$).

Acceptable ranges for serum copper are based on fixed criteria of $\pm 15\%$, or $\pm 95 \mu\text{g/L}$ below 635 $\mu\text{g/L}$. These criteria are consistent with those proposed by the OELM Network of EQAS organizers (1, 2) for trace elements in serum, and are slightly less stringent than those previously suggested for NYS ($\pm 10\%$).

Discussion. Based on the above criteria, 90.9% of test results reported were judged as satisfactory, with two of the 22 participant laboratories (9.1%) reporting 2 or more of the 5 results outside the acceptable ranges.

1. A. Taylor, J. Angerer, J. Arnaud, F. Claeys, R.L. Jones, O. Mazarrasa, E. Mairiaux, A. Menditto, P.J. Parsons, M. Patriarca, A. Pineau, S. Valkonen, J.-P. Weber and C. Weykamp Accreditation and Quality Assurance 2006 11 440-445.

2. J. Arnaud, J.-P. Weber, C.W. Weykamp, P.J. Parsons, J. Angerer, E. Mairiaux, O. Mazarrasa, S. Valkonen, A. Menditto, M. Patriarca, and A. Taylor Clinical Chemistry 2008 54 1892-1899.

**New York State Department of Health
Serum Copper Test Results, 2010 Event #3
ROBUST STATISTICAL SUMMARY**

TARGET VALUE ASSIGNMENT AND STATISTICS

Results ($\mu\text{g/L}$ serum)

	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
Robust Mean	1 0 5 1	2 8 3 6	1 3 0 1	1 6 4 8	2 1 0 9
Robust Standard Deviation	62	147	63	103	114
Standard Uncertainty	16	39	17	28	30
RSD (%)	5.9	5.2	4.9	6.3	5.4
Acceptable Range:					
Upper Limit:	1209	3261	1496	1895	2425
Lower Limit:	893	2411	1106	1401	1793

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Serum Copper Test Results, 2010 Event #3
PERFORMANCE OF PARTICIPATING LABORATORIES

Lab Code	Method	Results (µg/L serum)					Info Only
		SE10-11	SE10-12	SE10-13	SE10-14	SE10-15	
	Target Values:	1051	2836	1301	1648	2109	
107	DRC/CC-ICP-MS	1091	2905	1327	1631	2095	Info
110	ICP-MS	1115	2959	1377	1741	2178	
114	ICP-MS	1090	2960	1450	1790	2320	
147	ICP-MS	1055	2891	1328	1569	2166	Info
156	FAAS	1990 ↑	3780 ↑	2200 ↑	2580 ↑	2980 ↑	
159	ICP-AES/OES	1060	2740	1280	1640	2090	
160	ETAAS-Z	990	2710	1200	1560	1950	
164	ICP-MS	1075	2922	1284	1712	2110	
179	DRC/CC-ICP-MS	1050	2810	1290	1650	2110	
197	ICP-MS	1110	2920	1350	1720	2190	
200	FAAS	1048	2769	1270	1600	2038	Info
206	ICP-MS	1030	2750	1280	1520	1960	
293	ICP-MS	992	2817	1266	1647	2054	Info
305	ICP-MS	10 ↓	27 ↓	13 ↓	16 ↓	20 ↓	
324	ICP-MS	1053	2801	1298	1645	2089	Info
325	FAAS	1110	3120	1410	1770	2280	Info
360	FAAS	990	2740	1280	1650	2120	
362	ICP-MS	960	2700	1250	1530	2070	
366	ETAAS other	1047	2733	1294	1679	2094	Info
401	DRC/CC-ICP-MS	1100	2976	1336	1698	2232	Info
404	HR-ICP-MS	917	2557	1188	1454	1902	Info
457	ICP-AES/OES	1081	2965	1354	1706	2180	Info

Percent satisfactory results for all participants: 90.9 %

notes: ↑ reported outside upper limit
↓ reported outside lower limit

Info only: results included for informational purposes only.

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Serum Copper Test Results, 2010 Event #3
STATISTICAL SUMMARY BY METHOD

	Results ($\mu\text{g/L}$ serum)				
	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
DRC/CC-ICP-MS					
Number of Sample Measurements:	3	3	3	3	3
Mean:	1080	2897	1318	1660	2146
Standard Deviation:	27	83	24	35	75
RSD (%):	—	—	—	—	—
ETAAS other					
Number of Sample Measurements:	1	1	1	1	1
Mean:	1047	2733	1294	1679	2094
Standard Deviation:	?	?	?	?	?
RSD (%):	—	—	—	—	—
ETAAS-Z					
Number of Sample Measurements:	1	1	1	1	1
Mean:	990	2710	1200	1560	1950
Standard Deviation:	?	?	?	?	?
RSD (%):	—	—	—	—	—
FAAS					
Number of Sample Measurements:	4	4	4	4	4
Mean:	1285	3102	1540	1900	2355
Standard Deviation:	473	484	445	459	429
RSD (%):	36.8	15.6	28.9	24.2	18.2
HR-ICP-MS					
Number of Sample Measurements:	1	1	1	1	1
Mean:	917	2557	1188	1454	1902
Standard Deviation:	?	?	?	?	?
RSD (%):	—	—	—	—	—
ICP-AES/OES					
Number of Sample Measurements:	2	2	2	2	2
Mean:	1071	2853	1317	1673	2135
Standard Deviation:	15	159	52	47	64
RSD (%):	—	—	—	—	—
ICP-MS					
Number of Sample Measurements:	9	9	9	9	9
Mean:	1053	2858	1320	1653	2126
Standard Deviation:	52	94	64	96	102
RSD (%):	5.0	3.3	4.8	5.8	4.8
All Laboratories					
Number of Sample Measurements:	21	21	21	21	21
Mean:	1093	2882	1348	1690	2153
Standard Deviation:	212	242	205	221	215
RSD (%):	19.4	8.4	15.2	13.1	10.0

notes: ? Insufficient data for calculation.

New York State Department of Health
Event #3, 2010

Serum Selenium

The test materials for serum Se were prepared from human serum obtained from Tennessee Blood Services, Inc. The units were tested by FDA approved methods and found to be Non-reactive for Anti-HIV-1/2, Anti-HCV 3.0 and HBsAg. The serum has also been found to STS (RPR) Non-reactive and Negative for HIV-1 and HCV by PCR. Serum units were dispensed into acid-washed 500-mL polypropylene containers to make up five (5) serum pools. Each pool was spiked with a suite of additional trace elements including selenium as Se^{4+} at various concentrations.

The Target Value assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E** Statistical methods for use in proficiency testing by interlaboratory comparisons. Values for serum selenium range from 113 $\mu\text{g/L}$ (1.43 $\mu\text{mol/L}$) to 282 $\mu\text{g/L}$ (3.57 $\mu\text{mol/L}$).

Acceptable ranges for serum selenium are based on fixed criteria of $\pm 20\%$, or $\pm 2 \mu\text{g/L}$ below 10 $\mu\text{g/L}$. These criteria are a little less stringent than those proposed by the OELM Network of EQAS organizers ($\pm 15\%$ or $\pm 8 \mu\text{g/L}$ below 55 $\mu\text{g/L}$) (1, 2) for trace elements in serum. As performance for serum Se improves among NYS-permit laboratories, consideration will be given to adopting the OELM criteria.

Discussion. Based on the above criteria, 97.6% of test results reported were judged as satisfactory, with none of the 17 participant laboratories reporting 2 or more of the 5 results outside the acceptable ranges.

1. A. Taylor, J. Angerer, J. Arnaud, F. Claeys, R.L. Jones, O. Mazarrasa, E. Mairiaux, A. Menditto, P.J. Parsons, M. Patriarca, A. Pineau, S. Valkonen, J.-P. Weber and C. Weykamp Accreditation and Quality Assurance 2006 11 440-445.

2. J. Arnaud, J.-P. Weber, C.W. Weykamp, P.J. Parsons, J. Angerer, E. Mairiaux, O. Mazarrasa, S. Valkonen, A. Menditto, M. Patriarca, and A. Taylor Clinical Chemistry 2008 54 1892-1899.

New York State Department of Health
Serum Selenium Test Results, 2010 Event #3
ROBUST STATISTICAL SUMMARY

TARGET VALUE ASSIGNMENT AND STATISTICS

	Results ($\mu\text{g/L}$ serum)				
	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
Robust Mean	282	154	113	218	137
Robust Standard Deviation	21.9	10.9	9.9	17.4	12.4
Standard Uncertainty	6.6	3.3	3.0	5.3	3.8
RSD (%)	7.8	7.0	8.8	8.0	9.1
Acceptable Range:					
Upper Limit:	338	185	136	262	164
Lower Limit:	226	123	90	174	110

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Serum Selenium Test Results, 2010 Event #3
PERFORMANCE OF PARTICIPATING LABORATORIES

Lab Code	Method	Results (µg/L serum)					Info Only
		SE10-11	SE10-12	SE10-13	SE10-14	SE10-15	
Target Values:		282	154	113	218	137	
107	DRC/CC-ICP-MS	271	147	107	208	127	Info
110	DRC/CC-ICP-MS	267	148	106	206	128	
114	ICP-MS	249	141	101	168 ↓	121	
147	ICP-MS	250	143	104	203	125	
156	ICP-MS	288	156	112	222	146	
159	ETAAS-Z	288	162	116	227	141	Info
164	ICP-MS	280	148	109	213	133	
179	DRC/CC-ICP-MS	255	146	101	199	127	
197	ICP-MS	298	166	125	233	145	
200	DRC/CC-ICP-MS	269	145	104	205	126	
206	ICP-MS	310	196 ↑	122	230	149	Info
293	DRC/CC-ICP-MS	301	167	131	246	157	
305	ICP-MS	278	152	120	227	147	
324	ID-ICP-MS	277	148	107	208	126	
366	ETAAS-Z	306	166	115	220	145	
367	DRC/CC-ICP-MS	289	156	112	227	137	Info
401	DRC/CC-ICP-MS	313	167	125	249	150	Info

Percent satisfactory results for all participants: 97.6 %

notes: ↑ reported outside upper limit
↓ reported outside lower limit

Info only: results included for informational purposes only.

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Serum Selenium Test Results, 2010 Event #3
STATISTICAL SUMMARY BY METHOD

	Results ($\mu\text{g/L}$ serum)				
	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
DRC/CC-ICP-MS					
Number of Sample Measurements:	7	7	7	7	7
Mean:	281	154	112	220	136
Standard Deviation:	21	10	11	21	13
RSD (%):	7.4	6.3	10.1	9.4	9.3
ETAAS-Z					
Number of Sample Measurements:	2	2	2	2	2
Mean:	297	164	116	224	143
Standard Deviation:	13	3	1	5	3
RSD (%):	—	—	—	—	—
ICP-MS					
Number of Sample Measurements:	7	7	7	7	7
Mean:	279	157	113	214	138
Standard Deviation:	23	19	9	23	12
RSD (%):	8.2	12.1	8.2	10.6	8.4
ID-ICP-MS					
Number of Sample Measurements:	1	1	1	1	1
Mean:	277	148	107	208	126
Standard Deviation:	?	?	?	?	?
RSD (%):	—	—	—	—	—
All Laboratories					
Number of Sample Measurements:	17	17	17	17	17
Mean:	282	156	113	217	137
Standard Deviation:	20	14	9	19	11
RSD (%):	7.1	8.7	8.1	8.9	8.1

notes: ? Insufficient data for calculation.

New York State Department of Health
Event #3, 2010

Serum Zinc

The test materials for serum Zn were prepared from human serum obtained from Tennessee Blood Services, Inc. The units were tested by FDA approved methods and found to be Non-reactive for Anti-HIV-1/2, Anti-HCV 3.0 and HBsAg. The serum has also been found to STS (RPR) Non-reactive and Negative for HIV-1 and HCV by PCR. Serum units were dispensed into acid-washed 500-mL polypropylene containers to make up five (5) serum pools. Each pool was spiked with a suite of additional trace elements including zinc as Zn^{2+} at various concentrations.

The Target Value assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E** Statistical methods for use in proficiency testing by interlaboratory comparisons. Values for serum zinc range from 558 $\mu\text{g/L}$ (8.53 $\mu\text{mol/L}$) to 1754 $\mu\text{g/L}$ (26.82 $\mu\text{mol/L}$).

Acceptable ranges for serum zinc are based on fixed criteria of $\pm 15\%$, or $\pm 15 \mu\text{g/L}$ below 100 $\mu\text{g/L}$. These criteria are consistent with those proposed by the OELM network of EQAS organizers (1) for trace elements in serum.

Discussion. Based on the above criteria, 86.4% of test results reported were judged as satisfactory, with five out of 28 participant laboratories (17.8%) reporting 2 or more of the 5 results outside the acceptable ranges.

1. A. Taylor, J. Angerer, J. Arnaud, F. Claeys, R.L. Jones, O. Mazarrasa, E. Mairiaux, A. Menditto, P.J. Parsons, M. Patriarca, A. Pineau, S. Valkonen, J.-P. Weber and C. Weykamp Accreditation and Quality Assurance 2006 11 440-445.

2. J. Arnaud, J.-P. Weber, C.W. Weykamp, P.J. Parsons, J. Angerer, E. Mairiaux, O. Mazarrasa, S. Valkonen, A. Menditto, M. Patriarca, and A. Taylor Clinical Chemistry 2008 54 1892-1899.

New York State Department of Health
Serum Zinc Test Results, 2010 Event #3
ROBUST STATISTICAL SUMMARY

TARGET VALUE ASSIGNMENT AND STATISTICS

	Results ($\mu\text{g/L}$ serum)				
	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
Robust Mean	1 003	5 58	7 97	1 754	1 475
Robust Standard Deviation	69	46	70	126	110
Standard Uncertainty	16	11	16	30	26
RSD (%)	6.9	8.3	8.7	7.2	7.5
Acceptable Range:					
Upper Limit:	1154	642	917	2017	1696
Lower Limit:	853	474	678	1491	1254

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Serum Zinc Test Results, 2010 Event #3
PERFORMANCE OF PARTICIPATING LABORATORIES

Lab Code	Method	Results (µg/L serum)					Info Only
		SE10-11	SE10-12	SE10-13	SE10-14	SE10-15	
Target Values:		1003	558	797	1754	1475	
107	DRC/CC-ICP-MS	1100	592	851	1731	1497	Info
110	ICP-MS	1071	601	872	1949	1648	
114	ICP-MS	1090	690 ↑	970 ↑	1800	1590	
147	ICP-MS	948	528	752	1739	1458	Info
156	ICP-MS	971	546	750	1621	1225 ↓	
159	ICP-AES/OES	1030	610	830	1760	1500	
160	FAAS	940	450 ↓	700	1530	1140 ↓	
164	ICP-MS	1026	579	793	1856	1508	
179	DRC/CC-ICP-MS	1000	540	790	1810	1520	
197	ICP-MS	1030	540	810	1830	1540	
200	FAAS	1014	569	798	1753	1504	Info
206	ICP-MS	1000	570	810	1730	1470	
287	FAAS	990	540	750	1770	1460	
293	ICP-MS	850 ↓	485	706	1642	1347	Info
305	ICP-MS	9 ↓	5 ↓	7 ↓	16 ↓	13 ↓	
324	ID-ICP-MS	1033	562	840	1825	1543	Info
325	FAAS	1020	570	810	1650	1410	Info
355	ICP-MS	1065	598	883	1887	1559	
357	ICP-MS	940	520	750	1660	1350	
358	ICP-MS	994	575	808	1796	1497	
360	FAAS	1060	590	820	1850	1590	
362	ICP-MS	1020	610	860	1850	1450	
363	ICP-MS	1170 ↑	650 ↑	930 ↑	2050 ↑	1730 ↑	
366	FAAS	938	493	748	1653	1403	Info
401	DRC/CC-ICP-MS	1066	582	850	1896	1608	Info
404	HR-ICP-MS	801 ↓	392 ↓	643 ↓	1562	1285	Info
457	ICP-AES/OES	938	512	749	1665	1402	Info
458	FAAS	996	544	798	1747	1491	

Percent satisfactory results for all participants: 86.4 %

notes: ↑ reported outside upper limit
↓ reported outside lower limit

Info only: results included for informational purposes only.

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Serum Zinc Test Results, 2010 Event #3
STATISTICAL SUMMARY BY METHOD

	Results (µg/L serum)				
	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
DRC/CC-ICP-MS					
Number of Sample Measurements:	3	3	3	3	3
Mean:	1055	571	830	1812	1542
Standard Deviation:	51	28	35	83	59
RSD (%):	—	—	—	—	—
FAAS					
Number of Sample Measurements:	7	7	7	7	7
Mean:	994	537	775	1708	1428
Standard Deviation:	44	49	43	105	142
RSD (%):	4.4	9.2	5.6	6.1	9.9
HR-ICP-MS					
Number of Sample Measurements:	1	1	1	1	1
Mean:	801	392	643	1562	1285
Standard Deviation:	?	?	?	?	?
RSD (%):	—	—	—	—	—
ICP-AES/OES					
Number of Sample Measurements:	2	2	2	2	2
Mean:	984	561	790	1713	1451
Standard Deviation:	65	69	57	67	69
RSD (%):	—	—	—	—	—
ICP-MS					
Number of Sample Measurements:	13	13	13	13	13
Mean:	1013	576	823	1801	1490
Standard Deviation:	79	55	77	124	133
RSD (%):	7.8	9.6	9.4	6.9	9.0
ID-ICP-MS					
Number of Sample Measurements:	1	1	1	1	1
Mean:	1033	562	840	1825	1543
Standard Deviation:	?	?	?	?	?
RSD (%):	—	—	—	—	—
All Laboratories					
Number of Sample Measurements:	27	27	27	27	27
Mean:	1004	557	803	1763	1471
Standard Deviation:	76	60	70	118	127
RSD (%):	7.5	10.7	8.8	6.7	8.6

notes: ? Insufficient data for calculation.

New York State Department of Health
Event #3, 2010

Additional Trace Elements Reported in Serum

Participant laboratories reported their analytical results for any additional trace elements (other than Al, Cu, Se and Zn) that are routinely reported so that a more complete characterization can be recorded for these PT materials. Results for additional trace elements are reported here, but no target value is implied nor are any acceptable ranges provided. These data are provided solely for educational and informational purposes.

In addition to Al, Cu, Se and Zn, the serum pools were supplemented with additional trace elements as indicated below.

Additional Elements

Mn, Cr, V

**New York State Department of Health
Serum Additional Elements, 2010 Event #3
Page 1**

Serum Arsenic (µg/L)						
Lab Code	Method	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
147	ICP-MS	0.802	0.987	0.673	0.778	0.883
197	DRC/CC-ICP-MS	<10	<10	<10	<10	<10

Serum Barium (µg/L)						
Lab Code	Method	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
147	ICP-MS	1.634	0.891	1.483	2.307	1.14
197	ICP-MS	<2.0	<2.0	<2.0	2.6	<2.0

Serum Beryllium (µg/L)						
Lab Code	Method	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
197	ICP-MS	<0.2	<0.2	<0.2	<0.2	<0.2

Serum Cadmium (µg/L)						
Lab Code	Method	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
147	ICP-MS	0.038	0.049	0.023	0.043	0.006
197	DRC/CC-ICP-MS	<0.5	<0.5	<0.5	<0.5	<0.5

Serum Chromium (µg/L)						
Lab Code	Method	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
147	ICP-MS	0.262	5.065	2.361	1.04	7.176
164	ICP-MS	0.3	4.6	2.1	1	6.5
179	DRC/CC-ICP-MS	0.3	5.2	2.3	1.2	7.1
197	DRC/CC-ICP-MS	<1.0	5.1	2.5	1.2	7.3
Arithmetic Mean (n=4)		0.3	5.0	2.3	1.1	7.0
SD		0.02	0.3	0.2	0.1	0.4

Serum Cobalt (µg/L)						
Lab Code	Method	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
147	ICP-MS	0.4	0.695	0.444	0.286	0.267
179	ICP-MS	0.4	0.6	0.4	0.3	0.2
197	ICP-MS	<1.0	<1.0	<1.0	<1.0	<1.0
Arithmetic Mean (n=3)		0.4	0.6	0.4	0.3	0.2
SD		0.0	0.07	0.03	0.01	0.05

New York State Department of Health
Serum Additional Elements, 2010 Event #3
Page 2

Serum Iodine (µg/L)						
Lab Code	Method	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
147	ICP-MS	51.772	44.557	70.633	52.658	52.278

Serum Iron (µg/L)						
Lab Code	Method	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
457	ICP-AES/OES	1176	1878	1379	966	783

Serum Lead (µg/L)						
Lab Code	Method	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
197	DRC/CC-ICP-MS	<0.4	<0.4	<0.4	<0.4	<0.4

Serum Magnesium (µg/L)						
Lab Code	Method	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
457	ICP-AES/OES	16465	16200	16493	17580	17227

Serum Manganese (µg/L)						
Lab Code	Method	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
147	ICP-MS	5.258	1.852	9.176	18.956	2.769
179	DRC/CC-ICP-MS	5.4	2	9	19.4	2.7
197	DRC/CC-ICP-MS	5.4	2.3	9.4	19.1	2.7
305	ICP-MS	6	1.9	9.3	19.1	2.9
457	ICP-AES/OES	5	1	8	18	2
Arithmetic Mean (n=5)		5.4	1.8	9.0	18.9	2.6
SD		0.4	0.5	0.6	0.5	0.4

Serum Mercury (µg/L)						
Lab Code	Method	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
147	ICP-MS	0.178	0.231	0.211	0.279	0.194

**New York State Department of Health
Serum Additional Elements, 2010 Event #3
Page 3**

Serum Molybdenum (µg/L)

Lab Code	Method	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
147	ICP-MS	0.806	1.036	0.947	1.027	0.689

Serum Nickel (µg/L)

Lab Code	Method	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
147	ICP-MS	1.245	1.08	1.151	1.145	1.075

Serum Silver (µg/L)

Lab Code	Method	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
147	ICP-MS	0.138	0.143	0.109	0.091	0.087

Serum Tellurium (µg/L)

Lab Code	Method	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
197	ICP-MS	<1.0	<1.0	<1.0	NA	<1.0

Serum Thallium (µg/L)

Lab Code	Method	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
147	ICP-MS	2.453	0.235	0.564	1.619	1.006
197	ICP-MS	2.2	<1.0	<1.0	1.5	1

Serum Tin (µg/L)

Lab Code	Method	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
147	ICP-MS	0.336	0.281	0.166	0.227	1.176
197	ICP-MS	<5.0	<5.0	<5.0	<5.0	<5.0

Serum Vanadium (µg/L)

Lab Code	Method	SE10-11	SE10-12	SE10-13	SE10-14	SE10-15
179	DRC/CC-ICP-MS	4.5	1.3	2.3	0.6	1.0

New York State Department of Health
Trace Elements in Serum
METHOD NOTES

ATOMIC SPECTROMETRY METHODS

- A-1 ETAAS-Z (Electrothermal atomic absorption spectrometry with Zeeman background correction)
- A-2 ETAAS other (i.e., D₂, S-H background correction)
- A-3 FAAS (Flame atomic absorption spectrometry)
- A-4 CV-AAS (Cold vapor atomic absorption spectrometry)
- A-5 HG-AAS (Hydride generation atomic absorption spectrometry)
- A-6 AFS (Atomic fluorescence spectrometry)
- A-7 Other

INDUCTIVELY COUPLED PLASMA

- P-1 ICP-MS (Inductively coupled plasma - mass spectrometry)
- P-2 DRC/CC-ICP-MS (ICP-MS used in the Dynamic Reaction Cell or Collision Cell mode)
- P-3 ICP-AES/OES (ICP atomic/optical emission spectrometry)
- P-4 HR-ICP-MS (High resolution ICP-MS)
- P-5 ETV-ICP-MS (Electrothermal vaporization ICP-MS)
- P-6 ID-ICP-MS (Isotope dilution ICP-MS)
- P-7 Other

ELECTROCHEMICAL METHODS

- E-1 ASV (Anodic stripping voltammetry without digestion)
- E-2 ASV-LeadCare® (Anodic stripping voltammetry using the ESA LeadCare® system)
- E-3 Fluoride specific electrode
- E-4 Other

MOLECULAR FLUORIMETRY

- F-1 EtOAc (Ethyl acetate-acetic acid extraction method for determination of erythrocyte protoporphyrin)
- F-2 Aviv hematofluorometry (for determination of EP at hematocrit 35)
- F-3 Helena ZPP (for determination of zinc protoporphyrin in $\mu\text{mol ZPP/mol heme}$)
- F-4 Other

OTHER METHODS

If your method is not listed in the above list, please describe it briefly.
