

## **Participant Performance Summary - Initial (Screen) Test**

*Clinical Toxicology - September 2015 Test Event*

## **Participant Performance Summary - Initial (Screen) Test**

**Clinical Toxicology - September 2015 Test Event**

## Participant Performance Summary - Initial (Screen) Test

### Clinical Toxicology - September 2015 Test Event

Benzodiazepines Screen (ng/mL)			UTB 1			UTB 2			UTB 3			UTB 4					
Method Code	Cutoff	Response	No. labs	No. Resp.	%	Response	No. labs	No. Resp.	%	Response	No. labs	No. Res.	%	Response	No. labs	No. Resp.	%
Abbott Architect	150.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Abbott Architect	200.0	Negative	11	11	100	Negative	11	11	100	Negative	11	11	100	Positive	11	11	100
Abbott Architect	300.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Abbott Architect	75.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Alere (Biosite) Triage Meter/PRO	300.0	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100
Alere (BioSite) Triage Plus TCA	300.0	Negative	7	7	100	Negative	7	7	100	Negative	7	7	100	Positive	7	7	100
Alere (Biosite) Triage TOX Drug Screen	300.0	Negative	21	21	100	Negative	21	21	100	Negative	21	21	100	Positive	21	21	100
Ameditech ImmuTest Multi-Drug Screen	300.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Beckman Coulter	200.0	Negative	41	41	100	Negative	41	41	100	Negative	41	41	100	Positive	41	41	100
Beckman Coulter	300.0	Negative	9	9	100	Negative	9	9	100	Negative	9	9	100	Positive	9	9	100
Beckman Coulter	338.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Brana Medical Corporation QuickTox	300.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
BioRad Tox/See	300.0	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100
Diatron Pictus	200.0	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100
Immunalysis ELISA	100.0	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100
Immunalysis ELISA	200.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Immunalysis HEIA	200.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
MedTox Profile	150.0	Negative	14	14	100	Negative	14	14	100	Negative	14	14	100	Positive	14	14	100
MedTox Profile	300.0	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100
NexScreen Cup	300.0	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100
Ortho Vitros	200.0	Negative	14	14	100	Negative	14	14	100	Negative	14	14	100	Positive	14	14	100
Ortho Vitros	300.0	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100
Randox Laboratories Ltd	200.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Roche Cobas	100.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Roche Cobas	200.0	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100	Positive	5	5	100
Roche Cobas	300.0	Negative	6	6	100	Negative	6	6	100	Negative	6	6	100	Positive	6	6	100
Roche KIMS	100.0	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100
Roche KIMS	200.0	Negative	10	10	100	Negative	10	10	100	Negative	10	10	100	Positive	10	10	100
Roche KIMS	300.0	Negative	6	6	100	Negative	6	6	100	Negative	6	6	100	Positive	6	6	100
Roche ONLINE	300.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Siemens ADVIA	200.0	Negative	6	6	100	Negative	6	6	100	Negative	6	6	100	Positive	6	6	100
Siemens ADVIA	300.0	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100
Siemens Dimension	200.0	Negative	44	44	100	Negative	44	44	100	Negative	44	44	100	Positive	44	44	100
Siemens Dimension	300.0	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100
Siemens Syva Emit	200.0	Negative	13	13	100	Negative	13	13	100	Negative	13	13	100	Positive	13	13	100
Siemens Syva Emit	300.0	Negative	8	8	100	Negative	8	8	100	Negative	8	8	100	Positive	8	8	100
Thermo Scientific (Microgenics) CEDIA	100.0	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100
Thermo Scientific (Microgenics) CEDIA	200.0	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100	Positive	5	5	100
Thermo Scientific (Microgenics) CEDIA	300.0	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100
Thermo Scientific (Microgenics) DRI	100.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Thermo Scientific (Microgenics) DRI	200.0	Negative	14	14	100	Negative	14	13	92.86	Negative	14	14	100	Negative	14	1	7.14
						Positive	14	1	7.14					Positive	14	13	92.9
Thermo Scientific (Microgenics) DRI	300.0	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100
Thermo Scientific (Microgenics) DRI	40.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Thermo Scientific (Microgenics) DRI	50.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Other	200.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Other	500.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
(All Labs)	(All Labs)	Negative	282	282	100	Negative	282	281	99.65	Negative	282	282	100	Negative	282	1	0.35

**Participant Performance Summary - Initial (Screen) Test**

**Clinical Toxicology - September 2015 Test Event**

Positive	282	1	0.35	Positive	282	281	99.7
----------	-----	---	------	----------	-----	-----	------

## Participant Performance Summary - Initial (Screen) Test

### Clinical Toxicology - September 2015 Test Event

Cocaine/metab Screen (ng/mL)			UTB 1				UTB 2				UTB 3				UTB 4			
Method Code	Cutoff	Response	No. labs	No. Resp.	%	Response	No. labs	No. Resp.	%	Response	No. labs	No. Res.	%	Response	No. labs	No. Resp.	%	
Abbott Architect	150.0	Positive	6	6	100	Negative	6	6	100	Negative	6	6	100	Negative	6	6	100	
Abbott Architect	300.0	Positive	9	9	100	Negative	9	9	100	Negative	9	9	100	Negative	9	9	100	
Alere (Biosite) Triage Meter/PRO	300.0	Positive	3	3	100	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	
Alere (BioSite) Triage Plus TCA	300.0	Positive	7	7	100	Negative	7	7	100	Negative	7	7	100	Negative	7	7	100	
Alere (Biosite) Triage TOX Drug Screen	300.0	Positive	21	21	100	Negative	21	21	100	Negative	21	21	100	Negative	21	21	100	
Ameditech ImmuTest Multi-Drug Screen	300.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Beckman Coulter	150.0	Positive	7	7	100	Negative	7	7	100	Negative	7	7	100	Negative	7	7	100	
Beckman Coulter	300.0	Positive	42	42	100	Negative	42	42	100	Negative	42	42	100	Negative	42	42	100	
Beckman Coulter	391.1	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Branan Medical Corporation QuickTox	300.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
BioRad Tox/See	300.0	Positive	3	3	100	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	
Diatron Pictus	300.0	Positive	2	2	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	
E-Screen	300.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Immunalysis ELISA	100.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Immunalysis HEIA	150.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
MedTox Profile	150.0	Positive	14	14	100	Negative	14	14	100	Negative	14	14	100	Negative	14	14	100	
MedTox Profile	300.0	Positive	2	2	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	
NexScreen Cup	300.0	Positive	3	3	100	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	
Ortho Vitros	300.0	Positive	16	16	100	Negative	16	16	100	Negative	16	16	100	Negative	16	16	100	
Randox Laboratories Ltd	300.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Roche Cobas	150.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Roche Cobas	300.0	Positive	11	11	100	Negative	11	11	100	Negative	11	11	100	Negative	11	11	100	
Roche KIMS	300.0	Positive	19	19	100	Negative	19	19	100	Negative	19	19	100	Negative	19	19	100	
Roche ONLINE	300.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Siemens ADVIA	300.0	Positive	8	8	100	Negative	8	8	100	Negative	8	8	100	Negative	8	8	100	
Siemens Dimension	150.0	Positive	6	6	100	Negative	6	6	100	Negative	6	6	100	Negative	6	6	100	
Siemens Dimension	300.0	Positive	41	41	100	Negative	41	41	100	Negative	41	41	100	Negative	41	41	100	
Siemens Syva Emit	150.0	Positive	5	5	100	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100	
Siemens Syva Emit	300.0	Positive	12	12	100	Negative	12	12	100	Negative	12	12	100	Negative	12	12	100	
Thermo Scientific (Microgenics) CEDIA	100.0	Positive	5	5	100	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100	
Thermo Scientific (Microgenics) CEDIA	150.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Thermo Scientific (Microgenics) CEDIA	300.0	Positive	2	2	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	
Thermo Scientific (Microgenics) DRI	150.0	Positive	17	17	100	Negative	17	17	100	Negative	17	17	100	Negative	17	17	100	
Thermo Scientific (Microgenics) DRI	300.0	Positive	10	10	100	Negative	10	10	100	Negative	10	10	100	Negative	10	10	100	
Other	1000.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Other	150.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
(All Labs)	(All Labs)	Positive	283	283	100	Negative	283	283	100	Negative	283	283	100	Negative	283	283	100	

## Participant Performance Summary - Initial (Screen) Test

### Clinical Toxicology - September 2015 Test Event

Ethyl glucuronide Screen (ng/mL)			UTB 1					UTB 2					UTB 3					UTB 4				
Method Code	Cutoff	Response	No. labs	No. Resp.	%		Response	No. labs	No. Resp.	%		Response	No. labs	No. Res.	%		Response	No. labs	No. Resp.	%		
Beckman Coulter	500.0	Negative	4	4	100		Negative	4	4	100		Negative	4	4	100		Negative	4	4	100		
Diatron Pictus	500.0	Negative	1	1	100		Negative	1	1	100		Negative	1	1	100		Negative	1	1	100		
Immunalysis ELISA	100.0	Negative	1	1	100		Negative	1	1	100		Negative	1	1	100		Negative	1	1	100		
Immunalysis ELISA	1000.0	Negative	1	1	100		Negative	1	1	100		Negative	1	1	100		Negative	1	1	100		
Immunalysis HEIA	250.0	Negative	1	1	100		Negative	1	1	100		Negative	1	1	100		Negative	1	1	100		
Immunalysis HEIA	500.0	Negative	6	5	83.33		Negative	6	5	83.33		Negative	6	6	100		Negative	6	5	83.3		
	NR		6	1	16.67		Positive	6	1	16.67						Positive	6	1	16.7			
LC MS/MS	250.0	Negative	1	1	100		Negative	1	1	100		Negative	1	1	100		Negative	1	1	100		
LC MS/MS	500.0	Negative	5	5	100		Negative	5	5	100		Negative	5	5	100		Negative	5	5	100		
Siemens ADVIA	500.0	Negative	1	1	100		Negative	1	1	100		Negative	1	1	100		Negative	1	1	100		
Siemens Syva Emit	250.0	Negative	1	1	100		Negative	1	1	100		Negative	1	1	100		Negative	1	1	100		
Thermo Scientific (Microgenics) DRI	1000.0	Negative	1	1	100		Negative	1	1	100		Negative	1	1	100		Negative	1	1	100		
Thermo Scientific (Microgenics) DRI	250.0	Negative	2	2	100		Negative	2	2	100		Negative	2	2	100		Negative	2	2	100		
Thermo Scientific (Microgenics) DRI	500.0	Negative	15	15	100		Negative	15	15	100		Negative	15	15	100		Negative	15	15	100		
(All Labs)	(All Labs)	Negative	40	39	97.5		Negative	40	39	97.5		Negative	40	40	100		Negative	40	39	97.5		
		NR	40	1	2.5		Positive	40	1	2.5						Positive	40	1	2.5			

## Participant Performance Summary - Initial (Screen) Test

### Clinical Toxicology - September 2015 Test Event

Ethanol Screen (mg/dL)			UTB 1			UTB 2			UTB 3			UTB 4					
Method Code	Cutoff	Response	No. labs	No. Resp.	%	Response	No. labs	No. Resp.	%	Response	No. labs	No. Res.	%	Response	No. labs	No. Resp.	%
Abbott Architect	10.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Abbott Architect	13.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Abbott Architect	20.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Alcohol Dehydrogenases/UV or Visible S	10.0	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100	Positive	2	2	100
Alcohol Dehydrogenases/UV or Visible S	20.0	Negative	6	6	100	Positive	6	6	100	Negative	6	6	100	Positive	6	6	100
Alcohol Dehydrogenases/UV or Visible S	50.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Beckman Coulter	10.0	Negative	12	12	100	Positive	12	12	100	Negative	12	12	100	Positive	12	12	100
Beckman Coulter	100.0	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100
Beckman Coulter	104.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Beckman Coulter	20.0	Negative	8	8	100	Positive	8	8	100	Negative	8	8	100	Positive	8	8	100
Beckman Coulter	30.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Beckman Coulter	40.0	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100	Positive	2	2	100
Beckman Coulter	5.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Beckman Coulter	50.0	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100
Diatron Pictus	10.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
GC Flame Ionization	10.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Immunalysis HEIA	20.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Ortho Vitros	10.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Ortho Vitros	25.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Roche Cobas	10.0	Negative	6	6	100	Positive	6	6	100	Negative	6	6	100	Positive	6	6	100
Roche Cobas	10.1	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Roche Cobas	20.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Roche Cobas	50.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Roche KIMS	10.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Roche KIMS	20.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Roche KIMS	50.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Roche ONLINE	20.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Siemens ADVIA	10.0	Negative	4	4	100	Positive	4	4	100	Negative	4	4	100	Positive	4	4	100
Siemens ADVIA	20.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Siemens ADVIA	40.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Siemens Dimension	10.0	Negative	4	4	100	Positive	4	4	100	Negative	4	4	100	Positive	4	4	100
Siemens Dimension	20.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Siemens Dimension	5.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Siemens Dimension	50.0	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100	Positive	5	5	100
Siemens Syva Emit	10.0	Negative	3	3	100	Positive	3	3	100	Negative	3	3	100	Positive	3	3	100
Siemens Syva Emit	100.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Siemens Syva Emit	20.0	Negative	5	5	100	Positive	5	5	100	Negative	5	5	100	Positive	5	5	100
Siemens Syva Emit	40.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Thermo Scientific (Microgenics) CEDIA	20.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Thermo Scientific (Microgenics) DRI	100.0	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100
Thermo Scientific (Microgenics) DRI	20.0	Negative	13	13	100	Positive	13	13	100	Negative	13	13	100	Positive	13	13	100
Thermo Scientific (Microgenics) DRI	30.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Thermo Scientific (Microgenics) DRI	40.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
Thermo Scientific (Microgenics) DRI	50.0	Negative	5	5	100	Negative	5	4	80	Negative	5	5	100	Positive	5	5	100
Other	10.0	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Positive	1	1	100
(All Labs)	(All Labs)	Negative	110	110	100	Negative	110	20	18.18	Negative	110	110	100	Positive	110	110	100
						Positive	110	90	81.82								

## Participant Performance Summary - Initial (Screen) Test

### Clinical Toxicology - September 2015 Test Event

Fentanyl Screen (ng/mL)			UTB 1				UTB 2				UTB 3				UTB 4			
Method Code	Cutoff	Response	No. labs	No. Resp.	%	Response	No. labs	No. Resp.	%	Response	No. labs	No. Res.	%	Response	No. labs	No. Resp.	%	
Abbott Architect	2.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Beckman Coulter	0.5	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Beckman Coulter	2.0	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100	
Beckman Coulter	3.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Diatron Pictus	2.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
GC Mass Spectrometry	2.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Immunalysis ELISA	0.3	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100	
Immunalysis ELISA	1.0	Negative	7	7	100	Negative	7	7	100	Positive	7	7	100	Negative	7	7	100	
Immunalysis ELISA	2.0	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100	Negative	3	3	100	
Immunalysis ELISA	200.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Immunalysis HEIA	1.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Immunalysis HEIA	2.0	Negative	9	9	100	Negative	9	9	100	Positive	9	9	100	Negative	9	9	100	
LC MS/MS	1.0	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100	
LC MS/MS	10.0	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100	
LC MS/MS	2.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Randox Laboratories Ltd	1.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Thermo Scientific (Microgenics) DRI	2.0	Negative	5	5	100	Negative	5	5	100	Positive	5	5	100	Negative	5	5	100	
Other	10.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Other	2.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
(All Labs)	(All Labs)	Negative	43	43	100	Negative	43	43	100	Positive	43	43	100	Negative	43	43	100	

## **Participant Performance Summary - Initial (Screen) Test**

**Clinical Toxicology - September 2015 Test Event**

## **Participant Performance Summary - Initial (Screen) Test**

*Clinical Toxicology - September 2015 Test Event*

## Participant Performance Summary - Initial (Screen) Test

### Clinical Toxicology - September 2015 Test Event

Opiates Screen (ng/mL)			UTB 1				UTB 2				UTB 3				UTB 4			
Method Code	Cutoff	Response	No. labs	No. Resp.	%	Response	No. labs	No. Resp.	%	Response	No. labs	No. Res.	%	Response	No. labs	No. Resp.	%	
Abbott Architect	2000.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Abbott Architect	300.0	Negative	12	12	100	Negative	12	12	100	Positive	12	12	100	Negative	12	12	100	
Abbott Architect	50.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Alere (Biosite) Triage Meter/PRO	300.0	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100	Negative	3	3	100	
Alere (BioSite) Triage Plus TCA	300.0	Negative	7	7	100	Negative	7	7	100	Positive	7	7	100	Negative	7	7	100	
Alere (Biosite) Triage TOX Drug Screen	2000.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Alere (Biosite) Triage TOX Drug Screen	300.0	Negative	20	20	100	Negative	20	20	100	Positive	20	20	100	Negative	20	20	100	
Ameditech ImmuTest Multi-Drug Screen	300.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Beckman Coulter	2000.0	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100	Negative	3	3	100	
Beckman Coulter	300.0	Negative	46	46	100	Negative	46	45	97.83	Positive	46	46	100	Negative	46	46	100	
						Positive	46	1	2.17									
Beckman Coulter	310.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Branan Medical Corporation QuickTox	300.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
BioRad Tox/See	300.0	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100	Negative	3	3	100	
Diatron Pictus	300.0	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100	
E-Screen	2000.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Immunalysis ELISA	100.0	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100	Negative	3	3	100	
Immunalysis HEIA	300.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
MedTox Profile	100.0	Negative	14	14	100	Negative	14	14	100	Positive	14	14	100	Negative	14	13	92.9	
										Positive	14	1	7.14					
MedTox Profile	300.0	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100	
NexScreen Cup	300.0	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100	Negative	3	3	100	
Ortho Vitros	300.0	Negative	16	16	100	Negative	16	16	100	Positive	16	16	100	Negative	16	16	100	
Randox Laboratories Ltd	100.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Roche Cobas	300.0	Negative	12	12	100	Negative	12	12	100	Positive	12	12	100	Negative	12	12	100	
Roche KIMS	300.0	Negative	17	17	100	Negative	17	17	100	Positive	17	17	100	Negative	17	17	100	
Roche ONLINE	200.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Siemens ADVIA	300.0	Negative	8	8	100	Negative	8	8	100	Positive	8	8	100	Negative	8	8	100	
Siemens Dimension	200.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Siemens Dimension	2000.0	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100	
Siemens Dimension	300.0	Negative	44	44	100	Negative	44	44	100	Positive	44	44	100	Negative	44	44	100	
Siemens Syva Emit	100.0	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100	
Siemens Syva Emit	2000.0	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100	
Siemens Syva Emit	300.0	Negative	21	21	100	Negative	21	21	100	Positive	21	21	100	Negative	21	21	100	
Thermo Scientific (Microgenics) CEDIA	100.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Thermo Scientific (Microgenics) CEDIA	150.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Thermo Scientific (Microgenics) CEDIA	300.0	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100	
Thermo Scientific (Microgenics) DRI	150.0	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100	
Thermo Scientific (Microgenics) DRI	2000.0	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100	Negative	3	3	100	
Thermo Scientific (Microgenics) DRI	300.0	Negative	19	19	100	Negative	19	19	100	Positive	19	19	100	Negative	19	19	100	
Thermo Scientific (Microgenics) DRI	50.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Other	100.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
Other	1999.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	
(All Labs)	(All Labs)	Negative	284	284	100	Negative	284	283	99.65	Positive	284	284	100	Negative	284	283	99.7	
						Positive	284	1	0.35					Positive	284	1	0.35	

## Participant Performance Summary - Initial (Screen) Test

### Clinical Toxicology - September 2015 Test Event

Oxycodone Screen (ng/mL)			UTB 1			UTB 2			UTB 3			UTB 4					
Method Code	Cutoff	Response	No. labs	No. Resp.	%	Response	No. labs	No. Resp.	%	Response	No. labs	No. Res.	%	Response	No. labs	No. Resp.	%
Abbott Architect	100.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
Abbott Architect	300.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
Abbott Architect	50.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
Alere (Biosite) Triage TOX Drug Screen	300.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
Beckman Coulter	100.0	Negative	9	9	100	Negative	9	9	100	Positive	9	9	100	Negative	9	9	100
Beckman Coulter	300.0	Negative	4	4	100	Negative	4	4	100	Positive	4	4	100	Negative	4	4	100
BioRad Tox/See	100.0	Negative	4	4	100	Negative	4	4	100	Positive	4	4	100	Negative	4	4	100
Diatron Pictus	100.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
Immunalysis ELISA	100.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
Immunalysis HEIA	100.0	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100	Negative	3	3	100
Immunalysis HEIA	300.0	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100
Innovacon Oxycodone Single Drug Test	100.0	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100	Negative	3	3	100
LC MS/MS	100.0	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100
MedTox Profile	100.0	Negative	17	17	100	Negative	17	17	100	Positive	17	17	100	Negative	17	16	94.1
													Positive	17	1	5.88	
NexScreen Cup	100.0	Negative	3	3	100	Negative	3	3	100	Negative	3	1	33.33	Negative	3	3	100
										Positive	3	2	66.67				
Ortho Vitros	100.0	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100
Randox Laboratories Ltd	100.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
Roche Cobas	100.0	Negative	5	5	100	Negative	5	5	100	Positive	5	5	100	Negative	5	5	100
Roche Cobas	300.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
Roche KIMS	100.0	Negative	4	4	100	Negative	4	4	100	Positive	4	4	100	Negative	4	4	100
Siemens ADVIA	100.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
Siemens Dimension	100.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
Siemens Syva Emit	100.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
Siemens Syva Emit	300.0	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100	Negative	3	3	100
Thermo Scientific (Microgenics) CEDIA	100.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
Thermo Scientific (Microgenics) DRI	100.0	Negative	35	35	100	Negative	35	35	100	Positive	35	35	100	Negative	35	35	100
Thermo Scientific (Microgenics) DRI	300.0	Negative	6	6	100	Negative	6	6	100	Negative	6	1	16.67	Negative	6	6	100
										Positive	6	5	83.33				
Thermo Scientific (Microgenics) DRI	50.0	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
Other	100.0	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100	Negative	3	3	100
(All Labs)	(All Labs)	Negative	118	118	100	Negative	118	118	100	Negative	118	2	1.69	Negative	118	117	99.2
										Positive	118	116	98.31	Positive	118	1	0.85

## Participant Performance Summary - Initial (Screen) Test

### Clinical Toxicology - September 2015 Test Event

Phencyclidine (PCP) Screen (ng/mL)			UTB 1				UTB 2				UTB 3				UTB 4			
Method Code	Cutoff	Response	No. labs	No. Resp.	%	Response	No. labs	No. Resp.	%	Response	No. labs	No. Res.	%	Response	No. labs	No. Resp.	%	
Abbott Architect	25.0	Negative	13	13	100	Negative	13	13	100	Negative	13	13	100	Positive	13	13	100	
Alere (Biosite) Triage Meter/PRO	25.0	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100	
Alere (BioSite) Triage Plus TCA	25.0	Negative	7	7	100	Negative	7	7	100	Negative	7	7	100	Positive	7	7	100	
Alere (Biosite) Triage TOX Drug Screen	25.0	Negative	17	17	100	Negative	17	17	100	Negative	17	17	100	Positive	17	17	100	
Ameditech ImmuTest Multi-Drug Screen	25.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	
Beckman Coulter	25.0	Negative	45	45	100	Negative	45	45	100	Negative	45	45	100	Positive	45	45	100	
Beckman Coulter	470.6	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	
Branan Medical Corporation QuickTox	25.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	
BioRad Tox/See	25.0	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100	
Diatron Pictus	25.0	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	
E-Screen	25.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	
GC Mass Spectrometry	25.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	
Immunalysis ELISA	100.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	
Immunalysis HEIA	25.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	
MedTox Profile	25.0	Negative	16	16	100	Negative	16	16	100	Negative	16	16	100	Negative	16	1	6.25	
													Positive	16	15	93.8		
NexScreen Cup	25.0	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100	
Ortho Vitros	25.0	Negative	13	13	100	Negative	13	13	100	Negative	13	13	100	Positive	13	13	100	
Randox Laboratories Ltd	25.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	
Roche Cobas	25.0	Negative	12	12	100	Negative	12	12	100	Negative	12	12	100	Positive	12	12	100	
Roche KIMS	25.0	Negative	18	18	100	Negative	18	18	100	Negative	18	18	100	Positive	18	18	100	
Roche ONLINE	25.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	
Siemens ADVIA	25.0	Negative	7	7	100	Negative	7	7	100	Negative	7	7	100	Positive	7	7	100	
Siemens Dimension	25.0	Negative	43	43	100	Negative	43	43	100	Negative	43	43	100	Positive	43	43	100	
Siemens Syva Emit	25.0	Negative	15	15	100	Negative	15	15	100	Negative	15	15	100	Positive	15	15	100	
Thermo Scientific (Microgenics) CEDIA	25.0	Negative	7	7	100	Negative	7	7	100	Negative	7	7	100	Positive	7	7	100	
Thermo Scientific (Microgenics) DRI	25.0	Negative	30	29	96.67	Negative	30	29	96.67	Negative	30	29	96.67	Negative	30	1	3.33	
	NR	30	1	3.33		NR	30	1	3.33	NR	30	1	3.33	Positive	30	28	93.3	
													NR	30	1	3.33		
Other	25.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	
Other	50.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	
(All Labs)	(All Labs)	Negative	265	264	99.62	Negative	265	264	99.62	Negative	265	264	99.62	Negative	265	2	0.75	
		NR	265	1	0.38	NR	265	1	0.38	NR	265	1	0.38	Positive	265	262	98.9	
													NR	265	1	0.38		

## Participant Performance Summary - Initial (Screen) Test

### Clinical Toxicology - September 2015 Test Event

Propoxyphene Screen (ng/mL)			UTB 1				UTB 2				UTB 3				UTB 4			
Method Code	Cutoff	Response	No. labs	No. Resp.	%	Response	No. labs	No. Resp.	%	Response	No. labs	No. Res.	%	Response	No. labs	No. Resp.	%	
Abbott Architect	300.0	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100	
Beckman Coulter	300.0	Negative	22	22	100	Negative	22	22	100	Negative	22	22	100	Negative	22	22	100	
Beckman Coulter	314.7	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Diatron Pictus	300.0	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	
Immunalysis ELISA	100.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Immunalysis HEIA	300.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
HPLC Mass Spectrometry	300.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
MedTox Profile	300.0	Negative	14	14	100	Negative	14	13	92.86	Negative	14	14	100	Negative	14	13	92.9	
						Positive	14	1	7.14					Positive	14	1	7.14	
Ortho Vitros	300.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Randox Laboratories Ltd	300.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Roche Cobas	300.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Roche KIMS	300.0	Negative	8	8	100	Negative	8	8	100	Negative	8	8	100	Negative	8	8	100	
Siemens ADVIA	300.0	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	
Siemens Syva Emit	300.0	Negative	11	11	100	Negative	11	11	100	Negative	11	11	100	Negative	11	11	100	
Thermo Scientific (Microgenics) CEDIA	100.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Thermo Scientific (Microgenics) CEDIA	150.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Thermo Scientific (Microgenics) CEDIA	300.0	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100	
Thermo Scientific (Microgenics) DRI	180.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Thermo Scientific (Microgenics) DRI	200.0	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	
Thermo Scientific (Microgenics) DRI	300.0	Negative	19	19	100	Negative	19	19	100	Negative	19	19	100	Negative	19	19	100	
Other	2500.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
(All Labs)	(All Labs)	Negative	101	101	100	Negative	101	100	99.01	Negative	101	101	100	Negative	101	100	99	
						Positive	101	1	0.99					Positive	101	1	0.99	

## Participant Performance Summary - Initial (Screen) Test

### Clinical Toxicology - September 2015 Test Event

Cannabinoids (THCA) Screen (ng/mL)			UTB 1				UTB 2				UTB 3				UTB 4			
Method Code	Cutoff	Response	No. labs	No. Resp.	%	Response	No. labs	No. Resp.	%	Response	No. labs	No. Res.	%	Response	No. labs	No. Resp.	%	
Abbott Architect	20.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Abbott Architect	50.0	Positive	13	13	100	Negative	13	13	100	Negative	13	13	100	Negative	13	13	100	
Alere (Biosite) Triage Meter/PRO	50.0	Positive	3	3	100	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	
Alere (BioSite) Triage Plus TCA	50.0	Positive	7	7	100	Negative	7	7	100	Negative	7	7	100	Negative	7	7	100	
Alere (Biosite) Triage TOX Drug Screen	50.0	Positive	21	21	100	Negative	21	21	100	Negative	21	21	100	Negative	21	21	100	
Ameditech ImmuTest Multi-Drug Screen	50.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Beckman Coulter	100.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Beckman Coulter	20.0	Positive	6	6	100	Negative	6	6	100	Negative	6	6	100	Negative	6	6	100	
Beckman Coulter	381.4	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Beckman Coulter	50.0	Positive	43	43	100	Negative	43	43	100	Negative	43	43	100	Negative	43	43	100	
Branan Medical Corporation QuickTox	50.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
BioRad Tox/See	50.0	Positive	3	3	100	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	
Diatron Pictus	50.0	Positive	2	2	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	
E-Screen	50.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Immunalysis ELISA	100.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Immunalysis HEIA	50.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
MedTox Profile	50.0	Positive	16	16	100	Negative	16	16	100	Negative	16	16	100	Negative	16	16	100	
NexScreen Cup	50.0	Positive	3	3	100	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	
Ortho Vitros	50.0	Positive	16	16	100	Negative	16	16	100	Negative	16	16	100	Negative	16	16	100	
Randox Laboratories Ltd	50.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Roche Cobas	50.0	Positive	12	12	100	Negative	12	12	100	Negative	12	12	100	Negative	12	12	100	
Roche KIMS	20.0	Positive	2	2	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	
Roche KIMS	50.0	Positive	17	17	100	Negative	17	17	100	Negative	17	17	100	Negative	17	17	100	
Roche ONLINE	50.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Siemens ADVIA	50.0	Positive	8	8	100	Negative	8	8	100	Negative	8	8	100	Negative	8	8	100	
Siemens Dimension	25.0	Positive	3	3	100	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	
Siemens Dimension	300.0	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	
Siemens Dimension	50.0	Positive	42	42	100	Negative	42	42	100	Negative	42	42	100	Negative	42	42	100	
Siemens Syva Emit	20.0	Positive	7	7	100	Negative	7	7	100	Negative	7	7	100	Negative	7	7	100	
Siemens Syva Emit	50.0	Positive	14	14	100	Negative	14	14	100	Negative	14	14	100	Negative	14	14	100	
Thermo Scientific (Microgenics) CEDIA	20.0	Positive	3	3	100	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	
Thermo Scientific (Microgenics) CEDIA	25.0	Positive	3	3	100	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	
Thermo Scientific (Microgenics) CEDIA	50.0	Positive	2	2	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	
Thermo Scientific (Microgenics) DRI	20.0	Positive	3	3	100	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	
Thermo Scientific (Microgenics) DRI	50.0	Positive	22	22	100	Negative	22	22	100	Negative	22	22	100	Negative	22	22	100	
(All Labs)	(All Labs)	Positive	282	282	100	Negative	282	282	100	Negative	282	282	100	Negative	282	282	100	

## Participant Performance Summary - Initial (Screen) Test

### Clinical Toxicology - September 2015 Test Event

Tricyclic Antidepressants Screen (ng/ml)				UTB 1				UTB 2				UTB 3				UTB 4			
Method Code	Cutoff	Response	No. labs	No. Resp.	%	Response	No. labs	No. Resp.	%	Response	No. labs	No. Res.	%	Response	No. labs	No. Resp.	%		
Abbott Architect	1000.0	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	Negative	3	1	33.3		
														Positive	3	2	66.7		
Alere (Biosite) Triage Meter/PRO	1000.0	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100		
Alere (BioSite) Triage Plus TCA	1000.0	Negative	7	7	100	Negative	7	7	100	Negative	7	7	100	Positive	7	7	100		
Alere (Biosite) Triage TOX Drug Screen	1000.0	Negative	20	20	100	Negative	20	20	100	Negative	20	20	100	Positive	20	20	100		
Ameditech ImmuTest Multi-Drug Screen	1000.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100		
Beckman Coulter	1000.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100		
BioRad Tox/See	1000.0	Negative	4	4	100	Negative	4	4	100	Negative	4	4	100	Positive	4	4	100		
Diatron Pictus	300.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100		
DIPR - Diagnostic Products	110.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100		
GC Mass Spectrometry	100.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100		
GC Mass Spectrometry	400.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100		
Immunalysis ELISA	100.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100		
Immunalysis HEIA	300.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100		
LC MS/MS	100.0	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100		
LC MS/MS	200.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100		
MedTox Profile	300.0	Negative	15	15	100	Negative	15	15	100	Negative	15	15	100	Negative	15	2	13.3		
														Positive	15	13	86.7		
Medtox Verdict	30.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100		
NexScreen Cup	1000.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100		
Ortho Vitros	300.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100		
Randox Laboratories Ltd	100.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100		
Roche Cobas	500.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100		
Siemens Syva Emit	100.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100		
Siemens Syva Emit	300.0	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100		
Siemens Syva Rapid Test	1000.0	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100	Positive	5	5	100		
Thermo Scientific (Microgenics) DRI	100.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100		
Thermo Scientific (Microgenics) DRI	150.0	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100		
Thermo Scientific (Microgenics) DRI	300.0	Negative	10	10	100	Negative	10	10	100	Negative	10	10	100	Positive	10	10	100		
Thermo Scientific (Microgenics) DRI	500.0	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100		
Thermo Scientific (Microgenics) DRI	700.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100		
Thermo Scientific (Microgenics) DRI	999.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100		
Other	500.0	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100		
(All Labs)	(All Labs)	Negative	95	95	100	Negative	95	95	100	Negative	95	95	100	Negative	95	8	8.42		
														Positive	95	87	91.6		