

Nirav R. Shah, M.D., M.P.H.
Commissioner

Sue Kelly
Executive Deputy Commissioner

Clinical Chemistry Proficiency Testing – September 12, 2011

Enclosed are results from the clinical chemistry proficiency survey shipped September 12, 2011. Test samples were prepared in-house by the quantitative transfer of constituents to pooled human serum. The material was subsequently sterile filtered, dispensed into aliquots and stored frozen at -80 °C. Five specimens (**C61, C62, C63, C64, C65**) were distributed to each participant for analysis.

Outlined below is a description of the process utilized in the evaluation of your laboratory's proficiency test results. A summary of your laboratory's performance for the three most recent surveys is also included with your report.

Target Value: In general, targets utilized are derived from all-participant mean values calculated by robust statistical technique. In some cases, however, it is recognized that method, reagent, and/or instrument specific targets may be required and "peer group" specific targets are used where appropriate. Should an alternate target be required to evaluate your laboratory's data, an asterisk will be placed adjacent to the method, reagent, and/or instrument listing(s) corresponding to the peer group utilized.

Acceptable Range: Represents limits established using criteria specified by CLIA '88 regulations, allowing for rounding to appropriate significant digits. Results falling within this range are scored as 100%. Any result exceeding these limits is considered unsatisfactory and receives a score of 0%. Laboratories must achieve an overall analyte score $\geq 80\%$ in order to meet performance criteria for that analyte.

Range Plots: Plots relative distance of your laboratory's result (represented by an "x") from the target for each sample analyzed. Any result exceeding the high or low limit by $>20\%$ of the acceptable range is indicated by an asterisk (*).

Not Gradable: Results for graded analytes for a few laboratories using unique instrument, reagent, or instrument/reagent combinations were considered "not gradable". For these laboratories, pass credit (100%) has been issued. If a laboratory is unable to participate in the NYS clinical chemistry proficiency test event as a graded participant for any analyte, it is the responsibility of that laboratory to establish alternate means to verify the accuracy and precision of the test system for any non-graded analyte(s).

Troponin I, Troponin T, and Estimated Glomerular Filtration Rate: These analytes were included in the September 2011 test event for evaluation only. Although results were not graded, targets and acceptable ranges appropriate for your laboratory's methodology are provided.

The attached statistical report provides a summary of participant data for the five survey specimens. Results for individual instrument and reagent systems where the number of laboratories using those systems is three or greater are provided. Mean and Standard Deviation (1 SD) values shown are calculated by a robust statistical technique that does not assume a Gaussian distribution. These statistical reports are also available on the internet at:

<http://www.wadsworth.org/chemheme/chem/gencc/ptframes.htm>

Should you have any questions regarding these reports or wish to obtain an additional copy, please contact the Clinical Chemistry Section at (518) 474-5582.

Note: The use of brand and/or trade names in this report does not constitute an endorsement of the products on the part of the Wadsworth Center or the New York State Department of Health

Summary of Participant Performance (Mean and Standard Deviation)

Glucose (mg/dL)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
265.5 ± 6.43	44.3 ± 2.33	75.8 ± 2.36	192.7 ± 4.99	91.0 ± 2.61	n = 390	[---] All Methods & Instruments
263.3 ± 1.37	48.6 ± 1.02	80.0 ± 1.80	192.5 ± 1.86	94.3 ± 1.37	n = 3	<Instruments>
268.6 ± 2.87	42.4 ± 0.56	75.3 ± 0.74	192.9 ± 2.29	90.6 ± 1.05	n = 14	[AXA] Abaxis Piccolo
264.3 ± 7.68	43.0 ± 1.49	75.7 ± 1.94	190.8 ± 4.82	90.7 ± 2.79	n = 52	[ABJ] Abbott Architect c System
263.0 ± 9.89	43.2 ± 2.71	76.0 ± 3.55	192.4 ± 8.17	90.6 ± 4.12	n = 9	[OLC] Beckman Coulter AU Chemistry System
262.6 ± 8.23	43.8 ± 2.44	74.9 ± 3.03	189.7 ± 5.90	89.7 ± 2.47	n = 9	[BCS] Beckman Coulter CX
265.2 ± 5.78	44.9 ± 1.62	76.6 ± 1.39	192.7 ± 4.08	91.8 ± 2.50	n = 22	[BCX] Beckman Coulter LX-20
263.0 ± 5.04	43.8 ± 2.36	75.0 ± 2.05	191.0 ± 2.96	90.3 ± 1.75	n = 23	[BCG] Beckman Coulter UniCel DxC 600
273.0 ± 0.90	64.7 ± 3.37	95.3 ± 1.37	203.7 ± 0.51	108.2 ± 2.36	n = 3	[BCH] Beckman Coulter UniCel DxC 800
277.0 ± 0.90	67.5 ± 1.86	104.7 ± 2.26	213.0 ± 1.80	120.5 ± 5.40	n = 3	[HEB] HemoCue B-Glucose
272.8 ± 1.91	43.7 ± 1.85	77.5 ± 1.02	194.2 ± 0.91	92.5 ± 1.24	n = 6	[HEC] HemoCue Glucose 201
266.9 ± 5.30	43.9 ± 1.45	73.7 ± 2.53	197.6 ± 4.39	90.3 ± 2.33	n = 13	[IAA] i-STAT
265.5 ± 4.78	43.5 ± 0.89	73.1 ± 1.30	195.1 ± 3.81	88.7 ± 1.53	n = 22	[JJE] Ortho Vitros 250/350/950
264.9 ± 5.38	43.8 ± 1.51	72.8 ± 1.79	195.2 ± 4.10	88.9 ± 1.86	n = 11	[JJF] Ortho Vitros 5,1FS
264.8 ± 4.24	43.7 ± 0.98	76.1 ± 1.45	192.1 ± 3.56	91.1 ± 1.90	n = 18	[JGJ] Ortho Vitros 5600
262.9 ± 5.54	42.7 ± 0.74	75.4 ± 1.17	191.2 ± 3.65	89.6 ± 1.78	n = 13	[ROC] Roche cobas c501
266.5 ± 5.94	43.7 ± 1.09	76.3 ± 1.79	192.8 ± 4.74	91.6 ± 2.19	n = 36	[ROT] Roche Cobas INTEGRA
262.5 ± 6.17	43.0 ± 1.09	75.4 ± 1.89	191.1 ± 4.54	90.1 ± 2.19	n = 16	[ROD] Roche MODULAR D/P
259.3 ± 14.77	42.9 ± 2.86	74.2 ± 4.11	187.9 ± 10.08	89.0 ± 4.60	n = 3	[BYE] Siemens ADVIA 1800
269.2 ± 5.35	46.6 ± 0.96	77.0 ± 2.05	194.3 ± 4.69	92.0 ± 2.18	n = 12	[BYB] Siemens ADVIA 2400
268.5 ± 5.56	47.6 ± 1.32	78.1 ± 1.56	195.3 ± 4.10	93.4 ± 2.08	n = 28	[DUE] Siemens Dimension EXL
260.6 ± 6.17	46.2 ± 1.41	75.2 ± 2.17	189.2 ± 5.01	90.8 ± 2.69	n = 32	[DUR] Siemens Dimension RxL
265.4 ± 5.46	47.2 ± 1.62	77.1 ± 1.57	193.8 ± 3.57	92.3 ± 2.19	n = 21	[DUT] Siemens Dimension Vista
263.3 ± 1.37	48.6 ± 1.02	80.0 ± 1.80	192.5 ± 1.86	94.3 ± 1.37	n = 3	[DUX] Siemens Dimension Xpand
268.6 ± 2.87	42.4 ± 0.56	75.3 ± 0.74	192.9 ± 2.29	90.6 ± 1.05	n = 14	<Reagents>
263.8 ± 5.95	44.2 ± 2.17	75.6 ± 2.21	191.7 ± 4.48	90.8 ± 2.24	n = 60	[AB1] Abbott
264.0 ± 7.45	42.9 ± 1.31	75.6 ± 1.81	190.7 ± 4.67	90.7 ± 2.76	n = 48	[BC1] Beckman Coulter
261.4 ± 11.43	44.4 ± 1.69	75.6 ± 3.85	191.8 ± 9.36	89.5 ± 5.15	n = 6	[OL1] Beckman Coulter AU Series
275.0 ± 2.39	66.4 ± 2.64	99.9 ± 5.53	208.2 ± 5.41	113.6 ± 7.79	n = 6	[CR1] Carolina
272.8 ± 1.91	43.7 ± 1.85	77.5 ± 1.02	194.2 ± 0.91	92.5 ± 1.24	n = 6	[HE1] HemoCue
265.9 ± 5.13	43.6 ± 1.27	73.1 ± 1.77	195.8 ± 4.13	89.1 ± 1.97	n = 47	[IA1] i-STAT
264.6 ± 4.09	43.6 ± 0.96	76.1 ± 1.38	192.0 ± 3.43	91.1 ± 1.83	n = 19	[JJ1] Ortho Clinical Diagnostics
266.5 ± 5.80	43.7 ± 1.08	76.2 ± 1.85	192.8 ± 4.67	91.5 ± 2.22	n = 37	[RO4] Roche cobas c311/c501/c502/c701
262.9 ± 5.54	42.7 ± 0.74	75.4 ± 1.17	191.2 ± 3.65	89.6 ± 1.78	n = 13	[RO2] Roche Hitachi and Modular D/P
264.2 ± 7.05	43.3 ± 1.28	75.8 ± 2.08	192.0 ± 5.13	90.7 ± 2.62	n = 23	[RO1] Roche Integra and MIRA
265.4 ± 6.79	46.9 ± 1.53	76.9 ± 2.19	193.0 ± 5.19	92.2 ± 2.53	n = 92	[BY1] Siemens ADVIA/ADVIS Centaur
						[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Urea Nitrogen (mg/dL)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
30.3 ± 1.64	9.9 ± 0.73	19.4 ± 1.20	60.7 ± 3.26	23.2 ± 1.47	n = 372	[---] All Methods & Instruments
28.7 ± 0.51	10.7 ± 0.51	18.3 ± 0.51	60.0 ± 0.00	22.3 ± 0.51	n = 3	<Instruments>
30.1 ± 0.69	10.0 ± 0.00	19.4 ± 0.60	60.0 ± 1.62	23.0 ± 0.55	n = 13	[AXA] Abaxis Piccolo
30.7 ± 0.89	10.1 ± 0.45	19.9 ± 0.49	61.2 ± 1.49	23.6 ± 0.73	n = 50	[ABJ] Abbott Architect c System
31.2 ± 0.84	10.3 ± 0.80	20.0 ± 0.82	62.3 ± 2.04	23.6 ± 0.86	n = 9	[OLC] Beckman Coulter AU Chemistry System
29.0 ± 1.29	9.4 ± 0.73	18.5 ± 0.70	58.5 ± 2.16	22.0 ± 0.00	n = 9	[BCS] Beckman Coulter CX
30.7 ± 0.92	10.7 ± 0.53	19.7 ± 0.65	61.2 ± 1.35	23.8 ± 0.68	n = 21	[BCX] Beckman Coulter LX-20
28.5 ± 1.77	9.1 ± 1.24	18.3 ± 0.93	58.3 ± 1.95	21.7 ± 1.42	n = 24	[BCG] Beckman Coulter UniCel DxC 600
34.0 ± 0.90	9.0 ± 0.00	20.3 ± 0.51	69.7 ± 1.48	26.5 ± 1.02	n = 6	[BCH] Beckman Coulter UniCel DxC 800
27.2 ± 0.67	9.0 ± 0.00	17.0 ± 0.00	53.0 ± 1.19	20.0 ± 0.00	n = 13	[IAA] i-STAT
27.0 ± 0.65	9.0 ± 0.00	16.7 ± 0.50	52.8 ± 1.14	19.4 ± 0.57	n = 22	[JJE] Ortho Vitros 250/350/950
26.9 ± 1.05	8.7 ± 0.62	16.2 ± 1.02	52.7 ± 2.28	19.3 ± 1.05	n = 11	[JJF] Ortho Vitros 5,1FS
30.3 ± 0.51	10.0 ± 0.00	19.3 ± 0.61	60.6 ± 1.28	23.0 ± 0.00	n = 18	[JJG] Ortho Vitros 5600
30.2 ± 0.61	10.0 ± 0.00	19.2 ± 0.61	61.5 ± 2.35	23.2 ± 0.61	n = 10	[ROC] Roche cobas c501
30.8 ± 0.63	10.1 ± 0.41	19.8 ± 0.65	61.6 ± 1.41	23.6 ± 0.65	n = 35	[ROT] Roche Cobas INTEGRA
31.1 ± 0.99	10.3 ± 0.54	20.2 ± 0.66	61.1 ± 1.60	23.9 ± 0.74	n = 16	[ROD] Roche MODULAR D/P
30.7 ± 0.51	10.3 ± 0.51	20.0 ± 0.90	60.9 ± 2.05	24.0 ± 0.90	n = 3	[BYE] Siemens ADVIA 1800
30.9 ± 0.94	9.9 ± 0.44	19.6 ± 0.92	62.4 ± 2.49	23.5 ± 0.93	n = 11	[BYB] Siemens ADVIA 2400
31.4 ± 1.15	10.1 ± 0.76	20.1 ± 0.84	63.1 ± 1.96	23.9 ± 0.86	n = 29	[DUE] Siemens Dimension EXL
30.7 ± 0.97	10.0 ± 0.41	19.4 ± 0.72	61.4 ± 2.09	23.3 ± 0.66	n = 32	[DUR] Siemens Dimension RxL
31.1 ± 0.72	10.1 ± 0.56	19.7 ± 0.88	62.1 ± 1.94	23.5 ± 0.77	n = 20	[DUT] Siemens Dimension Vista
28.7 ± 0.51	10.7 ± 0.51	18.3 ± 0.51	60.0 ± 0.00	22.3 ± 0.51	n = 3	[DUX] Siemens Dimension Xpand
30.1 ± 0.69	10.0 ± 0.00	19.4 ± 0.60	60.0 ± 1.62	23.0 ± 0.55	n = 13	<Reagents>
29.7 ± 1.66	9.9 ± 1.08	19.0 ± 1.06	59.8 ± 2.40	22.7 ± 1.48	n = 59	[AB1] Abbott
30.6 ± 0.87	10.1 ± 0.40	19.8 ± 0.50	61.2 ± 1.52	23.6 ± 0.74	n = 49	[BC1] Beckman Coulter AU Series
30.6 ± 1.37	10.4 ± 1.09	19.6 ± 1.09	61.6 ± 2.61	23.4 ± 1.09	n = 5	[OL1] Beckman Coulter AU Series
34.0 ± 0.90	9.0 ± 0.00	20.3 ± 0.51	69.7 ± 1.48	26.5 ± 1.02	n = 6	[CR1] Carolina
27.1 ± 0.75	9.0 ± 0.00	16.8 ± 0.58	53.0 ± 1.32	19.7 ± 0.67	n = 48	[IA1] i-STAT
30.3 ± 0.50	10.0 ± 0.00	19.3 ± 0.60	60.5 ± 1.31	23.0 ± 0.00	n = 19	[JJ1] Ortho Clinical Diagnostics
30.8 ± 0.66	10.1 ± 0.39	19.8 ± 0.63	61.6 ± 1.41	23.7 ± 0.64	n = 36	[RO4] Roche cobas c311/c501/c502/c701
30.2 ± 0.61	10.0 ± 0.00	19.2 ± 0.61	61.5 ± 2.35	23.2 ± 0.61	n = 10	[RO2] Roche Hitachi and Modular D/P
31.2 ± 0.99	10.3 ± 0.52	20.2 ± 0.69	61.5 ± 1.74	24.0 ± 0.73	n = 23	[RO1] Roche Integra and MIRA
31.0 ± 1.00	10.0 ± 0.58	19.7 ± 0.85	62.2 ± 2.19	23.5 ± 0.82	n = 91	[BY1] Siemens ADVIA/ADVISIA Centaur
						[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Creatinine (mg/dL)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
3.02 ± 0.18	1.07 ± 0.13	1.10 ± 0.09	4.62 ± 0.17	1.36 ± 0.11	n = 380	[---] All Methods & Instruments
3.01 ± 0.20	1.06 ± 0.13	1.09 ± 0.10	4.67 ± 0.32	1.35 ± 0.12	n = 199	[---] All IDMS Traceable Methods
3.03 ± 0.17	1.09 ± 0.12	1.11 ± 0.08	4.64 ± 0.16	1.37 ± 0.11	n = 178	[---] All Non-IDMS Traceable Methods
3.02 ± 0.15	1.09 ± 0.11	1.10 ± 0.07	4.64 ± 0.15	1.36 ± 0.10	n = 156	[-G-] Alkaline picrate/Jaffe
2.95 ± 0.14	1.07 ± 0.14	1.07 ± 0.08	4.59 ± 0.15	1.31 ± 0.09	n = 136	[-H-] Alkaline picrate/Jaffe - IDMS calibration
3.12 ± 0.24	1.06 ± 0.14	1.17 ± 0.12	5.13 ± 0.51	1.42 ± 0.16	n = 22	[-I-] Enzymatic
3.16 ± 0.24	1.04 ± 0.10	1.16 ± 0.10	5.19 ± 0.54	1.43 ± 0.14	n = 63	[-J-] Enzymatic - IDMS-traceable calibration
2.98 ± 0.06	1.04 ± 0.08	1.11 ± 0.09	4.64 ± 0.36	1.33 ± 0.10	n = 3	[-Z-] Other
<Instruments>						
2.83 ± 0.05	0.90 ± 0.18	1.02 ± 0.09	4.57 ± 0.18	1.24 ± 0.05	n = 3	[AXA] Abaxis Piccolo
3.36 ± 0.12	1.24 ± 0.08	1.12 ± 0.05	4.94 ± 0.12	1.50 ± 0.00	n = 13	[ABJ] Abbott Architect c System
2.95 ± 0.07	1.10 ± 0.02	1.08 ± 0.03	4.53 ± 0.10	1.32 ± 0.04	n = 52	[OLC] Beckman Coulter AU Chemistry System
2.88 ± 0.09	1.10 ± 0.09	1.12 ± 0.06	4.55 ± 0.14	1.35 ± 0.06	n = 9	[BCS] Beckman Coulter CX
2.91 ± 0.03	0.92 ± 0.02	1.01 ± 0.03	4.64 ± 0.07	1.25 ± 0.04	n = 9	[BCX] Beckman Coulter LX-20
2.81 ± 0.06	0.95 ± 0.07	1.00 ± 0.00	4.54 ± 0.13	1.23 ± 0.07	n = 22	[BCG] Beckman Coulter UniCel DxC 600
2.88 ± 0.08	0.94 ± 0.06	1.04 ± 0.05	4.62 ± 0.08	1.27 ± 0.05	n = 24	[BCH] Beckman Coulter UniCel DxC 800
3.40 ± 0.15	1.04 ± 0.06	1.16 ± 0.06	5.44 ± 0.30	1.45 ± 0.08	n = 5	[IAA] i-STAT
3.22 ± 0.11	1.13 ± 0.05	1.26 ± 0.05	5.42 ± 0.21	1.55 ± 0.05	n = 13	[JJE] Ortho Vitros 250/350/950
3.33 ± 0.08	1.08 ± 0.06	1.20 ± 0.00	5.58 ± 0.12	1.50 ± 0.01	n = 22	[JJF] Ortho Vitros 5,1FS
3.29 ± 0.08	1.07 ± 0.06	1.20 ± 0.01	5.58 ± 0.14	1.50 ± 0.06	n = 11	[JJG] Ortho Vitros 5600
2.93 ± 0.10	1.01 ± 0.12	1.07 ± 0.06	4.60 ± 0.14	1.30 ± 0.08	n = 20	[ROC] Roche cobas c501
2.91 ± 0.09	1.04 ± 0.10	1.10 ± 0.00	4.59 ± 0.18	1.30 ± 0.00	n = 11	[ROT] Roche Cobas INTEGRA
3.09 ± 0.18	1.17 ± 0.18	1.14 ± 0.08	4.73 ± 0.14	1.41 ± 0.12	n = 35	[ROD] Roche MODULAR D/P
3.06 ± 0.08	1.21 ± 0.07	1.10 ± 0.00	4.62 ± 0.06	1.40 ± 0.00	n = 16	[BYE] Siemens ADVIA 1800
2.98 ± 0.13	1.14 ± 0.10	1.02 ± 0.08	4.51 ± 0.15	1.30 ± 0.07	n = 3	[BYB] Siemens ADVIA 2400
3.02 ± 0.09	1.06 ± 0.06	1.10 ± 0.00	4.59 ± 0.09	1.34 ± 0.06	n = 11	[DUE] Siemens Dimension EXL
3.05 ± 0.11	1.10 ± 0.11	1.12 ± 0.09	4.66 ± 0.12	1.38 ± 0.09	n = 30	[DUR] Siemens Dimension RxL
3.05 ± 0.13	1.05 ± 0.10	1.10 ± 0.08	4.65 ± 0.12	1.36 ± 0.10	n = 33	[DUT] Siemens Dimension Vista
2.99 ± 0.11	1.05 ± 0.13	1.08 ± 0.10	4.60 ± 0.11	1.34 ± 0.11	n = 20	[DUX] Siemens Dimension Xpand
<Reagents>						
2.83 ± 0.05	0.90 ± 0.18	1.02 ± 0.09	4.57 ± 0.18	1.24 ± 0.05	n = 3	[AX1] Abaxis
3.37 ± 0.11	1.23 ± 0.10	1.13 ± 0.05	4.96 ± 0.12	1.50 ± 0.00	n = 14	[AB1] Abbott
2.86 ± 0.08	0.94 ± 0.06	1.02 ± 0.06	4.60 ± 0.10	1.26 ± 0.06	n = 60	[BC1] Beckman Coulter
2.96 ± 0.07	1.10 ± 0.02	1.08 ± 0.03	4.53 ± 0.10	1.32 ± 0.04	n = 49	[OL1] Beckman Coulter AU Series
2.94 ± 0.10	1.13 ± 0.09	1.12 ± 0.05	4.55 ± 0.18	1.34 ± 0.05	n = 6	[CR1] Carolina
3.40 ± 0.18	1.05 ± 0.06	1.15 ± 0.06	5.53 ± 0.27	1.43 ± 0.09	n = 4	[IA1] i-STAT
3.29 ± 0.10	1.09 ± 0.07	1.23 ± 0.05	5.55 ± 0.16	1.52 ± 0.06	n = 48	[JJ1] Ortho Clinical Diagnostics
2.93 ± 0.10	1.02 ± 0.12	1.08 ± 0.06	4.60 ± 0.14	1.31 ± 0.08	n = 21	[RO4] Roche cobas c311/c501/c502/c701
3.09 ± 0.18	1.16 ± 0.18	1.14 ± 0.08	4.73 ± 0.13	1.41 ± 0.12	n = 36	[RO2] Roche Hitachi and Modular D/P
2.91 ± 0.09	1.04 ± 0.10	1.10 ± 0.00	4.59 ± 0.18	1.30 ± 0.00	n = 11	[R01] Roche Integra and MIRA
3.06 ± 0.09	1.20 ± 0.08	1.10 ± 0.00	4.62 ± 0.07	1.38 ± 0.04	n = 23	[BY1] Siemens ADVIA/ADVIA Centaur
3.03 ± 0.12	1.07 ± 0.11	1.10 ± 0.09	4.63 ± 0.12	1.36 ± 0.10	n = 92	[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Estimated Glomerular Filtration Rate (mL/min/1.73 m²)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
22.0 ± 1.79	71.1 ± 11.39	70.2 ± 8.76	13.2 ± 1.18	54.8 ± 5.39	n = 281	[---] All Methods & Instruments
21.4 ± 1.98	68.9 ± 11.65	66.9 ± 8.97	12.8 ± 1.42	53.2 ± 5.75	n = 160	[-A-] IDMS-traceable MDRD Study Equation
22.6 ± 1.26	72.2 ± 10.18	72.5 ± 6.01	13.7 ± 0.68	56.2 ± 4.09	n = 102	[-B-] Original MDRD Study Equation (4-variable)
22.3 ± 1.38	73.7 ± 8.50	75.7 ± 4.70	13.2 ± 0.68	58.8 ± 3.75	n = 15	[-F-] CKD-EPI Equation
20.5 ± 1.86	85.5 ± 1.71	76.5 ± 8.55	12.0 ± 0.90	55.1 ± 5.22	n = 3	[-Z-] Other

Target values and allowable ranges for Estimated Glomerular Filtration Rate (eGFR) :

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Method
22 (18-25)	72 (61-90)	70 (59-88)	13 (11-15)	55 (46-63)	IDMS-traceable MDRD Study Equation
23 (19-27)	74 (63-93)	73 (61-91)	14 (11-16)	57 (48-66)	Original MDRD Study Equation
22 (18-26)	78 (65-97)	75 (63-94)	13 (11-15)	58 (49-67)	CKD-EPI Equation
31 (26-36)	88 (66-111)	86 (64-108)	21 (17-24)	70 (59-87)	Cockcroft-Gault Equation

Laboratories were asked to report Estimated Glomerular Filtration Rate (eGFR) for samples C61-C65 for a 57-year-old non-African American man weighing 82 kg.

Target values for eGFR calculated by the MDRD Study Equations were derived from participant mean values for serum creatinine for both conventional and isotope dilution mass spectroscopy (IDMS) calibration methods and application of the appropriate 4-variable MDRD Study equation.

Target values for eGFR calculated by the CKD-EPI equation were derived from participant mean values for IDMS-traceable serum creatinine methods and application of the CKD-EPI equation.

Target values for eGFR calculated by the Cockcroft-Gault equation were derived from all-method mean values for serum creatinine and application of the Cockcroft-Gault formula.

Allowable ranges are ±/- 15% of the target eGFR for eGFR ≤ 59 mL/min; ±/- 25% of the target eGFR for eGFR > 80 mL/min; and a range of -15% to ±25% of the target eGFR for eGFR = 60-80 mL/min.

Note: the NKDEP recommends reporting estimated GFR values greater than or equal to 60 mL/min/1.73 m² as "> 60 mL/min/1.73 m²" and not as an exact number. However, ranges exceeding 60 mL/min are provided as a relative indicator of acceptability for laboratories that report numeric results above that threshold. Note that some laboratories reported results > 60 mL/min/1.73 m² for samples C62, C63, and C65. These data were removed from the calculations of mean and standard deviation since their inclusion would have skewed results. Participant results for specimens C62, C63, and C65 reported as >60 mL/min/1.73 m² were considered acceptable performance.

Summary of Participant Performance (Mean and Standard Deviation)

Uric Acid (mg/dL)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
2.73 ± 0.20	9.91 ± 0.49	5.73 ± 0.30	4.08 ± 0.21	6.85 ± 0.35	n = 334	[---] All Methods & Instruments
2.98 ± 0.10	10.24 ± 0.07	5.98 ± 0.10	4.29 ± 0.08	7.14 ± 0.12	n = 12	<Instruments>
3.05 ± 0.09	10.90 ± 0.35	6.47 ± 0.16	4.51 ± 0.12	7.67 ± 0.18	n = 48	[ABJ] Abbott Architect c System
2.64 ± 0.12	9.48 ± 0.25	5.71 ± 0.20	4.10 ± 0.16	6.87 ± 0.26	n = 9	[OLC] Beckman Coulter AU Chemistry System
2.74 ± 0.09	9.58 ± 0.14	5.83 ± 0.10	4.20 ± 0.07	6.89 ± 0.14	n = 9	[BCS] Beckman Coulter CX
2.63 ± 0.06	9.59 ± 0.21	5.71 ± 0.13	4.12 ± 0.09	6.82 ± 0.12	n = 18	[BCX] Beckman Coulter LX-20
2.61 ± 0.05	9.59 ± 0.11	5.72 ± 0.09	4.10 ± 0.08	6.78 ± 0.09	n = 22	[BCG] Beckman Coulter UniCel DxC 600
2.52 ± 0.07	9.79 ± 0.17	5.60 ± 0.11	4.05 ± 0.06	6.73 ± 0.10	n = 11	[BCH] Beckman Coulter UniCel DxC 800
2.57 ± 0.06	9.71 ± 0.19	5.57 ± 0.13	3.99 ± 0.10	6.65 ± 0.15	n = 22	[JJE] Ortho Vitros 250/350/950
2.54 ± 0.07	9.77 ± 0.22	5.64 ± 0.12	4.01 ± 0.09	6.70 ± 0.14	n = 11	[JJF] Ortho Vitros 5,1FS
2.69 ± 0.06	10.29 ± 0.20	5.91 ± 0.14	4.14 ± 0.10	7.07 ± 0.13	n = 18	[JJG] Ortho Vitros 5600
2.55 ± 0.06	9.86 ± 0.21	5.70 ± 0.13	3.95 ± 0.09	6.82 ± 0.14	n = 8	[ROC] Roche cobas c501
2.60 ± 0.05	10.07 ± 0.15	5.75 ± 0.11	4.00 ± 0.08	6.90 ± 0.11	n = 34	[ROT] Roche Cobas INTEGRA
2.62 ± 0.08	9.97 ± 0.15	5.74 ± 0.09	4.08 ± 0.08	6.89 ± 0.09	n = 16	[ROD] Roche MODULAR D/P
2.67 ± 0.05	10.13 ± 0.31	5.82 ± 0.15	4.10 ± 0.09	6.90 ± 0.18	n = 3	[BYE] Siemens ADVIA 1800
2.83 ± 0.11	9.90 ± 0.16	5.61 ± 0.14	4.04 ± 0.11	6.72 ± 0.08	n = 11	[BYB] Siemens ADVIA 2400
2.88 ± 0.12	9.90 ± 0.21	5.59 ± 0.15	4.02 ± 0.13	6.73 ± 0.16	n = 26	[DUE] Siemens Dimension EXL
2.79 ± 0.10	9.30 ± 0.19	5.36 ± 0.11	3.80 ± 0.10	6.36 ± 0.11	n = 31	[DUR] Siemens Dimension RxL
2.87 ± 0.08	9.92 ± 0.22	5.63 ± 0.14	4.10 ± 0.00	6.74 ± 0.11	n = 14	[DUT] Siemens Dimension Vista
2.98 ± 0.10	10.24 ± 0.07	5.98 ± 0.10	4.29 ± 0.08	7.14 ± 0.12	n = 12	[DUX] Siemens Dimension Xpand
2.64 ± 0.08	9.58 ± 0.17	5.73 ± 0.12	4.12 ± 0.09	6.81 ± 0.12	n = 55	<Reagents>
3.05 ± 0.09	10.91 ± 0.34	6.47 ± 0.15	4.51 ± 0.11	7.67 ± 0.17	n = 47	[BC1] Beckman Coulter
2.60 ± 0.10	9.53 ± 0.14	5.71 ± 0.24	4.07 ± 0.16	6.97 ± 0.38	n = 5	[OL1] Beckman Coulter AU Series
2.55 ± 0.07	9.74 ± 0.19	5.59 ± 0.13	4.01 ± 0.10	6.68 ± 0.14	n = 44	[CR1] Carolina
2.69 ± 0.06	10.29 ± 0.20	5.91 ± 0.14	4.14 ± 0.10	7.07 ± 0.13	n = 18	[JJ1] Ortho Clinical Diagnostics
2.60 ± 0.04	10.07 ± 0.15	5.75 ± 0.11	4.00 ± 0.08	6.90 ± 0.11	n = 35	[RO4] Roche cobas c311/c501/c502/c701
2.55 ± 0.06	9.86 ± 0.21	5.70 ± 0.13	3.95 ± 0.09	6.82 ± 0.14	n = 8	[RO2] Roche Hitachi and Modular D/P
2.64 ± 0.08	10.03 ± 0.20	5.77 ± 0.11	4.08 ± 0.08	6.92 ± 0.11	n = 23	[RO1] Roche Integra and MIRA
2.84 ± 0.11	9.69 ± 0.38	5.51 ± 0.19	3.95 ± 0.17	6.59 ± 0.23	n = 81	[BY1] Siemens ADVIA/ADVIS Centaur
						[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Bilirubin (mg/dL)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
1.65 ± 0.18	0.67 ± 0.17	1.01 ± 0.17	3.13 ± 0.22	1.27 ± 0.19	n = 361	[---] All Methods & Instruments
1.60 ± 0.00	0.63 ± 0.05	1.00 ± 0.00	3.10 ± 0.00	1.20 ± 0.09	n = 3	<Instruments>
1.80 ± 0.09	0.72 ± 0.07	1.13 ± 0.12	3.41 ± 0.12	1.41 ± 0.11	n = 13	[AXA] Abaxis Piccolo
1.67 ± 0.07	0.78 ± 0.05	1.12 ± 0.08	3.07 ± 0.12	1.34 ± 0.10	n = 50	[ABJ] Abbott Architect c System
1.78 ± 0.21	0.86 ± 0.23	1.17 ± 0.14	3.29 ± 0.18	1.49 ± 0.18	n = 9	[OLC] Beckman Coulter AU Chemistry System
1.96 ± 0.10	0.96 ± 0.09	1.23 ± 0.09	3.42 ± 0.12	1.53 ± 0.18	n = 9	[BCS] Beckman Coulter CX
1.82 ± 0.18	0.91 ± 0.16	1.17 ± 0.18	3.38 ± 0.16	1.43 ± 0.19	n = 21	[BCX] Beckman Coulter LX-20
1.80 ± 0.12	0.91 ± 0.11	1.12 ± 0.13	3.39 ± 0.15	1.40 ± 0.14	n = 24	[BCG] Beckman Coulter UniCel DxC 600
1.86 ± 0.08	0.72 ± 0.08	1.15 ± 0.10	3.15 ± 0.10	1.47 ± 0.08	n = 12	[BCH] Beckman Coulter UniCel DxC 800
1.74 ± 0.15	0.57 ± 0.13	1.01 ± 0.16	3.08 ± 0.18	1.28 ± 0.17	n = 22	[JJE] Ortho Vitros 250/350/950
1.74 ± 0.13	0.60 ± 0.07	1.03 ± 0.07	3.07 ± 0.19	1.33 ± 0.11	n = 11	[JJF] Ortho Vitros 5,1FS
1.38 ± 0.06	0.45 ± 0.06	0.79 ± 0.08	2.86 ± 0.11	1.04 ± 0.09	n = 17	[JJG] Ortho Vitros 5600
1.31 ± 0.10	0.45 ± 0.06	0.81 ± 0.11	2.77 ± 0.12	1.02 ± 0.13	n = 12	[ROC] Roche cobas c501
1.45 ± 0.10	0.50 ± 0.07	0.84 ± 0.11	2.99 ± 0.14	1.07 ± 0.12	n = 35	[ROT] Roche Cobas INTEGRA
1.70 ± 0.06	0.69 ± 0.06	1.05 ± 0.10	3.30 ± 0.15	1.32 ± 0.13	n = 16	[ROD] Roche MODULAR D/P
1.70 ± 0.09	0.67 ± 0.05	1.04 ± 0.10	3.34 ± 0.10	1.27 ± 0.14	n = 3	[BYE] Siemens ADVIA 1800
1.62 ± 0.05	0.63 ± 0.06	0.96 ± 0.08	3.12 ± 0.08	1.25 ± 0.06	n = 11	[BYB] Siemens ADVIA 2400
1.60 ± 0.09	0.62 ± 0.08	0.97 ± 0.08	3.13 ± 0.12	1.22 ± 0.10	n = 29	[DUE] Siemens Dimension EXL
1.61 ± 0.07	0.67 ± 0.07	0.99 ± 0.08	3.10 ± 0.11	1.26 ± 0.10	n = 32	[DUR] Siemens Dimension RxL
1.65 ± 0.09	0.64 ± 0.08	0.97 ± 0.08	3.12 ± 0.13	1.23 ± 0.11	n = 20	[DUT] Siemens Dimension Vista
1.60 ± 0.00	0.63 ± 0.05	1.00 ± 0.00	3.10 ± 0.00	1.20 ± 0.09	n = 3	[DUX] Siemens Dimension Xpand
1.80 ± 0.09	0.72 ± 0.07	1.13 ± 0.12	3.41 ± 0.12	1.41 ± 0.11	n = 13	<Reagents>
1.83 ± 0.15	0.92 ± 0.13	1.17 ± 0.14	3.39 ± 0.16	1.45 ± 0.18	n = 60	[BC1] Abbott
1.67 ± 0.07	0.79 ± 0.04	1.12 ± 0.08	3.07 ± 0.12	1.35 ± 0.10	n = 48	[OL1] Beckman Coulter AU Series
1.60 ± 0.16	0.65 ± 0.14	1.02 ± 0.14	3.21 ± 0.11	1.32 ± 0.14	n = 6	[CR1] Carolina
1.77 ± 0.14	0.61 ± 0.14	1.06 ± 0.15	3.11 ± 0.17	1.34 ± 0.17	n = 47	[JJ1] Ortho Clinical Diagnostics
1.38 ± 0.07	0.45 ± 0.06	0.78 ± 0.08	2.85 ± 0.12	1.03 ± 0.10	n = 18	[RO4] Roche cobas c311/c501/c502/c701
1.45 ± 0.10	0.50 ± 0.07	0.84 ± 0.11	2.99 ± 0.14	1.07 ± 0.12	n = 35	[RO2] Roche Hitachi and Modular D/P
1.31 ± 0.10	0.45 ± 0.06	0.81 ± 0.11	2.77 ± 0.12	1.02 ± 0.13	n = 12	[RO1] Roche Integra and MIRA
1.70 ± 0.07	0.69 ± 0.05	1.05 ± 0.09	3.31 ± 0.14	1.31 ± 0.13	n = 23	[BY1] Siemens ADVIA/ADVIS Centaur
1.62 ± 0.08	0.64 ± 0.08	0.98 ± 0.08	3.12 ± 0.12	1.24 ± 0.10	n = 91	[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Phosphorus (mg/dL)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
2.84 ± 0.16	3.97 ± 0.18	4.86 ± 0.17	2.51 ± 0.16	5.82 ± 0.22	n = 338	[---] All Methods & Instruments
2.82 ± 0.07	3.95 ± 0.08	4.83 ± 0.07	2.45 ± 0.08	5.75 ± 0.07	n = 13	<Instruments>
2.75 ± 0.08	3.84 ± 0.11	4.72 ± 0.11	2.44 ± 0.09	5.62 ± 0.13	n = 49	[ABJ] Abbott Architect c System
3.08 ± 0.20	4.09 ± 0.25	4.90 ± 0.19	2.64 ± 0.16	5.85 ± 0.20	n = 8	[OLC] Beckman Coulter AU Chemistry System
2.93 ± 0.09	4.00 ± 0.07	4.96 ± 0.07	2.68 ± 0.07	5.99 ± 0.08	n = 9	[BCS] Beckman Coulter CX
2.83 ± 0.12	3.98 ± 0.14	4.85 ± 0.15	2.52 ± 0.13	5.81 ± 0.21	n = 19	[BCX] Beckman Coulter LX-20
2.92 ± 0.08	4.02 ± 0.08	4.97 ± 0.08	2.66 ± 0.06	5.98 ± 0.08	n = 23	[BCG] Beckman Coulter UniCel DxC 600
3.14 ± 0.09	4.28 ± 0.15	5.11 ± 0.12	2.84 ± 0.09	6.20 ± 0.15	n = 11	[BCH] Beckman Coulter UniCel DxC 800
3.18 ± 0.11	4.33 ± 0.13	5.14 ± 0.14	2.85 ± 0.13	6.24 ± 0.20	n = 22	[JJE] Ortho Vitros 250/350/950
3.11 ± 0.11	4.26 ± 0.12	5.04 ± 0.16	2.76 ± 0.13	6.09 ± 0.16	n = 11	[JJF] Ortho Vitros 5,1FS
2.85 ± 0.06	3.99 ± 0.07	4.89 ± 0.08	2.49 ± 0.08	5.83 ± 0.09	n = 19	[JJG] Ortho Vitros 5600
2.84 ± 0.09	3.98 ± 0.09	4.86 ± 0.12	2.47 ± 0.05	5.80 ± 0.14	n = 9	[ROC] Roche cobas c501
2.84 ± 0.09	4.01 ± 0.11	4.89 ± 0.13	2.52 ± 0.10	5.85 ± 0.12	n = 34	[ROT] Roche Cobas INTEGRA
2.86 ± 0.09	4.02 ± 0.13	4.89 ± 0.10	2.51 ± 0.10	5.86 ± 0.14	n = 16	[ROD] Roche MODULAR D/P
2.76 ± 0.10	3.93 ± 0.14	4.75 ± 0.19	2.46 ± 0.10	5.78 ± 0.15	n = 3	[BYE] Siemens ADVIA 1800
2.72 ± 0.05	3.85 ± 0.06	4.78 ± 0.06	2.40 ± 0.10	5.73 ± 0.07	n = 10	[BYB] Siemens ADVIA 2400
2.80 ± 0.10	3.92 ± 0.10	4.87 ± 0.12	2.46 ± 0.09	5.81 ± 0.16	n = 26	[DUE] Siemens Dimension EXL
2.65 ± 0.09	3.78 ± 0.13	4.69 ± 0.13	2.34 ± 0.09	5.64 ± 0.15	n = 31	[DUR] Siemens Dimension RxL
2.79 ± 0.06	3.91 ± 0.08	4.81 ± 0.08	2.43 ± 0.08	5.76 ± 0.10	n = 17	[DUT] Siemens Dimension Vista
						[DUX] Siemens Dimension Xpand
2.82 ± 0.07	3.95 ± 0.08	4.83 ± 0.07	2.45 ± 0.08	5.75 ± 0.07	n = 13	<Reagents>
2.90 ± 0.11	4.00 ± 0.11	4.93 ± 0.12	2.62 ± 0.10	5.94 ± 0.15	n = 55	[AB1] Abbott
2.75 ± 0.08	3.85 ± 0.11	4.72 ± 0.10	2.44 ± 0.08	5.63 ± 0.12	n = 48	[BC1] Beckman Coulter
3.15 ± 0.24	4.18 ± 0.24	4.96 ± 0.11	2.65 ± 0.16	5.85 ± 0.16	n = 5	[OL1] Beckman Coulter AU Series
3.15 ± 0.11	4.30 ± 0.14	5.11 ± 0.14	2.83 ± 0.13	6.19 ± 0.19	n = 44	[CR1] Carolina
2.85 ± 0.06	3.99 ± 0.07	4.89 ± 0.08	2.49 ± 0.08	5.83 ± 0.09	n = 19	[JJ1] Ortho Clinical Diagnostics
2.84 ± 0.09	4.01 ± 0.11	4.89 ± 0.13	2.52 ± 0.10	5.85 ± 0.12	n = 34	[RO4] Roche cobas c311/c501/c502/c701
2.84 ± 0.09	3.98 ± 0.09	4.86 ± 0.12	2.47 ± 0.05	5.80 ± 0.14	n = 9	[RO2] Roche Hitachi and Modular D/P
2.85 ± 0.09	4.00 ± 0.12	4.88 ± 0.12	2.51 ± 0.10	5.86 ± 0.13	n = 23	[RO1] Roche Integra and MIRA
2.74 ± 0.11	3.87 ± 0.12	4.79 ± 0.13	2.41 ± 0.10	5.73 ± 0.15	n = 83	[BY1] Siemens ADVIA/ADVIS Centaur
						[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Calcium (mg/dL)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
8.91 ± 0.23	10.18 ± 0.28	7.12 ± 0.19	13.80 ± 0.39	8.46 ± 0.22	n = 367	[---] All Methods & Instruments
9.03 ± 0.32	10.16 ± 0.26	7.22 ± 0.15	13.45 ± 0.54	8.44 ± 0.26	n = 3	<Instruments>
9.00 ± 0.13	10.41 ± 0.15	7.19 ± 0.14	13.92 ± 0.13	8.50 ± 0.17	n = 13	[AXA] Abaxis Piccolo
8.89 ± 0.17	10.23 ± 0.16	7.08 ± 0.17	13.72 ± 0.24	8.42 ± 0.16	n = 51	[ABJ] Abbott Architect c System
8.92 ± 0.27	10.23 ± 0.39	7.25 ± 0.13	13.57 ± 0.31	8.48 ± 0.32	n = 9	[OLC] Beckman Coulter AU Chemistry System
8.92 ± 0.17	10.03 ± 0.12	7.20 ± 0.10	13.62 ± 0.09	8.47 ± 0.11	n = 9	[BCS] Beckman Coulter CX
8.79 ± 0.10	9.89 ± 0.13	7.10 ± 0.10	13.50 ± 0.19	8.38 ± 0.11	n = 22	[BCX] Beckman Coulter LX-20
8.84 ± 0.10	9.98 ± 0.10	7.16 ± 0.08	13.55 ± 0.20	8.43 ± 0.10	n = 24	[BCG] Beckman Coulter UniCel DxC 600
9.03 ± 0.24	10.24 ± 0.24	6.98 ± 0.22	13.82 ± 0.23	8.53 ± 0.22	n = 13	[BCH] Beckman Coulter UniCel DxC 800
9.04 ± 0.18	10.14 ± 0.18	6.97 ± 0.15	13.60 ± 0.22	8.45 ± 0.17	n = 22	[JJE] Ortho Vitros 250/350/950
8.96 ± 0.15	10.12 ± 0.17	6.91 ± 0.14	13.64 ± 0.23	8.39 ± 0.16	n = 11	[JJF] Ortho Vitros 5,1FS
9.11 ± 0.23	10.50 ± 0.20	7.17 ± 0.17	14.62 ± 0.23	8.66 ± 0.22	n = 19	[JJG] Ortho Vitros 5600
8.85 ± 0.16	10.30 ± 0.20	7.01 ± 0.15	14.43 ± 0.27	8.43 ± 0.20	n = 11	[ROC] Roche cobas c501
9.16 ± 0.24	10.52 ± 0.25	7.29 ± 0.21	14.26 ± 0.29	8.75 ± 0.23	n = 35	[ROT] Roche Cobas INTEGRA
9.02 ± 0.25	10.34 ± 0.26	7.21 ± 0.26	13.91 ± 0.27	8.59 ± 0.22	n = 16	[ROD] Roche MODULAR D/P
8.98 ± 0.50	10.22 ± 0.32	7.12 ± 0.24	13.58 ± 0.41	8.55 ± 0.36	n = 3	[BYE] Siemens ADVIA 1800
8.74 ± 0.18	10.04 ± 0.21	7.07 ± 0.14	13.79 ± 0.21	8.33 ± 0.10	n = 11	[BYB] Siemens ADVIA 2400
8.81 ± 0.16	10.10 ± 0.20	7.16 ± 0.16	13.80 ± 0.37	8.40 ± 0.16	n = 29	[DUE] Siemens Dimension EXL
8.69 ± 0.16	9.97 ± 0.17	7.05 ± 0.19	13.66 ± 0.21	8.34 ± 0.15	n = 32	[DUR] Siemens Dimension RxL
8.84 ± 0.16	10.12 ± 0.20	7.15 ± 0.16	13.96 ± 0.30	8.44 ± 0.18	n = 20	[DUT] Siemens Dimension Vista
						[DUX] Siemens Dimension Xpand
9.03 ± 0.32	10.16 ± 0.26	7.22 ± 0.15	13.45 ± 0.54	8.44 ± 0.26	n = 3	<Reagents>
9.00 ± 0.13	10.41 ± 0.15	7.19 ± 0.14	13.92 ± 0.13	8.50 ± 0.17	n = 13	[AX1] Abaxis
8.83 ± 0.13	9.95 ± 0.15	7.16 ± 0.10	13.54 ± 0.18	8.41 ± 0.13	n = 60	[AB1] Abbott
8.89 ± 0.16	10.22 ± 0.16	7.08 ± 0.17	13.71 ± 0.24	8.41 ± 0.15	n = 50	[BC1] Beckman Coulter
8.93 ± 0.09	10.33 ± 0.34	7.18 ± 0.13	13.70 ± 0.32	8.49 ± 0.28	n = 5	[OL1] Beckman Coulter AU Series
9.03 ± 0.20	10.17 ± 0.20	6.96 ± 0.18	13.67 ± 0.24	8.46 ± 0.19	n = 47	[CR1] Carolina
9.12 ± 0.18	10.52 ± 0.14	7.17 ± 0.14	14.61 ± 0.23	8.66 ± 0.13	n = 17	[JJ1] Ortho Clinical Diagnostics
9.16 ± 0.24	10.52 ± 0.25	7.29 ± 0.21	14.26 ± 0.29	8.75 ± 0.23	n = 35	[RO4] Roche cobas c311/c501/c502/c701
8.85 ± 0.16	10.30 ± 0.20	7.01 ± 0.15	14.43 ± 0.27	8.43 ± 0.20	n = 11	[RO2] Roche Hitachi and Modular D/P
9.06 ± 0.29	10.38 ± 0.27	7.24 ± 0.27	13.88 ± 0.31	8.63 ± 0.25	n = 23	[RO1] Roche Integra and MIRA
8.77 ± 0.18	10.05 ± 0.21	7.11 ± 0.18	13.78 ± 0.30	8.37 ± 0.16	n = 91	[BY1] Siemens ADVIA/ADVISIA Centaur
						[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Magnesium (mg/dL)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
1.57 ± 0.09	2.74 ± 0.11	1.94 ± 0.09	3.43 ± 0.15	2.32 ± 0.10	n = 331	[---] All Methods & Instruments
1.60 ± 0.08	2.71 ± 0.06	1.98 ± 0.04	3.36 ± 0.12	2.34 ± 0.08	n = 12	<Instruments>
1.53 ± 0.07	2.66 ± 0.11	1.89 ± 0.07	3.31 ± 0.12	2.26 ± 0.07	n = 47	[ABJ] Abbott Architect c System
1.63 ± 0.05	2.83 ± 0.11	2.00 ± 0.00	3.52 ± 0.10	2.43 ± 0.05	n = 8	[OLC] Beckman Coulter AU Chemistry System
1.60 ± 0.00	2.80 ± 0.08	1.98 ± 0.04	3.50 ± 0.08	2.39 ± 0.06	n = 8	[BCS] Beckman Coulter CX
1.63 ± 0.07	2.84 ± 0.07	2.00 ± 0.07	3.53 ± 0.11	2.42 ± 0.05	n = 21	[BCX] Beckman Coulter LX-20
1.62 ± 0.06	2.82 ± 0.09	1.98 ± 0.07	3.48 ± 0.12	2.39 ± 0.08	n = 23	[BCG] Beckman Coulter UniCel DxC 600
1.54 ± 0.06	2.71 ± 0.11	1.92 ± 0.07	3.53 ± 0.11	2.30 ± 0.05	n = 7	[BCH] Beckman Coulter UniCel DxC 800
1.54 ± 0.06	2.68 ± 0.07	1.91 ± 0.07	3.45 ± 0.09	2.28 ± 0.05	n = 22	[JJE] Ortho Vitros 250/350/950
1.56 ± 0.06	2.69 ± 0.07	1.91 ± 0.05	3.46 ± 0.08	2.28 ± 0.05	n = 10	[JJF] Ortho Vitros 5,1FS
1.58 ± 0.06	2.61 ± 0.08	1.92 ± 0.06	3.23 ± 0.09	2.30 ± 0.00	n = 17	[JJG] Ortho Vitros 5600
1.57 ± 0.09	2.66 ± 0.07	1.91 ± 0.05	3.30 ± 0.08	2.30 ± 0.07	n = 9	[ROC] Roche cobas c501
1.60 ± 0.00	2.77 ± 0.07	1.99 ± 0.08	3.47 ± 0.09	2.37 ± 0.06	n = 32	[ROT] Roche Cobas INTEGRA
1.67 ± 0.08	2.84 ± 0.11	2.02 ± 0.09	3.58 ± 0.13	2.44 ± 0.06	n = 16	[ROD] Roche MODULAR D/P
1.67 ± 0.05	2.90 ± 0.09	2.07 ± 0.05	3.63 ± 0.14	2.43 ± 0.05	n = 3	[BYE] Siemens ADVIA 1800
1.47 ± 0.08	2.73 ± 0.09	1.86 ± 0.07	3.37 ± 0.12	2.25 ± 0.07	n = 9	[BYB] Siemens ADVIA 2400
1.52 ± 0.07	2.78 ± 0.07	1.90 ± 0.00	3.44 ± 0.09	2.29 ± 0.07	n = 28	[DUE] Siemens Dimension EXL
1.48 ± 0.09	2.73 ± 0.09	1.87 ± 0.10	3.42 ± 0.13	2.27 ± 0.12	n = 32	[DUR] Siemens Dimension RxL
1.50 ± 0.00	2.72 ± 0.09	1.90 ± 0.00	3.41 ± 0.10	2.30 ± 0.08	n = 18	[DUT] Siemens Dimension Vista
1.60 ± 0.08	2.71 ± 0.06	1.98 ± 0.04	3.36 ± 0.12	2.34 ± 0.08	n = 12	[DUX] Siemens Dimension Xpand
1.62 ± 0.06	2.83 ± 0.08	1.99 ± 0.06	3.51 ± 0.10	2.40 ± 0.06	n = 56	<Reagents>
1.54 ± 0.07	2.67 ± 0.11	1.89 ± 0.07	3.32 ± 0.13	2.27 ± 0.08	n = 47	[BC1] Beckman Coulter
1.68 ± 0.04	2.75 ± 0.12	2.02 ± 0.04	3.42 ± 0.13	2.42 ± 0.04	n = 4	[OL1] Beckman Coulter AU Series
1.55 ± 0.06	2.69 ± 0.08	1.91 ± 0.06	3.47 ± 0.10	2.28 ± 0.05	n = 39	[CR1] Carolina
1.58 ± 0.06	2.61 ± 0.08	1.92 ± 0.06	3.23 ± 0.09	2.30 ± 0.00	n = 17	[JJ1] Ortho Clinical Diagnostics
1.60 ± 0.00	2.77 ± 0.08	1.98 ± 0.08	3.46 ± 0.10	2.36 ± 0.06	n = 32	[RO4] Roche cobas c311/c501/c502/c701
1.57 ± 0.09	2.67 ± 0.09	1.92 ± 0.07	3.31 ± 0.11	2.31 ± 0.08	n = 9	[RO2] Roche Hitachi and Modular D/P
1.67 ± 0.07	2.86 ± 0.10	2.05 ± 0.09	3.60 ± 0.14	2.44 ± 0.07	n = 24	[RO1] Roche Integra and MIRA
1.50 ± 0.08	2.75 ± 0.09	1.90 ± 0.08	3.42 ± 0.11	2.28 ± 0.09	n = 85	[BY1] Siemens ADVIA/ADVISIA Centaur
						[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Iron ($\mu\text{g/dL}$)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
109.6 ± 12.34	132.2 ± 15.06	94.7 ± 4.12	120.1 ± 7.55	113.8 ± 5.90	n = 269	[---] All Methods & Instruments
64.7 ± 1.38	86.4 ± 23.77	73.8 ± 1.45	87.2 ± 1.92	87.5 ± 1.51	n = 13	<Instruments>
110.2 ± 3.52	132.9 ± 3.77	95.5 ± 2.79	121.4 ± 3.79	115.1 ± 3.23	n = 45	[ABJ] Abbott Architect c System
90.8 ± 5.77	109.7 ± 6.90	87.5 ± 6.24	110.7 ± 6.52	104.9 ± 3.71	n = 6	[OLC] Beckman Coulter AU Chemistry System
95.0 ± 6.99	112.6 ± 6.43	92.3 ± 3.71	113.6 ± 5.73	109.7 ± 5.08	n = 8	[BCS] Beckman Coulter CX
92.3 ± 3.86	111.3 ± 3.92	92.7 ± 3.16	114.1 ± 3.98	110.6 ± 4.23	n = 12	[BCX] Beckman Coulter LX-20
92.5 ± 3.17	110.9 ± 3.16	91.2 ± 2.42	112.5 ± 3.08	108.1 ± 2.39	n = 17	[BCG] Beckman Coulter UniCel DxC 600
132.8 ± 8.58	161.4 ± 6.91	98.1 ± 3.94	139.8 ± 8.42	124.9 ± 3.58	n = 5	[BCH] Beckman Coulter UniCel DxC 800
131.4 ± 3.26	160.7 ± 4.18	99.0 ± 5.09	137.3 ± 5.25	124.0 ± 4.15	n = 20	[JJE] Ortho Vitros 250/350/950
131.2 ± 4.05	161.5 ± 4.52	98.4 ± 3.51	135.3 ± 6.69	123.6 ± 3.34	n = 11	[JJF] Ortho Vitros 5,1FS
116.6 ± 2.07	138.6 ± 2.41	97.7 ± 2.58	123.3 ± 2.50	117.2 ± 2.67	n = 10	[ROC] Roche cobas c501
115.4 ± 2.12	139.4 ± 1.68	99.3 ± 3.40	124.3 ± 1.77	117.3 ± 2.08	n = 8	[ROT] Roche Cobas INTEGRA
112.9 ± 1.85	135.7 ± 2.11	95.1 ± 1.66	121.3 ± 1.82	113.9 ± 1.85	n = 33	[ROD] Roche MODULAR D/P
113.4 ± 3.25	136.7 ± 2.72	96.2 ± 3.07	122.4 ± 3.23	115.3 ± 3.15	n = 15	[BYE] Siemens ADVIA 1800
111.1 ± 6.08	134.9 ± 6.58	93.4 ± 5.58	120.7 ± 5.97	113.0 ± 6.31	n = 3	[BYB] Siemens ADVIA 2400
105.0 ± 1.60	129.6 ± 2.07	92.4 ± 0.94	117.0 ± 1.92	111.2 ± 1.22	n = 6	[DUE] Siemens Dimension EXL
104.2 ± 1.92	127.8 ± 2.37	91.8 ± 1.65	116.1 ± 2.05	110.5 ± 1.87	n = 17	[DUR] Siemens Dimension RxL
104.8 ± 2.62	128.4 ± 3.07	92.6 ± 2.92	116.6 ± 3.10	110.9 ± 3.16	n = 27	[DUT] Siemens Dimension Vista
105.8 ± 1.27	129.0 ± 0.00	92.2 ± 0.41	116.7 ± 0.82	111.7 ± 1.51	n = 4	[DUX] Siemens Dimension Xpand
118.5 ± 2.74	141.5 ± 2.74	97.0 ± 2.70	121.7 ± 2.26	116.0 ± 2.70	n = 3	<Reagents>
64.8 ± 1.36	80.1 ± 1.70	73.9 ± 1.45	87.2 ± 1.92	87.5 ± 1.50	n = 10	[AB3] Abbott-Iron/6K95
92.4 ± 4.18	110.9 ± 4.17	91.7 ± 2.83	113.2 ± 3.78	108.8 ± 3.67	n = 39	[AB2] Abbott-Iron/7D68
110.3 ± 3.69	133.1 ± 4.05	95.6 ± 2.98	122.0 ± 3.75	115.6 ± 3.38	n = 36	[BC1] Beckman Coulter
93.6 ± 6.11	113.9 ± 8.27	85.4 ± 6.83	108.6 ± 6.59	104.3 ± 4.21	n = 4	[OL1] Beckman Coulter AU Series
111.2 ± 4.89	134.6 ± 4.72	96.2 ± 4.10	121.4 ± 6.14	116.5 ± 4.61	n = 3	[CR1] Carolina
110.2 ± 2.70	132.1 ± 2.96	95.5 ± 2.10	119.4 ± 3.05	113.4 ± 2.41	n = 7	[DG1] Diagnostic Chemicals Ltd - Endpoint
131.3 ± 4.26	160.9 ± 4.74	98.7 ± 4.52	137.0 ± 6.09	124.0 ± 3.81	n = 36	[GZ1] Genzyme
116.6 ± 2.07	138.6 ± 2.41	97.7 ± 2.58	123.3 ± 2.50	117.2 ± 2.67	n = 10	[JJ1] Ortho Clinical Diagnostics
113.0 ± 1.85	135.7 ± 2.11	95.2 ± 1.67	121.3 ± 1.85	113.9 ± 1.89	n = 33	[RO4] Roche cobas c311/c501/c502/c701
115.4 ± 2.12	139.4 ± 1.68	99.3 ± 3.40	124.3 ± 1.77	117.3 ± 2.08	n = 8	[RO2] Roche Hitachi and Modular D/P
113.7 ± 3.03	137.2 ± 2.68	96.4 ± 2.72	122.7 ± 2.96	115.5 ± 2.99	n = 21	[RO1] Roche Integra and MIRA
104.7 ± 2.27	128.4 ± 2.76	92.2 ± 2.17	116.5 ± 2.53	110.9 ± 2.44	n = 51	[BY1] Siemens ADVIA/ADVISIA Centaur
						[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Sodium (mmol/L)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
138.0 ± 2.22	135.9 ± 2.57	128.7 ± 1.67	147.1 ± 2.18	153.3 ± 2.17	n = 374	[---] All Methods & Instruments
138.0 ± 0.90	134.0 ± 0.90	128.3 ± 0.51	142.0 ± 0.00	152.0 ± 0.00	n = 3	<Instruments>
138.2 ± 1.70	136.7 ± 0.79	127.6 ± 1.61	147.4 ± 1.80	153.3 ± 1.47	n = 13	[AXA] Abaxis Piccolo
137.4 ± 1.12	135.7 ± 1.42	128.0 ± 0.93	146.6 ± 1.17	152.5 ± 1.43	n = 51	[ABJ] Abbott Architect c System
137.3 ± 1.57	135.7 ± 1.30	128.9 ± 1.37	146.4 ± 2.03	153.1 ± 1.24	n = 9	[OLC] Beckman Coulter AU Chemistry System
137.2 ± 1.47	135.7 ± 1.49	128.9 ± 1.50	145.8 ± 1.66	153.0 ± 1.58	n = 9	[BCS] Beckman Coulter CX
137.3 ± 0.84	135.7 ± 0.95	128.7 ± 0.97	146.6 ± 1.11	153.1 ± 0.83	n = 22	[BCX] Beckman Coulter LX-20
137.4 ± 1.17	136.0 ± 1.06	129.1 ± 1.16	146.5 ± 1.17	153.3 ± 1.31	n = 24	[BCG] Beckman Coulter UniCel DxC 600
135.8 ± 0.47	131.2 ± 0.47	127.0 ± 0.00	145.2 ± 0.66	151.3 ± 0.74	n = 7	[BCH] Beckman Coulter UniCel DxC 800
143.2 ± 1.44	141.9 ± 1.99	132.0 ± 2.06	153.5 ± 2.83	160.5 ± 2.20	n = 13	[IAA] i-STAT
142.8 ± 1.62	141.4 ± 1.51	131.1 ± 1.49	152.4 ± 1.99	159.1 ± 1.71	n = 22	[JJE] Ortho Vitros 250/350/950
143.3 ± 2.11	141.8 ± 2.13	131.6 ± 1.75	153.6 ± 2.31	159.7 ± 2.08	n = 11	[JJF] Ortho Vitros 5,1FS
136.6 ± 1.18	135.1 ± 1.07	127.2 ± 1.21	146.5 ± 1.15	153.1 ± 1.05	n = 18	[JJG] Ortho Vitros 5600
136.7 ± 1.07	134.5 ± 0.83	127.6 ± 0.56	145.7 ± 0.85	152.5 ± 0.69	n = 10	[ROC] Roche cobas c501
138.9 ± 1.47	136.8 ± 1.06	128.2 ± 1.24	147.7 ± 1.44	154.0 ± 1.54	n = 34	[ROT] Roche Cobas INTEGRA
139.8 ± 1.17	137.6 ± 1.15	130.0 ± 1.12	148.7 ± 1.02	154.7 ± 1.02	n = 16	[ROD] Roche MODULAR D/P
138.7 ± 0.51	136.3 ± 0.51	128.3 ± 0.51	147.7 ± 0.51	153.3 ± 0.51	n = 3	[BYE] Siemens ADVIA 1800
137.6 ± 1.69	135.7 ± 1.57	128.7 ± 1.57	147.1 ± 1.99	153.3 ± 1.98	n = 11	[BYB] Siemens ADVIA 2400
137.0 ± 1.80	134.1 ± 1.86	127.8 ± 1.65	146.2 ± 1.86	152.0 ± 1.69	n = 29	[DUE] Siemens Dimension EXL
136.4 ± 1.36	132.5 ± 1.50	129.2 ± 1.23	146.3 ± 1.46	152.0 ± 1.35	n = 32	[DUR] Siemens Dimension RxL
137.9 ± 1.53	135.6 ± 1.57	129.1 ± 0.90	147.1 ± 1.22	153.0 ± 1.44	n = 20	[DUT] Siemens Dimension Vista
138.0 ± 0.90	134.0 ± 0.90	128.3 ± 0.51	142.0 ± 0.00	152.0 ± 0.00	n = 3	[DUX] Siemens Dimension Xpand
138.0 ± 1.83	136.0 ± 1.95	127.5 ± 1.50	147.3 ± 1.81	153.2 ± 1.48	n = 14	<Reagents>
137.3 ± 1.15	135.8 ± 1.17	128.9 ± 1.19	146.5 ± 1.25	153.2 ± 1.12	n = 60	[AB1] Abbott
137.4 ± 1.13	135.8 ± 1.35	128.0 ± 0.95	146.6 ± 1.18	152.5 ± 1.45	n = 50	[BC1] Beckman Coulter AU Series
137.6 ± 1.09	135.8 ± 1.07	128.9 ± 1.27	147.2 ± 2.55	153.1 ± 1.83	n = 5	[OL1] Beckman Coulter AU Series
135.7 ± 0.51	131.0 ± 0.00	127.0 ± 0.00	145.0 ± 0.55	151.2 ± 0.73	n = 6	[CR1] Carolina
137.5 ± 1.86	137.3 ± 1.37	130.2 ± 2.36	151.9 ± 2.05	157.5 ± 1.86	n = 3	[IA1] i-STAT
143.1 ± 1.70	141.6 ± 1.80	131.5 ± 1.86	153.0 ± 2.34	159.6 ± 2.02	n = 47	[IL1] Instrumentation Lab
136.7 ± 1.33	135.2 ± 1.17	127.3 ± 1.28	146.6 ± 1.19	153.1 ± 1.04	n = 19	[JJ1] Ortho Clinical Diagnostics
138.9 ± 1.47	136.8 ± 1.06	128.2 ± 1.24	147.7 ± 1.44	154.0 ± 1.54	n = 34	[RO4] Roche cobas c311/c501/c502/c701
136.7 ± 1.07	134.5 ± 0.83	127.6 ± 0.56	145.7 ± 0.85	152.5 ± 0.69	n = 10	[RO2] Roche Hitachi and Modular D/P
139.6 ± 1.19	137.3 ± 1.14	129.6 ± 1.23	148.5 ± 1.09	154.4 ± 1.17	n = 22	[RO1] Roche Integra and MIRA
137.0 ± 1.70	134.1 ± 2.16	128.7 ± 1.51	146.5 ± 1.64	152.3 ± 1.65	n = 91	[BY1] Siemens ADVIA/ADVISIA Centaur
						[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Potassium (mmol/L)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
4.39 ± 0.10	3.47 ± 0.11	5.27 ± 0.11	2.97 ± 0.10	6.35 ± 0.13	n = 374	[---] All Methods & Instruments
4.51 ± 0.20	3.67 ± 0.05	5.25 ± 0.19	3.26 ± 0.10	6.40 ± 0.18	n = 3	<Instruments>
4.43 ± 0.08	3.52 ± 0.06	5.26 ± 0.08	3.00 ± 0.00	6.33 ± 0.08	n = 13	[AXA] Abaxis Piccolo
4.40 ± 0.00	3.50 ± 0.00	5.22 ± 0.06	3.02 ± 0.05	6.27 ± 0.08	n = 51	[ABJ] Abbott Architect c System
4.39 ± 0.08	3.50 ± 0.00	5.27 ± 0.05	2.90 ± 0.00	6.34 ± 0.09	n = 9	[OLC] Beckman Coulter AU Chemistry System
4.36 ± 0.06	3.43 ± 0.05	5.26 ± 0.07	2.90 ± 0.00	6.40 ± 0.07	n = 9	[BCS] Beckman Coulter CX
4.34 ± 0.06	3.42 ± 0.05	5.24 ± 0.06	2.91 ± 0.04	6.35 ± 0.07	n = 22	[BCX] Beckman Coulter LX-20
4.36 ± 0.06	3.41 ± 0.05	5.26 ± 0.06	2.90 ± 0.00	6.39 ± 0.08	n = 24	[BCG] Beckman Coulter UniCel DxC 600
4.30 ± 0.00	3.34 ± 0.06	5.20 ± 0.00	2.90 ± 0.00	6.20 ± 0.00	n = 7	[BCH] Beckman Coulter UniCel DxC 800
4.64 ± 0.09	3.72 ± 0.07	5.53 ± 0.11	3.17 ± 0.06	6.64 ± 0.13	n = 13	[IAA] i-STAT
4.64 ± 0.07	3.70 ± 0.04	5.49 ± 0.07	3.10 ± 0.00	6.59 ± 0.09	n = 22	[JJE] Ortho Vitros 250/350/950
4.62 ± 0.09	3.69 ± 0.07	5.48 ± 0.11	3.10 ± 0.07	6.59 ± 0.11	n = 11	[JJF] Ortho Vitros 5,1FS
4.30 ± 0.07	3.40 ± 0.00	5.20 ± 0.00	2.85 ± 0.07	6.28 ± 0.06	n = 17	[JJG] Ortho Vitros 5600
4.40 ± 0.00	3.50 ± 0.00	5.32 ± 0.05	3.00 ± 0.00	6.38 ± 0.05	n = 10	[ROC] Roche cobas c501
4.34 ± 0.09	3.42 ± 0.07	5.21 ± 0.07	2.92 ± 0.09	6.29 ± 0.09	n = 34	[ROT] Roche Cobas INTEGRA
4.48 ± 0.06	3.58 ± 0.07	5.35 ± 0.09	3.08 ± 0.06	6.44 ± 0.11	n = 16	[ROD] Roche MODULAR D/P
4.43 ± 0.05	3.50 ± 0.00	5.30 ± 0.00	3.03 ± 0.05	6.33 ± 0.05	n = 3	[BYE] Siemens ADVIA 1800
4.35 ± 0.07	3.42 ± 0.04	5.29 ± 0.09	2.89 ± 0.04	6.34 ± 0.08	n = 11	[BYB] Siemens ADVIA 2400
4.36 ± 0.06	3.40 ± 0.00	5.22 ± 0.05	2.90 ± 0.00	6.29 ± 0.08	n = 29	[DUE] Siemens Dimension EXL
4.38 ± 0.05	3.43 ± 0.06	5.29 ± 0.06	2.99 ± 0.04	6.36 ± 0.08	n = 32	[DUR] Siemens Dimension RxL
4.35 ± 0.06	3.40 ± 0.00	5.29 ± 0.05	2.90 ± 0.00	6.33 ± 0.05	n = 20	[DUT] Siemens Dimension Vista
4.51 ± 0.20	3.67 ± 0.05	5.25 ± 0.19	3.26 ± 0.10	6.40 ± 0.18	n = 3	[DUX] Siemens Dimension Xpand
4.42 ± 0.09	3.51 ± 0.07	5.25 ± 0.07	2.99 ± 0.05	6.32 ± 0.09	n = 14	<Reagents>
4.36 ± 0.06	3.43 ± 0.06	5.25 ± 0.06	2.90 ± 0.00	6.38 ± 0.07	n = 60	[BC1] Beckman Coulter
4.40 ± 0.00	3.50 ± 0.00	5.22 ± 0.06	3.02 ± 0.05	6.27 ± 0.08	n = 50	[OL1] Beckman Coulter AU Series
4.42 ± 0.08	3.50 ± 0.06	5.30 ± 0.00	2.95 ± 0.08	6.30 ± 0.09	n = 5	[CR1] Carolina
4.30 ± 0.00	3.33 ± 0.05	5.20 ± 0.00	2.90 ± 0.00	6.20 ± 0.00	n = 6	[IA1] i-STAT
4.33 ± 0.05	3.40 ± 0.00	5.30 ± 0.09	2.90 ± 0.00	6.40 ± 0.09	n = 3	[IL1] Instrumentation Lab
4.64 ± 0.08	3.70 ± 0.06	5.50 ± 0.09	3.13 ± 0.07	6.60 ± 0.11	n = 47	[JJ1] Ortho Clinical Diagnostics
4.31 ± 0.08	3.40 ± 0.00	5.20 ± 0.07	2.86 ± 0.08	6.28 ± 0.07	n = 18	[RO4] Roche cobas c311/c501/c502/c701
4.34 ± 0.09	3.42 ± 0.07	5.21 ± 0.07	2.92 ± 0.09	6.29 ± 0.09	n = 34	[RO2] Roche Hitachi and Modular D/P
4.40 ± 0.00	3.50 ± 0.00	5.32 ± 0.05	3.00 ± 0.00	6.38 ± 0.05	n = 11	[RO1] Roche Integra and MIRA
4.47 ± 0.06	3.57 ± 0.07	5.33 ± 0.08	3.07 ± 0.07	6.41 ± 0.11	n = 22	[BY1] Siemens ADVIA/ADVIa Centaur
4.36 ± 0.06	3.41 ± 0.04	5.27 ± 0.07	2.92 ± 0.06	6.33 ± 0.08	n = 91	[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Chloride (mmol/L)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
97.4 ± 2.29	86.9 ± 2.35	97.4 ± 2.83	107.8 ± 2.80	117.4 ± 2.87	n = 369	[---] All Methods & Instruments
95.9 ± 2.86	85.9 ± 2.86	95.9 ± 2.86	107.8 ± 4.10	115.7 ± 3.37	n = 3	<Instruments>
98.5 ± 1.33	87.1 ± 1.26	99.1 ± 1.19	109.4 ± 1.14	119.2 ± 0.90	n = 12	[AXA] Abaxis Piccolo
95.9 ± 0.92	85.0 ± 0.97	96.5 ± 1.06	106.4 ± 1.12	115.6 ± 1.23	n = 50	[ABJ] Abbott Architect c System
99.7 ± 2.04	89.3 ± 1.25	100.2 ± 1.13	109.9 ± 1.27	119.6 ± 1.84	n = 9	[OLC] Beckman Coulter AU Chemistry System
98.3 ± 1.21	88.2 ± 1.55	100.1 ± 1.20	109.7 ± 1.79	119.9 ± 1.50	n = 9	[BCS] Beckman Coulter CX
98.2 ± 1.02	87.7 ± 0.83	99.4 ± 0.93	109.5 ± 1.05	119.3 ± 1.15	n = 22	[BCX] Beckman Coulter LX-20
98.8 ± 1.09	88.2 ± 0.95	100.2 ± 1.18	110.2 ± 1.47	120.0 ± 1.15	n = 24	[BCG] Beckman Coulter UniCel DxC 600
104.2 ± 0.73	95.0 ± 0.90	102.8 ± 0.73	117.8 ± 0.91	125.1 ± 1.13	n = 6	[BCH] Beckman Coulter UniCel DxC 800
99.5 ± 1.45	88.7 ± 1.37	100.1 ± 1.82	111.5 ± 2.36	120.9 ± 1.91	n = 13	[IAA] i-STAT
99.9 ± 0.77	88.1 ± 0.95	99.9 ± 1.02	110.9 ± 1.14	120.3 ± 0.93	n = 22	[JJE] Ortho Vitros 250/350/950
100.0 ± 1.13	88.7 ± 1.10	100.3 ± 0.94	111.5 ± 1.62	120.7 ± 1.22	n = 11	[JJF] Ortho Vitros 5,1FS
92.0 ± 1.36	81.3 ± 1.03	91.9 ± 1.38	103.1 ± 1.51	113.2 ± 1.27	n = 18	[JJG] Ortho Vitros 5600
97.5 ± 0.69	87.5 ± 0.89	97.3 ± 0.99	107.8 ± 0.65	117.1 ± 0.49	n = 10	[ROC] Roche cobas c501
95.1 ± 1.24	83.7 ± 1.16	94.6 ± 1.16	105.0 ± 1.10	114.9 ± 1.30	n = 34	[ROT] Roche Cobas INTEGRA
97.6 ± 0.80	86.2 ± 0.86	97.1 ± 0.74	107.6 ± 0.56	117.2 ± 1.26	n = 16	[ROD] Roche MODULAR D/P
97.3 ± 0.51	85.0 ± 0.90	96.3 ± 0.51	106.4 ± 1.02	116.2 ± 1.54	n = 3	[BYE] Siemens ADVIA 1800
96.7 ± 0.87	88.7 ± 1.03	95.8 ± 0.84	106.5 ± 1.42	115.6 ± 1.29	n = 11	[BYB] Siemens ADVIA 2400
96.5 ± 1.89	86.8 ± 1.48	94.4 ± 1.16	106.3 ± 1.28	115.9 ± 1.57	n = 29	[DUE] Siemens Dimension EXL
97.9 ± 1.19	87.0 ± 0.89	97.9 ± 1.15	108.3 ± 1.45	118.1 ± 1.68	n = 32	[DUR] Siemens Dimension RxL
96.5 ± 0.90	88.4 ± 1.09	95.5 ± 1.23	106.1 ± 0.82	115.2 ± 0.91	n = 20	[DUT] Siemens Dimension Vista
95.9 ± 2.86	85.9 ± 2.86	95.9 ± 2.86	107.8 ± 4.10	115.7 ± 3.37	n = 3	[DUX] Siemens Dimension Xpand
98.5 ± 1.27	87.2 ± 1.24	99.2 ± 1.16	109.5 ± 1.07	119.2 ± 0.84	n = 13	<Reagents>
98.5 ± 1.24	88.0 ± 1.04	99.9 ± 1.13	109.9 ± 1.39	119.8 ± 1.24	n = 60	[AB1] Abbott
96.0 ± 0.94	85.0 ± 0.98	96.5 ± 1.08	106.4 ± 1.13	115.6 ± 1.22	n = 49	[BC1] Beckman Coulter
99.2 ± 1.66	89.2 ± 0.80	99.2 ± 1.91	109.6 ± 1.09	118.3 ± 1.99	n = 5	[OL1] Beckman Coulter AU Series
104.2 ± 0.73	95.0 ± 0.90	102.8 ± 0.73	117.8 ± 0.91	125.1 ± 1.13	n = 6	[CR1] Carolina
99.8 ± 1.12	88.4 ± 1.13	100.0 ± 1.30	111.1 ± 1.70	120.5 ± 1.38	n = 47	[IA1] i-STAT
91.9 ± 1.34	81.3 ± 0.98	91.8 ± 1.35	102.9 ± 1.57	113.1 ± 1.35	n = 19	[JJ1] Ortho Clinical Diagnostics
95.1 ± 1.24	83.7 ± 1.16	94.6 ± 1.16	105.0 ± 1.10	114.9 ± 1.30	n = 34	[RO2] Roche Hitachi and Modular D/P
97.5 ± 0.69	87.5 ± 0.89	97.3 ± 0.99	107.8 ± 0.65	117.1 ± 0.49	n = 10	[RO1] Roche Integra and MIRA
97.8 ± 0.95	86.2 ± 0.93	97.1 ± 0.79	107.5 ± 0.59	117.2 ± 1.26	n = 22	[BY1] Siemens ADVIA/ADVISIA Centaur
97.0 ± 1.45	87.5 ± 1.42	96.0 ± 1.87	106.9 ± 1.56	116.4 ± 1.85	n = 91	[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Albumin (g/dL)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
4.20 ± 0.20	5.19 ± 0.24	3.71 ± 0.17	4.80 ± 0.18	4.43 ± 0.17	n = 360	[---] All Methods & Instruments
4.26 ± 0.10	5.13 ± 0.05	3.80 ± 0.00	4.70 ± 0.09	4.46 ± 0.10	n = 3	[AXA] Abaxis Piccolo
3.99 ± 0.16	4.87 ± 0.13	3.59 ± 0.12	4.58 ± 0.14	4.24 ± 0.18	n = 13	[ABJ] Abbott Architect c System
4.21 ± 0.10	5.13 ± 0.11	3.75 ± 0.08	4.80 ± 0.13	4.42 ± 0.11	n = 52	[OLC] Beckman Coulter AU Chemistry System
3.99 ± 0.15	4.95 ± 0.13	3.58 ± 0.14	4.65 ± 0.17	4.26 ± 0.16	n = 9	[BCS] Beckman Coulter CX
4.02 ± 0.09	5.00 ± 0.07	3.60 ± 0.08	4.63 ± 0.05	4.23 ± 0.08	n = 9	[BCX] Beckman Coulter LX-20
3.94 ± 0.07	4.96 ± 0.07	3.54 ± 0.06	4.52 ± 0.12	4.21 ± 0.11	n = 21	[BCG] Beckman Coulter UniCel DxC 600
4.00 ± 0.00	5.00 ± 0.04	3.58 ± 0.05	4.60 ± 0.06	4.24 ± 0.06	n = 24	[BCH] Beckman Coulter UniCel DxC 800
3.99 ± 0.12	5.10 ± 0.18	3.42 ± 0.13	4.68 ± 0.13	4.34 ± 0.16	n = 12	[JJE] Ortho Vitros 250/350/950
4.04 ± 0.12	5.11 ± 0.12	3.43 ± 0.09	4.73 ± 0.12	4.33 ± 0.12	n = 21	[JJF] Ortho Vitros 5,1FS
4.03 ± 0.15	5.11 ± 0.20	3.42 ± 0.10	4.68 ± 0.28	4.35 ± 0.20	n = 11	[JJG] Ortho Vitros 5600
4.36 ± 0.08	5.29 ± 0.08	3.85 ± 0.09	4.91 ± 0.09	4.56 ± 0.08	n = 17	[ROC] Roche cobas c501
4.23 ± 0.12	5.11 ± 0.15	3.75 ± 0.14	4.70 ± 0.11	4.37 ± 0.16	n = 10	[ROT] Roche Cobas INTEGRA
4.38 ± 0.11	5.27 ± 0.12	3.85 ± 0.10	4.90 ± 0.12	4.54 ± 0.09	n = 35	[ROD] Roche MODULAR D/P
4.23 ± 0.10	5.08 ± 0.09	3.76 ± 0.07	4.78 ± 0.08	4.48 ± 0.07	n = 16	[BYE] Siemens ADVIA 1800
4.24 ± 0.10	5.10 ± 0.18	3.74 ± 0.10	4.77 ± 0.14	4.47 ± 0.14	n = 3	[BYB] Siemens ADVIA 2400
4.30 ± 0.07	5.47 ± 0.12	3.75 ± 0.07	4.91 ± 0.10	4.54 ± 0.09	n = 11	[DUE] Siemens Dimension EXL
4.31 ± 0.10	5.47 ± 0.12	3.78 ± 0.08	4.91 ± 0.09	4.52 ± 0.10	n = 29	[DUR] Siemens Dimension RxL
4.35 ± 0.09	5.48 ± 0.11	3.81 ± 0.10	4.94 ± 0.10	4.57 ± 0.10	n = 32	[DUT] Siemens Dimension Vista
4.35 ± 0.09	5.48 ± 0.11	3.81 ± 0.07	4.93 ± 0.08	4.55 ± 0.11	n = 20	[DUX] Siemens Dimension Xpand
<Reagents>						
4.26 ± 0.10	5.13 ± 0.05	3.80 ± 0.00	4.70 ± 0.09	4.46 ± 0.10	n = 3	[AX1] Abaxis
3.99 ± 0.16	4.87 ± 0.13	3.59 ± 0.12	4.58 ± 0.14	4.24 ± 0.18	n = 13	[AB1] Abbott
3.97 ± 0.07	4.98 ± 0.07	3.56 ± 0.07	4.57 ± 0.09	4.22 ± 0.08	n = 57	[BC1] Beckman Coulter
4.21 ± 0.11	5.14 ± 0.11	3.75 ± 0.08	4.80 ± 0.13	4.42 ± 0.11	n = 51	[OL1] Beckman Coulter AU Series
4.11 ± 0.18	5.10 ± 0.32	3.68 ± 0.17	4.76 ± 0.17	4.35 ± 0.22	n = 5	[CR1] Carolina
4.02 ± 0.13	5.10 ± 0.16	3.42 ± 0.11	4.70 ± 0.15	4.34 ± 0.15	n = 45	[JJ1] Ortho Clinical Diagnostics
4.36 ± 0.08	5.28 ± 0.08	3.85 ± 0.09	4.90 ± 0.09	4.56 ± 0.08	n = 18	[RO4] Roche cobas c311/c501/c502/c701
4.38 ± 0.11	5.27 ± 0.12	3.85 ± 0.10	4.90 ± 0.12	4.55 ± 0.09	n = 35	[RO2] Roche Hitachi and Modular D/P
4.23 ± 0.12	5.11 ± 0.15	3.75 ± 0.14	4.70 ± 0.11	4.37 ± 0.16	n = 10	[RO1] Roche Integra and MIRA
4.25 ± 0.11	5.11 ± 0.10	3.77 ± 0.07	4.80 ± 0.09	4.50 ± 0.08	n = 23	[BY1] Siemens ADVIA/ADVIS Centaur
4.33 ± 0.09	5.48 ± 0.11	3.79 ± 0.09	4.93 ± 0.09	4.55 ± 0.11	n = 91	[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Total Protein (g/dL)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
7.00 ± 0.20	7.95 ± 0.21	6.08 ± 0.17	7.60 ± 0.22	7.27 ± 0.20	n = 361	[---] All Methods & Instruments
7.03 ± 0.23	7.90 ± 0.18	6.04 ± 0.10	7.50 ± 0.18	7.24 ± 0.10	n = 3	<Instruments>
6.98 ± 0.08	7.87 ± 0.10	6.00 ± 0.00	7.51 ± 0.09	7.18 ± 0.08	n = 13	[AXA] Abaxis Piccolo
6.93 ± 0.11	7.83 ± 0.15	6.03 ± 0.12	7.49 ± 0.12	7.18 ± 0.12	n = 51	[ABJ] Abbott Architect c System
6.96 ± 0.20	7.93 ± 0.08	6.16 ± 0.22	7.65 ± 0.23	7.35 ± 0.18	n = 9	[OLC] Beckman Coulter AU Chemistry System
6.74 ± 0.15	7.74 ± 0.10	5.89 ± 0.11	7.34 ± 0.17	7.07 ± 0.14	n = 10	[BCS] Beckman Coulter CX
6.97 ± 0.13	7.93 ± 0.15	6.06 ± 0.15	7.60 ± 0.17	7.23 ± 0.18	n = 20	[BCX] Beckman Coulter LX-20
6.65 ± 0.15	7.67 ± 0.11	5.85 ± 0.10	7.33 ± 0.14	6.98 ± 0.16	n = 24	[BCG] Beckman Coulter UniCel DxC 600
6.90 ± 0.13	8.08 ± 0.14	6.06 ± 0.13	7.69 ± 0.16	7.35 ± 0.12	n = 13	[BCH] Beckman Coulter UniCel DxC 800
6.93 ± 0.15	8.02 ± 0.19	6.03 ± 0.17	7.64 ± 0.18	7.19 ± 0.16	n = 22	[JJE] Ortho Vitros 250/350/950
7.01 ± 0.13	8.12 ± 0.20	6.03 ± 0.13	7.69 ± 0.22	7.31 ± 0.14	n = 11	[JJF] Ortho Vitros 5,1FS
6.97 ± 0.06	7.87 ± 0.11	6.07 ± 0.08	7.49 ± 0.12	7.23 ± 0.09	n = 17	[JJG] Ortho Vitros 5600
6.87 ± 0.14	7.80 ± 0.10	5.97 ± 0.13	7.41 ± 0.13	7.15 ± 0.13	n = 10	[ROC] Roche cobas c501
6.99 ± 0.10	7.90 ± 0.08	6.10 ± 0.08	7.53 ± 0.11	7.26 ± 0.09	n = 35	[ROT] Roche Cobas INTEGRA
7.04 ± 0.13	7.95 ± 0.14	6.08 ± 0.09	7.65 ± 0.13	7.30 ± 0.10	n = 16	[ROD] Roche MODULAR D/P
7.02 ± 0.24	7.99 ± 0.29	6.12 ± 0.24	7.65 ± 0.27	7.32 ± 0.24	n = 3	[BYE] Siemens ADVIA 1800
7.26 ± 0.12	8.23 ± 0.16	6.24 ± 0.10	7.81 ± 0.13	7.51 ± 0.08	n = 11	[BYB] Siemens ADVIA 2400
7.20 ± 0.11	8.15 ± 0.12	6.23 ± 0.10	7.80 ± 0.11	7.47 ± 0.11	n = 29	[DUE] Siemens Dimension EXL
7.20 ± 0.13	8.11 ± 0.11	6.22 ± 0.08	7.78 ± 0.12	7.43 ± 0.10	n = 32	[DUR] Siemens Dimension RxL
7.21 ± 0.15	8.17 ± 0.14	6.24 ± 0.12	7.81 ± 0.15	7.52 ± 0.21	n = 20	[DUT] Siemens Dimension Vista
7.03 ± 0.23	7.90 ± 0.18	6.04 ± 0.10	7.50 ± 0.18	7.24 ± 0.10	n = 3	[DUX] Siemens Dimension Xpand
6.98 ± 0.08	7.87 ± 0.10	6.00 ± 0.00	7.51 ± 0.09	7.18 ± 0.08	n = 13	<Reagents>
6.80 ± 0.22	7.78 ± 0.18	5.94 ± 0.16	7.43 ± 0.20	7.10 ± 0.20	n = 59	[AB1] Abbott
6.93 ± 0.10	7.82 ± 0.15	6.02 ± 0.12	7.49 ± 0.13	7.18 ± 0.13	n = 50	[BC1] Beckman Coulter
6.88 ± 0.19	7.89 ± 0.36	6.19 ± 0.32	7.75 ± 0.28	7.45 ± 0.18	n = 5	[OL1] Beckman Coulter AU Series
6.94 ± 0.15	8.05 ± 0.19	6.04 ± 0.15	7.67 ± 0.18	7.27 ± 0.17	n = 46	[CR1] Carolina
6.96 ± 0.07	7.86 ± 0.12	6.06 ± 0.09	7.49 ± 0.12	7.22 ± 0.10	n = 17	[JJ1] Ortho Clinical Diagnostics
6.99 ± 0.10	7.90 ± 0.08	6.10 ± 0.07	7.53 ± 0.11	7.26 ± 0.09	n = 36	[RO4] Roche cobas c311/c501/c502/c701
6.87 ± 0.14	7.80 ± 0.10	5.97 ± 0.13	7.41 ± 0.13	7.15 ± 0.13	n = 10	[RO2] Roche Hitachi and Modular D/P
7.07 ± 0.14	7.99 ± 0.16	6.12 ± 0.13	7.68 ± 0.15	7.34 ± 0.13	n = 23	[RO1] Roche Integra and MIRA
7.21 ± 0.13	8.15 ± 0.13	6.23 ± 0.09	7.79 ± 0.13	7.46 ± 0.11	n = 90	[BY1] Siemens ADVIA/ADVISIA Centaur
7.21 ± 0.13	8.15 ± 0.13	6.23 ± 0.09	7.79 ± 0.13	7.46 ± 0.11	n = 90	[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Cholesterol (mg/dL)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
239.6 ± 7.49	190.3 ± 7.80	136.0 ± 4.72	200.6 ± 7.07	162.4 ± 5.51	n = 333	[---] All Methods & Instruments
239.3 ± 1.37	190.7 ± 0.51	136.2 ± 1.54	203.3 ± 1.37	164.4 ± 1.02	n = 3	<Instruments>
244.6 ± 1.31	196.4 ± 1.42	139.6 ± 1.07	202.8 ± 1.36	166.6 ± 0.69	n = 13	[AXA] Abaxis Piccolo
236.2 ± 5.52	188.1 ± 4.02	133.1 ± 3.01	195.8 ± 4.23	158.9 ± 3.28	n = 55	[ABJ] Abbott Architect c System
235.8 ± 5.60	187.0 ± 6.42	135.8 ± 4.75	199.9 ± 5.74	162.1 ± 4.39	n = 9	[OLC] Beckman Coulter AU Chemistry System
240.2 ± 3.16	187.8 ± 2.29	134.4 ± 1.30	202.7 ± 1.28	160.0 ± 1.78	n = 7	[BCS] Beckman Coulter CX
238.2 ± 5.14	187.3 ± 4.54	134.4 ± 3.82	202.3 ± 5.13	160.4 ± 3.30	n = 18	[BCX] Beckman Coulter LX-20
238.8 ± 0.79	186.2 ± 3.07	133.2 ± 1.72	200.6 ± 3.10	160.1 ± 2.95	n = 20	[BCG] Beckman Coulter UniCel DxC 600
250.4 ± 5.56	205.0 ± 4.24	145.5 ± 0.83	219.5 ± 3.86	177.3 ± 0.90	n = 5	[BCH] Beckman Coulter UniCel DxC 800
254.0 ± 6.29	207.9 ± 5.36	145.3 ± 3.77	217.8 ± 5.48	175.7 ± 4.02	n = 21	[JJE] Ortho Vitros 250/350/950
253.9 ± 6.33	207.1 ± 1.59	144.8 ± 2.03	219.1 ± 2.60	175.8 ± 2.89	n = 11	[JJF] Ortho Vitros 5,1FS
244.5 ± 4.00	195.0 ± 2.33	139.7 ± 2.38	203.3 ± 3.08	165.2 ± 2.63	n = 15	[JJG] Ortho Vitros 5600
241.5 ± 4.72	193.7 ± 4.29	136.9 ± 3.51	199.7 ± 3.89	163.3 ± 3.15	n = 12	[ROC] Roche cobas c501
241.9 ± 5.21	194.4 ± 4.11	137.8 ± 2.96	202.0 ± 4.77	164.3 ± 3.64	n = 36	[ROT] Roche Cobas INTEGRA
237.1 ± 3.97	189.2 ± 2.75	135.3 ± 2.57	202.2 ± 4.00	161.4 ± 2.75	n = 16	[ROD] Roche MODULAR D/P
237.4 ± 7.34	187.7 ± 3.16	133.5 ± 2.74	201.0 ± 5.48	160.4 ± 3.87	n = 3	[BYE] Siemens ADVIA 1800
237.2 ± 6.70	183.2 ± 5.05	134.6 ± 3.54	197.5 ± 4.05	161.5 ± 3.88	n = 11	[BYB] Siemens ADVIA 2400
237.5 ± 5.51	184.9 ± 3.99	135.2 ± 2.75	199.6 ± 3.48	163.4 ± 2.87	n = 22	[DUE] Siemens Dimension EXL
233.9 ± 5.50	186.3 ± 6.02	133.2 ± 3.11	194.8 ± 5.55	159.5 ± 3.73	n = 29	[DUR] Siemens Dimension RxL
236.9 ± 6.44	183.6 ± 6.36	133.6 ± 4.32	197.2 ± 5.96	161.2 ± 4.33	n = 15	[DUT] Siemens Dimension Vista
						[DUX] Siemens Dimension Xpand
239.3 ± 1.37	190.7 ± 0.51	136.2 ± 1.54	203.3 ± 1.37	164.4 ± 1.02	n = 3	<Reagents>
244.6 ± 1.31	196.4 ± 1.42	139.6 ± 1.07	202.8 ± 1.36	166.6 ± 0.69	n = 13	[AX1] Abaxis
238.3 ± 4.26	187.0 ± 3.53	133.8 ± 2.85	201.5 ± 3.71	160.5 ± 3.18	n = 50	[AB1] Abbott
236.6 ± 5.49	188.5 ± 3.86	133.2 ± 3.08	195.9 ± 4.43	158.9 ± 3.43	n = 52	[BC1] Beckman Coulter
231.7 ± 2.90	182.5 ± 4.49	136.3 ± 4.42	197.3 ± 6.29	158.9 ± 3.00	n = 5	[OL1] Beckman Coulter AU Series
253.4 ± 6.38	207.5 ± 4.67	145.2 ± 2.84	218.7 ± 4.63	175.9 ± 3.49	n = 37	[CR1] Carolina
244.5 ± 4.00	195.0 ± 2.33	139.7 ± 2.38	203.3 ± 3.08	165.2 ± 2.63	n = 15	[JJ1] Ortho Clinical Diagnostics
242.1 ± 5.85	194.5 ± 4.38	137.9 ± 3.13	202.1 ± 4.86	164.4 ± 3.97	n = 37	[RO4] Roche cobas c311/c501/c502/c701
241.5 ± 4.72	193.7 ± 4.29	136.9 ± 3.51	199.7 ± 3.89	163.3 ± 3.15	n = 12	[RO2] Roche Hitachi and Modular D/P
237.9 ± 4.99	189.5 ± 3.14	135.4 ± 2.80	202.7 ± 4.36	161.8 ± 3.14	n = 23	[RO1] Roche Integra and MIRA
235.9 ± 6.06	184.8 ± 5.43	134.1 ± 3.47	197.2 ± 5.26	161.3 ± 4.14	n = 76	[BY1] Siemens ADVIA/ADVISIA Centaur
						[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

HDL-Cholesterol (mg/dL)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
53.8 ± 7.49	38.7 ± 6.09	35.1 ± 3.91	69.6 ± 7.22	42.6 ± 5.24	n = 316	[---] All Methods & Instruments
55.5 ± 9.09	40.6 ± 7.40	35.9 ± 5.31	72.5 ± 10.40	44.5 ± 6.60	n = 15	[---] All Precipitation Methods
53.7 ± 7.39	38.7 ± 6.00	35.1 ± 3.84	69.6 ± 7.09	42.5 ± 5.15	n = 301	[---] All Homogeneous (Direct) Methods
55.8 ± 2.88	40.1 ± 2.31	37.6 ± 2.05	72.7 ± 3.61	44.9 ± 2.27	n = 12	[AB1] Abbott
44.9 ± 2.86	28.6 ± 2.56	25.7 ± 0.51	54.1 ± 2.05	32.0 ± 1.80	n = 3	[AX1] Abaxis
66.0 ± 2.35	47.7 ± 1.82	39.0 ± 1.67	75.9 ± 3.05	47.9 ± 1.81	n = 45	[BC1] Beckman Coulter
56.9 ± 1.93	40.8 ± 1.59	37.9 ± 1.35	74.2 ± 2.38	45.8 ± 1.50	n = 35	[OL1] Beckman Coulter AU Series
64.7 ± 4.06	50.1 ± 3.72	44.7 ± 0.51	81.7 ± 3.16	55.5 ± 1.86	n = 3	[CR1] Carolina
57.0 ± 2.72	41.0 ± 2.72	38.0 ± 2.12	73.6 ± 3.98	46.5 ± 2.89	n = 5	[EQ1/GZ1] Equal/Genzyme
61.3 ± 1.63	45.2 ± 1.49	38.9 ± 0.98	81.5 ± 2.90	48.9 ± 1.52	n = 30	[JJ1] Ortho Clinical Diagnostics
48.6 ± 1.29	35.3 ± 1.06	32.4 ± 1.00	64.2 ± 1.83	38.9 ± 1.06	n = 13	[RO4] Roche cobas c311/c501/c502/c701
51.4 ± 2.53	37.5 ± 2.25	33.7 ± 1.47	66.6 ± 2.43	40.3 ± 1.97	n = 33	[RO2] Roche Hitachi and Modular D/P
50.2 ± 1.39	36.0 ± 1.56	33.5 ± 0.89	66.0 ± 2.70	39.6 ± 0.90	n = 10	[RO1] Roche Integra and MIRA
43.9 ± 1.71	28.6 ± 1.24	29.4 ± 0.96	63.9 ± 1.67	36.0 ± 1.09	n = 22	[BY1] Siemens ADVIA/ADVIA Centaur
48.7 ± 1.94	34.8 ± 1.63	32.5 ± 1.18	63.9 ± 2.01	38.6 ± 1.14	n = 68	[DA5] Siemens Dimension HDL DF48B or K3048A
52.8 ± 3.85	36.5 ± 3.31	35.6 ± 3.14	71.4 ± 4.34	43.3 ± 4.00	n = 5	[DA7] Siemens Dimension HDL DF48A or K3048

Summary of Participant Performance (Mean and Standard Deviation)

LDL-Cholesterol (mg/dL)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
150.2 ± 15.26	124.1 ± 14.76	80.0 ± 12.06	105.9 ± 12.19	95.8 ± 13.31	n = 303	[---] All Methods
156.9 ± 9.89	132.0 ± 9.21	86.5 ± 5.72	110.9 ± 8.44	102.8 ± 6.86	n = 160	[-A-] All Calculated Results Friedewald formula [LDL=TC-HDL-(Trigs÷5)]
140.7 ± 16.46	113.5 ± 13.83	69.8 ± 11.08	98.6 ± 12.80	85.3 ± 12.83	n = 141	[---] All Homogeneous (Direct) Methods
131.6 ± 1.90	108.0 ± 2.45	65.8 ± 1.46	90.9 ± 1.13	79.7 ± 1.51	n = 4	[AB1] Abbott
128.5 ± 4.60	105.3 ± 3.98	61.9 ± 2.68	88.6 ± 3.46	75.8 ± 3.40	n = 23	[BC1] Beckman Coulter
121.4 ± 6.08	97.4 ± 5.46	59.7 ± 4.06	85.1 ± 4.80	72.0 ± 4.39	n = 16	[OL1] Beckman Coulter AU Series
128.5 ± 10.98	104.6 ± 10.71	62.7 ± 7.36	87.5 ± 8.37	77.2 ± 9.49	n = 14	[EQ1/GZ1] Equal/Genzyme
156.3 ± 7.48	119.8 ± 6.38	67.3 ± 3.40	107.5 ± 5.14	86.6 ± 4.57	n = 13	[JJ1] Ortho Clinical Diagnostics
163.0 ± 1.14	130.5 ± 1.71	90.0 ± 1.14	118.5 ± 1.71	108.0 ± 2.28	n = 2	[RO4] Roche cobas c311/c501/c502/c701
163.9 ± 4.75	134.3 ± 3.91	92.5 ± 3.34	121.3 ± 3.93	110.7 ± 2.75	n = 16	[RO2] Roche Hitachi and Modular D/P
148.8 ± 4.04	115.8 ± 3.66	69.0 ± 2.57	105.6 ± 3.86	86.0 ± 3.03	n = 11	[BY1] Siemens ADVIA/ADVISIA Centaur
145.2 ± 5.61	120.5 ± 5.84	76.6 ± 3.84	101.4 ± 4.30	91.9 ± 4.97	n = 33	[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Triglycerides (mg/dL)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
154.2 ± 7.50	110.2 ± 6.67	77.4 ± 4.40	108.9 ± 5.46	94.8 ± 5.04	n = 326	[---] All Methods & Instruments
171.7 ± 2.26	124.2 ± 2.36	84.7 ± 0.51	118.4 ± 1.02	104.3 ± 0.51	n = 3	<Instruments>
149.5 ± 2.73	106.6 ± 1.47	77.6 ± 1.95	109.9 ± 2.69	94.2 ± 1.62	n = 14	[AXA] Abaxis Piccolo
150.6 ± 4.34	107.1 ± 3.08	75.1 ± 1.95	105.1 ± 3.35	91.4 ± 2.68	n = 53	[ABJ] Abbott Architect c System
168.3 ± 10.41	123.3 ± 10.40	87.6 ± 6.31	118.6 ± 7.99	103.3 ± 5.37	n = 9	[OLC] Beckman Coulter AU Chemistry System
161.0 ± 5.20	119.1 ± 5.23	78.2 ± 3.77	111.8 ± 4.10	96.7 ± 2.60	n = 7	[BCS] Beckman Coulter CX
162.0 ± 4.11	117.0 ± 4.25	78.0 ± 3.00	111.6 ± 2.85	97.7 ± 3.38	n = 15	[BCX] Beckman Coulter LX-20
161.1 ± 3.97	116.6 ± 2.62	77.6 ± 3.29	111.5 ± 3.73	97.0 ± 3.40	n = 17	[BCG] Beckman Coulter UniCel DxC 600
165.6 ± 4.93	122.6 ± 3.94	82.8 ± 2.32	115.5 ± 3.68	102.6 ± 2.84	n = 5	[BCH] Beckman Coulter UniCel DxC 800
159.2 ± 6.56	116.4 ± 4.96	79.0 ± 3.00	110.5 ± 4.61	97.4 ± 3.96	n = 21	[JJE] Ortho Vitros 250/350/950
156.8 ± 4.77	116.3 ± 2.94	79.0 ± 1.85	109.2 ± 3.99	97.0 ± 2.07	n = 11	[JJF] Ortho Vitros 5,1FS
154.9 ± 2.84	109.3 ± 1.87	80.9 ± 2.50	109.0 ± 2.10	97.8 ± 1.49	n = 16	[JJG] Ortho Vitros 5600
149.6 ± 4.79	105.8 ± 1.74	78.1 ± 2.07	103.7 ± 1.84	93.5 ± 2.72	n = 11	[ROC] Roche cobas c501
151.5 ± 4.27	106.6 ± 2.79	77.6 ± 2.98	108.0 ± 3.25	94.7 ± 2.71	n = 36	[ROT] Roche Cobas INTEGRA
151.0 ± 1.85	107.5 ± 1.31	76.3 ± 0.74	109.5 ± 1.82	93.0 ± 1.00	n = 16	[ROD] Roche MODULAR D/P
152.5 ± 1.86	110.0 ± 0.00	77.3 ± 0.51	111.0 ± 0.00	94.7 ± 0.51	n = 3	[BYE] Siemens ADVIA 1800
147.5 ± 2.67	103.3 ± 1.53	71.9 ± 1.57	105.0 ± 2.63	88.5 ± 2.14	n = 11	[BYB] Siemens ADVIA 2400
149.4 ± 3.77	104.7 ± 4.15	72.6 ± 2.66	106.4 ± 4.52	90.4 ± 3.59	n = 22	[DUE] Siemens Dimension EXL
163.8 ± 4.43	116.5 ± 3.83	83.1 ± 3.23	117.6 ± 4.59	101.0 ± 4.19	n = 31	[DUR] Siemens Dimension RxL
147.7 ± 2.19	104.3 ± 2.37	71.2 ± 2.16	103.8 ± 2.21	88.9 ± 1.85	n = 12	[DUT] Siemens Dimension Vista
171.7 ± 2.26	124.2 ± 2.36	84.7 ± 0.51	118.4 ± 1.02	104.3 ± 0.51	n = 3	[DUX] Siemens Dimension Xpand
149.5 ± 2.73	106.6 ± 1.47	77.6 ± 1.95	109.9 ± 2.69	94.2 ± 1.62	n = 14	<Reagents>
162.0 ± 5.29	117.5 ± 4.23	78.2 ± 3.57	111.8 ± 3.92	97.5 ± 3.67	n = 46	[AB1] Abbott
150.4 ± 3.91	107.0 ± 2.72	75.1 ± 1.56	105.2 ± 3.09	91.3 ± 2.27	n = 49	[BC1] Beckman Coulter
160.6 ± 13.00	117.5 ± 12.11	86.0 ± 6.46	112.9 ± 11.26	101.2 ± 5.28	n = 5	[OL1] Beckman Coulter AU Series
159.3 ± 6.41	117.2 ± 4.74	79.5 ± 2.84	110.7 ± 4.80	97.9 ± 3.88	n = 37	[CR1] Carolina
154.9 ± 2.84	109.3 ± 1.87	80.9 ± 2.50	109.0 ± 2.10	97.8 ± 1.49	n = 16	[JJ1] Ortho Clinical Diagnostics
151.4 ± 4.28	106.4 ± 2.82	77.6 ± 2.98	107.9 ± 3.19	94.7 ± 2.71	n = 36	[RO4] Roche cobas c311/c501/c502/c701
149.9 ± 4.22	106.0 ± 1.80	78.0 ± 1.97	104.1 ± 4.49	93.7 ± 2.49	n = 12	[RO2] Roche Hitachi and Modular D/P
151.8 ± 2.27	108.3 ± 1.74	76.8 ± 1.10	110.3 ± 1.99	93.6 ± 1.57	n = 23	[RO1] Roche Integra and MIRA
154.3 ± 9.22	108.9 ± 7.61	75.9 ± 6.74	109.8 ± 7.80	93.5 ± 7.21	n = 74	[BY1] Siemens ADVIA/ADVISIA Centaur
						[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Homocysteine ($\mu\text{mol/L}$)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
19.38 ± 1.75	12.70 ± 1.43	9.20 ± 0.99	26.10 ± 2.11	11.18 ± 1.20	n = 122	[---] All Methods & Instruments
20.30 ± 0.92	13.75 ± 0.58	10.06 ± 0.45	27.21 ± 0.84	12.11 ± 0.67	n = 8	<Instruments>
19.78 ± 1.10	13.47 ± 0.73	9.83 ± 0.28	25.75 ± 1.49	11.79 ± 0.37	n = 6	[ABH] Abbott Architect i System
20.33 ± 1.12	13.38 ± 1.16	9.68 ± 0.67	26.78 ± 1.58	11.74 ± 0.71	n = 19	[ABB] Abbott AxSym
21.01 ± 1.21	13.78 ± 0.80	10.17 ± 0.40	29.01 ± 0.11	12.29 ± 0.41	n = 5	[OLC] Beckman Coulter AU Chemistry System
19.03 ± 1.30	13.03 ± 1.02	8.90 ± 0.86	25.34 ± 1.88	10.90 ± 0.93	n = 4	[BCH] Beckman Coulter UniCel DxC 800
19.84 ± 1.28	13.94 ± 1.27	10.06 ± 1.18	26.35 ± 1.05	12.10 ± 0.96	n = 6	[JJG] Ortho Vitros 5600
17.34 ± 0.96	11.58 ± 0.67	8.36 ± 0.47	24.08 ± 1.23	10.17 ± 0.73	n = 23	[ROD] Roche MODULAR D/P
18.32 ± 1.00	11.10 ± 0.58	8.95 ± 0.76	25.23 ± 1.67	10.30 ± 0.62	n = 7	[COB] Siemens ADVIA Centaur
19.56 ± 1.26	11.86 ± 0.61	8.68 ± 0.78	26.59 ± 1.92	10.56 ± 0.88	n = 21	[DUT] Siemens Dimension Vista
21.06 ± 1.63	13.79 ± 1.08	9.61 ± 0.28	27.56 ± 2.46	12.27 ± 1.32	n = 6	[DPD] Siemens Immulite 2000
20.08 ± 1.04	13.65 ± 0.66	9.97 ± 0.41	26.61 ± 1.32	11.94 ± 0.59	n = 14	[DPE] Siemens Immulite 2500
20.48 ± 0.90	14.22 ± 0.94	9.78 ± 0.78	27.50 ± 1.86	12.18 ± 0.90	n = 8	<Reagents>
21.39 ± 1.63	14.11 ± 1.07	9.95 ± 0.63	28.24 ± 1.38	12.27 ± 0.77	n = 3	[AB1] Abbott
19.95 ± 1.31	13.07 ± 0.96	9.69 ± 0.76	26.69 ± 1.60	11.66 ± 0.79	n = 24	[CR1] Carolina
19.01 ± 1.06	13.16 ± 0.77	8.69 ± 0.85	25.22 ± 1.61	10.74 ± 0.91	n = 6	[CC1] Catch, Inc
17.34 ± 0.96	11.58 ± 0.67	8.36 ± 0.47	24.08 ± 1.23	10.17 ± 0.73	n = 23	[DZ1] Diazyme
18.54 ± 0.91	11.22 ± 0.53	8.91 ± 0.84	25.60 ± 1.56	10.44 ± 0.53	n = 6	[JJ1] Ortho Clinical Diagnostics
19.85 ± 1.47	12.27 ± 1.17	8.92 ± 0.82	26.76 ± 2.05	10.84 ± 1.13	n = 27	[BY1] Siemens ADVIA/ADVISIA Centaur
						[DA5] Siemens Dimension
						[DP5] Siemens Immulite

Summary of Participant Performance (Mean and Standard Deviation)

Troponin I ($\mu\text{g/L}$)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
1.043 \pm 0.226	0.520 \pm 0.180	0.017 \pm 0.017	0.276 \pm 0.083	0.018 \pm 0.018	n = 224	[---] All Methods & Instruments
6.875 \pm 0.296	3.466 \pm 0.170	0.010 \pm 0.008	2.046 \pm 0.063	0.010 \pm 0.008	n = 16	<Instruments>
0.854 \pm 0.072	0.444 \pm 0.042	0.010 \pm 0.000	0.238 \pm 0.024	0.010 \pm 0.000	n = 33	[ABH] Abbott Architect i System
0.291 \pm 0.045	0.118 \pm 0.041	0.050 \pm 0.000	0.065 \pm 0.018	0.050 \pm 0.000	n = 9	[SAA] Beckman Coulter ACCESS
4.445 \pm 0.203	1.889 \pm 0.138	0.042 \pm 0.037	1.147 \pm 0.079	0.034 \pm 0.045	n = 4	[BSA] BioSite Triage
4.009 \pm 0.177	2.032 \pm 0.100	0.010 \pm 0.000	1.199 \pm 0.064	0.010 \pm 0.000	n = 9	[IAA] i-STAT
4.137 \pm 0.125	2.115 \pm 0.087	0.010 \pm 0.000	1.216 \pm 0.059	0.010 \pm 0.000	n = 14	[JJG] Ortho Vitros 5600
1.068 \pm 0.073	0.472 \pm 0.041	0.009 \pm 0.007	0.245 \pm 0.025	0.009 \pm 0.007	n = 42	[JJC] Ortho Vitros ECi/ECiQ
0.936 \pm 0.053	0.400 \pm 0.050	0.015 \pm 0.023	0.226 \pm 0.018	0.015 \pm 0.023	n = 4	[COB] Siemens ADVIA Centaur
1.326 \pm 0.108	0.807 \pm 0.054	0.053 \pm 0.048	0.403 \pm 0.035	0.053 \pm 0.048	n = 9	[BYP] Siemens ADVIA Centaur CP
1.023 \pm 0.087	0.519 \pm 0.048	0.040 \pm 0.000	0.287 \pm 0.042	0.039 \pm 0.016	n = 22	[DUE] Siemens Dimension EXL
1.299 \pm 0.123	0.765 \pm 0.077	0.020 \pm 0.000	0.391 \pm 0.047	0.020 \pm 0.000	n = 30	[DUR] Siemens Dimension RxL
1.006 \pm 0.081	0.517 \pm 0.042	0.029 \pm 0.023	0.284 \pm 0.026	0.026 \pm 0.022	n = 11	[DUT] Siemens Dimension Vista
2.138 \pm 0.220	1.158 \pm 0.055	0.200 \pm 0.000	0.794 \pm 0.044	0.200 \pm 0.000	n = 5	[DUX] Siemens Dimension Xpand
7.891 \pm 0.340	4.461 \pm 0.421	0.060 \pm 0.000	2.496 \pm 0.262	0.060 \pm 0.000	n = 6	[DPD] Siemens Immulite 2000
6.875 \pm 0.297	3.466 \pm 0.170	0.011 \pm 0.008	2.046 \pm 0.064	0.009 \pm 0.008	n = 21	[TOM] Tosoh Bioscience
0.849 \pm 0.071	0.442 \pm 0.039	0.010 \pm 0.000	0.236 \pm 0.024	0.010 \pm 0.000	n = 35	<Reagents>
0.291 \pm 0.045	0.118 \pm 0.041	0.050 \pm 0.000	0.065 \pm 0.018	0.050 \pm 0.000	n = 9	[BC1] Beckman Coulter
4.087 \pm 0.167	2.086 \pm 0.101	0.010 \pm 0.000	1.209 \pm 0.062	0.010 \pm 0.000	n = 23	[BS1] Biosite Diagnostics
0.733 \pm 0.051	0.303 \pm 0.005	0.303 \pm 0.005	0.303 \pm 0.005	0.303 \pm 0.005	n = 3	[JJ1] Ortho Clinical Diagnostics
1.058 \pm 0.083	0.468 \pm 0.046	0.009 \pm 0.006	0.243 \pm 0.025	0.009 \pm 0.006	n = 46	[RO3] Roche Elecsys/Modular E/e601/e411
1.033 \pm 0.105	0.522 \pm 0.055	0.033 \pm 0.016	0.292 \pm 0.042	0.033 \pm 0.019	n = 36	[BY1] Siemens ADVIA/ADVIS Centaur
1.310 \pm 0.130	0.780 \pm 0.079	0.028 \pm 0.032	0.398 \pm 0.047	0.028 \pm 0.032	n = 35	[DA5] Siemens Dimension
2.071 \pm 0.258	1.141 \pm 0.067	0.200 \pm 0.000	0.780 \pm 0.055	0.200 \pm 0.000	n = 6	[DA6] Siemens Dimension LOCI
8.194 \pm 0.726	4.570 \pm 0.447	0.060 \pm 0.000	2.557 \pm 0.261	0.060 \pm 0.000	n = 5	[DP5] Siemens Immulite
						[TO2] Tosoh ST AIA

Summary of Participant Performance (Mean and Standard Deviation)

Troponin T (µg/L)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
0.885 ± 0.072	0.527 ± 0.042	0.010 ± 0.000	0.314 ± 0.020	0.010 ± 0.000	n = 33	[---] All Methods & Instruments
0.875 ± 0.019	0.530 ± 0.018	0.010 ± 0.000	0.315 ± 0.019	0.010 ± 0.000	n = 3	<Instruments>
0.872 ± 0.066	0.529 ± 0.049	0.010 ± 0.000	0.315 ± 0.020	0.010 ± 0.000	n = 8	[ROF] Roche cobas e411
0.947 ± 0.051	0.558 ± 0.031	0.010 ± 0.000	0.326 ± 0.021	0.010 ± 0.000	n = 12	[ROA] Roche cobas e601
0.837 ± 0.040	0.501 ± 0.020	0.010 ± 0.000	0.303 ± 0.012	0.010 ± 0.000	n = 9	[BME] Roche Elecsys
0.880 ± 0.066	0.525 ± 0.041	0.010 ± 0.000	0.314 ± 0.022	0.010 ± 0.000	n = 29	[ROE] Roche MODULAR E
						<Reagents>
						[RO3] Roche Elecsys/Modular E/e601/e411

Summary of Participant Performance (Mean and Standard Deviation)

Alanine Aminotransferase (U/L 37°C)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
89.5 ± 7.36	231.8 ± 12.51	50.7 ± 7.78	175.4 ± 9.58	59.7 ± 7.53	n = 363	[---] All Methods & Instruments
79.7 ± 3.16	202.0 ± 5.48	44.5 ± 1.86	153.4 ± 3.87	52.7 ± 0.51	n = 3	<Instruments>
87.1 ± 2.59	234.3 ± 6.25	46.4 ± 0.80	175.5 ± 4.08	55.7 ± 1.15	n = 13	[AXA] Abaxis Piccolo
78.5 ± 2.36	206.8 ± 5.63	41.9 ± 1.65	154.9 ± 4.63	50.0 ± 1.70	n = 51	[ABJ] Abbott Architect c System
85.0 ± 2.23	221.1 ± 6.25	45.2 ± 1.03	167.0 ± 4.55	54.7 ± 0.54	n = 9	[OLC] Beckman Coulter AU Chemistry System
85.6 ± 2.62	223.7 ± 4.72	47.3 ± 1.22	170.5 ± 3.57	56.2 ± 1.60	n = 10	[BCS] Beckman Coulter CX
85.7 ± 2.11	224.6 ± 3.64	47.4 ± 1.45	170.1 ± 2.49	56.4 ± 1.68	n = 20	[BCX] Beckman Coulter LX-20
85.9 ± 2.21	226.1 ± 3.76	47.1 ± 1.39	171.4 ± 3.64	56.8 ± 1.53	n = 23	[BCG] Beckman Coulter UniCel DxC 600
100.3 ± 3.66	240.5 ± 6.77	63.3 ± 3.53	179.2 ± 4.49	70.2 ± 3.09	n = 13	[BCH] Beckman Coulter UniCel DxC 800
100.7 ± 5.65	237.9 ± 7.70	64.4 ± 4.73	177.5 ± 6.28	69.8 ± 5.25	n = 22	[JJE] Ortho Vitros 250/350/950
99.3 ± 6.06	242.1 ± 7.93	63.4 ± 4.07	177.4 ± 5.83	69.4 ± 5.21	n = 11	[JJF] Ortho Vitros 5,1FS
89.5 ± 2.79	240.0 ± 4.65	48.3 ± 2.07	180.3 ± 5.03	57.9 ± 2.07	n = 17	[JJG] Ortho Vitros 5600
87.3 ± 3.33	232.9 ± 6.34	46.9 ± 2.06	175.9 ± 5.62	55.9 ± 2.14	n = 12	[ROC] Roche cobas c501
89.3 ± 3.12	238.0 ± 6.31	48.3 ± 2.53	178.0 ± 4.73	57.6 ± 2.22	n = 35	[ROT] Roche Cobas INTEGRA
92.1 ± 3.01	244.4 ± 5.47	50.6 ± 2.59	184.3 ± 4.95	60.2 ± 2.87	n = 16	[ROD] Roche MODULAR D/P
88.2 ± 4.89	238.2 ± 14.99	47.7 ± 3.07	178.9 ± 11.62	56.6 ± 4.72	n = 3	[BYE] Siemens ADVIA 1800
95.9 ± 1.78	237.2 ± 3.34	59.5 ± 1.29	181.2 ± 4.00	69.0 ± 1.53	n = 12	[BYB] Siemens ADVIA 2400
96.2 ± 3.09	235.2 ± 6.81	60.5 ± 2.29	181.9 ± 4.36	69.9 ± 2.59	n = 28	[DUE] Siemens Dimension EXL
90.5 ± 2.49	235.9 ± 4.88	51.0 ± 1.36	179.1 ± 3.38	60.2 ± 1.63	n = 32	[DUR] Siemens Dimension RxL
96.3 ± 2.74	234.9 ± 5.53	59.4 ± 2.11	182.4 ± 4.48	68.9 ± 1.84	n = 19	[DUT] Siemens Dimension Vista
79.7 ± 3.16	202.0 ± 5.48	44.5 ± 1.86	153.4 ± 3.87	52.7 ± 0.51	n = 3	[DUX] Siemens Dimension Xpand
87.1 ± 2.59	234.3 ± 6.25	46.4 ± 0.80	175.5 ± 4.08	55.7 ± 1.15	n = 13	<Reagents>
85.6 ± 2.29	224.6 ± 4.39	47.1 ± 1.55	170.2 ± 3.52	56.4 ± 1.70	n = 60	[AB1] Abbott
78.5 ± 2.36	206.7 ± 5.63	41.9 ± 1.62	154.7 ± 4.54	49.9 ± 1.67	n = 49	[BC1] Beckman Coulter
85.7 ± 2.15	225.3 ± 10.68	45.2 ± 1.96	169.6 ± 6.96	54.0 ± 1.66	n = 5	[OL1] Beckman Coulter AU Series
100.2 ± 5.18	239.4 ± 7.70	63.7 ± 4.27	177.8 ± 5.75	69.6 ± 4.75	n = 47	[CR1] Carolina
89.6 ± 2.72	240.0 ± 4.25	48.3 ± 1.95	180.2 ± 4.84	57.8 ± 1.97	n = 18	[JJ1] Ortho Clinical Diagnostics
89.1 ± 3.27	237.7 ± 6.41	48.2 ± 2.53	177.8 ± 4.85	57.5 ± 2.34	n = 36	[RO4] Roche cobas c311/c501/c502/c701
87.3 ± 3.33	232.9 ± 6.34	46.9 ± 2.06	175.9 ± 5.62	55.9 ± 2.14	n = 12	[RO2] Roche Hitachi and Modular D/P
90.7 ± 3.83	242.2 ± 8.10	49.7 ± 2.85	182.3 ± 6.87	59.5 ± 3.19	n = 23	[RO1] Roche Integra and MIRA
94.1 ± 4.06	235.6 ± 5.66	57.0 ± 5.30	180.7 ± 4.22	66.3 ± 5.36	n = 82	[BY1] Siemens ADVIA/ADVISIA Centaur
95.0 ± 3.23	235.6 ± 5.06	57.2 ± 4.10	182.6 ± 4.03	66.9 ± 4.04	n = 7	[DA5] Siemens Dimension
						[DA8] Siemens Dimension IFCC Standardized

Summary of Participant Performance (Mean and Standard Deviation)

Aspartate Aminotransferase (U/L 37°C)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
81.7 ± 4.80	554.3 ± 49.13	58.3 ± 4.44	184.9 ± 9.71	69.8 ± 4.99	n = 363	[---] All Methods & Instruments
83.5 ± 1.86	535.0 ± 5.48	58.0 ± 1.80	184.4 ± 3.87	69.2 ± 2.36	n = 3	<Instruments>
80.8 ± 1.89	550.0 ± 11.42	56.4 ± 1.31	184.4 ± 3.61	67.2 ± 1.30	n = 13	[AXA] Abaxis Piccolo
74.2 ± 2.47	496.1 ± 17.14	51.1 ± 1.86	166.1 ± 5.59	61.2 ± 2.15	n = 51	[ABJ] Abbott Architect c System
82.0 ± 4.91	535.1 ± 70.59	57.7 ± 5.01	183.9 ± 8.00	69.0 ± 4.52	n = 9	[OLC] Beckman Coulter AU Chemistry System
81.5 ± 2.48	531.9 ± 22.14	56.0 ± 0.69	182.4 ± 4.32	67.5 ± 1.86	n = 9	[BCS] Beckman Coulter CX
81.1 ± 2.24	531.3 ± 20.29	55.9 ± 1.05	183.2 ± 2.98	67.0 ± 1.35	n = 21	[BCX] Beckman Coulter LX-20
82.4 ± 1.73	530.7 ± 22.21	56.2 ± 1.29	185.5 ± 3.51	67.6 ± 1.54	n = 23	[BCG] Beckman Coulter UniCel DxC 600
85.8 ± 3.04	683.6 ± 27.20	62.3 ± 1.86	191.9 ± 5.94	74.8 ± 2.14	n = 13	[BCH] Beckman Coulter UniCel DxC 800
84.0 ± 2.57	666.1 ± 30.84	60.6 ± 1.81	186.4 ± 3.59	72.6 ± 1.95	n = 22	[JJE] Ortho Vitros 250/350/950
83.9 ± 3.55	673.4 ± 31.10	60.4 ± 1.83	186.1 ± 9.36	73.4 ± 2.73	n = 11	[JJF] Ortho Vitros 5,1FS
84.3 ± 1.96	578.6 ± 16.26	59.2 ± 3.57	191.1 ± 6.10	70.7 ± 4.04	n = 17	[JJG] Ortho Vitros 5600
82.9 ± 1.25	573.2 ± 7.34	58.4 ± 1.41	190.8 ± 2.79	69.4 ± 1.25	n = 12	[ROC] Roche cobas c501
84.0 ± 3.32	567.2 ± 18.72	59.5 ± 2.68	190.6 ± 6.28	71.0 ± 2.84	n = 35	[ROT] Roche Cobas INTEGRA
89.8 ± 3.26	602.0 ± 12.33	64.5 ± 2.72	203.4 ± 4.72	76.3 ± 2.88	n = 16	[ROD] Roche MODULAR D/P
87.5 ± 6.32	587.6 ± 32.02	60.8 ± 4.11	198.0 ± 10.94	73.0 ± 4.51	n = 3	[BYE] Siemens ADVIA 1800
80.4 ± 1.80	551.7 ± 9.89	59.7 ± 2.12	182.3 ± 2.51	72.1 ± 1.81	n = 11	[BYB] Siemens ADVIA 2400
80.6 ± 2.95	548.3 ± 15.13	61.4 ± 2.54	183.3 ± 5.06	73.0 ± 2.57	n = 29	[DUE] Siemens Dimension EXL
77.6 ± 4.34	547.7 ± 14.22	56.8 ± 2.86	181.5 ± 5.29	69.2 ± 3.68	n = 32	[DUR] Siemens Dimension RxL
82.7 ± 2.59	549.5 ± 21.64	61.4 ± 1.53	185.3 ± 5.29	73.0 ± 2.42	n = 19	[DUT] Siemens Dimension Vista
83.5 ± 1.86	535.0 ± 5.48	58.0 ± 1.80	184.4 ± 3.87	69.2 ± 2.36	n = 3	[DUX] Siemens Dimension Xpand
80.8 ± 1.89	550.0 ± 11.42	56.4 ± 1.31	184.4 ± 3.61	67.2 ± 1.30	n = 13	<Reagents>
81.5 ± 2.29	530.2 ± 22.01	56.0 ± 1.28	183.9 ± 4.00	67.4 ± 1.66	n = 60	[AB1] Abbott
74.2 ± 2.36	496.1 ± 17.00	51.1 ± 1.79	166.1 ± 5.46	61.2 ± 2.12	n = 49	[BC1] Beckman Coulter
85.7 ± 3.56	588.0 ± 28.03	61.2 ± 3.60	190.9 ± 6.47	72.0 ± 2.64	n = 5	[OL1] Beckman Coulter AU Series
84.5 ± 3.04	674.0 ± 31.70	61.0 ± 2.03	187.9 ± 5.50	73.4 ± 2.35	n = 47	[CR1] Carolina
84.5 ± 2.19	579.6 ± 16.18	59.3 ± 3.36	191.5 ± 6.11	70.7 ± 3.80	n = 18	[JJ1] Ortho Clinical Diagnostics
83.9 ± 3.30	567.1 ± 18.19	59.4 ± 2.62	190.6 ± 6.14	70.9 ± 2.77	n = 36	[RO4] Roche cobas c311/c501/c502/c701
82.9 ± 1.32	574.0 ± 7.17	58.2 ± 1.16	191.0 ± 2.87	69.3 ± 1.16	n = 11	[RO2] Roche Hitachi and Modular D/P
88.9 ± 3.83	595.5 ± 19.37	63.7 ± 3.35	201.0 ± 7.18	75.4 ± 3.31	n = 23	[RO1] Roche Integra and MIRA
80.3 ± 3.96	548.1 ± 15.40	59.8 ± 3.45	182.8 ± 5.26	71.7 ± 3.43	n = 86	[BY1] Siemens ADVIA/ADVISIA Centaur
79.5 ± 1.86	558.4 ± 10.30	58.1 ± 2.05	185.0 ± 3.61	71.8 ± 1.54	n = 3	[DA5] Siemens Dimension
						[DA8] Siemens Dimension IFCC Standardized

Summary of Participant Performance (Mean and Standard Deviation)

 α -Amylase (U/L 37°C)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
271.1 ± 45.33	53.8 ± 7.79	122.8 ± 16.45	80.3 ± 8.05	146.3 ± 20.14	n = 319	[---] All Methods & Instruments
293.5 ± 5.72	54.0 ± 1.27	131.0 ± 2.58	85.7 ± 2.04	156.3 ± 2.75	n = 10	<Instruments>
233.3 ± 12.36	41.5 ± 2.47	103.8 ± 6.61	66.8 ± 4.02	123.7 ± 7.31	n = 42	[ABJ] Abbott Architect c System
368.8 ± 92.69	64.7 ± 3.77	162.2 ± 36.87	104.2 ± 20.34	192.2 ± 38.51	n = 4	[OLC] Beckman Coulter AU Chemistry System
287.5 ± 8.42	60.4 ± 2.10	130.1 ± 2.65	87.4 ± 2.47	156.5 ± 4.12	n = 9	[BCS] Beckman Coulter CX
283.9 ± 10.21	60.2 ± 1.77	129.9 ± 4.14	86.5 ± 2.90	154.5 ± 4.99	n = 17	[BCX] Beckman Coulter IX-20
289.2 ± 4.73	61.1 ± 1.45	130.8 ± 4.45	87.9 ± 1.76	156.1 ± 4.43	n = 23	[BCG] Beckman Coulter UniCel DxC 600
166.7 ± 9.41	54.5 ± 5.31	80.1 ± 3.60	64.7 ± 4.32	95.0 ± 5.45	n = 10	[BCH] Beckman Coulter UniCel DxC 800
168.9 ± 9.77	57.2 ± 5.42	82.2 ± 6.93	68.9 ± 5.31	98.3 ± 8.31	n = 21	[JJE] Ortho Vitros 250/350/950
169.6 ± 4.97	57.0 ± 5.38	81.4 ± 7.12	66.8 ± 4.84	96.5 ± 5.67	n = 11	[JJF] Ortho Vitros 5,1FS
264.1 ± 4.36	58.6 ± 1.07	122.0 ± 1.84	83.2 ± 1.40	145.9 ± 3.04	n = 17	[JTG] Ortho Vitros 5600
262.4 ± 4.32	58.8 ± 0.65	121.4 ± 1.29	82.7 ± 0.55	144.1 ± 1.20	n = 9	[ROC] Roche cobas c501
262.0 ± 2.83	58.5 ± 1.55	121.5 ± 1.86	83.4 ± 1.60	145.4 ± 2.56	n = 31	[ROT] Roche Cobas INTEGRA
274.3 ± 5.08	59.3 ± 1.16	124.9 ± 2.51	85.1 ± 1.89	150.4 ± 3.14	n = 15	[ROD] Roche MODULAR D/P
257.7 ± 2.26	55.3 ± 0.51	117.0 ± 0.90	80.0 ± 0.90	140.3 ± 0.51	n = 3	[BYE] Siemens ADVIA 1800
319.1 ± 5.52	49.4 ± 0.98	136.2 ± 2.41	83.0 ± 2.46	163.0 ± 2.75	n = 9	[BYB] Siemens ADVIA 2400
316.4 ± 7.50	49.5 ± 1.47	136.9 ± 3.52	83.1 ± 1.71	163.2 ± 3.39	n = 29	[DUE] Siemens Dimension EXL
303.1 ± 5.61	45.7 ± 1.06	131.0 ± 1.83	78.4 ± 1.74	155.9 ± 3.33	n = 31	[DUR] Siemens Dimension RxL
319.8 ± 6.79	49.4 ± 1.26	137.1 ± 3.51	83.2 ± 1.60	163.1 ± 3.82	n = 17	[DUT] Siemens Dimension Vista
293.5 ± 5.72	54.0 ± 1.27	131.0 ± 2.58	85.7 ± 2.04	156.3 ± 2.75	n = 10	[DUX] Siemens Dimension Xpand
287.1 ± 8.10	60.7 ± 1.83	130.4 ± 3.86	87.3 ± 2.53	155.9 ± 4.49	n = 52	<Reagents>
232.9 ± 12.36	41.5 ± 2.38	103.6 ± 6.45	66.6 ± 3.88	123.4 ± 7.21	n = 41	[BC2] Beckman Coulter IFCC Standardized
168.9 ± 8.97	56.8 ± 6.06	81.3 ± 6.48	67.7 ± 5.94	97.1 ± 7.62	n = 44	[OL1] Beckman Coulter AU Series
263.9 ± 4.31	58.5 ± 1.03	122.0 ± 1.78	83.2 ± 1.34	145.8 ± 2.98	n = 18	[JJ1] Ortho Clinical Diagnostics
262.1 ± 3.04	58.5 ± 1.51	121.6 ± 1.83	83.5 ± 1.60	145.5 ± 2.68	n = 32	[RO4] Roche cobas c311/c501/c502/c701
262.4 ± 4.32	58.8 ± 0.65	121.4 ± 1.29	82.7 ± 0.55	144.1 ± 1.20	n = 9	[RO2] Roche Hitachi and Modular D/P
269.8 ± 9.39	58.2 ± 2.15	122.8 ± 4.35	83.7 ± 2.89	147.7 ± 5.44	n = 22	[RO1] Roche Integra and MIRA
312.7 ± 10.09	48.2 ± 2.31	134.7 ± 4.23	81.4 ± 3.08	160.7 ± 4.96	n = 85	[BY1] Siemens ADVIA/ADVISIA Centaur
						[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Alkaline Phosphatase (U/L 37°C)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
187.8 ± 19.97	44.7 ± 8.39	235.7 ± 20.30	93.2 ± 14.63	285.0 ± 24.77	n = 360	[---] All Methods & Instruments
162.5 ± 5.43	32.1 ± 2.86	207.7 ± 1.37	77.9 ± 6.08	251.9 ± 4.38	n = 3	<Instruments>
196.2 ± 5.73	45.3 ± 1.53	252.3 ± 8.54	95.3 ± 3.40	301.4 ± 10.26	n = 13	[AXA] Abaxis Piccolo
172.7 ± 7.91	39.2 ± 2.49	222.9 ± 11.03	83.4 ± 4.21	266.3 ± 13.07	n = 50	[ABJ] Abbott Architect c System
167.3 ± 4.57	39.0 ± 1.38	219.0 ± 7.50	81.8 ± 3.41	259.9 ± 11.91	n = 9	[OLC] Beckman Coulter AU Chemistry System
170.9 ± 6.36	37.4 ± 0.73	222.3 ± 8.72	82.0 ± 3.11	264.4 ± 9.77	n = 9	[BCS] Beckman Coulter CX
169.4 ± 6.09	39.1 ± 2.14	218.9 ± 8.02	81.8 ± 3.48	262.3 ± 9.30	n = 21	[BCX] Beckman Coulter LX-20
172.7 ± 5.83	40.2 ± 1.91	221.7 ± 8.06	83.5 ± 3.37	266.3 ± 6.76	n = 24	[BCG] Beckman Coulter UniCel DxC 600
214.7 ± 10.28	64.3 ± 3.08	232.8 ± 13.94	126.1 ± 6.39	296.6 ± 16.16	n = 12	[BCH] Beckman Coulter UniCel DxC 800
218.5 ± 6.18	64.3 ± 2.57	232.2 ± 8.60	127.7 ± 3.69	294.2 ± 10.47	n = 22	[JJE] Ortho Vitros 250/350/950
213.6 ± 6.13	63.5 ± 4.39	225.6 ± 6.88	125.6 ± 5.03	289.5 ± 17.60	n = 11	[JJF] Ortho Vitros 5,1FS
188.4 ± 4.76	43.8 ± 1.26	243.7 ± 6.01	92.2 ± 2.61	290.4 ± 7.01	n = 17	[JJG] Ortho Vitros 5600
186.6 ± 4.86	42.5 ± 1.23	242.4 ± 3.61	90.7 ± 3.24	288.9 ± 4.39	n = 10	[ROC] Roche cobas c501
182.3 ± 4.82	42.9 ± 1.52	234.6 ± 6.54	90.2 ± 2.90	280.3 ± 8.48	n = 34	[ROT] Roche Cobas INTEGRA
198.1 ± 6.96	44.6 ± 1.75	256.1 ± 8.21	95.5 ± 3.41	307.0 ± 10.55	n = 16	[ROD] Roche MODULAR D/P
189.9 ± 2.05	42.0 ± 0.90	243.2 ± 2.36	90.5 ± 1.86	291.7 ± 1.37	n = 3	[BYE] Siemens ADVIA 1800
220.0 ± 9.73	59.5 ± 6.22	283.4 ± 6.01	112.3 ± 4.72	336.0 ± 9.52	n = 11	[BYB] Siemens ADVIA 2400
211.8 ± 10.79	55.7 ± 7.44	278.2 ± 15.13	110.6 ± 8.49	328.6 ± 13.22	n = 29	[DUE] Siemens Dimension EXL
177.7 ± 10.05	42.8 ± 2.09	229.0 ± 11.38	88.3 ± 4.26	274.2 ± 13.03	n = 32	[DUR] Siemens Dimension RxL
201.1 ± 4.57	51.8 ± 3.65	261.2 ± 4.03	103.2 ± 3.64	314.8 ± 7.86	n = 19	[DUT] Siemens Dimension Vista
162.5 ± 5.43	32.1 ± 2.86	207.7 ± 1.37	77.9 ± 6.08	251.9 ± 4.38	n = 3	[DUX] Siemens Dimension Xpand
196.2 ± 5.73	45.3 ± 1.53	252.3 ± 8.54	95.3 ± 3.40	301.4 ± 10.26	n = 13	<Reagents>
170.6 ± 5.95	39.2 ± 2.05	220.3 ± 7.86	82.5 ± 3.48	263.9 ± 8.92	n = 58	[AB1] Abbott
172.7 ± 7.53	39.1 ± 2.39	222.7 ± 10.43	83.4 ± 4.04	266.1 ± 12.27	n = 49	[BC1] Beckman Coulter
153.9 ± 31.07	40.4 ± 4.70	199.4 ± 40.04	75.3 ± 14.32	241.8 ± 42.85	n = 4	[OL1] Beckman Coulter AU Series
216.1 ± 8.17	64.1 ± 3.16	230.7 ± 10.26	126.8 ± 4.96	293.3 ± 13.23	n = 46	[CR1] Carolina
188.2 ± 4.64	43.8 ± 1.23	243.3 ± 5.97	92.1 ± 2.54	290.1 ± 6.89	n = 18	[JJ1] Ortho Clinical Diagnostics
182.2 ± 4.67	43.0 ± 1.54	234.6 ± 6.29	90.2 ± 2.86	280.3 ± 8.16	n = 36	[RO4] Roche cobas c311/c501/c502/c701
185.9 ± 5.19	42.4 ± 1.17	241.8 ± 4.30	90.3 ± 3.04	288.2 ± 4.84	n = 11	[RO2] Roche Hitachi and Modular D/P
194.9 ± 8.36	43.9 ± 2.09	251.8 ± 10.66	93.8 ± 4.20	301.6 ± 13.20	n = 23	[RO1] Roche Integra and MIRA
199.9 ± 20.73	50.4 ± 8.57	259.5 ± 27.50	101.6 ± 12.72	309.6 ± 31.02	n = 90	[BY1] Siemens ADVIA/ADVISIA Centaur
						[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

 γ -Glutamyltransferase (U/L 37°C)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
47.0 ± 11.19	99.6 ± 23.50	70.0 ± 15.96	24.2 ± 5.85	83.8 ± 18.96	n = 305	[---] All Methods & Instruments
48.1 ± 2.35	102.4 ± 4.53	71.9 ± 3.35	24.6 ± 1.64	86.6 ± 3.80	n = 12	<Instruments>
37.1 ± 1.99	76.9 ± 3.93	54.5 ± 2.81	20.6 ± 1.22	65.4 ± 3.58	n = 48	[ABJ] Abbott Architect c System
36.4 ± 4.53	76.8 ± 14.51	54.5 ± 9.11	20.3 ± 3.01	64.8 ± 11.69	n = 8	[OLC] Beckman Coulter AU Chemistry System
41.2 ± 1.86	94.9 ± 3.38	66.2 ± 3.18	19.6 ± 1.34	78.9 ± 3.29	n = 7	[BCS] Beckman Coulter CX
41.7 ± 1.87	95.9 ± 3.17	66.9 ± 1.83	19.3 ± 0.60	80.4 ± 3.22	n = 16	[BCX] Beckman Coulter LX-20
42.7 ± 1.45	98.1 ± 2.46	68.0 ± 2.17	19.9 ± 2.38	81.8 ± 3.06	n = 18	[BCG] Beckman Coulter UniCel DxC 600
64.8 ± 1.18	153.2 ± 3.84	107.8 ± 2.87	28.2 ± 1.50	129.2 ± 2.76	n = 9	[BCH] Beckman Coulter UniCel DxC 800
65.4 ± 2.15	151.7 ± 4.64	106.9 ± 3.63	28.9 ± 1.55	127.2 ± 4.26	n = 21	[JJE] Ortho Vitros 250/350/950
64.6 ± 1.51	149.0 ± 0.82	105.1 ± 2.30	28.2 ± 1.24	126.4 ± 2.60	n = 10	[JJF] Ortho Vitros 5,1FS
39.7 ± 0.66	84.9 ± 2.14	60.5 ± 1.28	20.6 ± 0.64	72.8 ± 1.19	n = 15	[JG] Ortho Vitros 5600
38.3 ± 1.27	82.5 ± 1.67	58.2 ± 1.47	19.7 ± 0.85	69.6 ± 1.71	n = 9	[ROC] Roche cobas c501
40.2 ± 1.34	87.0 ± 2.22	61.7 ± 1.93	20.5 ± 0.97	73.8 ± 1.82	n = 34	[ROT] Roche Cobas INTEGRA
44.5 ± 1.21	92.9 ± 2.17	66.2 ± 1.51	23.7 ± 0.99	79.5 ± 2.16	n = 16	[ROD] Roche MODULAR D/P
41.0 ± 0.90	87.9 ± 2.05	61.2 ± 1.54	22.2 ± 1.54	76.3 ± 3.37	n = 3	[BYE] Siemens ADVIA 1800
59.0 ± 1.62	116.2 ± 2.84	81.9 ± 1.83	34.2 ± 2.23	97.8 ± 2.60	n = 10	[BYB] Siemens ADVIA 2400
58.4 ± 2.76	116.6 ± 3.18	81.6 ± 2.52	33.7 ± 2.01	96.8 ± 3.23	n = 22	[DUE] Siemens Dimension EXL
56.9 ± 2.08	118.8 ± 2.61	82.3 ± 2.63	31.6 ± 2.34	98.1 ± 2.41	n = 29	[DUR] Siemens Dimension RxL
58.7 ± 1.75	117.3 ± 2.23	81.9 ± 2.22	34.4 ± 1.12	97.7 ± 2.23	n = 9	[DUT] Siemens Dimension Vista
						[DUX] Siemens Dimension Xpand
47.9 ± 1.94	101.9 ± 4.12	71.6 ± 2.84	24.4 ± 1.39	86.3 ± 3.18	n = 11	<Reagents>
41.9 ± 2.07	96.7 ± 3.47	67.2 ± 2.46	19.5 ± 1.50	80.6 ± 3.55	n = 44	[AB1] Abbott
37.0 ± 2.00	76.7 ± 3.94	54.5 ± 2.74	20.5 ± 1.22	65.4 ± 3.54	n = 46	[BC1] Beckman Coulter
34.7 ± 0.90	67.5 ± 1.07	48.8 ± 0.80	21.6 ± 3.50	58.2 ± 4.50	n = 5	[OL1] Beckman Coulter AU Series
65.0 ± 1.82	151.7 ± 4.34	106.7 ± 3.36	28.5 ± 1.50	127.4 ± 3.81	n = 40	[CR1] Carolina
39.7 ± 0.66	84.9 ± 2.14	60.5 ± 1.28	20.6 ± 0.64	72.8 ± 1.19	n = 15	[JJ1] Ortho Clinical Diagnostics
40.2 ± 1.35	87.0 ± 2.22	61.7 ± 1.93	20.5 ± 0.99	73.8 ± 1.82	n = 35	[RO4] Roche cobas c311/c501/c502/c701
38.3 ± 1.27	82.5 ± 1.67	58.2 ± 1.47	19.7 ± 0.85	69.6 ± 1.71	n = 9	[RO2] Roche Hitachi and Modular D/P
43.7 ± 2.16	91.7 ± 3.18	65.1 ± 2.77	23.2 ± 1.41	78.7 ± 2.90	n = 23	[RO1] Roche Integra and MIRA
58.0 ± 2.31	117.6 ± 2.96	82.0 ± 2.50	33.2 ± 2.37	97.7 ± 2.80	n = 66	[BY1] Siemens ADVIA/ADVISIA Centaur
55.8 ± 2.36	116.0 ± 0.90	81.3 ± 0.51	31.4 ± 2.56	97.0 ± 0.90	n = 3	[DA5] Siemens Dimension
						[DA8] Siemens Dimension IFCC Standardized

Summary of Participant Performance (Mean and Standard Deviation)

Creatine Kinase (U/L 37°C)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
226.4 ± 18.94	277.6 ± 25.52	79.1 ± 7.76	154.0 ± 10.41	95.4 ± 8.04	n = 328	[---] All Methods & Instruments
246.6 ± 4.85	304.7 ± 7.92	84.3 ± 1.80	167.8 ± 2.45	102.9 ± 1.67	n = 11	<Instruments>
195.4 ± 9.57	241.2 ± 15.03	68.0 ± 2.83	135.4 ± 5.55	82.2 ± 3.59	n = 45	[ABJ] Abbott Architect c System
195.2 ± 19.78	232.1 ± 26.69	71.2 ± 7.47	135.5 ± 11.38	87.9 ± 9.44	n = 7	[OLC] Beckman Coulter AU Chemistry System
227.7 ± 6.49	288.3 ± 14.61	82.8 ± 2.28	156.3 ± 3.50	99.1 ± 2.47	n = 9	[BCS] Beckman Coulter CX
224.4 ± 7.93	279.7 ± 14.80	81.6 ± 1.50	154.3 ± 3.80	98.2 ± 2.98	n = 18	[BCX] Beckman Coulter LX-20
226.4 ± 6.43	280.9 ± 11.78	83.2 ± 2.30	154.7 ± 3.79	99.1 ± 2.72	n = 23	[BCG] Beckman Coulter UniCel DxC 600
242.8 ± 22.25	293.8 ± 26.68	92.3 ± 6.87	162.8 ± 11.92	107.4 ± 5.78	n = 10	[BCH] Beckman Coulter UniCel DxC 800
242.3 ± 12.02	286.2 ± 18.78	95.3 ± 3.71	161.2 ± 7.68	107.1 ± 4.17	n = 20	[JJE] Ortho Vitros 250/350/950
242.0 ± 14.97	293.5 ± 21.60	96.0 ± 4.43	163.7 ± 9.66	107.8 ± 5.38	n = 10	[JJF] Ortho Vitros 5,1FS
222.4 ± 8.51	280.6 ± 14.68	76.1 ± 2.25	151.9 ± 3.98	92.0 ± 2.54	n = 17	[JGJ] Ortho Vitros 5600
206.3 ± 7.60	254.4 ± 11.15	68.5 ± 3.46	138.6 ± 4.90	83.4 ± 4.02	n = 8	[ROC] Roche cobas c501
230.9 ± 5.42	281.1 ± 11.63	79.3 ± 1.59	155.9 ± 3.43	95.6 ± 1.93	n = 34	[ROT] Roche Cobas INTEGRA
216.8 ± 10.30	266.5 ± 20.28	76.8 ± 1.97	154.9 ± 4.78	93.4 ± 2.45	n = 16	[ROD] Roche MODULAR D/P
207.4 ± 12.55	248.0 ± 26.21	73.8 ± 2.36	146.5 ± 8.19	87.8 ± 2.36	n = 3	[BYE] Siemens ADVIA 1800
237.9 ± 8.24	290.1 ± 16.15	79.1 ± 2.10	158.0 ± 5.01	95.9 ± 2.83	n = 10	[BYB] Siemens ADVIA 2400
238.3 ± 6.97	294.8 ± 10.00	79.9 ± 2.44	159.1 ± 3.84	97.0 ± 2.97	n = 30	[DUE] Siemens Dimension EXL
237.1 ± 5.44	284.8 ± 22.27	80.8 ± 1.40	155.5 ± 8.26	97.3 ± 1.96	n = 31	[DUR] Siemens Dimension RxL
236.7 ± 6.09	293.2 ± 12.35	78.7 ± 2.08	157.8 ± 4.02	95.6 ± 2.12	n = 16	[DUT] Siemens Dimension Vista
246.6 ± 4.85	304.7 ± 7.92	84.3 ± 1.80	167.8 ± 2.45	102.9 ± 1.67	n = 11	[DUX] Siemens Dimension Xpand
225.4 ± 7.23	281.3 ± 14.21	82.5 ± 2.38	154.7 ± 4.25	98.6 ± 2.83	n = 53	<Reagents>
195.2 ± 9.62	241.2 ± 14.37	67.9 ± 2.79	135.2 ± 5.51	82.1 ± 3.57	n = 43	[BC1] Beckman Coulter
184.6 ± 10.35	229.4 ± 8.79	68.2 ± 3.11	130.0 ± 5.88	82.2 ± 5.30	n = 5	[OL1] Beckman Coulter AU Series
242.7 ± 14.39	289.7 ± 21.24	95.0 ± 4.52	162.3 ± 8.99	107.4 ± 4.70	n = 40	[CR1] Carolina
221.7 ± 8.72	277.9 ± 18.83	76.1 ± 2.16	152.0 ± 3.81	92.0 ± 2.43	n = 18	[JJ1] Ortho Clinical Diagnostics
230.7 ± 5.29	280.6 ± 11.56	79.3 ± 1.62	155.8 ± 3.47	95.5 ± 1.86	n = 33	[RO4] Roche cobas c311/c501/c502/c701
206.3 ± 7.60	254.4 ± 11.15	68.5 ± 3.46	138.6 ± 4.90	83.4 ± 4.02	n = 8	[RO2] Roche Hitachi and Modular D/P
214.6 ± 10.50	263.0 ± 20.65	76.1 ± 2.30	153.3 ± 5.98	92.2 ± 3.26	n = 23	[RO1] Roche Integra and MIRA
230.6 ± 18.92	277.3 ± 31.92	78.6 ± 5.50	154.0 ± 9.84	94.6 ± 6.47	n = 31	[BY1] Siemens ADVIA/ADVISIA Centaur
237.6 ± 6.71	293.4 ± 11.53	79.8 ± 2.50	158.7 ± 4.07	96.7 ± 2.75	n = 54	[DA5] Siemens Dimension
						[DA8] Siemens Dimension IFCC Standardized

Summary of Participant Performance (Mean and Standard Deviation)

Creatine Kinase-MB (ng/mL, U/L 37°C, %)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
27.79 ± 4.39	63.97 ± 10.43	0.94 ± 0.35	0.98 ± 0.36	1.08 ± 0.40	n = 196	[-A-] All Methods - Results reported in ng/mL
27.80 ± 2.93	58.84 ± 8.24	0.86 ± 0.15	0.94 ± 0.16	1.04 ± 0.21	n = 18	[AB1] Abbott
31.39 ± 1.18	76.76 ± 4.05	2.10 ± 1.96	2.19 ± 1.90	2.27 ± 1.84	n = 22	[SAA] Beckman Coulter ACCESS
32.19 ± 1.34	77.82 ± 3.45	1.14 ± 0.06	1.28 ± 0.06	1.32 ± 0.07	n = 14	[BC-] Beckman Coulter DxC 600/DxI 800
19.76 ± 0.85	45.68 ± 5.46	1.00 ± 0.00	1.00 ± 0.00	1.00 ± 0.00	n = 5	[BS1] Biosite Diagnostics
20.63 ± 1.27	49.23 ± 3.19	0.78 ± 0.10	0.84 ± 0.12	0.91 ± 0.11	n = 20	[JJ1] Ortho Clinical Diagnostics
31.79 ± 1.30	64.79 ± 2.27	1.64 ± 0.13	1.50 ± 0.07	1.71 ± 0.08	n = 25	[RO3] Roche Elecsys/Modular E/e601/e411
26.66 ± 1.02	61.36 ± 2.13	0.75 ± 0.17	0.76 ± 0.16	0.85 ± 0.15	n = 33	[BY1] Siemens ADVIA/ADVIA Centaur
26.86 ± 2.28	69.74 ± 5.83	0.65 ± 0.28	0.62 ± 0.23	0.70 ± 0.27	n = 32	[DA5] Siemens Dimension
24.10 ± 0.80	56.06 ± 1.69	0.92 ± 0.12	0.84 ± 0.17	0.97 ± 0.18	n = 18	[DA6] Siemens Dimension LOCI
29.30 ± 0.36	65.22 ± 2.14	1.99 ± 0.37	1.96 ± 0.47	1.97 ± 0.23	n = 3	[DP5] Siemens Immulite
35.93 ± 0.68	86.53 ± 1.95	1.37 ± 0.05	1.53 ± 0.05	1.63 ± 0.05	n = 3	[TO2] Tosoh ST AIA
-----	-----	-----	-----	-----	-----	-----
29.05 ± 4.44	67.22 ± 9.45	1.37 ± 0.99	2.11 ± 2.48	1.56 ± 1.01	n = 13	[-B-] All Methods - Results reported in U/L
-----	-----	-----	-----	-----	-----	-----
11.00 ± 1.76	20.31 ± 2.65	0.00 ± 0.00	0.00 ± 0.00	0.00 ± 0.00	n = 4	[-P-] All Methods - Results reported as %
10.53 ± 1.86	20.93 ± 2.86	0.00 ± 0.00	0.00 ± 0.00	0.00 ± 0.00	n = 3	[HL1] Helena Laboratories

Summary of Participant Performance (Mean and Standard Deviation)

Lactate Dehydrogenase (U/L 37°C)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
178.3 ± 19.59	237.7 ± 23.50	94.7 ± 9.76	172.1 ± 18.63	114.4 ± 11.61	n = 316	[---] All Methods & Instruments
178.7 ± 19.48	238.2 ± 23.35	94.9 ± 9.70	172.5 ± 18.56	114.6 ± 11.55	n = 268	[A-] All Methods - Lactate to Pyruvate
476.3 ± 16.52	641.1 ± 21.62	262.6 ± 23.04	482.8 ± 22.62	337.9 ± 24.38	n = 44	[B-] All Methods - Pyruvate to Lactate
						<Instruments>
176.4 ± 7.92	235.8 ± 12.09	97.2 ± 3.68	172.4 ± 6.71	117.2 ± 4.30	n = 13	[ABJ] Abbott Architect c System
163.2 ± 6.98	220.8 ± 10.23	87.4 ± 4.05	156.8 ± 7.09	104.7 ± 4.81	n = 46	[OLC] Beckman Coulter AU Chemistry System
152.9 ± 7.14	205.0 ± 9.23	82.3 ± 6.14	151.4 ± 11.96	102.9 ± 6.56	n = 7	[BCS] Beckman Coulter CX
154.2 ± 5.17	210.6 ± 6.64	82.4 ± 3.14	150.7 ± 4.13	100.0 ± 2.78	n = 8	[BCX] Beckman Coulter LX-20
153.6 ± 4.87	208.8 ± 4.63	83.4 ± 1.90	148.3 ± 4.78	99.6 ± 3.13	n = 19	[BCG] Beckman Coulter UniCel DxC 600
154.7 ± 5.91	209.5 ± 5.80	82.8 ± 3.47	150.3 ± 4.45	101.0 ± 2.99	n = 21	[BCH] Beckman Coulter UniCel DxC 800
489.6 ± 10.55	650.2 ± 15.30	266.9 ± 18.80	497.6 ± 17.31	348.1 ± 23.22	n = 10	[JJE] Ortho Vitros 250/350/950
472.7 ± 17.40	638.0 ± 26.86	261.6 ± 26.75	478.3 ± 21.29	331.9 ± 23.48	n = 22	[JJF] Ortho Vitros 5,1FS
472.1 ± 13.29	641.1 ± 18.05	259.1 ± 13.14	479.0 ± 18.60	337.2 ± 17.89	n = 11	[JJG] Ortho Vitros 5600
188.8 ± 3.83	249.4 ± 4.89	99.3 ± 2.28	181.5 ± 3.45	119.2 ± 2.93	n = 17	[ROC] Roche cobas c501
189.9 ± 3.25	248.3 ± 6.11	99.7 ± 2.78	181.3 ± 4.46	120.3 ± 2.24	n = 8	[ROT] Roche Cobas INTEGRA
187.9 ± 4.98	248.0 ± 6.53	98.6 ± 2.66	180.3 ± 3.98	118.9 ± 3.18	n = 32	[ROD] Roche MODULAR D/P
190.8 ± 6.27	253.6 ± 8.34	101.6 ± 3.73	183.7 ± 5.80	122.3 ± 4.11	n = 16	[BYE] Siemens ADVIA 1800
187.0 ± 3.61	245.8 ± 5.90	99.7 ± 3.16	179.5 ± 5.43	120.2 ± 3.23	n = 3	[BYB] Siemens ADVIA 2400
191.9 ± 3.79	254.3 ± 5.36	98.8 ± 2.92	186.8 ± 7.29	122.8 ± 3.97	n = 9	[DUE] Siemens Dimension EXL
197.3 ± 8.96	260.0 ± 11.23	104.1 ± 5.26	191.3 ± 12.42	124.1 ± 7.37	n = 21	[DUR] Siemens Dimension RxL
194.7 ± 13.53	256.5 ± 17.58	102.7 ± 7.76	185.6 ± 14.11	123.2 ± 9.75	n = 32	[DUT] Siemens Dimension Vista
193.0 ± 9.12	258.1 ± 8.02	100.0 ± 6.20	187.0 ± 8.45	123.0 ± 7.19	n = 12	[DUX] Siemens Dimension Xpand
						<Reagents>
176.4 ± 7.92	235.8 ± 12.09	97.2 ± 3.68	172.4 ± 6.71	117.2 ± 4.30	n = 13	[AB1] Abbott
154.0 ± 5.26	209.1 ± 5.48	82.8 ± 2.95	149.7 ± 4.46	100.3 ± 2.95	n = 51	[BC1] Beckman Coulter
163.2 ± 6.95	220.8 ± 10.10	87.4 ± 3.99	156.8 ± 7.03	104.6 ± 4.74	n = 45	[OL1] Beckman Coulter AU Series
153.6 ± 10.55	204.6 ± 13.21	85.8 ± 6.35	157.7 ± 19.18	104.8 ± 9.15	n = 4	[CR1] Carolina
476.7 ± 16.90	642.7 ± 21.66	261.4 ± 22.58	483.2 ± 21.50	337.0 ± 23.39	n = 43	[JJ1] Ortho Clinical Diagnostics
188.8 ± 3.83	249.4 ± 4.89	99.3 ± 2.28	181.5 ± 3.45	119.2 ± 2.93	n = 17	[RO4] Roche cobas c311/c501/c502/c701
188.0 ± 4.88	248.1 ± 6.42	98.6 ± 2.60	180.4 ± 4.00	118.9 ± 3.10	n = 33	[RO2] Roche Hitachi and Modular D/P
189.9 ± 3.25	248.3 ± 6.11	99.7 ± 2.78	181.3 ± 4.46	120.3 ± 2.24	n = 8	[RO1] Roche Integra and MIRA
188.9 ± 6.06	250.7 ± 8.50	100.6 ± 3.68	181.9 ± 5.93	121.3 ± 3.97	n = 23	[BY1] Siemens ADVIA/ADVIS Centaur
186.8 ± 19.43	246.0 ± 24.12	98.7 ± 9.40	179.5 ± 19.78	119.3 ± 11.75	n = 25	[DA5] Siemens Dimension
196.3 ± 7.53	259.8 ± 8.24	103.4 ± 5.22	189.6 ± 8.29	124.7 ± 6.06	n = 46	[DA8] Siemens Dimension IFCC Standardized

Summary of Participant Performance (Mean and Standard Deviation)

LDH Isoenzyme 1 (%)

Specimen: C61	Specimen: C62	Specimen: C63	Specimen: C64	Specimen: C65	Number	[Code] Instrument or Reagent System
52.4 ± 1.39	21.0 ± 3.67	30.7 ± 4.90	35.1 ± 3.18	31.6 ± 4.25	n = 8	[-P-] All Methods - Results reported as %
53.0 ± 0.75	24.0 ± 0.00	34.3 ± 1.58	37.8 ± 1.46	34.2 ± 0.41	n = 4	<Instruments>
51.7 ± 1.58	17.5 ± 1.23	26.6 ± 3.39	32.5 ± 1.22	27.8 ± 3.59	n = 4	[HLS] Helena SPIFE [SEE] Sebia Electrophoresis
53.0 ± 0.75	24.0 ± 0.00	34.3 ± 1.58	37.8 ± 1.46	34.2 ± 0.41	n = 4	<Reagents>
51.7 ± 1.58	17.5 ± 1.23	26.6 ± 3.39	32.5 ± 1.22	27.8 ± 3.59	n = 4	[HL1] Helena Laboratories [SE1] Sebia