



Department of Health

ANDREW M. CUOMO
Governor

HOWARD A. ZUCKER, M.D., J.D.
Commissioner

SALLY DRESLIN, M.S., R.N.
Executive Deputy Commissioner

Statistical Report: Quantitative Urine Clinical Chemistry Mail out date: October 26, 2015

This report summarizes data from the educational quantitative urine clinical chemistry proficiency test of October 26, 2015. Individual evaluation reports are provided for your laboratory for this testing.

Results for the quantitative urine clinical chemistry are listed as the mean \pm 1SD for each instrument and reagent system as well as overall results. For albumin, creatinine, and albumin/creatinine ratio units shown are those most frequently used by participants; results from laboratories using different units were converted to the units shown. Please keep this in mind when comparing results reported by your laboratory. Individual laboratory reports were evaluated using ranges appropriate for units reported.

The following criteria were used for the educational quantitative urine clinical chemistry: Albumin (\pm 25%); Creatinine (\pm 17%); Albumin-Creatinine Ratio (\pm 20%); alpha-Amylase (not evaluated); Calcium (\pm 15%); Chloride (\pm 20%); Glucose (\pm 15%); Magnesium (\pm 20%); Sodium (\pm 15%); Phosphorus (\pm 20%); Potassium (\pm 15%); Total Protein (\pm 30%); Uric Acid (\pm 20%); Urea Nitrogen (\pm 15%). At low analyte concentrations, ranges were based on the dispersion of results obtained. Overall mean values were used for calculation of target concentrations for all analytes.

The use of brand and/or trade names in this report does not constitute an endorsement of the products on the part of the Wadsworth Center or the New York State Department of Health.

New York State Department of Health – Wadsworth Center
 Quantitative Urine Chemistry – Educational Proficiency Testing – 26 October 2015

Summary of Participant Performance (Mean and Standard Deviation)

Albumin

The following albumin results are summaries with all results converted to mg/L:

Specimen: U36	Specimen: U37	Specimen: U38	Specimen: U39	Specimen: U40	Number	Instrument or Reagent System
2072.2 ± 183.45	347.3 ± 26.85	2584.7 ± 239.15	742.9 ± 57.18	997.5 ± 80.02	n = 176	All Methods & Instruments
2049.6 ± 184.44	344.8 ± 28.82	2580.9 ± 233.77	731.3 ± 56.93	982.2 ± 84.10	n = 82	<Method Principles>
2092.7 ± 173.33	350.3 ± 24.81	2595.1 ± 233.11	752.2 ± 55.15	1007.7 ± 73.24	n = 76	Reporting in mg/dL
2023.4 ± 318.00	343.4 ± 25.79	2459.8 ± 485.53	756.3 ± 55.35	1018.1 ± 83.36	n = 18	Reporting in mg/L
						Reporting in ug/mL
2020.9 ± 168.93	358.7 ± 21.85	2363.9 ± 500.96	727.8 ± 18.43	998.3 ± 36.27	n = 9	<Instruments>
2072.9 ± 143.23	339.2 ± 25.31	2578.8 ± 186.43	715.1 ± 40.56	969.1 ± 67.70	n = 37	Abbott Architect c System
1920.9 ± 83.14	345.1 ± 12.03	2378.7 ± 173.40	746.6 ± 18.18	1023.6 ± 29.90	n = 3	Beckman Coulter AU Chemistry System
2113.3 ± 270.06	358.0 ± 4.90	2650.5 ± 222.61	785.2 ± 8.40	1032.9 ± 50.26	n = 5	Beckman Coulter Immage
1945.8 ± 458.39	342.8 ± 2.11	2325.1 ± 559.72	756.7 ± 29.40	979.1 ± 85.03	n = 5	Beckman Coulter UniCel DxC 600
2355.0 ± 129.06	374.6 ± 14.33	2903.0 ± 150.47	849.4 ± 59.42	1202.9 ± 50.11	n = 4	Beckman Coulter UniCel DxC 800
1726.7 ± 424.06	338.0 ± 38.58	1879.7 ± 650.60	741.9 ± 39.19	1029.5 ± 44.47	n = 3	Nephelometer
1474.9 ± 555.93	372.3 ± 12.39	1610.0 ± 797.90	765.3 ± 64.15	1072.3 ± 44.62	n = 13	Ortho Vitros 5,1FS
2092.0 ± 137.09	347.6 ± 10.48	2663.5 ± 165.31	731.1 ± 43.01	980.2 ± 69.66	n = 18	Ortho Vitros 5600
1928.0 ± 39.12	329.1 ± 15.16	2392.8 ± 90.83	716.8 ± 22.71	945.2 ± 21.50	n = 5	Roche cobas c501
2006.4 ± 65.76	347.6 ± 10.56	2526.3 ± 66.29	757.7 ± 25.76	991.1 ± 45.08	n = 12	Roche c701
1914.6 ± 80.95	306.4 ± 22.36	2383.9 ± 65.06	661.0 ± 22.32	919.5 ± 37.90	n = 12	Roche MODULAR D/P
1927.6 ± 62.03	325.7 ± 8.42	2424.6 ± 50.31	684.6 ± 27.53	940.9 ± 40.49	n = 3	Siemens ADVIA 1800
2160.5 ± 152.30	360.2 ± 13.64	2707.7 ± 210.18	774.9 ± 23.43	992.2 ± 127.35	n = 6	Siemens ADVIA 2400
1834.2 ± 247.77	363.1 ± 25.29	2342.8 ± 670.81	781.6 ± 23.72	1001.7 ± 63.82	n = 3	Siemens Dimension EXL
2209.9 ± 111.70	359.8 ± 31.60	2719.3 ± 150.30	795.8 ± 46.04	1054.0 ± 59.39	n = 27	Siemens Dimension RxL
						Siemens Dimension Vista
2057.4 ± 134.75	361.5 ± 16.76	2354.8 ± 544.31	724.2 ± 17.04	1003.6 ± 27.47	n = 8	<Reagents>
2042.1 ± 242.42	345.5 ± 15.80	2523.1 ± 236.79	760.5 ± 31.07	1020.4 ± 49.75	n = 13	Abbott
2108.0 ± 121.45	344.3 ± 23.06	2623.4 ± 168.88	725.7 ± 36.18	985.5 ± 57.79	n = 30	Beckman Coulter AU Series
1922.3 ± 86.03	313.6 ± 9.16	2408.7 ± 107.63	674.2 ± 12.90	927.3 ± 21.38	n = 5	Beckman Coulter Immage
1524.1 ± 554.33	369.4 ± 12.60	1669.9 ± 804.52	764.9 ± 18.36	1063.7 ± 30.55	n = 12	Kamiya
1523.9 ± 524.11	298.2 ± 52.66	1653.5 ± 721.28	682.1 ± 99.15	943.3 ± 120.02	n = 4	Ortho Clinical Diagnostics
2043.1 ± 137.45	345.3 ± 12.56	2602.7 ± 188.62	726.0 ± 36.85	969.7 ± 57.67	n = 25	Randox
2006.4 ± 65.76	347.6 ± 10.56	2526.3 ± 66.29	757.7 ± 25.76	991.1 ± 45.08	n = 12	Roche cobas c501/c311/c502/c701
2165.1 ± 189.08	359.6 ± 6.37	2436.3 ± 111.58	774.3 ± 5.61	1033.6 ± 14.44	n = 3	Roche Hitachi and Modular D/P
1925.4 ± 77.75	313.0 ± 19.91	2401.5 ± 71.55	667.8 ± 27.36	927.1 ± 41.08	n = 16	Roche Integra and MIRA S
2191.3 ± 131.68	361.9 ± 28.05	2726.0 ± 170.36	792.6 ± 39.95	1048.8 ± 72.30	n = 36	Siemens ADVIA/ADVIA Centaur
						Siemens Dimension

New York State Department of Health – Wadsworth Center
 Quantitative Urine Chemistry – Educational Proficiency Testing – 26 October 2015

Summary of Participant Performance (Mean and Standard Deviation)

Creatinine

The following creatinine results are summaries with all results converted to mg/dL:

Specimen: U36	Specimen: U37	Specimen: U38	Specimen: U39	Specimen: U40	Number	Instrument or Reagent System
140.9 ± 9.46	71.0 ± 5.26	88.3 ± 6.58	46.4 ± 3.82	94.9 ± 5.40	n = 256	All Methods & Instruments
141.0 ± 9.51	71.0 ± 5.29	88.3 ± 6.59	46.4 ± 3.85	94.9 ± 5.43	n = 252	<Method Principles> Reporting in mg/dL
134.7 ± 2.93	67.8 ± 1.90	83.7 ± 1.97	44.3 ± 1.20	91.2 ± 2.48	n = 17	<Instruments>
154.3 ± 4.52	80.1 ± 2.29	99.0 ± 3.41	53.0 ± 1.78	102.4 ± 3.34	n = 52	Abbott Architect c System
148.6 ± 3.71	74.5 ± 3.39	92.9 ± 1.74	50.2 ± 0.84	101.3 ± 2.66	n = 9	Beckman Coulter AU Chemistry System
143.4 ± 5.48	73.1 ± 1.47	90.8 ± 2.40	47.9 ± 0.76	96.8 ± 1.82	n = 9	Beckman Coulter UniCel DxC 600
134.4 ± 3.72	67.5 ± 2.05	84.4 ± 3.22	44.1 ± 1.36	94.0 ± 2.84	n = 7	Beckman Coulter UniCel DxC 800
131.5 ± 2.96	66.2 ± 1.84	81.1 ± 1.71	42.7 ± 0.75	92.7 ± 2.45	n = 18	Ortho Vitros 5,1FS
140.7 ± 4.76	72.0 ± 2.05	88.6 ± 2.66	47.0 ± 1.65	96.2 ± 3.01	n = 29	Ortho Vitros 5600
131.8 ± 8.94	66.5 ± 4.18	82.3 ± 5.28	43.7 ± 2.40	88.9 ± 6.72	n = 7	Roche cobas c501
137.9 ± 4.15	69.4 ± 2.39	84.9 ± 3.36	44.8 ± 1.39	92.5 ± 1.46	n = 5	Roche cobas c701
135.3 ± 4.37	69.1 ± 2.36	85.4 ± 2.53	45.0 ± 1.56	92.6 ± 2.78	n = 10	Roche Cobas INTEGRA
133.4 ± 4.18	67.7 ± 2.50	83.9 ± 2.86	44.4 ± 1.48	90.4 ± 3.37	n = 18	Siemens ADVIA 1800
128.8 ± 3.06	66.5 ± 1.78	81.2 ± 4.62	42.7 ± 1.81	90.6 ± 3.02	n = 3	Siemens ADVIA 2400
141.5 ± 3.27	69.6 ± 2.41	87.8 ± 2.76	45.9 ± 2.78	95.1 ± 4.51	n = 10	Siemens Dimension EXL
142.2 ± 3.91	71.5 ± 2.12	88.3 ± 2.97	46.2 ± 1.18	96.3 ± 3.02	n = 3	Siemens Dimension RxL
144.4 ± 3.84	70.8 ± 2.25	89.3 ± 2.49	45.4 ± 1.78	93.5 ± 3.10	n = 43	Siemens Dimension Vista
134.7 ± 2.93	67.8 ± 1.90	83.7 ± 1.97	44.3 ± 1.20	91.2 ± 2.48	n = 17	<Reagents>
146.3 ± 5.30	73.9 ± 2.87	92.1 ± 2.38	48.9 ± 1.85	99.1 ± 3.40	n = 18	Abbott
154.7 ± 4.06	80.3 ± 1.98	99.4 ± 2.95	53.2 ± 1.57	102.7 ± 2.90	n = 47	Beckman Coulter
131.9 ± 3.50	66.4 ± 1.98	81.6 ± 2.50	42.9 ± 1.15	92.9 ± 2.49	n = 26	Beckman Coulter AU Series
139.4 ± 6.26	71.3 ± 2.93	87.9 ± 3.67	46.5 ± 2.19	95.4 ± 4.06	n = 36	Ortho Clinical Diagnostics
135.3 ± 4.37	69.1 ± 2.36	85.4 ± 2.53	45.0 ± 1.56	92.6 ± 2.78	n = 18	Roche cobas c501/c311/c502/c701
136.8 ± 4.38	69.1 ± 2.20	84.6 ± 3.03	44.7 ± 1.24	92.1 ± 1.68	n = 6	Roche Hitachi and Modular D/P
132.8 ± 4.20	67.5 ± 2.30	83.9 ± 3.23	44.3 ± 1.55	90.6 ± 3.24	n = 21	Roche Integra and MIRA S
143.7 ± 4.15	70.6 ± 2.38	88.9 ± 2.63	45.5 ± 1.90	93.7 ± 3.37	n = 58	Siemens ADVIA/ADVIS Centaur
						Siemens Dimension

New York State Department of Health – Wadsworth Center
 Quantitative Urine Chemistry – Educational Proficiency Testing – 26 October 2015

Summary of Participant Performance (Mean and Standard Deviation)

Albumin-Creatinine Ratio

The following albumin-creatinine ratio results are summaries with all results converted to $\mu\text{g}/\text{mg}$:

Specimen: U36	Specimen: U37	Specimen: U38	Specimen: U39	Specimen: U40	Number	Instrument or Reagent System
1467.3 \pm 146.49	481.1 \pm 55.31	2886.1 \pm 331.10	1594.6 \pm 183.78	1048.6 \pm 102.63	n = 152	All Methods & Instruments
1479.5 \pm 143.33	484.6 \pm 51.19	2889.4 \pm 302.63	1597.6 \pm 168.94	1051.5 \pm 103.68	n = 99	<Method Principles>
1442.9 \pm 153.93	472.5 \pm 61.78	2845.9 \pm 443.41	1583.6 \pm 208.43	1041.0 \pm 100.57	n = 50	Reporting in mg/g
1435.0 \pm 131.90	442.9 \pm 54.47	2847.2 \pm 312.91	1473.2 \pm 154.90	1016.0 \pm 80.95	n = 12	<Instruments>
1541.5 \pm 121.76	470.7 \pm 32.61	2972.5 \pm 237.00	1515.0 \pm 149.68	1013.3 \pm 76.03	n = 4	Clinical Analyzer calculation
1469.6 \pm 146.25	485.6 \pm 54.74	2890.5 \pm 333.97	1609.7 \pm 179.87	1054.1 \pm 103.88	n = 135	Manual Calculation
1435.0 \pm 131.90	442.9 \pm 54.47	2847.2 \pm 312.91	1473.2 \pm 154.90	1016.0 \pm 80.95	n = 12	<Reagents>
1541.5 \pm 121.76	470.7 \pm 32.61	2972.5 \pm 237.00	1515.0 \pm 149.68	1013.3 \pm 76.03	n = 4	Lab Information System
						Clinical Analyzer calculation
						Manual Calculation

New York State Department of Health – Wadsworth Center
Quantitative Urine Chemistry – Educational Proficiency Testing – 26 October 2015

Summary of Participant Performance (Mean and Standard Deviation)

Calcium (mg/dL):

Specimen: U36	Specimen: U37	Specimen: U38	Specimen: U39	Specimen: U40	Number	Instrument or Reagent System
4.91 ± 0.50	9.49 ± 0.42	3.23 ± 0.89	7.66 ± 0.36	10.31 ± 0.41	n = 160	All Methods & Instruments
5.01 ± 0.32	9.59 ± 0.38	4.12 ± 1.23	7.72 ± 0.32	10.38 ± 0.33	n = 62	<Method Principles>
4.89 ± 0.58	9.49 ± 0.42	3.06 ± 0.22	7.63 ± 0.41	10.39 ± 0.49	n = 67	o-Cresolphthalein Arsenazo dye
4.48 ± 0.51	8.96 ± 0.27	2.46 ± 0.45	7.44 ± 0.23	9.88 ± 0.15	n = 12	Ion selective electrode
4.81 ± 0.62	9.48 ± 0.28	2.81 ± 0.10	7.71 ± 0.26	10.18 ± 0.17	n = 15	5-nitro-5-methyl-BAPTA
<Instruments>						
4.80 ± 0.40	9.07 ± 0.16	2.90 ± 0.10	7.35 ± 0.29	9.73 ± 0.31	n = 10	Abbott Architect c System
4.89 ± 0.61	9.58 ± 0.41	3.14 ± 0.15	7.67 ± 0.38	10.42 ± 0.32	n = 26	Beckman Coulter AU Chemistry System
4.02 ± 0.51	8.81 ± 0.11	2.28 ± 0.13	7.27 ± 0.16	9.77 ± 0.16	n = 5	Beckman Coulter UniCel DxC 600
4.77 ± 0.21	8.95 ± 0.50	2.53 ± 0.52	7.58 ± 0.22	9.90 ± 0.00	n = 6	Beckman Coulter UniCel DxC 800
5.25 ± 0.48	9.83 ± 0.34	3.11 ± 0.28	8.00 ± 0.13	10.86 ± 0.26	n = 6	Ortho Vitros 5,1FS
4.75 ± 0.75	9.49 ± 0.42	3.14 ± 0.28	7.48 ± 0.44	10.72 ± 0.28	n = 14	Ortho Vitros 5600
4.72 ± 0.57	9.64 ± 0.33	2.82 ± 0.10	7.82 ± 0.26	10.29 ± 0.15	n = 14	Roche cobas c501
4.74 ± 0.45	9.38 ± 0.21	2.80 ± 0.15	7.67 ± 0.20	10.08 ± 0.08	n = 5	Roche cobas c701
4.72 ± 0.38	9.68 ± 0.29	3.03 ± 0.19	7.74 ± 0.27	10.61 ± 0.32	n = 11	Roche MODULAR D/P
4.87 ± 0.36	9.58 ± 0.31	2.97 ± 0.17	7.74 ± 0.35	10.26 ± 0.40	n = 13	Siemens ADVIA 1800
4.83 ± 0.68	9.48 ± 0.30	2.94 ± 0.05	7.80 ± 0.15	10.06 ± 0.26	n = 3	Siemens ADVIA 2400
5.16 ± 0.15	9.87 ± 0.18	<5	7.97 ± 0.14	10.76 ± 0.20	n = 5	Siemens Dimension EXL
5.12 ± 0.18	9.47 ± 0.39	<5	7.65 ± 0.34	10.26 ± 0.29	n = 34	Siemens Dimension Vista
<Reagents>						
4.80 ± 0.40	9.07 ± 0.16	2.90 ± 0.10	7.35 ± 0.29	9.73 ± 0.31	n = 10	Abbott
4.48 ± 0.51	8.96 ± 0.27	2.46 ± 0.45	7.44 ± 0.23	9.88 ± 0.15	n = 12	Beckman Coulter
4.87 ± 0.62	9.55 ± 0.38	3.14 ± 0.16	7.65 ± 0.35	10.39 ± 0.28	n = 25	Beckman Coulter AU Series
4.92 ± 0.73	9.58 ± 0.44	3.13 ± 0.28	7.61 ± 0.46	10.76 ± 0.29	n = 20	Ortho Clinical Diagnostics
4.74 ± 0.52	9.55 ± 0.32	2.81 ± 0.13	7.77 ± 0.25	10.22 ± 0.18	n = 20	Roche cobas c501/c311/c502/c701
4.70 ± 0.40	9.71 ± 0.28	3.06 ± 0.17	7.78 ± 0.24	10.66 ± 0.27	n = 10	Roche Hitachi and Modular D/P
4.84 ± 0.51	9.57 ± 0.34	2.97 ± 0.16	7.76 ± 0.38	10.25 ± 0.38	n = 18	Siemens ADVIA/ADVIS Centaur
5.13 ± 0.18	9.52 ± 0.40	<5	7.71 ± 0.34	10.33 ± 0.34	n = 39	Siemens Dimension

New York State Department of Health – Wadsworth Center
Quantitative Urine Chemistry – Educational Proficiency Testing – 26 October 2015

Summary of Participant Performance (Mean and Standard Deviation)

Chloride (mmol/L):

Specimen: U36	Specimen: U37	Specimen: U38	Specimen: U39	Specimen: U40	Number	Instrument or Reagent System
45.18 ± 3.55	141.14 ± 5.25	223.55 ± 4.81	107.89 ± 5.86	165.44 ± 5.26	n = 182	All Methods & Instruments
45.07 ± 3.56	141.35 ± 5.36	223.59 ± 4.78	108.03 ± 6.01	165.56 ± 5.35	n = 166	<Method Principles>
46.78 ± 2.59	139.47 ± 2.72	223.64 ± 5.19	107.28 ± 3.25	164.80 ± 3.63	n = 14	Ion selective electrode (diluted) Ion selective electrode (undiluted)
46.06 ± 0.55	139.42 ± 1.83	222.93 ± 2.54	107.08 ± 1.99	163.96 ± 1.98	n = 15	<Instruments>
46.55 ± 0.87	141.68 ± 2.09	225.53 ± 3.88	108.22 ± 1.35	165.72 ± 2.24	n = 31	Abbott Architect c System
44.48 ± 1.86	140.77 ± 0.79	218.49 ± 3.71	108.88 ± 0.71	163.12 ± 2.17	n = 8	Beckman Coulter AU Chemistry System
45.54 ± 1.37	140.68 ± 1.22	217.76 ± 3.67	108.85 ± 0.85	164.18 ± 1.89	n = 8	Beckman Coulter UniCel DxC 600
50.00 ± 0.00	137.44 ± 1.02	179.92 ± 38.54	106.44 ± 1.02	153.78 ± 8.59	n = 3	Beckman Coulter UniCel DxC 800
38.30 ± 1.52	134.08 ± 1.89	226.84 ± 2.89	98.31 ± 1.60	160.69 ± 1.99	n = 20	Ortho Vitros 5600
39.18 ± 0.91	128.24 ± 1.93	214.23 ± 2.31	94.26 ± 1.83	152.19 ± 0.73	n = 6	Roche cobas c501
39.80 ± 1.70	135.12 ± 1.58	227.36 ± 5.32	99.40 ± 1.24	161.10 ± 1.32	n = 14	Roche cobas c701
45.67 ± 0.74	141.18 ± 0.98	218.74 ± 1.92	107.90 ± 1.22	164.07 ± 1.38	n = 14	Roche MODULAR D/P
46.00 ± 0.90	142.74 ± 1.37	221.00 ± 3.61	109.00 ± 1.80	165.00 ± 1.80	n = 3	Siemens ADVIA 1800
55.00 ± 3.92	147.33 ± 2.23	224.17 ± 1.72	116.08 ± 2.80	170.93 ± 1.65	n = 9	Siemens ADVIA 2400
45.95 ± 1.20	146.85 ± 1.97	224.78 ± 2.76	113.04 ± 1.79	171.83 ± 2.13	n = 41	Siemens Dimension EXL
46.06 ± 0.55	139.42 ± 1.83	222.93 ± 2.54	107.08 ± 1.99	163.96 ± 1.98	n = 15	Siemens Dimension Vista
45.15 ± 1.73	140.89 ± 1.02	218.71 ± 3.94	108.73 ± 0.83	163.98 ± 1.92	n = 18	<Reagents>
46.53 ± 0.92	141.72 ± 2.16	225.53 ± 4.11	108.22 ± 1.43	165.75 ± 2.36	n = 29	Beckman Coulter
49.39 ± 1.64	137.87 ± 1.13	191.16 ± 37.12	106.70 ± 0.82	156.43 ± 7.54	n = 4	Beckman Coulter AU Series
38.50 ± 1.44	132.95 ± 3.13	224.31 ± 6.03	97.74 ± 2.56	159.06 ± 4.20	n = 26	Ortho Clinical Diagnostics
39.80 ± 1.70	135.12 ± 1.58	227.36 ± 5.32	99.40 ± 1.24	161.10 ± 1.32	n = 14	Roche cobas c501/c311/c502/c701
45.70 ± 0.76	141.14 ± 1.52	218.70 ± 2.55	107.87 ± 1.40	163.97 ± 1.72	n = 19	Roche Hitachi and Modular D/P
46.46 ± 2.20	146.88 ± 2.01	224.56 ± 2.68	113.49 ± 2.12	171.60 ± 2.01	n = 54	Siemens ADVIA Centaur
						Siemens Dimension

New York State Department of Health – Wadsworth Center
 Quantitative Urine Chemistry – Educational Proficiency Testing – 26 October 2015

Summary of Participant Performance (Mean and Standard Deviation)

Glucose (mg/dL):

Specimen: U36	Specimen: U37	Specimen: U38	Specimen: U39	Specimen: U40	Number	Instrument or Reagent System
461.6 ± 15.80	3.8 ± 6.05	295.5 ± 8.19	3.8 ± 6.04	206.5 ± 5.55	n = 117	All Methods & Instruments
459.8 ± 14.99	1.3 ± 0.96	295.1 ± 8.10	1.3 ± 0.95	206.1 ± 5.04	n = 69	<Method Principles>
456.2 ± 15.44	1.0 ± 0.50	291.9 ± 8.82	1.0 ± 0.62	204.1 ± 8.30	n = 24	Hexokinase, UV
478.9 ± 8.68	<20	300.7 ± 4.72	<20	209.8 ± 3.04	n = 16	Hexokinase, colorimetric
456.4 ± 6.55	7.8 ± 3.90	295.0 ± 6.86	7.5 ± 4.33	206.4 ± 3.26	n = 8	Glucose oxidase, colorimetric
						Glucose oxidase, O ₂ electrode
478.3 ± 9.94	<1	301.2 ± 4.78	<1	208.6 ± 4.51	n = 6	<Instruments>
452.8 ± 9.20	6.8 ± 5.20	289.5 ± 5.22	6.8 ± 5.17	202.4 ± 3.66	n = 16	Abbott Architect c System
453.9 ± 5.18	5.0 ± 3.76	289.4 ± 6.41	5.0 ± 3.76	207.1 ± 1.28	n = 4	Beckman Coulter AU Chemistry System
458.3 ± 6.01	8.9 ± 2.86	297.6 ± 8.34	8.6 ± 3.68	205.7 ± 4.89	n = 4	Beckman Coulter UniCel DxC 600
475.7 ± 9.31	<20	299.9 ± 2.16	<20	208.3 ± 1.99	n = 5	Beckman Coulter UniCel DxC 800
478.6 ± 7.67	<20	300.4 ± 4.77	<20	210.2 ± 3.77	n = 11	Ortho Vitros 5,1FS
466.3 ± 11.88	<2	299.4 ± 6.64	<2	209.6 ± 4.24	n = 11	Ortho Vitros 5600
462.7 ± 4.41	<2	298.2 ± 1.27	<2	206.7 ± 0.90	n = 4	Roche cobas c501
471.8 ± 6.97	1.6 ± 0.64	301.6 ± 4.37	1.8 ± 0.43	210.7 ± 4.09	n = 11	Roche cobas c701
460.1 ± 7.83	2.2 ± 1.72	294.9 ± 4.35	2.2 ± 1.72	206.5 ± 2.66	n = 10	Siemens ADVIA 1800
460.0 ± 6.42	1.5 ± 1.86	294.6 ± 4.72	1.5 ± 1.86	206.1 ± 3.72	n = 3	Siemens ADVIA 2400
465.4 ± 7.25	0.2 ± 0.41	298.3 ± 4.29	0.2 ± 0.41	207.2 ± 3.73	n = 4	Siemens Dimension EXL
444.0 ± 9.40	<1	287.3 ± 6.48	<1	200.1 ± 5.68	n = 24	Siemens Dimension Vista
478.3 ± 9.94	<1	301.2 ± 4.78	<1	208.6 ± 4.51	n = 6	<Reagents>
457.2 ± 6.51	7.4 ± 3.83	293.7 ± 7.72	7.2 ± 4.18	206.8 ± 3.16	n = 9	Abbott
452.8 ± 9.14	7.3 ± 4.91	289.5 ± 5.09	7.3 ± 4.89	202.4 ± 3.64	n = 15	Beckman Coulter
478.4 ± 8.39	<20	300.7 ± 4.52	<20	209.6 ± 3.22	n = 17	Beckman Coulter AU Series
465.7 ± 9.61	<2	299.2 ± 5.04	<2	208.6 ± 3.67	n = 17	Ortho Clinical Diagnostics
471.8 ± 6.97	1.6 ± 0.64	301.6 ± 4.37	1.8 ± 0.43	210.7 ± 4.09	n = 11	Roche Hitachi and Modular D/P
460.3 ± 7.19	2.2 ± 1.81	294.8 ± 4.18	2.2 ± 1.81	206.4 ± 2.80	n = 14	Siemens ADVIA/ADVISIA Centaur
446.5 ± 11.73	<1	288.7 ± 7.33	<1	201.3 ± 6.17	n = 28	Siemens Dimension

New York State Department of Health – Wadsworth Center
 Quantitative Urine Chemistry – Educational Proficiency Testing – 26 October 2015

Summary of Participant Performance (Mean and Standard Deviation)

Magnesium (mg/dL):

Specimen: U36	Specimen: U37	Specimen: U38	Specimen: U39	Specimen: U40	Number	Instrument or Reagent System
2.29 ± 0.27	7.32 ± 0.44	10.17 ± 0.52	4.89 ± 0.28	13.09 ± 0.62	n = 110	All Methods & Instruments
2.25 ± 0.38	7.03 ± 0.40	9.86 ± 0.71	4.65 ± 0.31	12.74 ± 0.61	n = 12	<Method Principles>
2.34 ± 0.09	7.43 ± 0.20	10.30 ± 0.28	4.94 ± 0.12	13.22 ± 0.30	n = 29	Calmagite
2.32 ± 0.19	7.44 ± 0.36	10.13 ± 0.42	4.94 ± 0.28	13.25 ± 0.54	n = 45	Methylthymol blue
1.20 ± 0.00	6.54 ± 0.16	10.46 ± 0.72	3.14 ± 0.89	13.02 ± 0.17	n = 9	Xylydyl blue (Magon)
2.35 ± 0.33	7.44 ± 0.44	10.05 ± 0.81	4.83 ± 0.60	13.07 ± 1.12	n = 5	Formazon dye
1.89 ± 0.20	5.88 ± 0.80	8.39 ± 0.73	3.53 ± 0.52	11.61 ± 0.63	n = 6	Chlorophosphonazo III
2.23 ± 0.16	7.03 ± 0.58	10.00 ± 0.43	4.61 ± 0.27	12.31 ± 0.92	n = 4	Arsenazo dye
						Other
1.84 ± 0.11	5.68 ± 0.58	8.19 ± 0.52	3.44 ± 0.30	11.47 ± 0.62	n = 5	<Instruments>
2.28 ± 0.10	7.24 ± 0.11	9.88 ± 0.26	4.80 ± 0.15	13.24 ± 0.50	n = 14	Abbott Architect c System
2.43 ± 0.20	7.25 ± 0.24	9.95 ± 0.63	4.65 ± 0.19	13.10 ± 0.46	n = 4	Beckman Coulter AU Chemistry System
2.29 ± 0.39	6.95 ± 0.39	9.70 ± 0.42	4.73 ± 0.27	12.50 ± 0.45	n = 5	Beckman Coulter UniCel DxC 600
1.28 ± 0.15	5.87 ± 1.34	10.07 ± 0.97	2.89 ± 1.20	12.40 ± 0.92	n = 3	Beckman Coulter UniCel DxC 800
1.29 ± 0.17	6.55 ± 0.18	10.72 ± 0.26	3.40 ± 0.33	13.05 ± 0.18	n = 7	Ortho Vitros 5,1FS
2.45 ± 0.22	7.89 ± 0.10	10.46 ± 0.30	5.19 ± 0.13	13.91 ± 0.30	n = 9	Ortho Vitros 5600
2.76 ± 0.29	7.41 ± 0.23	10.18 ± 0.20	5.08 ± 0.21	13.01 ± 0.11	n = 4	Roche cobas c501
2.43 ± 0.48	7.42 ± 0.18	10.01 ± 0.38	4.80 ± 0.15	12.66 ± 0.25	n = 9	Roche cobas c701
2.34 ± 0.15	7.39 ± 0.49	10.42 ± 1.10	4.96 ± 0.33	13.16 ± 0.37	n = 9	Roche MODULAR D/P
2.62 ± 0.32	7.56 ± 0.29	10.24 ± 0.24	5.12 ± 0.30	13.28 ± 0.43	n = 3	Siemens ADVIA 1800
2.34 ± 0.09	7.42 ± 0.20	10.30 ± 0.31	4.94 ± 0.13	13.22 ± 0.30	n = 28	Siemens ADVIA 2400
						Siemens Dimension Vista
1.84 ± 0.11	5.68 ± 0.58	8.19 ± 0.52	3.44 ± 0.30	11.47 ± 0.62	n = 5	<Reagents>
2.31 ± 0.35	7.03 ± 0.40	9.68 ± 0.62	4.65 ± 0.27	12.64 ± 0.64	n = 10	Abbott
2.27 ± 0.09	7.24 ± 0.12	9.91 ± 0.23	4.79 ± 0.14	13.31 ± 0.43	n = 13	Beckman Coulter
2.17 ± 0.14	6.78 ± 0.32	9.83 ± 0.14	4.50 ± 0.09	11.97 ± 0.05	n = 3	Beckman Coulter AU Series
1.29 ± 0.16	6.57 ± 0.16	10.68 ± 0.22	3.45 ± 0.28	13.00 ± 0.17	n = 10	In-House
2.52 ± 0.27	7.82 ± 0.22	10.41 ± 0.30	5.18 ± 0.16	13.66 ± 0.47	n = 15	Ortho Clinical Diagnostics
2.43 ± 0.48	7.42 ± 0.18	10.01 ± 0.38	4.80 ± 0.15	12.66 ± 0.25	n = 9	Roche Hitachi and Modular D/P
2.35 ± 0.15	7.42 ± 0.42	10.30 ± 0.58	4.98 ± 0.32	13.16 ± 0.39	n = 13	Siemens ADVIA/ADVIa Centaur
2.33 ± 0.08	7.43 ± 0.20	10.31 ± 0.30	4.93 ± 0.12	13.21 ± 0.29	n = 29	Siemens Dimension

New York State Department of Health – Wadsworth Center
 Quantitative Urine Chemistry – Educational Proficiency Testing – 26 October 2015

Summary of Participant Performance (Mean and Standard Deviation)

Sodium (mmol/L):

Specimen: U36	Specimen: U37	Specimen: U38	Specimen: U39	Specimen: U40	Number	Instrument or Reagent System
52.8 ± 2.18	138.6 ± 6.83	232.2 ± 5.81	106.5 ± 5.09	172.3 ± 6.79	n = 211	All Methods & Instruments
52.7 ± 2.09	137.8 ± 6.34	231.8 ± 5.06	106.0 ± 5.26	171.5 ± 5.79	n = 183	<Method Principles>
54.4 ± 2.84	145.6 ± 6.85	242.8 ± 15.65	109.2 ± 3.34	182.6 ± 10.33	n = 27	Ion selective electrode (diluted) Ion selective electrode (undiluted)
52.0 ± 0.91	136.4 ± 1.71	227.4 ± 3.02	106.4 ± 0.98	170.8 ± 2.02	n = 16	<Instruments>
53.4 ± 0.84	140.1 ± 1.49	234.8 ± 3.66	107.8 ± 1.07	173.9 ± 2.40	n = 33	Abbott Architect c System
53.9 ± 1.88	139.1 ± 2.05	227.3 ± 3.90	108.1 ± 2.33	171.0 ± 2.48	n = 9	Beckman Coulter AU Chemistry System
55.6 ± 1.13	139.6 ± 1.13	223.7 ± 4.34	109.2 ± 0.76	170.6 ± 0.79	n = 7	Beckman Coulter UniCel DxC 600
56.7 ± 2.14	151.4 ± 3.19	261.7 ± 8.68	112.3 ± 1.99	190.1 ± 4.16	n = 7	Beckman Coulter UniCel DxC 800
54.6 ± 2.62	150.1 ± 3.47	248.2 ± 4.30	111.0 ± 4.13	189.9 ± 4.61	n = 15	Ortho Vitros 5,1FS
53.1 ± 1.89	140.8 ± 1.77	233.8 ± 2.84	108.1 ± 1.34	175.2 ± 1.48	n = 22	Ortho Vitros 5600
53.2 ± 1.72	143.2 ± 1.59	234.5 ± 1.02	109.5 ± 2.74	175.2 ± 2.53	n = 6	Roche cobas c501
54.1 ± 1.01	141.3 ± 1.31	234.4 ± 2.44	108.3 ± 1.43	175.1 ± 1.54	n = 16	Roche cobas c701
52.8 ± 0.82	141.1 ± 1.14	235.4 ± 2.92	108.1 ± 1.18	175.3 ± 1.63	n = 14	Siemens ADVIA 1800
54.0 ± 0.90	141.6 ± 1.02	233.7 ± 6.76	110.2 ± 1.54	175.3 ± 2.26	n = 3	Siemens ADVIA 2400
51.9 ± 0.79	134.1 ± 1.22	228.3 ± 2.23	102.9 ± 0.92	168.1 ± 1.10	n = 9	Siemens Dimension EXL
50.2 ± 1.36	126.8 ± 2.54	229.6 ± 2.30	95.5 ± 2.21	163.1 ± 2.37	n = 42	Siemens Dimension Vista
52.0 ± 0.91	136.4 ± 1.71	227.4 ± 3.02	106.4 ± 0.98	170.8 ± 2.02	n = 16	<Reagents>
54.8 ± 1.70	139.4 ± 1.75	226.0 ± 4.80	108.5 ± 1.90	171.0 ± 2.18	n = 19	Abbott
53.4 ± 0.79	140.1 ± 1.55	234.8 ± 3.82	107.8 ± 1.06	173.9 ± 2.37	n = 31	Beckman Coulter
55.2 ± 2.91	150.5 ± 3.80	252.9 ± 10.19	111.3 ± 3.61	189.9 ± 4.80	n = 24	Beckman Coulter AU Series
53.1 ± 1.85	141.2 ± 1.96	234.0 ± 2.46	108.3 ± 1.71	175.1 ± 1.79	n = 28	Ortho Clinical Diagnostics
54.1 ± 1.01	141.3 ± 1.31	234.4 ± 2.44	108.3 ± 1.43	175.1 ± 1.54	n = 16	Roche cobas c501/c311/c502/c701
52.9 ± 0.97	141.0 ± 1.48	235.3 ± 3.54	108.1 ± 1.63	175.3 ± 1.65	n = 19	Roche Hitachi and Modular D/P
50.7 ± 1.65	128.6 ± 4.25	229.2 ± 2.44	97.4 ± 4.18	164.3 ± 3.19	n = 55	Siemens ADVIA/ADVIS Centaur
						Siemens Dimension

New York State Department of Health – Wadsworth Center
Quantitative Urine Chemistry – Educational Proficiency Testing – 26 October 2015
Summary of Participant Performance (Mean and Standard Deviation)

Phosphorus (mg/dL):

Specimen: U36	Specimen: U37	Specimen: U38	Specimen: U39	Specimen: U40	Number	Instrument or Reagent System
71.52 ± 3.17	129.78 ± 6.56	43.18 ± 2.42	140.68 ± 7.34	101.99 ± 6.06	n = 138	All Methods & Instruments
71.11 ± 2.31	129.71 ± 4.65	42.95 ± 1.64	140.51 ± 4.78	101.88 ± 3.96	n = 77	<Method Principles>
74.77 ± 0.59	133.35 ± 3.15	45.81 ± 1.68	144.00 ± 2.41	106.01 ± 3.23	n = 4	Phosphomolybdate – no reduction
72.50 ± 4.80	129.62 ± 9.71	43.97 ± 3.90	141.50 ± 12.15	102.29 ± 9.36	n = 42	Phosphomolybdate reduction (ANS)
72.68 ± 2.76	125.59 ± 21.22	43.05 ± 2.20	133.54 ± 22.90	100.59 ± 8.35	n = 4	Phosphomolybdate reduction (PMA phe)
71.84 ± 2.18	131.50 ± 7.27	44.20 ± 0.67	140.99 ± 7.92	103.64 ± 4.34	n = 8	Phosphomolybdate reduct.-ascorbic/malon
						Other
70.21 ± 1.44	125.57 ± 2.19	41.43 ± 0.64	136.16 ± 2.30	99.71 ± 1.56	n = 7	<Instruments>
71.23 ± 2.03	130.07 ± 2.13	43.25 ± 1.35	141.20 ± 2.35	102.35 ± 1.01	n = 23	Abbott Architect c System
74.41 ± 0.92	138.29 ± 4.39	44.65 ± 1.04	147.50 ± 5.01	107.44 ± 2.17	n = 5	Beckman Coulter AU Chemistry System
71.13 ± 3.07	128.50 ± 7.12	42.91 ± 1.50	141.63 ± 5.57	101.89 ± 4.63	n = 6	Beckman Coulter UniCel DxC 600
76.75 ± 0.39	137.96 ± 4.46	48.34 ± 1.01	155.29 ± 4.62	110.88 ± 2.22	n = 4	Beckman Coulter UniCel DxC 800
76.19 ± 2.41	137.43 ± 5.07	47.31 ± 1.36	148.70 ± 5.91	111.58 ± 3.56	n = 13	Ortho Vitros 5,1FS
71.87 ± 0.75	130.47 ± 1.45	43.09 ± 1.05	140.23 ± 2.58	101.84 ± 2.96	n = 13	Ortho Vitros 5600
70.96 ± 2.96	127.10 ± 7.86	42.24 ± 2.79	139.81 ± 8.63	102.25 ± 6.15	n = 4	Roche cobas c501
69.67 ± 1.99	128.21 ± 4.58	42.39 ± 1.04	139.07 ± 3.84	101.00 ± 2.83	n = 12	Roche cobas c701
71.47 ± 2.70	133.41 ± 5.15	43.55 ± 0.75	143.16 ± 3.89	104.37 ± 4.37	n = 10	Roche MODULAR D/P
70.36 ± 3.95	128.77 ± 3.91	43.56 ± 2.53	139.21 ± 3.72	104.69 ± 4.77	n = 3	Siemens ADVIA 1800
69.86 ± 2.40	123.68 ± 5.26	41.25 ± 1.22	133.18 ± 8.38	95.48 ± 4.70	n = 30	Siemens ADVIA 2400
						Siemens Dimension Vista
70.21 ± 1.44	125.57 ± 2.19	41.43 ± 0.64	136.16 ± 2.30	99.71 ± 1.56	n = 7	<Reagents>
73.05 ± 2.62	133.46 ± 7.75	43.77 ± 1.57	144.27 ± 5.96	104.81 ± 4.46	n = 11	Abbott
71.30 ± 2.09	130.26 ± 1.86	43.32 ± 1.34	141.47 ± 1.97	102.44 ± 0.93	n = 22	Beckman Coulter
76.31 ± 2.07	137.55 ± 4.92	47.58 ± 1.36	150.41 ± 6.36	111.32 ± 3.07	n = 17	Beckman Coulter AU Series
71.74 ± 1.86	130.96 ± 2.90	43.13 ± 1.58	140.56 ± 4.44	102.38 ± 3.87	n = 19	Ortho Clinical Diagnostics
69.67 ± 1.99	128.21 ± 4.58	42.39 ± 1.04	139.07 ± 3.84	101.00 ± 2.83	n = 12	Roche cobas c501/c311/c502/c701
71.33 ± 2.97	131.71 ± 5.44	43.70 ± 1.54	141.82 ± 4.26	104.06 ± 4.50	n = 14	Roche Hitachi and Modular D/P
69.94 ± 2.60	123.74 ± 6.46	41.27 ± 1.29	133.96 ± 11.37	95.92 ± 5.34	n = 32	Siemens ADVIA Centaur
						Siemens Dimension

New York State Department of Health – Wadsworth Center
 Quantitative Urine Chemistry – Educational Proficiency Testing – 26 October 2015

Summary of Participant Performance (Mean and Standard Deviation)

Potassium (mmol/L):

Specimen: U36	Specimen: U37	Specimen: U38	Specimen: U39	Specimen: U40	Number	Instrument or Reagent System
30.53 ± 1.82	57.48 ± 3.69	16.34 ± 0.72	65.12 ± 4.28	44.40 ± 2.44	n = 210	All Methods & Instruments
30.33 ± 1.88	57.63 ± 3.78	16.29 ± 0.75	64.61 ± 4.15	44.07 ± 2.25	n = 182	<Method Principles>
31.76 ± 0.71	56.48 ± 2.89	16.63 ± 0.42	68.95 ± 3.06	47.08 ± 2.02	n = 27	Ion selective electrode (diluted) Ion selective electrode (undiluted)
<Instruments>						
30.65 ± 0.34	57.70 ± 1.13	16.25 ± 0.30	64.94 ± 0.82	44.21 ± 0.38	n = 16	Abbott Architect c System
31.70 ± 0.54	60.72 ± 1.43	16.85 ± 0.27	67.53 ± 1.39	45.47 ± 1.01	n = 32	Beckman Coulter AU Chemistry System
30.04 ± 0.20	58.13 ± 0.80	16.67 ± 0.09	63.99 ± 0.72	44.12 ± 0.36	n = 9	Beckman Coulter UniCel DxC 600
30.20 ± 0.14	58.72 ± 0.52	16.93 ± 0.14	64.31 ± 0.34	44.54 ± 0.31	n = 8	Beckman Coulter UniCel DxC 800
31.83 ± 0.58	54.66 ± 1.00	16.67 ± 0.34	70.43 ± 1.00	48.29 ± 0.37	n = 7	Ortho Vitros 5,1FS
32.14 ± 0.39	54.85 ± 0.71	16.61 ± 0.26	70.93 ± 0.94	48.47 ± 0.71	n = 15	Ortho Vitros 5600
30.59 ± 0.58	59.45 ± 1.43	16.32 ± 0.30	65.63 ± 1.65	44.64 ± 0.90	n = 22	Roche cobas c501
31.38 ± 0.67	60.78 ± 1.43	16.57 ± 0.39	66.14 ± 1.23	45.21 ± 1.09	n = 6	Roche cobas c701
30.69 ± 0.81	58.32 ± 1.42	16.36 ± 0.32	63.96 ± 2.08	44.02 ± 1.22	n = 15	Roche MODULAR D/P
31.41 ± 0.51	60.85 ± 1.01	16.95 ± 0.11	67.19 ± 1.24	45.42 ± 0.73	n = 15	Siemens ADVIA 1800
31.47 ± 0.67	61.82 ± 1.50	17.05 ± 0.36	67.90 ± 1.56	45.60 ± 1.00	n = 3	Siemens ADVIA 2400
30.33 ± 0.34	55.63 ± 1.32	15.99 ± 0.15	63.86 ± 0.54	43.26 ± 0.39	n = 9	Siemens Dimension EXL
26.85 ± 0.70	52.22 ± 1.39	15.08 ± 0.18	57.82 ± 1.65	40.66 ± 0.84	n = 42	Siemens Dimension Vista
<Reagents>						
30.65 ± 0.34	57.70 ± 1.13	16.25 ± 0.30	64.94 ± 0.82	44.21 ± 0.38	n = 16	Abbott
30.19 ± 0.33	58.59 ± 0.87	16.78 ± 0.19	64.25 ± 0.71	44.38 ± 0.45	n = 19	Beckman Coulter
31.69 ± 0.52	60.71 ± 1.44	16.85 ± 0.29	67.50 ± 1.41	45.45 ± 1.00	n = 30	Beckman Coulter AU Series
31.98 ± 0.54	54.83 ± 0.90	16.60 ± 0.28	70.76 ± 0.94	48.37 ± 0.78	n = 24	Ortho Clinical Diagnostics
30.74 ± 0.67	59.73 ± 1.52	16.37 ± 0.34	65.75 ± 1.58	44.75 ± 0.98	n = 28	Roche cobas c501/c311/c502/c701
30.69 ± 0.81	58.32 ± 1.42	16.36 ± 0.32	63.96 ± 2.08	44.02 ± 1.22	n = 15	Roche Hitachi and Modular D/P
31.33 ± 0.56	60.82 ± 1.24	16.92 ± 0.26	67.11 ± 1.33	45.36 ± 0.72	n = 20	Siemens ADVIA/ADVISIA Centaur
27.36 ± 1.53	52.80 ± 1.88	15.23 ± 0.43	58.86 ± 3.08	41.13 ± 1.37	n = 55	Siemens Dimension

New York State Department of Health – Wadsworth Center
 Quantitative Urine Chemistry – Educational Proficiency Testing – 26 October 2015
 Summary of Participant Performance (Mean and Standard Deviation)

Total Protein (mg/dL):

Specimen: U36	Specimen: U37	Specimen: U38	Specimen: U39	Specimen: U40	Number	Instrument or Reagent System
231.06 ± 13.33	43.42 ± 5.28	283.00 ± 26.44	86.90 ± 5.43	116.89 ± 5.25	n = 214	All Methods & Instruments
229.77 ± 12.58	43.31 ± 4.99	278.93 ± 23.80	87.08 ± 5.05	117.02 ± 5.42	n = 60	<Method Principles>
236.10 ± 8.52	39.26 ± 1.58	293.07 ± 9.73	82.22 ± 2.42	114.28 ± 3.60	n = 42	Biuret (alkaline cupric sulfate)
229.86 ± 14.85	44.67 ± 4.02	277.03 ± 22.37	88.78 ± 4.13	118.67 ± 5.85	n = 94	Turbidimetric/Benzethonium Chloride
370.52 ± 17.77	56.70 ± 2.32	367.52 ± 24.99	114.26 ± 5.98	189.16 ± 8.56	n = 14	Pyrogallol red
215.48 ± 12.05	39.31 ± 2.16	255.42 ± 28.31	83.08 ± 1.42	113.12 ± 2.85	n = 3	Pyrocatechol Violet
						Other
240.92 ± 8.38	41.46 ± 1.62	301.82 ± 9.97	84.47 ± 1.56	117.75 ± 1.95	n = 14	<Instruments>
235.96 ± 12.63	42.80 ± 1.53	292.90 ± 16.45	87.70 ± 2.77	116.97 ± 4.95	n = 39	Abbott Architect c System
214.82 ± 51.67	45.25 ± 1.55	265.68 ± 41.21	98.36 ± 2.65	135.49 ± 2.52	n = 8	Beckman Coulter AU Chemistry System
224.79 ± 38.25	45.19 ± 1.49	273.36 ± 37.27	97.56 ± 4.51	133.65 ± 7.27	n = 7	Beckman Coulter UniCel DxC 600
366.00 ± 25.24	56.87 ± 1.87	364.26 ± 26.39	113.17 ± 5.20	187.35 ± 3.14	n = 6	Beckman Coulter UniCel DxC 800
372.07 ± 20.82	56.38 ± 3.14	367.67 ± 38.47	113.79 ± 11.28	181.26 ± 21.83	n = 15	Ortho Vitros 5,1FS
234.05 ± 9.92	38.96 ± 1.55	292.62 ± 8.51	82.61 ± 2.41	114.37 ± 3.33	n = 23	Ortho Vitros 5600
231.82 ± 5.38	38.35 ± 1.42	289.35 ± 6.01	81.46 ± 2.07	112.40 ± 1.67	n = 4	Roche cobas c501
230.68 ± 11.09	38.54 ± 0.93	287.11 ± 7.90	78.56 ± 1.98	108.35 ± 1.63	n = 3	Roche cobas c502
236.01 ± 6.38	38.50 ± 0.45	285.33 ± 7.95	80.36 ± 1.36	111.67 ± 1.83	n = 3	Roche cobas c701
232.00 ± 11.38	39.04 ± 0.35	290.26 ± 5.01	81.42 ± 1.48	112.54 ± 3.32	n = 16	Roche Cobas INTEGRA
208.48 ± 2.42	37.26 ± 1.45	261.85 ± 12.01	83.68 ± 1.43	113.84 ± 1.97	n = 15	Roche MODULAR D/P
215.04 ± 1.47	38.80 ± 2.36	271.48 ± 1.87	85.44 ± 1.02	114.81 ± 1.81	n = 3	Siemens ADVIA 1800
231.47 ± 6.77	48.01 ± 1.23	280.28 ± 8.43	88.82 ± 1.61	118.29 ± 1.90	n = 8	Siemens ADVIA 2400
230.66 ± 5.05	47.56 ± 1.21	265.00 ± 13.24	89.78 ± 1.89	120.04 ± 3.12	n = 42	Siemens Dimension EXL
						Siemens Dimension Vista
239.87 ± 9.85	41.48 ± 1.55	300.10 ± 12.01	84.54 ± 1.51	117.91 ± 1.95	n = 15	<Reagents>
223.44 ± 40.62	44.87 ± 1.78	272.27 ± 36.89	97.15 ± 4.49	132.94 ± 7.29	n = 17	Abbott
235.36 ± 12.80	42.77 ± 1.57	292.90 ± 17.24	87.85 ± 2.82	117.02 ± 5.11	n = 36	Beckman Coulter AU Series
372.22 ± 19.05	56.31 ± 2.82	365.90 ± 32.03	114.75 ± 6.42	189.18 ± 6.28	n = 21	Ortho Clinical Diagnostics
233.76 ± 8.58	38.77 ± 1.36	291.46 ± 8.15	82.02 ± 2.42	113.68 ± 3.42	n = 29	Roche cobas c501/c311/c502/c701
232.00 ± 11.38	39.04 ± 0.35	290.26 ± 5.01	81.42 ± 1.48	112.54 ± 3.32	n = 16	Roche Hitachi and Modular D/P
231.17 ± 11.24	38.28 ± 0.54	281.87 ± 9.10	79.57 ± 1.97	110.64 ± 2.63	n = 4	Roche Integra and MIRA S
210.40 ± 5.55	37.39 ± 1.48	264.99 ± 10.81	84.04 ± 1.51	114.16 ± 2.01	n = 19	Siemens ADVIA/ADVISIA Centaur
230.72 ± 5.50	47.65 ± 1.26	268.46 ± 13.97	89.71 ± 2.05	119.56 ± 3.10	n = 53	Siemens Dimension

New York State Department of Health – Wadsworth Center
Quantitative Urine Chemistry – Educational Proficiency Testing – 26 October 2015

Summary of Participant Performance (Mean and Standard Deviation)

Uric Acid (mg/dL):

Specimen: U36	Specimen: U37	Specimen: U38	Specimen: U39	Specimen: U40	Number	Instrument or Reagent System
39.15 ± 2.22	21.06 ± 1.26	33.95 ± 3.24	27.45 ± 1.61	17.19 ± 1.05	n = 142	All Methods & Instruments
39.24 ± 2.36	20.44 ± 1.53	34.83 ± 2.01	26.89 ± 1.96	16.56 ± 1.45	n = 6	<Method Principles>
37.73 ± 2.15	20.58 ± 1.16	33.50 ± 2.78	26.65 ± 1.42	16.83 ± 0.99	n = 48	Uricase (NAD-NADH reaction)
39.87 ± 1.71	21.37 ± 1.17	34.19 ± 3.51	27.90 ± 1.43	17.41 ± 0.95	n = 86	Uricase/allantoin (differential abs)
						Uricase/peroxidase (colorimetric)
37.81 ± 0.65	20.73 ± 0.33	31.56 ± 3.33	26.52 ± 0.89	16.99 ± 0.29	n = 8	<Instruments>
41.22 ± 1.36	22.69 ± 0.71	36.06 ± 3.51	29.19 ± 0.82	18.20 ± 0.61	n = 23	Abbott Architect c System
40.69 ± 1.39	22.51 ± 0.58	37.07 ± 1.46	28.91 ± 0.83	18.78 ± 0.60	n = 5	Beckman Coulter AU Chemistry System
40.85 ± 1.44	22.44 ± 0.83	36.35 ± 1.18	29.18 ± 0.73	18.63 ± 0.55	n = 6	Beckman Coulter UniCel DxC 600
40.27 ± 1.04	20.95 ± 0.55	35.25 ± 0.82	27.51 ± 0.62	17.19 ± 0.39	n = 5	Beckman Coulter UniCel DxC 800
39.98 ± 0.87	20.71 ± 0.55	33.14 ± 5.93	27.35 ± 0.88	17.08 ± 0.69	n = 11	Ortho Vitros 5,1FS
39.25 ± 1.15	20.95 ± 0.74	33.91 ± 1.77	27.05 ± 0.92	16.95 ± 0.18	n = 14	Ortho Vitros 5600
38.91 ± 2.08	20.61 ± 0.87	32.81 ± 0.58	27.04 ± 1.71	16.61 ± 0.94	n = 4	Roche cobas c501
37.21 ± 1.32	19.73 ± 0.55	32.73 ± 1.52	25.76 ± 1.00	15.82 ± 0.32	n = 11	Roche cobas c701
39.85 ± 0.74	21.59 ± 0.57	32.01 ± 5.40	28.42 ± 0.81	17.75 ± 0.69	n = 12	Roche MODULAR D/P
41.00 ± 0.55	21.95 ± 0.19	35.93 ± 0.41	28.13 ± 0.69	18.22 ± 0.50	n = 3	Siemens ADVIA 1800
41.00 ± 2.09	22.48 ± 0.64	37.37 ± 0.94	28.95 ± 1.07	18.32 ± 0.78	n = 4	Siemens ADVIA 2400
36.61 ± 0.92	20.03 ± 0.57	33.07 ± 0.95	26.04 ± 0.65	16.44 ± 0.55	n = 30	Siemens Dimension EXL
						Siemens Dimension Vista
37.81 ± 0.65	20.73 ± 0.33	31.56 ± 3.33	26.52 ± 0.89	16.99 ± 0.29	n = 8	<Reagents>
40.77 ± 1.42	22.45 ± 0.73	36.69 ± 1.44	29.06 ± 0.79	18.71 ± 0.58	n = 11	Abbott
41.15 ± 1.36	22.66 ± 0.73	36.41 ± 2.75	29.14 ± 0.75	18.21 ± 0.63	n = 22	Beckman Coulter
40.09 ± 0.91	20.76 ± 0.54	34.25 ± 3.86	27.40 ± 0.79	17.10 ± 0.55	n = 17	Beckman Coulter AU Series
39.29 ± 1.44	20.96 ± 0.79	33.83 ± 1.76	27.12 ± 1.09	16.85 ± 0.55	n = 20	Ortho Clinical Diagnostics
37.21 ± 1.32	19.73 ± 0.55	32.73 ± 1.52	25.76 ± 1.00	15.82 ± 0.32	n = 11	Roche Hitachi and Modular D/P
40.04 ± 0.97	21.62 ± 0.62	34.82 ± 1.46	28.28 ± 0.86	17.78 ± 0.72	n = 16	Siemens ADVIA/ADVISIA Centaur
36.70 ± 1.03	20.12 ± 0.77	33.14 ± 2.62	26.11 ± 0.77	16.54 ± 0.68	n = 34	Siemens Dimension

New York State Department of Health – Wadsworth Center
 Quantitative Urine Chemistry – Educational Proficiency Testing – 26 October 2015

Summary of Participant Performance (Mean and Standard Deviation)

Urea Nitrogen (mg/dL):

Specimen: U36	Specimen: U37	Specimen: U38	Specimen: U39	Specimen: U40	Number	Instrument or Reagent System
736.1 ± 29.85	821.5 ± 32.75	698.9 ± 26.70	1128.3 ± 38.58	1206.8 ± 52.13	n = 154	All Methods & Instruments
736.4 ± 31.65	821.8 ± 34.37	699.5 ± 27.21	1126.2 ± 40.09	1197.2 ± 47.54	n = 129	<Method Principles>
736.7 ± 37.93	827.8 ± 37.26	701.0 ± 33.48	1149.3 ± 51.06	1239.7 ± 61.41	n = 7	Urease w/glutamate dehydrogenase
735.0 ± 12.05	815.3 ± 14.86	692.4 ± 13.83	1135.4 ± 17.52	1257.8 ± 18.77	n = 16	Urease, conductivity rate
						Urease with indicator dye
714.8 ± 6.92	790.6 ± 18.05	670.6 ± 9.46	1108.3 ± 17.07	1179.9 ± 18.86	n = 8	<Instruments>
734.8 ± 22.44	818.2 ± 23.25	701.7 ± 16.54	1124.2 ± 26.96	1191.2 ± 33.58	n = 28	Abbott Architect c System
742.5 ± 33.65	831.0 ± 21.69	705.6 ± 30.70	1152.7 ± 16.21	1223.7 ± 21.90	n = 6	Beckman Coulter AU Chemistry System
724.5 ± 35.30	813.1 ± 41.87	694.7 ± 32.74	1133.5 ± 78.60	1229.8 ± 78.15	n = 5	Beckman Coulter UniCel DxC 600
736.6 ± 8.09	822.9 ± 12.35	703.2 ± 22.24	1136.8 ± 15.23	1262.4 ± 6.64	n = 5	Beckman Coulter UniCel DxC 800
732.6 ± 18.81	815.3 ± 19.53	690.1 ± 13.33	1133.5 ± 20.90	1256.6 ± 27.01	n = 13	Ortho Vitros 5,1FS
729.3 ± 27.76	818.5 ± 27.90	698.9 ± 21.13	1125.2 ± 31.98	1192.7 ± 40.34	n = 16	Ortho Vitros 5600
711.7 ± 29.03	789.8 ± 43.98	682.6 ± 32.33	1097.3 ± 56.80	1151.7 ± 72.57	n = 6	Roche cobas c501
725.4 ± 14.44	813.3 ± 20.87	688.4 ± 13.02	1115.4 ± 27.77	1183.4 ± 21.25	n = 12	Roche MODULAR D/P
792.1 ± 11.36	885.2 ± 21.03	750.1 ± 17.84	1191.7 ± 62.26	1266.1 ± 60.71	n = 11	Siemens ADVIA 1800
794.4 ± 20.76	856.9 ± 44.28	728.9 ± 22.36	1120.5 ± 34.13	1225.3 ± 35.88	n = 3	Siemens ADVIA 2400
806.5 ± 4.60	877.2 ± 10.81	764.1 ± 9.54	1228.2 ± 31.53	1312.2 ± 24.91	n = 4	Siemens Dimension EXL
739.0 ± 19.85	819.0 ± 28.61	697.3 ± 20.44	1118.4 ± 35.79	1189.5 ± 45.72	n = 31	Siemens Dimension Vista
714.8 ± 6.92	790.6 ± 18.05	670.6 ± 9.46	1108.3 ± 17.07	1179.9 ± 18.86	n = 8	<Reagents>
737.2 ± 34.22	828.8 ± 26.47	702.4 ± 30.23	1144.7 ± 44.90	1223.7 ± 47.90	n = 12	Abbott
735.0 ± 23.02	817.8 ± 23.98	701.8 ± 17.15	1125.4 ± 26.77	1191.8 ± 34.31	n = 27	Beckman Coulter
734.1 ± 15.30	818.0 ± 17.15	692.6 ± 15.36	1135.2 ± 18.81	1257.4 ± 21.20	n = 19	Beckman Coulter AU Series
725.9 ± 29.24	815.2 ± 36.49	696.2 ± 20.52	1123.7 ± 42.96	1188.9 ± 53.86	n = 23	Ortho Clinical Diagnostics
725.4 ± 14.44	813.3 ± 20.87	688.4 ± 13.02	1115.4 ± 27.77	1183.4 ± 21.25	n = 12	Roche cobas c501/c311/c502/c701
791.2 ± 19.80	877.3 ± 34.51	743.7 ± 25.06	1171.1 ± 64.84	1254.5 ± 54.91	n = 15	Roche Hitachi and Modular D/P
742.3 ± 27.00	825.0 ± 33.01	701.0 ± 26.40	1125.5 ± 45.35	1198.4 ± 56.93	n = 36	Siemens ADVIA/ADVISIA Centaur
						Siemens Dimension