



**Department
of Health**

**Wadsworth
Center**

TRACE ELEMENTS IN SERUM

Proficiency Test Report

Event #2, 2015

July 2nd, 2015



Department of Health

ANDREW M. CUOMO
Governor

HOWARD A. ZUCKER, M.D., J.D.
Commissioner

SALLY DRESLIN, M.S., R.N.
Executive Deputy Commissioner

July 2, 2015

Trace Elements in Serum Event #2, 2015

Dear Laboratory Director:

Results from the second proficiency test (PT) event for 2015 in the category Trace Elements in Serum have been tabulated and are summarized. Target values for aluminum, copper, selenium and zinc have been established along with acceptable ranges. Results are graded using element-specific criteria as indicated in each narrative section. A laboratory with an unacceptable significant analytical bias relative to the target value will be expected to investigate the source of the error. A confidential three-digit code number assigned by the PT program identifies participant laboratories.

PT Materials

Test materials were prepared from human serum obtained from Tennessee Blood Services, Inc. Serum units were spiked with a suite of additional trace elements as described in each narrative. In addition to aluminum (Al), copper (Cu), selenium (Se) and zinc (Zn), serum pools were supplemented with the trace elements arsenic (As), cadmium (Cd), chromium (Cr), cobalt, (Co), lead (Pb), manganese (Mn), mercury (Hg), molybdenum (Mo), nickel (Ni), thallium (Tl), tin (Sn), titanium (Ti), tungsten (W) and vanadium (V).

The next PT event for trace elements in serum is scheduled to be mailed Wednesday, September 16th, 2015. Please inform our laboratory staff at (518) 474-7161 if the test materials have not arrived within five days of the scheduled mail out date. **The deadline for reporting results is Wednesday, October 14th, 2015.**

Thank you for your participation.

Sincerely,

Patrick J. Parsons, Ph.D.
Chief, Laboratory of Inorganic and Nuclear Chemistry
Deputy Director, Division of Environmental Health

Mary Frances Verostek, Ph.D.
Assistant Section Head
PT Program for Blood Lead /Trace Elements
Biggs Laboratory - Wadsworth Center
New York State Department of Health
PO Box 509
Albany NY 12201-0509
www.wadsworth.org/testing/lead/index.htm

New York State Department of Health
Event #2, 2015

Serum Aluminum

The test materials for serum Al were prepared from human serum obtained from Tennessee Blood Services, Inc. The units were tested by FDA approved methods and found to be negative for HIV 1/2 and HIV-1 RNA, and non-reactive to HBsAg, HCV3 and STS. Serum was dispensed into acid-washed 500-mL polypropylene containers to make up five (5) serum pools. Each pool was spiked with a suite of additional trace elements including aluminum as Al³⁺ at various concentrations.

The Target Value assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E** Statistical methods for use in proficiency testing by interlaboratory comparisons. Values for serum aluminum range from 24 µg/L (0.89 µmol/L) to 120 µg/L (4.45 µmol/L).

Acceptable ranges for serum aluminum are based on fixed criteria of ±20%, or ±5 µg/L below 25 µg/L. These criteria are based on consensus recommendations from several EQAS organizers (1).

Discussion. Based on the above criteria, 98.2% of test results reported were judged as satisfactory, with none out of 22 participant laboratories reporting 2 or more of the 5 results outside the acceptable ranges.

1. Taylor, A., Angerer, J., Claeys, F., Kristiansen, J., Mazarrasa, O., Menditto, A., Patriarca, M., Pineau, A., Schoeters, I., Sykes, C., Valkonen, S. and Weykamp, C. Comparison of procedures for evaluating laboratory performance in external quality assessment schemes for lead in blood and aluminum in serum demonstrates the need for common quality specifications. Clinical Chemistry 2002 48 2000-2007.

**New York State Department of Health
Serum Aluminum Test Results, 2015 Event #2
ROBUST STATISTICAL SUMMARY**

TARGET VALUE ASSIGNMENT AND STATISTICS

Results ($\mu\text{g/L}$ serum)

| | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
|-------------------------------|------------|-----------|-----------|-----------|-----------|
| Robust Mean | 120 | 50 | 67 | 24 | 40 |
| Robust Standard Deviation | 3 | 3 | 3 | 2 | 3 |
| Standard Uncertainty | 0.9 | 0.7 | 0.8 | 0.5 | 0.7 |
| RSD (%) | 2.6 | 5.3 | 4.3 | 8.0 | 6.3 |
| Number of Sample Measurements | 21 | 22 | 22 | 22 | 22 |
| Acceptable Range: | | | | | |
| Upper Limit: | 144 | 60 | 80 | 29 | 48 |
| Lower Limit: | 96 | 40 | 54 | 19 | 32 |

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Serum Aluminum Test Results, 2015 Event #2
PERFORMANCE OF PARTICIPATING LABORATORIES

| Lab Code | Method | Results ($\mu\text{g/L}$ serum) | | | | | Info Only |
|----------|----------------|----------------------------------|---------|---------|---------------|---------|-----------|
| | | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 | |
| | Target Values: | 120 | 50 | 67 | 24 | 40 | |
| 114 | ICP-MS | 124 | 49 | 66 | 21 | 39 | |
| 147 | FAAS | 122 | 53 | 71 | 26 | 43 | Info |
| 156 | ICP-MS | 120 | 52 | 68 | 24 | 41 | |
| 160 | ICP-MS | 112 | 47 | 63 | 23 | 38 | |
| 164 | ICP-MS | 120 | 51 | 69 | 25 | 41 | |
| 179 | DRC/CC-ICP-MS | 121 | 52 | 69 | 24 | 42 | |
| 197 | ICP-MS | 121 | 51 | 69 | 26 | 43 | |
| 200 | DRC/CC-ICP-MS | 120 | 52 | 73 | 31 \uparrow | 40 | Info |
| 206 | DRC/CC-ICP-MS | >100 | 51 | 67 | 26 | 44 | |
| 287 | ETAAS-Z | 125 | 53 | 70 | 26 | 43 | |
| 293 | ICP-MS | 125 | 53 | 69 | 26 | 43 | Info |
| 305 | ICP-MS | 121 | 48 | 65 | 24 | 40 | |
| 324 | HR-ICP-MS | 123 | 47 | 69 | 23 | 39 | Info |
| 325 | ETAAS-Z | 114 | 51 | 67 | 26 | 40 | Info |
| 355 | ICP-MS | 123 | 50 | 69 | 24 | 40 | |
| 357 | ICP-MS | 121 | 51 | 68 | 24 | 40 | |
| 358 | ICP-MS | 110 | 47 | 65 | 22 | 38 | |
| 363 | ICP-MS | 106 | 47 | 61 | 22 | 38 | |
| 366 | ETAAS-Z | 93 \downarrow | 45 | 58 | 22 | 32 | Info |
| 401 | DRC/CC-ICP-MS | 119 | 49 | 65 | 22 | 38 | Info |
| 458 | ETAAS other | 119 | 46 | 61 | 23 | 38 | |
| 485 | HR-ICP-MS | 121.6 | 53.36 | 69.07 | 23.50 | 44.83 | Info |

Percent satisfactory results for all participants: 98.2 %

notes: \uparrow reported outside upper limit
 \downarrow reported outside lower limit

Info only: results included for informational purposes only.

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

\blacktriangle : Result not reported

New York State Department of Health
Serum Aluminum Test Results, 2015 Event #2
STATISTICAL SUMMARY BY METHOD

| | Results ($\mu\text{g/L}$ serum) | | | | |
|--------------------------------|----------------------------------|---------|---------|---------|---------|
| | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| DRC/CC-ICP-MS | | | | | |
| Number of Sample Measurements: | 3 | 4 | 4 | 4 | 4 |
| Mean: | 120 | 51 | 69 | 26 | 41 |
| Standard Deviation: | 1 | 1 | 3 | 4 | 3 |
| RSD (%): | — | 2.8 | 5.0 | 15.0 | 6.3 |
| ETAAS other | | | | | |
| Number of Sample Measurements: | 1 | 1 | 1 | 1 | 1 |
| Mean: | 119 | 46 | 61 | 23 | 38 |
| Standard Deviation: | ? | ? | ? | ? | ? |
| RSD (%): | — | — | — | — | — |
| ETAAS-Z | | | | | |
| Number of Sample Measurements: | 3 | 3 | 3 | 3 | 3 |
| Mean: | 111 | 50 | 65 | 25 | 38 |
| Standard Deviation: | 16 | 4 | 6 | 2 | 6 |
| RSD (%): | — | — | — | — | — |
| FAAS | | | | | |
| Number of Sample Measurements: | 1 | 1 | 1 | 1 | 1 |
| Mean: | 122 | 53 | 71 | 26 | 43 |
| Standard Deviation: | ? | ? | ? | ? | ? |
| RSD (%): | — | — | — | — | — |
| HR-ICP-MS | | | | | |
| Number of Sample Measurements: | 2 | 2 | 2 | 2 | 2 |
| Mean: | 122 | 50 | 69 | 23 | 42 |
| Standard Deviation: | 1 | 4 | 0 | 0 | 4 |
| RSD (%): | — | — | — | — | — |
| ICP-MS | | | | | |
| Number of Sample Measurements: | 11 | 11 | 11 | 11 | 11 |
| Mean: | 118 | 50 | 67 | 24 | 40 |
| Standard Deviation: | 6 | 2 | 3 | 2 | 2 |
| RSD (%): | 5.2 | 4.3 | 4.2 | 6.8 | 4.5 |
| All Laboratories | | | | | |
| Number of Sample Measurements: | 21 | 22 | 22 | 22 | 22 |
| Mean: | 118 | 50 | 67 | 24 | 40 |
| Standard Deviation: | 8 | 3 | 4 | 2 | 3 |
| RSD (%): | 6.4 | 5.1 | 5.4 | 9.0 | 7.0 |

notes: ? Insufficient data for calculation.

New York State Department of Health
Event #2, 2015

Serum Copper

The test materials for serum Cu were prepared from human serum obtained from Tennessee Blood Services, Inc. The units were tested by FDA approved methods and found to be Non-reactive for Anti-HIV-1/2, Anti-HCV 3.0 and HBsAg. The serum has also been found to be STS (RPR) Non-reactive and Negative for HIV-1 and HCV by PCR. Serum units were dispensed into acid-washed 500-mL polypropylene containers to make up five (5) serum pools. Each pool was spiked with a suite of additional trace elements including copper as Cu²⁺ at various concentrations.

The Target Value assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E** Statistical methods for use in proficiency testing by interlaboratory comparisons. Values for serum copper range from 1156 µg/L (18.19 µmol/L) to 3171 µg/L (49.90 µmol/L).

Acceptable ranges for serum copper are based on fixed criteria of ±15%, or ±95 µg/L below 635 µg/L. These criteria are consistent with those proposed by the OELM Network of EQAS organizers (1, 2) for trace elements in serum, and are slightly less stringent than those previously suggested for NYS (±10%).

Discussion. Based on the above criteria, 93.6% of test results reported were judged as satisfactory, with two out of 22 participant laboratories (9.1%) reporting 2 or more of the 5 results outside the acceptable ranges.

1. A. Taylor, J. Angerer, J. Arnaud, F. Claeys, R.L. Jones, O. Mazarrasa, E. Mairiaux, A. Menditto, P.J. Parsons, M. Patriarca, A. Pineau, S. Valkonen, J.-P. Weber and C. Weykamp Accreditation and Quality Assurance 2006 **11** 440-445.

2. J. Arnaud, J.-P. Weber, C.W. Weykamp, P.J. Parsons, J. Angerer, E. Mairiaux, O. Mazarrasa, S. Valkonen, A. Menditto, M. Patriarca, and A. Taylor Clinical Chemistry 2008 **54** 1892-1899.

**New York State Department of Health
Serum Copper Test Results, 2015 Event #2
ROBUST STATISTICAL SUMMARY**

TARGET VALUE ASSIGNMENT AND STATISTICS

Results ($\mu\text{g/L}$ serum)

| | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
|-------------------------------|-------------|-------------|-------------|-------------|-------------|
| Robust Mean | 1814 | 1445 | 3171 | 1821 | 1156 |
| Robust Standard Deviation | 96 | 94 | 233 | 105 | 76 |
| Standard Uncertainty | 26 | 25 | 62 | 28 | 20 |
| RSD (%) | 5.3 | 6.5 | 7.4 | 5.8 | 6.6 |
| Number of Sample Measurements | 22 | 22 | 22 | 22 | 22 |
| Acceptable Range: | | | | | |
| Upper Limit: | 2086 | 1662 | 3647 | 2094 | 1329 |
| Lower Limit: | 1542 | 1228 | 2695 | 1548 | 983 |

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Serum Copper Test Results, 2015 Event #2
PERFORMANCE OF PARTICIPATING LABORATORIES

| Lab Code | Method | Results (µg/L serum) | | | | | Info Only |
|----------|----------------|----------------------|---------|---------|---------|---------|-----------|
| | | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 | |
| | Target Values: | 1814 | 1445 | 3171 | 1821 | 1156 | |
| 107 | DRC/CC-ICP-MS | 1800 | 1400 | 3600 | 1800 | 1200 | Info |
| 110 | ICP-MS | 1942 | 1560 | 3505 | 1937 | 1227 | |
| 114 | ICP-MS | 1820 | 1450 | 3010 | 1810 | 1170 | |
| 147 | ICP-MS | 1747 | 1385 | 2986 | 1760 | 1093 | Info |
| 156 | ICP-AES/OES | 1800 | 1500 | 3300 | 1900 | 1100 | |
| 160 | ICP-MS | 1740 | 1370 | 2950 | 1710 | 1070 | |
| 164 | ICP-MS | 1833 | 1467 | 3178 | 1841 | 1163 | |
| 179 | DRC/CC-ICP-MS | 1860 | 1500 | 3270 | 1870 | 1180 | |
| 197 | ICP-MS | 1800 | 1400 | 3190 | 1790 | 1110 | |
| 200 | ICP-MS | 1949 | 1562 | 3429 | 1988 | 1251 | Info |
| 206 | ICP-MS | 1750 | 1385 | 3052 | 1812 | 1118 | |
| 293 | ICP-MS | 1844 | 1526 | 3319 | 1850 | 1195 | Info |
| 305 | ICP-MS | 1540 ↓ | 1280 | 2930 | 1810 | 1080 | |
| 324 | HR-ICP-MS | 1748 | 1358 | 3000 | 1720 | 1098 | Info |
| 325 | ICP-MS | 2370 ↑ | 2090 ↑ | 3330 | 2200 ↑ | 1550 ↑ | Info |
| 359 | ICP-MS | 1893 | 1463 | 3407 | 1847 | 1196 | |
| 366 | ETAAS-Z | 1820 | 1450 | 2931 | 1690 | 1184 | Info |
| 388 | ICP-MS | 2136 ↑ | 1661 | 3414 | 2074 | 1409 ↑ | |
| 401 | DRC/CC-ICP-MS | 1703 | 1379 | 2923 | 1709 | 1087 | Info |
| 457 | ICP-AES/OES | 1845 | 1448 | 3135 | 1857 | 1214 | Info |
| 483 | DRC/CC-ICP-MS | 1733 | 1367 | 2949 | 1714 | 1075 | Info |
| 484 | ICP-MS | 1704 | 1351 | 3037 | 1705 | 1076 | |

Percent satisfactory results for all participants: 93.6 %

notes: ↑ reported outside upper limit
↓ reported outside lower limit

Info only: results included for informational purposes only.

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

▲: Result not reported

**New York State Department of Health
Serum Copper Test Results, 2015 Event #2
STATISTICAL SUMMARY BY METHOD**

| | Results ($\mu\text{g/L}$ serum) | | | | |
|--------------------------------|----------------------------------|---------|---------|---------|---------|
| | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| DRC/CC-ICP-MS | | | | | |
| Number of Sample Measurements: | 4 | 4 | 4 | 4 | 4 |
| Mean: | 1774 | 1412 | 3186 | 1773 | 1136 |
| Standard Deviation: | 70 | 61 | 318 | 77 | 64 |
| RSD (%): | 4.0 | 4.3 | 10.0 | 4.3 | 5.6 |
| ETAAS-Z | | | | | |
| Number of Sample Measurements: | 1 | 1 | 1 | 1 | 1 |
| Mean: | 1820 | 1450 | 2931 | 1690 | 1184 |
| Standard Deviation: | ? | ? | ? | ? | ? |
| RSD (%): | — | — | — | — | — |
| HR-ICP-MS | | | | | |
| Number of Sample Measurements: | 1 | 1 | 1 | 1 | 1 |
| Mean: | 1748 | 1358 | 3000 | 1720 | 1098 |
| Standard Deviation: | ? | ? | ? | ? | ? |
| RSD (%): | — | — | — | — | — |
| ICP-AES/OES | | | | | |
| Number of Sample Measurements: | 2 | 2 | 2 | 2 | 2 |
| Mean: | 1823 | 1474 | 3218 | 1879 | 1157 |
| Standard Deviation: | 32 | 37 | 117 | 30 | 81 |
| RSD (%): | — | — | — | — | — |
| ICP-MS | | | | | |
| Number of Sample Measurements: | 14 | 14 | 14 | 14 | 14 |
| Mean: | 1862 | 1496 | 3196 | 1867 | 1193 |
| Standard Deviation: | 201 | 198 | 202 | 139 | 137 |
| RSD (%): | 10.8 | 13.2 | 6.3 | 7.5 | 11.5 |
| All Laboratories | | | | | |
| Number of Sample Measurements: | 22 | 22 | 22 | 22 | 22 |
| Mean: | 1835 | 1471 | 3175 | 1836 | 1175 |
| Standard Deviation: | 165 | 163 | 213 | 127 | 116 |
| RSD (%): | 9.0 | 11.1 | 6.7 | 6.9 | 9.8 |

notes: ? Insufficient data for calculation.

New York State Department of Health
Event #2, 2015

Serum Selenium

The test materials for serum Se were prepared from human serum obtained from Tennessee Blood Services, Inc. The units were tested by FDA approved methods and found to be Non-reactive for Anti-HIV-1/2, Anti-HCV 3.0 and HBsAg. The serum has also been found to be STS (RPR) Non-reactive and Negative for HIV-1 and HCV by PCR. Serum units were dispensed into acid-washed 500-mL polypropylene containers to make up five (5) serum pools. Each pool was spiked with a suite of additional trace elements including selenium as Se⁴⁺ at various concentrations.

The Target Value assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E** Statistical methods for use in proficiency testing by interlaboratory comparisons. Values for serum selenium range from 91 µg/L (1.15 µmol/L) to 269 µg/L (3.41 µmol/L).

Acceptable ranges for serum selenium are based on fixed criteria of ±20%, or ±2 µg/L below 10 µg/L. These criteria are a little less stringent than those proposed by the OELM Network of EQAS organizers (±15% or ±8 µg/L below 55 µg/L) (1, 2) for trace elements in serum. As performance for serum Se improves among NYS-permit laboratories, consideration will be given to adopting the OELM criteria.

Discussion. Based on the above criteria, 100% of test results reported were judged as satisfactory, with none of the 19 participant laboratories reporting 2 or more of the 5 results outside the acceptable ranges.

1. A. Taylor, J. Angerer, J. Arnaud, F. Claeys, R.L. Jones, O. Mazarrasa, E. Mairiaux, A. Menditto, P.J. Parsons, M. Patriarca, A. Pineau, S. Valkonen, J.-P. Weber and C. Weykamp Accreditation and Quality Assurance 2006 11 440-445.

2. J. Arnaud, J.-P. Weber, C.W. Weykamp, P.J. Parsons, J. Angerer, E. Mairiaux, O. Mazarrasa, S. Valkonen, A. Menditto, M. Patriarca, and A. Taylor Clinical Chemistry 2008 54 1892-1899.

**New York State Department of Health
Serum Selenium Test Results, 2015 Event #2
ROBUST STATISTICAL SUMMARY**

TARGET VALUE ASSIGNMENT AND STATISTICS

Results ($\mu\text{g/L}$ serum)

| | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
|-------------------------------|------------|------------|------------|-----------|------------|
| Robust Mean | 269 | 134 | 223 | 91 | 116 |
| Robust Standard Deviation | 15 | 11 | 14 | 6 | 8 |
| Standard Uncertainty | 4 | 3 | 4 | 2 | 2 |
| RSD (%) | 5.5 | 8.0 | 6.3 | 7.0 | 6.7 |
| Number of Sample Measurements | 19 | 19 | 19 | 19 | 19 |
| Acceptable Range: | | | | | |
| Upper Limit: | 323 | 161 | 268 | 109 | 139 |
| Lower Limit: | 215 | 107 | 178 | 73 | 93 |

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Serum Selenium Test Results, 2015 Event #2
PERFORMANCE OF PARTICIPATING LABORATORIES

| Lab Code | Method | Results ($\mu\text{g/L}$ serum) | | | | | Info Only |
|----------------|---------------|----------------------------------|---------|---------|---------|---------|-----------|
| | | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 | |
| Target Values: | | 269 | 134 | 223 | 91 | 116 | |
| 103 | DRC/CC-ICP-MS | 281 | 136 | 228 | 93.5 | 118 | Info |
| 107 | DRC/CC-ICP-MS | 260 | 120 | 210 | 84 | 110 | Info |
| 110 | DRC/CC-ICP-MS | 261 | 132 | 218 | 90 | 114 | |
| 114 | ICP-MS | 269 | 137 | 224 | 95 | 116 | |
| 147 | ICP-MS | 242 | 119 | 207 | 85 | 107 | Info |
| 156 | DRC/CC-ICP-MS | 280 | 130 | 230 | 93 | 110 | |
| 160 | ICP-MS | 264 | 130 | 223 | 93 | 118 | |
| 164 | DRC/CC-ICP-MS | 306 | 146 | 258 | 98 | 129 | |
| 179 | DRC/CC-ICP-MS | 268 | 133 | 219 | 84 | 120 | |
| 200 | DRC/CC-ICP-MS | 261 | 133 | 213 | 88 | 115 | Info |
| 206 | DRC/CC-ICP-MS | 245 | 127 | 203 | 88 | 109 | |
| 293 | DRC/CC-ICP-MS | 291 | 148 | 235 | 100 | 125 | Info |
| 305 | ICP-MS | 276 | 138 | 224 | 96 | 120 | |
| 324 | HR-ICP-MS | 259 | 120 | 211 | 85 | 109 | Info |
| 325 | ETAAS-Z | 260 | 155 | 240 | 85 | 115 | Info |
| 366 | ETAAS-Z | 283 | 143 | 234 | 96 | 129 | Info |
| 388 | ICP-MS | 275 | 140 | 227 | 100 | 119 | |
| 401 | DRC/CC-ICP-MS | 289 | 145 | 238 | 97 | 124 | Info |
| 483 | DRC/CC-ICP-MS | 250 | 124 | 205 | 86 | 105 | Info |

Percent satisfactory results for all participants: 100.0 %

notes: ↑ reported outside upper limit
↓ reported outside lower limit

Info only: results included for informational purposes only.

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

▲: Result not reported

New York State Department of Health
Serum Selenium Test Results, 2015 Event #2
STATISTICAL SUMMARY BY METHOD

| | Results ($\mu\text{g/L}$ serum) | | | | |
|--------------------------------|----------------------------------|---------|---------|---------|---------|
| | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| DRC/CC-ICP-MS | | | | | |
| Number of Sample Measurements: | 11 | 11 | 11 | 11 | 11 |
| Mean: | 272 | 134 | 223 | 91 | 116 |
| Standard Deviation: | 19 | 9 | 16 | 6 | 8 |
| RSD (%): | 6.9 | 6.8 | 7.4 | 6.2 | 6.6 |
| ETAAS-Z | | | | | |
| Number of Sample Measurements: | 2 | 2 | 2 | 2 | 2 |
| Mean: | 272 | 149 | 237 | 91 | 122 |
| Standard Deviation: | 16 | 8 | 4 | 8 | 10 |
| RSD (%): | — | — | — | — | — |
| HR-ICP-MS | | | | | |
| Number of Sample Measurements: | 1 | 1 | 1 | 1 | 1 |
| Mean: | 259 | 120 | 211 | 85 | 109 |
| Standard Deviation: | ? | ? | ? | ? | ? |
| RSD (%): | — | — | — | — | — |
| ICP-MS | | | | | |
| Number of Sample Measurements: | 5 | 5 | 5 | 5 | 5 |
| Mean: | 265 | 133 | 221 | 94 | 116 |
| Standard Deviation: | 14 | 9 | 8 | 6 | 5 |
| RSD (%): | 5.2 | 6.5 | 3.6 | 5.9 | 4.5 |
| All Laboratories | | | | | |
| Number of Sample Measurements: | 19 | 19 | 19 | 19 | 19 |
| Mean: | 269 | 135 | 224 | 91 | 116 |
| Standard Deviation: | 16 | 10 | 14 | 6 | 7 |
| RSD (%): | 6.1 | 7.5 | 6.3 | 6.2 | 6.1 |

notes: ? Insufficient data for calculation.

New York State Department of Health
Event #2, 2015

Serum Zinc

The test materials for serum Zn were prepared from human serum obtained from Tennessee Blood Services, Inc. The units were tested by FDA approved methods and found to be Non-reactive for Anti-HIV-1/2, Anti-HCV 3.0 and HBsAg. The serum has also been found to be STS (RPR) Non-reactive and Negative for HIV-1 and HCV by PCR. Serum units were dispensed into acid-washed 500-mL polypropylene containers to make up five (5) serum pools. Each pool was spiked with a suite of additional trace elements including zinc as Zn²⁺ at various concentrations.

The Target Value assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E** Statistical methods for use in proficiency testing by interlaboratory comparisons. Values for serum zinc range from 597 µg/L (9.13 µmol/L) to 2827 µg/L (43.23 µmol/L).

Acceptable ranges for serum zinc are based on fixed criteria of ±15%, or ±15 µg/L below 100 µg/L. These criteria are consistent with those proposed by the OELM network of EQAS organizers (1) for trace elements in serum.

Discussion. Based on the above criteria, 89.3% of test results reported were judged as satisfactory, with four out of 28 participant laboratories (14.3%) reporting 2 or more of the 5 results outside the acceptable ranges.

1. A. Taylor, J. Angerer, J. Arnaud, F. Claeys, R.L. Jones, O. Mazarrasa, E. Mairiaux, A. Menditto, P.J. Parsons, M. Patriarca, A. Pineau, S. Valkonen, J.-P. Weber and C. Weykamp Accreditation and Quality Assurance 2006 11 440-445.

2. J. Arnaud, J.-P. Weber, C.W. Weykamp, P.J. Parsons, J. Angerer, E. Mairiaux, O. Mazarrasa, S. Valkonen, A. Menditto, M. Patriarca, and A. Taylor Clinical Chemistry 2008 54 1892-1899.

**New York State Department of Health
Serum Zinc Test Results, 2015 Event #2
ROBUST STATISTICAL SUMMARY**

TARGET VALUE ASSIGNMENT AND STATISTICS

Results ($\mu\text{g/L}$ serum)

| | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
|-------------------------------|-------------|-------------|-------------|------------|------------|
| Robust Mean | 2827 | 1657 | 2017 | 597 | 679 |
| Robust Standard Deviation | 245 | 125 | 152 | 55 | 63 |
| Standard Uncertainty | 58 | 30 | 36 | 13 | 15 |
| RSD (%) | 8.7 | 7.5 | 7.5 | 9.3 | 9.3 |
| Number of Sample Measurements | 28 | 28 | 28 | 28 | 28 |
| Acceptable Range: | | | | | |
| Upper Limit: | 3251 | 1906 | 2320 | 687 | 781 |
| Lower Limit: | 2403 | 1408 | 1714 | 507 | 577 |

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Serum Zinc Test Results, 2015 Event #2
PERFORMANCE OF PARTICIPATING LABORATORIES

| Lab Code | Method | Results (µg/L serum) | | | | | Info Only |
|----------|----------------|----------------------|---------|---------|---------|---------|-----------|
| | | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 | |
| | Target Values: | 2827 | 1657 | 2017 | 597 | 679 | |
| 107 | DRC/CC-ICP-MS | 3200 | 1900 | 2200 | 630 | 720 | Info |
| 110 | ICP-MS | 3200 | 1967 ↑ | 2321 ↑ | 689 ↑ | 795 ↑ | |
| 114 | ICP-MS | 2610 | 1610 | 1910 | 630 | 690 | |
| 147 | ICP-MS | 2614 | 1556 | 1902 | 601 | 620 | Info |
| 156 | ICP-AES/OES | 2600 | 1600 | 1900 | 570 | 640 | |
| 160 | ICP-MS | 2710 | 1580 | 1900 | 550 | 640 | |
| 164 | ICP-MS | 2795 | 1661 | 2007 | 580 | 668 | |
| 179 | DRC/CC-ICP-MS | 2780 | 1610 | 2000 | 560 | 640 | |
| 197 | ICP-MS | 2840 | 1590 | 1980 | 570 | 630 | |
| 200 | ICP-MS | 3041 | 1799 | 2217 | 634 | 739 | Info |
| 206 | ICP-MS | 2694 | 1581 | 2005 | 601 | 668 | |
| 287 | FAAS | 2980 | 1790 | 2180 | 700 ↑ | 730 | |
| 293 | ICP-MS | 2752 | 1654 | 2007 | 562 | 667 | Info |
| 305 | ICP-MS | 2280 ↓ | 1400 ↓ | 1740 | 560 | 610 | |
| 324 | HR-ICP-MS | 2306 ↓ | 1326 ↓ | 1647 ↓ | 474 ↓ | 560 ↓ | Info |
| 325 | ICP-MS | 3150 | 2000 ↑ | 2175 | 740 ↑ | 875 ↑ | Info |
| 355 | ICP-MS | 2881 | 1665 | 2038 | 591 | 671 | |
| 357 | ICP-MS | 2507 | 1486 | 1831 | 523 | 619 | |
| 358 | ICP-MS | 3110 | 1770 | 2170 | 640 | 740 | |
| 359 | ICP-MS | 2955 | 1691 | 2107 | 613 | 711 | |
| 363 | ICP-MS | 3020 | 1750 | 2150 | 600 | 700 | |
| 366 | FAAS | 2960 | 1760 | 2100 | 645 | 717 | Info |
| 388 | ICP-MS | 3101 | 1765 | 2083 | 652 | 768 | |
| 401 | DRC/CC-ICP-MS | 2844 | 1622 | 2027 | 569 | 654 | Info |
| 457 | ICP-AES/OES | 2664 | 1575 | 1897 | 573 | 640 | Info |
| 458 | FAAS | 2908 | 1773 | 2072 | 663 | 757 | |
| 483 | DRC/CC-ICP-MS | 2680 | 1562 | 1846 | 523 | 589 | Info |
| 484 | ICP-MS | 2680 | 1570 | 1940 | 530 | 670 | |

Percent satisfactory results for all participants: 89.3 %

notes: ↑ reported outside upper limit
↓ reported outside lower limit

Info only: results included for informational purposes only.

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

▲: Result not reported

**New York State Department of Health
Serum Zinc Test Results, 2015 Event #2
STATISTICAL SUMMARY BY METHOD**

| | Results ($\mu\text{g/L}$ serum) | | | | |
|--------------------------------|----------------------------------|---------|---------|---------|---------|
| | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| DRC/CC-ICP-MS | | | | | |
| Number of Sample Measurements: | 4 | 4 | 4 | 4 | 4 |
| Mean: | 2876 | 1674 | 2018 | 571 | 651 |
| Standard Deviation: | 226 | 153 | 145 | 44 | 54 |
| RSD (%): | 7.9 | 9.2 | 7.2 | 7.8 | 8.3 |
| FAAS | | | | | |
| Number of Sample Measurements: | 3 | 3 | 3 | 3 | 3 |
| Mean: | 2949 | 1774 | 2117 | 669 | 735 |
| Standard Deviation: | 37 | 15 | 56 | 28 | 20 |
| RSD (%): | — | — | — | — | — |
| HR-ICP-MS | | | | | |
| Number of Sample Measurements: | 1 | 1 | 1 | 1 | 1 |
| Mean: | 2306 | 1326 | 1647 | 474 | 560 |
| Standard Deviation: | ? | ? | ? | ? | ? |
| RSD (%): | — | — | — | — | — |
| ICP-AES/OES | | | | | |
| Number of Sample Measurements: | 2 | 2 | 2 | 2 | 2 |
| Mean: | 2632 | 1588 | 1899 | 572 | 640 |
| Standard Deviation: | 45 | 18 | 2 | 2 | 0 |
| RSD (%): | — | — | — | — | — |
| ICP-MS | | | | | |
| Number of Sample Measurements: | 18 | 18 | 18 | 18 | 18 |
| Mean: | 2830 | 1672 | 2027 | 604 | 693 |
| Standard Deviation: | 249 | 153 | 147 | 55 | 69 |
| RSD (%): | 8.8 | 9.1 | 7.2 | 9.1 | 10.0 |
| All Laboratories | | | | | |
| Number of Sample Measurements: | 28 | 28 | 28 | 28 | 28 |
| Mean: | 2817 | 1665 | 2013 | 599 | 683 |
| Standard Deviation: | 244 | 153 | 153 | 59 | 68 |
| RSD (%): | 8.7 | 9.2 | 7.6 | 9.9 | 9.9 |

notes: ? Insufficient data for calculation.

New York State Department of Health
Event #2, 2015

Additional Trace Elements Reported in Serum

Participant laboratories reported their analytical results for any additional trace elements (other than Al, Cu, Se and Zn) that are routinely reported so that a more complete characterization can be recorded for these PT materials. Results for additional trace elements are reported here, but no target value is implied nor are any acceptable ranges provided. These data are provided solely for educational and informational purposes.

In addition to Al, Cu, Se and Zn, the serum pools were supplemented with additional trace elements as indicated below.

Additional Elements

As, Cd, Cr, Co, Hg, Pb, Mn, Mo, Ni, Sn, Ti, Tl, W and V

New York State Department of Health
 Serum Additional Elements, 2015 Event #2
 Page 1

| Serum Antimony (µg/L) | | | | | | |
|-----------------------|---------------|---------|---------|---------|---------|---------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 103 | DRC/CC-ICP-MS | 0.0145 | <0.0101 | 0.0130 | 0.0132 | <0.0101 |
| 110 | ICP-MS | <0.2 | <0.2 | <0.2 | <0.2 | <0.2 |
| 147 | ICP-MS | <0.03 | <0.03 | <0.03 | <0.03 | <0.03 |

| Serum Arsenic (µg/L) | | | | | | |
|----------------------|---------------|---------|---------|---------|---------|---------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 103 | DRC/CC-ICP-MS | 4.82 | 16.90 | 6.25 | 4.47 | 11.4 |
| 147 | ICP-MS | 4.22 | 14.5 | 5.44 | 3.87 | 9.81 |
| 197 | DRC/CC-ICP-MS | <10 | 16 | <10 | <10 | 10 |
| Arithmetic mean | | - | 15.8 | - | - | 10.4 |
| SD | | - | 1.2 | - | - | 0.9 |
| n | | 2 | 3 | 2 | 2 | 3 |

| Serum Barium (µg/L) | | | | | | |
|---------------------|--------|---------|---------|---------|---------|---------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 110 | ICP-MS | 1.3 | 0.8 | 1.2 | 0.4 | 0.9 |
| 147 | ICP-MS | 0.924 | 0.614 | 0.777 | 0.313 | 0.814 |
| 197 | ICP-MS | <2.0 | <2.0 | <2.0 | <2.0 | <2.0 |

| Serum Beryllium (µg/L) | | | | | | |
|------------------------|--------|---------|---------|---------|---------|---------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 110 | ICP-MS | 0.7 | <0.3 | 0.33 | 0.30 | 0.7 |
| 147 | ICP-MS | <0.36 | <0.36 | <0.36 | <0.36 | <0.36 |
| 197 | ICP-MS | <0.2 | <0.2 | <0.2 | <0.2 | <0.2 |

| Serum Bismuth (µg/L) | | | | | | |
|----------------------|--------|---------|---------|---------|---------|---------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 197 | ICP-MS | <1.0 | <1.0 | <1.0 | <1.0 | <1.0 |

| Serum Cadmium (µg/L) | | | | | | |
|----------------------|---------------|---------|---------|---------|---------|---------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 103 | DRC/CC-ICP-MS | 1.13 | 2.93 | 2.88 | 5.30 | 10.7 |
| 110 | ICP-MS | 1.1 | 3.3 | 3.2 | 5.7 | 11.3 |
| 147 | ICP-MS | 1.00 | 2.87 | 2.80 | 5.20 | 10.5 |
| 197 | DRC/CC-ICP-MS | 1.1 | 3.0 | 2.9 | 5.4 | 10.6 |
| 324 | HR-ICP-MS | 0.95 | 2.6 | 2.6 | 4.7 | 9.6 |
| Arithmetic mean | | 1.06 | 2.9 | 2.9 | 5.3 | 10.5 |
| SD | | 0.08 | 0.3 | 0.2 | 0.4 | 0.6 |
| n | | 5 | 5 | 5 | 5 | 5 |

New York State Department of Health
 Serum Additional Elements, 2015 Event #2
 Page 2

| Serum Chromium (µg/L) | | | | | | |
|---------------------------------|---------------|-----------|------------|------------|------------|------------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 103 | DRC/CC-ICP-MS | 17.5 | 6.90 | 3.55 | 9.86 | 1.10 |
| 114 | ICP-MS | 16.8 | 7.0 | 3.1 | *1.4 | *9.7 |
| 147 | DRC/CC-ICP-MS | 16.1 | 6.50 | 2.75 | 8.84 | 1.14 |
| 156 | DRC/CC-ICP-MS | 16 | 6.6 | 2.8 | 9.2 | <1 |
| 160 | ICP-MS | 16 | 6 | 3 | 8 | 1 |
| 164 | DRC/CC-ICP-MS | 14.9 | 5.5 | 2.5 | 8.0 | 0.8 |
| 179 | DRC/CC-ICP-MS | 18.1 | 7.3 | 3.1 | 10.1 | 1.1 |
| 197 | DRC/CC-ICP-MS | 17.4 | 7.0 | 3.0 | 9.5 | 1.3 |
| 206 | DRC/CC-ICP-MS | 16.3 | 6.7 | 2.9 | 9.0 | 1.1 |
| 305 | ICP-MS | 18.4 | 7.0 | 3.0 | 10.2 | 1.0 |
| 324 | HR-ICP-MS | 17.6382 | 6.5147 | 3.0527 | 9.3048 | 1.1236 |
| 485 | HR-ICP-MS | 17.02 | 6.90 | 3.09 | 9.22 | 1.08 |
| *Outlier Arithmetic mean | | 17 | 6.7 | 3.0 | 9.2 | 1.1 |
| SD | | 1 | 0.5 | 0.3 | 0.7 | 0.1 |
| n | | 12 | 12 | 12 | 11 | 10 |

| Serum Cobalt (µg/L) | | | | | | |
|------------------------|---------------|------------|------------|------------|-------------|-------------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 103 | DRC/CC-ICP-MS | 2.23 | 6.83 | 4.00 | 18.6 | 11.0 |
| 110 | ICP-MS | 2.5 | 7.3 | 4.3 | 18.6 | 11.1 |
| 114 | ICP-MS | 2.3 | 6.8 | 4.0 | 18.6 | 10.6 |
| 147 | ICP-MS | 2.11 | 6.48 | 3.83 | 17.7 | 10.2 |
| 156 | DRC/CC-ICP-MS | 2.2 | 7.1 | 4.2 | 18 | 11 |
| 164 | ICP-MS | 2.1 | 6.6 | 3.9 | 17.6 | 9.9 |
| 179 | DRC/CC-ICP-MS | 2.2 | 6.8 | 4.2 | 18 | 11 |
| 197 | ICP-MS | 1.9 | 6.3 | 3.6 | 16.8 | 10.1 |
| 206 | ICP-MS | 2.5 | 6.9 | 4.2 | 18.2 | 11.1 |
| 324 | HR-ICP-MS | 2.5 | 7.1 | 4.3 | 18.9 | 11.3 |
| 485 | HR-ICP-MS | 2.15 | 7.34 | 4.28 | 18.08 | 11.07 |
| Arithmetic mean | | 2.2 | 6.9 | 4.1 | 18.1 | 10.8 |
| SD | | 0.2 | 0.3 | 0.2 | 0.6 | 0.5 |
| n | | 11 | 11 | 11 | 11 | 11 |

| Serum Iodine (µg/L) | | | | | | |
|------------------------|---------------|-----------|-----------|-----------|-----------|-----------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 114 | ICP-MS | 56.3 | 59.1 | 70.2 | 68.5 | 78.9 |
| 147 | ICP-MS | 53.2 | 56.8 | 68.1 | 75.1 | 74.1 |
| 156 | DRC/CC-ICP-MS | 54 | 61 | 73 | 78 | 79 |
| 164 | ICP-MS | 50 | 55 | 67 | 71 | 72 |
| 179 | ICP-MS | 57 | 61 | 73 | 79 | 80 |
| 197 | ICP-MS | 51.8 | 54.1 | 65.7 | 71.3 | 73.1 |
| 206 | ICP-MS | 55.6 | 59.8 | 69.0 | 78.3 | 77.3 |
| Arithmetic mean | | 54 | 58 | 69 | 74 | 76 |
| SD | | 3 | 3 | 3 | 4 | 3 |
| n | | 7 | 7 | 7 | 7 | 7 |

New York State Department of Health
Serum Additional Elements, 2015 Event #2
Page 3

| Serum Iron (µg/L) | | | | | | |
|--------------------------|-------------|---------|---------|---------|---------|---------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 457 | ICP-AES/OES | 739 | 693 | 487 | 1650 | 215 |

| Serum Lead (µg/L) | | | | | | |
|--------------------------|------------------------|------------|------------|-------------|-------------|------------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 103 | DRC/CC-ICP-MS | 5.48 | 8.80 | 2.18 | 14.2 | 2.11 |
| 110 | ICP-MS | 5.78 | 9.56 | 2.22 | 15.40 | 2.55 |
| 147 | ICP-MS | 5.99 | 9.10 | 2.18 | 15.2 | 2.26 |
| 197 | DRC/CC-ICP-MS | 5.2 | 8.6 | *2.6 | 13.9 | 2.0 |
| <i>*Outlier</i> | Arithmetic mean | 5.6 | 9.0 | 2.19 | 14.7 | 2.2 |
| | SD | 0.3 | 0.4 | 0.02 | 0.7 | 0.2 |
| | n | 4 | 4 | 3 | 4 | 4 |

| Serum Lithium (µg/L) | | | | | | |
|-----------------------------|--------|---------|---------|---------|---------|---------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 147 | ICP-MS | 0.694 | 0.467 | 0.472 | 0.393 | 0.325 |

| Serum Manganese (µg/L) | | | | | | |
|-------------------------------|------------------------|-------------|------------|------------|-------------|-------------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 103 | DRC/CC-ICP-MS | 12.8 | 6.93 | 4.56 | 15.0 | 22.0 |
| 114 | ICP-MS | 13.2 | 7.4 | 4.7 | 15.2 | 21.2 |
| 147 | ICP-MS | 12.0 | 6.54 | 4.27 | 14.2 | 20.4 |
| 179 | DRC/CC-ICP-MS | 13.1 | 7.4 | 4.9 | 15.3 | 22.2 |
| 197 | DRC/CC-ICP-MS | 12.8 | 6.8 | 4.4 | 14.8 | 21.7 |
| 206 | ICP-MS | 12.8 | 7.2 | 4.4 | *13.3 | 20.9 |
| 305 | ICP-MS | 12.6 | 7.1 | 4.6 | 15.1 | 21.3 |
| 324 | ICP-MS | 12.7517 | 6.7826 | 4.5232 | 14.8123 | 22.4821 |
| <i>*Outlier</i> | Arithmetic mean | 12.8 | 7.0 | 4.5 | 14.9 | 21.5 |
| | SD | 0.4 | 0.3 | 0.2 | 0.4 | 0.7 |
| | n | 8 | 8 | 8 | 7 | 8 |

| Serum Mercury (µg/L) | | | | | | |
|-----------------------------|------------------------|---------|---------|------------|-------------|------------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 103 | DRC/CC-ICP-MS | 3.29 | 4.55 | 5.77 | 15.0 | 9.08 |
| 147 | ICP-MS | 3.07 | 4.09 | 5.40 | 14.9 | 8.73 |
| 197 | ICP-MS | <5 | <5 | 7 | 18 | 10 |
| | Arithmetic mean | - | - | 6.1 | 16.0 | 9.3 |
| | SD | - | - | 0.8 | 1.8 | 0.7 |
| | n | 2 | 2 | 3 | 3 | 3 |

New York State Department of Health
 Serum Additional Elements, 2015 Event #2
 Page 4

| Serum Molybdenum (µg/L) | | | | | | |
|--------------------------------|------------------------|----------------|----------------|----------------|----------------|----------------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 103 | DRC/CC-ICP-MS | 3.60 | 16.8 | 5.67 | 12.2 | 8.03 |
| 110 | ICP-MS | 3.7 | 16.7 | 5.7 | 12.3 | 8.5 |
| 147 | ICP-MS | 3.45 | 15.9 | 5.36 | 11.4 | 7.34 |
| 179 | ICP-MS | 3.5 | 17 | 5.9 | 12 | 8.2 |
| 197 | ICP-MS | 3.6 | 18.5 | 5.8 | 12.8 | 8.5 |
| 485 | HR-ICP-MS | 3.44 | 16.49 | 5.84 | 11.41 | 8.12 |
| | Arithmetic mean | 3.5 | 16.9 | 5.7 | 12.0 | 8.1 |
| | SD | 0.1 | 0.9 | 0.2 | 0.5 | 0.4 |
| | n | 6 | 6 | 6 | 6 | 6 |

| Serum Nickel (µg/L) | | | | | | |
|----------------------------|------------------------|----------------|----------------|----------------|----------------|----------------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 114 | ICP-MS | 4.3 | 13.8 | 5.5 | 19.8 | 8.1 |
| 147 | ICP-MS | 1.99 | 11.9 | 3.50 | 17.7 | 6.58 |
| 179 | DRC/CC-ICP-MS | 2.5 | 13 | 4.0 | 19 | 7.0 |
| 197 | ICP-MS | 2.9 | 13.8 | 4.0 | 19.4 | 7.8 |
| 206 | ICP-MS | <10.0 | 12.9 | <10.0 | 17.8 | <10.0 |
| 485 | HR-ICP-MS | 2.27 | 12.55 | 4.08 | 18.08 | 7.65 |
| | Arithmetic mean | 2.8 | 13.0 | 4.2 | 18.6 | 7.4 |
| | SD | 0.9 | 0.7 | 0.8 | 0.9 | 0.6 |
| | n | 5 | 6 | 5 | 6 | 5 |

| Serum Platinum (µg/L) | | | | | | |
|------------------------------|---------------|----------------|----------------|----------------|----------------|----------------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 110 | ICP-MS | <0.6 | <0.6 | <0.6 | <0.6 | <0.6 |
| 179 | ICP-MS | <10 | <10 | <10 | <10 | <10 |

| Serum Silver (µg/L) | | | | | | |
|----------------------------|---------------|----------------|----------------|----------------|----------------|----------------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 147 | ICP-MS | 0.247 | 0.114 | 0.0947 | 0.288 | 0.169 |
| 179 | ICP-MS | <0.2 | <0.2 | <0.2 | 0.2 | <0.2 |
| 197 | ICP-MS | <1.0 | <1.0 | <1.0 | <1.0 | <1.0 |

| Serum Strontium (µg/L) | | | | | | |
|-------------------------------|---------------|----------------|----------------|----------------|----------------|----------------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 103 | DRC/CC-ICP-MS | 34.1 | 34.6 | 29.0 | 26.3 | 20.7 |

| Serum Tellurium (µg/L) | | | | | | |
|-------------------------------|---------------|----------------|----------------|----------------|----------------|----------------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 147 | ICP-MS | <0.077 | <0.077 | <0.077 | <0.077 | <0.077 |

New York State Department of Health
Serum Additional Elements, 2015 Event #2
Page 5

| Serum Thallium (µg/L) | | | | | | |
|------------------------|---------------|------------|-------------|------------|------------|------------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 103 | DRC/CC-ICP-MS | 1.06 | 11.0 | 1.82 | 1.71 | 5.51 |
| 110 | ICP-MS | 1.2 | 12.0 | 2.0 | 1.9 | 6.0 |
| 147 | ICP-MS | 0.848 | 10.1 | 1.73 | 1.47 | 5.09 |
| 197 | ICP-MS | 1.0 | 10.9 | 1.8 | 1.6 | 5.4 |
| 324 | HR-ICP-MS | 1.0979 | 11.7492 | 1.9617 | 1.8216 | 6.0017 |
| Arithmetic mean | | 1.0 | 11.1 | 1.9 | 1.7 | 5.6 |
| SD | | 0.1 | 0.8 | 0.1 | 0.2 | 0.4 |
| n | | 5 | 5 | 5 | 5 | 5 |

| Serum Thorium (µg/L) | | | | | | |
|----------------------|--------|---------|---------|---------|---------|---------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 147 | ICP-MS | <0.0070 | <0.0070 | <0.0070 | <0.0070 | <0.0070 |

| Serum Tin (µg/L) | | | | | | |
|------------------------|--------|-------------|----------|-------------|------------|----------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 110 | ICP-MS | 13.4 | 2.1 | 10.5 | 5.5 | 3.4 |
| 147 | ICP-MS | 12.1 | 1.94 | 9.62 | 5.02 | 2.98 |
| 197 | ICP-MS | 13.6 | <5.0 | 10.3 | 5.6 | <5.0 |
| Arithmetic mean | | 13.0 | - | 10.1 | 5.4 | - |
| SD | | 0.8 | - | 0.5 | 0.3 | - |
| n | | 3 | 2 | 3 | 3 | 2 |

| Serum Titanium (µg/L) | | | | | | |
|-----------------------|-----------|---------|---------|---------|---------|---------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 485 | HR-ICP-MS | 17.02 | 2.41 | 8.52 | 1.84 | 4.23 |

| Serum Tungsten (µg/L) | | | | | | |
|-----------------------|--------|---------|---------|---------|---------|---------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 147 | ICP-MS | 2.00 | 10.2 | 2.85 | 7.76 | 1.97 |

| Serum Uranium (µg/L) | | | | | | |
|----------------------|---------------|---------|---------|---------|---------|---------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 103 | DRC/CC-ICP-MS | 0.00231 | <0.0012 | <0.0012 | <0.0012 | 0.00138 |
| 110 | ICP-MS | <0.05 | <0.05 | <0.05 | <0.05 | <0.05 |
| 147 | ICP-MS | <0.0071 | <0.0071 | <0.0071 | <0.0071 | <0.0071 |

| Serum Vanadium (µg/L) | | | | | | |
|------------------------|---------------|-------------|-------------|------------|------------|------------|
| Lab Code | Method | SE15-06 | SE15-07 | SE15-08 | SE15-09 | SE15-10 |
| 147 | DRC/CC-ICP-MS | 14.5 | 9.13 | 2.65 | 1.68 | 4.37 |
| 179 | DRC/CC-ICP-MS | 17.4 | 11.3 | 3.5 | 2.2 | 5.3 |
| 485 | HR-ICP-MS | 15.00 | 9.93 | 3.15 | 1.76 | 4.92 |
| Arithmetic mean | | 15.6 | 10.1 | 3.1 | 1.9 | 4.9 |
| SD | | 1.6 | 1.1 | 0.4 | 0.3 | 0.5 |
| n | | 3 | 3 | 3 | 3 | 3 |

**New York State Department of Health
Trace Elements in Serum
METHOD NOTES**

ATOMIC SPECTROMETRY METHODS

- A-1 ETAAS-Z (Electrothermal atomic absorption spectrometry with Zeeman background correction)
- A-2 ETAAS Other (i.e., D₂, S-H background correction)
- A-3 FAAS (Flame atomic absorption spectrometry)
- A-4 CV-AAS (Cold vapor atomic absorption spectrometry)
- A-5 HG-AAS (Hydride generation atomic absorption spectrometry)
- A-6 AFS (Atomic fluorescence spectrometry)

INDUCTIVELY COUPLED PLASMA

- P-1 ICP-MS (Inductively coupled plasma - mass spectrometry)
- P-2 DRC/CC-ICP-MS (ICP-MS used in the Dynamic Reaction Cell or Collision Cell mode)
- P-3 ICP-AES/OES (ICP atomic/optical emission spectrometry)
- P-4 HR-ICP-MS (High resolution ICP-MS)
- P-5 ETV-ICP-MS (Electrothermal vaporization ICP-MS)
- P-6 ID-ICP-MS (Isotope dilution ICP-MS)

ELECTROCHEMICAL METHODS

- E-1 ASV (Anodic stripping voltammetry without digestion)
- E-2 ASV-LeadCare® Blood Lead Testing System
- E-5 ASV-LeadCare® II Blood Lead Testing System
- E-6 ASV-LeadCare® Ultra™ Blood Lead Testing System
- E-3 Fluoride specific electrode

MOLECULAR FLUORIMETRY

- F-1 EtOAc (Ethyl acetate-acetic acid extraction method for determination of erythrocyte protoporphyrin)
- F-2 Aviv hematofluorometry (for determination of EP at hematocrit 35)
- F-3 Helena ZPP (for determination of zinc protoporphyrin in $\mu\text{mol ZPP/mol heme}$)

OTHER METHODS

If your method is not listed in the above list, please describe it briefly.
