

### Clinical Chemistry Proficiency Testing – May 6, 2013

Enclosed are results from the clinical chemistry proficiency survey shipped May 6, 2013. Test samples were prepared in-house by the quantitative transfer of constituents to pooled human serum. The material was subsequently sterile filtered, dispensed into aliquots and stored frozen at -80 °C. Five specimens (**C86, C87, C88, C89, C90**) were distributed to each participant for analysis.

Outlined below is a description of the process utilized in the evaluation of your laboratory's proficiency test results. A summary of your laboratory's performance for the three most recent surveys is also included with your report.

**Target Value:** In general, targets utilized are derived from all-participant mean values calculated by robust statistical technique. In some cases, however, it is recognized that method, reagent, and/or instrument specific targets may be required and "peer group" specific targets are used where appropriate. Should an alternate target be required to evaluate your laboratory's data, an asterisk will be placed adjacent to the method, reagent, and/or instrument listing(s) corresponding to the peer group utilized.

**Acceptable Range:** Represents limits established using criteria specified by CLIA '88 regulations, allowing for rounding to appropriate significant digits. Results falling within this range are scored as 100%. Any result exceeding these limits is considered unsatisfactory and receives a score of 0%. Laboratories must achieve an overall analyte score  $\geq 80\%$  in order to meet performance criteria for that analyte.

**Range Plots:** Plots relative distance of your laboratory's result (represented by an "x") from the target for each sample analyzed. Any result exceeding the high or low limit by  $>20\%$  of the acceptable range is indicated by a pound sign (#).

**Not Gradable:** Results for graded analytes for a few laboratories using unique instrument, reagent, or instrument/reagent combinations were considered "not gradable". For these laboratories, pass credit (100%) has been issued. **If a laboratory is unable to participate in the NYS clinical chemistry proficiency test event as a graded participant for any analyte, it is the responsibility of that laboratory to establish alternate means to verify the accuracy and precision of the test system for any non-graded analyte(s).**

**Troponin I, Troponin T, and Estimated Glomerular Filtration Rate:** These analytes were included in the May 2013 test event for evaluation only. Although results were not graded, targets and acceptable ranges appropriate for your laboratory's methodology are provided.

The attached statistical report provides a summary of participant data for the five survey specimens. Results for individual instrument and reagent systems where the number of laboratories using those systems is three or greater are provided. Mean and Standard Deviation (1 SD) values shown are calculated by a robust statistical technique that does not assume a Gaussian distribution. These statistical reports are also available on the internet at: <http://www.wadsworth.org/chemheme>

Should you have any questions regarding these reports or wish to obtain an additional copy, please contact the Clinical Chemistry Section at (518) 474-5582.

Note: The use of brand and/or trade names in this report does not constitute an endorsement of the products on the part of the Wadsworth Center or the New York State Department of Health

Summary of Participant Performance (Mean and Standard Deviation)

Glucose (mg/dL)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
205.1 ± 5.57	65.4 ± 1.87	43.2 ± 2.66	306.9 ± 7.13	173.9 ± 4.37	n = 387	[---] All Methods & Instruments
<Instruments>						
208.8 ± 1.27	72.0 ± 0.75	49.5 ± 1.22	310.4 ± 1.64	177.7 ± 1.58	n = 4	[AXA] Abaxis Piccolo
207.5 ± 2.36	64.5 ± 0.80	41.4 ± 0.59	314.2 ± 4.25	174.2 ± 1.93	n = 22	[ABJ] Abbott Architect c System
204.4 ± 4.50	64.6 ± 1.52	41.7 ± 1.11	301.4 ± 5.68	171.6 ± 3.99	n = 64	[OLC] Beckman Coulter AU Chemistry System
203.9 ± 2.86	64.6 ± 1.02	42.2 ± 2.36	302.7 ± 3.37	173.5 ± 3.63	n = 3	[BCX] Beckman Coulter LX-20
205.4 ± 4.04	64.7 ± 2.24	42.6 ± 2.78	306.6 ± 6.38	173.6 ± 4.06	n = 17	[BCG] Beckman Coulter UniCel DxC 600
205.7 ± 3.32	63.5 ± 2.15	42.6 ± 1.81	307.9 ± 5.75	173.2 ± 2.82	n = 8	[BCH] Beckman Coulter UniCel DxC 800
248.6 ± 1.02	97.0 ± 1.80	66.3 ± 3.16	362.8 ± 4.11	202.5 ± 1.86	n = 3	[HEC] HemoCue Glucose 201
208.1 ± 3.40	65.2 ± 0.73	38.2 ± 1.00	309.9 ± 4.88	169.2 ± 2.31	n = 6	[IAA] i-STAT
198.8 ± 5.14	66.6 ± 2.41	42.8 ± 1.54	314.0 ± 7.86	174.2 ± 4.83	n = 8	[JJE] Ortho Vitros 250/350/950
194.9 ± 3.37	65.2 ± 1.26	42.0 ± 1.08	310.0 ± 2.96	172.0 ± 2.75	n = 19	[JJF] Ortho Vitros 5,1FS
195.6 ± 3.62	65.2 ± 0.97	42.4 ± 1.03	310.3 ± 5.38	172.4 ± 2.94	n = 19	[JJG] Ortho Vitros 5600
205.8 ± 2.36	65.0 ± 0.90	42.3 ± 0.51	303.9 ± 4.38	173.1 ± 2.86	n = 3	[ROK] Roche cobas c111
206.9 ± 2.42	66.2 ± 0.78	42.5 ± 0.75	308.3 ± 4.33	174.7 ± 2.23	n = 20	[ROC] Roche cobas c501
204.7 ± 3.39	64.8 ± 1.66	41.7 ± 1.38	305.9 ± 5.43	170.3 ± 0.90	n = 5	[ROH] Roche cobas c701
206.6 ± 3.09	65.3 ± 0.72	42.0 ± 0.00	305.7 ± 3.71	173.2 ± 2.57	n = 6	[ROS] Roche Cobas INTEGRA 400
206.0 ± 6.31	64.3 ± 1.37	41.3 ± 1.37	302.5 ± 5.40	170.5 ± 3.63	n = 3	[ROT] Roche Cobas INTEGRA 800
206.6 ± 3.54	65.3 ± 1.40	42.6 ± 1.03	306.5 ± 5.69	174.6 ± 2.91	n = 32	[ROD] Roche MODULAR D/P
203.8 ± 3.55	64.7 ± 1.26	42.2 ± 0.95	302.6 ± 5.86	171.5 ± 2.83	n = 22	[BYE] Siemens ADVIA 1800
206.8 ± 2.36	66.0 ± 0.90	43.0 ± 0.90	308.7 ± 0.51	174.8 ± 1.54	n = 3	[BYB] Siemens ADVIA 2400
208.8 ± 2.97	67.5 ± 1.65	47.4 ± 1.39	311.4 ± 5.16	178.8 ± 3.68	n = 21	[DUE] Siemens Dimension EXL
209.6 ± 3.55	67.8 ± 1.68	47.9 ± 1.82	310.9 ± 6.69	179.7 ± 3.69	n = 22	[DUR] Siemens Dimension RxL
202.8 ± 5.62	65.4 ± 2.03	45.3 ± 1.35	301.9 ± 7.43	173.5 ± 5.14	n = 41	[DUT] Siemens Dimension Vista
208.2 ± 4.81	67.5 ± 1.18	47.5 ± 0.99	309.0 ± 6.75	177.6 ± 3.21	n = 16	[DUX] Siemens Dimension Xpand
<Reagents>						
208.8 ± 1.27	72.0 ± 0.75	49.5 ± 1.22	310.4 ± 1.64	177.7 ± 1.58	n = 4	[AX1] Abaxis
207.5 ± 2.36	64.5 ± 0.80	41.4 ± 0.59	314.2 ± 4.25	174.2 ± 1.93	n = 22	[AB1] Abbott
205.2 ± 3.91	64.2 ± 2.10	42.4 ± 2.53	305.9 ± 6.65	173.3 ± 3.83	n = 30	[BC1] Beckman Coulter
204.4 ± 4.07	64.6 ± 1.41	41.7 ± 0.97	301.4 ± 5.43	171.6 ± 3.85	n = 59	[OL1] Beckman Coulter AU Series
204.8 ± 4.11	65.6 ± 1.02	46.0 ± 0.00	302.5 ± 6.32	172.5 ± 3.63	n = 3	[CR1] Carolina
243.1 ± 8.68	91.0 ± 8.91	61.8 ± 6.65	358.2 ± 7.25	196.6 ± 9.16	n = 5	[HE1] HemoCue
208.1 ± 3.40	65.2 ± 0.73	38.2 ± 1.00	309.9 ± 4.88	169.2 ± 2.31	n = 6	[IA1] i-STAT
195.9 ± 3.97	65.3 ± 1.39	42.3 ± 1.16	310.9 ± 5.02	172.5 ± 3.28	n = 49	[JJ1] Ortho Clinical Diagnostics
205.8 ± 2.36	65.0 ± 0.90	42.3 ± 0.51	303.9 ± 4.38	173.1 ± 2.86	n = 3	[RO8] Roche cobas c111
206.7 ± 2.94	66.0 ± 1.09	42.4 ± 0.89	308.0 ± 4.87	174.3 ± 3.01	n = 27	[RO4] Roche cobas c311/c501/c502/c701/c702
206.7 ± 3.46	65.3 ± 1.37	42.6 ± 1.01	306.5 ± 5.56	174.6 ± 2.87	n = 33	[RO2] Roche Hitachi and Modular D/P
206.4 ± 4.27	65.1 ± 1.10	42.0 ± 0.82	304.6 ± 4.68	172.5 ± 3.18	n = 9	[RO1] Roche Integra and MIRA
204.3 ± 3.46	64.9 ± 1.29	42.3 ± 0.95	303.9 ± 5.90	172.1 ± 2.98	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
206.9 ± 5.50	66.8 ± 2.09	46.7 ± 1.88	307.4 ± 7.85	177.0 ± 5.14	n = 100	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Urea Nitrogen (mg/dL)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
12.3 ± 0.82	22.0 ± 0.98	39.0 ± 2.02	18.5 ± 1.08	26.9 ± 1.60	n = 369	[---] All Methods & Instruments
<Instruments>						
12.0 ± 0.00	20.7 ± 0.90	37.0 ± 0.75	17.8 ± 0.41	25.0 ± 1.14	n = 4	[AXA] Abaxis Piccolo
12.3 ± 0.54	22.0 ± 0.00	39.9 ± 0.52	18.8 ± 0.55	27.3 ± 0.66	n = 21	[ABJ] Abbott Architect c System
12.6 ± 0.60	22.2 ± 0.77	39.2 ± 1.26	18.6 ± 0.67	27.1 ± 0.87	n = 61	[OLC] Beckman Coulter AU Chemistry System
11.0 ± 0.90	19.7 ± 0.51	37.2 ± 1.54	16.0 ± 0.90	24.3 ± 1.37	n = 3	[BCX] Beckman Coulter LX-20
13.0 ± 0.00	23.0 ± 0.00	40.0 ± 0.90	19.7 ± 0.50	27.9 ± 0.65	n = 17	[BCG] Beckman Coulter UniCel DxC 600
11.8 ± 1.29	20.1 ± 0.60	36.9 ± 2.37	16.7 ± 1.49	24.7 ± 2.00	n = 8	[BCH] Beckman Coulter UniCel DxC 800
12.5 ± 0.57	23.8 ± 0.73	45.8 ± 1.45	21.3 ± 0.51	31.2 ± 0.73	n = 6	[IAA] i-STAT
11.0 ± 0.00	21.4 ± 0.56	34.3 ± 0.85	17.0 ± 0.00	23.6 ± 0.56	n = 9	[JJE] Ortho Vitros 250/350/950
11.0 ± 0.00	21.0 ± 0.00	34.2 ± 0.51	16.8 ± 0.43	23.2 ± 0.59	n = 19	[JJF] Ortho Vitros 5,1FS
11.0 ± 0.00	20.9 ± 0.60	33.7 ± 0.87	16.4 ± 0.61	23.2 ± 0.51	n = 19	[JJG] Ortho Vitros 5600
12.0 ± 0.00	21.3 ± 0.51	38.3 ± 0.51	18.3 ± 0.51	26.7 ± 0.51	n = 3	[ROK] Roche cobas c111
12.2 ± 0.56	21.8 ± 0.76	38.9 ± 0.98	18.3 ± 0.60	26.6 ± 0.77	n = 19	[ROC] Roche cobas c501
12.0 ± 0.00	21.4 ± 0.55	38.6 ± 0.55	18.4 ± 0.55	26.4 ± 0.55	n = 5	[ROH] Roche cobas c701
12.8 ± 0.41	21.8 ± 0.41	39.8 ± 0.41	18.8 ± 0.41	27.0 ± 0.75	n = 4	[ROS] Roche Cobas INTEGRA 400
12.0 ± 0.00	21.3 ± 0.51	38.3 ± 0.51	18.0 ± 0.90	26.7 ± 0.51	n = 3	[ROT] Roche Cobas INTEGRA 800
12.6 ± 0.63	22.2 ± 0.75	39.5 ± 0.83	18.9 ± 0.55	27.1 ± 0.69	n = 31	[ROD] Roche MODULAR D/P
12.8 ± 0.61	22.6 ± 0.72	39.6 ± 0.78	18.8 ± 0.60	27.7 ± 0.83	n = 22	[BYE] Siemens ADVIA 1800
13.3 ± 0.51	23.3 ± 0.51	41.7 ± 1.37	19.7 ± 0.51	28.3 ± 0.51	n = 3	[BYB] Siemens ADVIA 2400
12.6 ± 0.61	22.7 ± 0.65	39.9 ± 0.98	18.8 ± 0.53	27.5 ± 0.68	n = 21	[DUE] Siemens Dimension EXL
12.4 ± 0.66	22.3 ± 0.69	39.4 ± 1.07	18.8 ± 0.55	27.6 ± 0.88	n = 21	[DUR] Siemens Dimension RxL
12.2 ± 0.58	21.9 ± 0.82	39.2 ± 1.31	18.6 ± 0.70	27.0 ± 1.05	n = 41	[DUT] Siemens Dimension Vista
12.6 ± 0.83	22.4 ± 1.08	40.1 ± 1.11	18.9 ± 1.01	27.2 ± 1.20	n = 15	[DUX] Siemens Dimension Xpand
<Reagents>						
12.0 ± 0.00	20.7 ± 0.90	37.0 ± 0.75	17.8 ± 0.41	25.0 ± 1.14	n = 4	[AX1] Abaxis
12.3 ± 0.54	22.0 ± 0.00	39.9 ± 0.52	18.8 ± 0.55	27.3 ± 0.66	n = 21	[AB1] Abbott
12.8 ± 0.59	21.6 ± 1.62	39.0 ± 1.67	18.6 ± 1.97	27.0 ± 1.76	n = 30	[BC1] Beckman Coulter
12.7 ± 0.60	22.3 ± 0.75	39.3 ± 1.25	18.6 ± 0.65	27.2 ± 0.86	n = 57	[OL1] Beckman Coulter AU Series
12.5 ± 0.57	23.8 ± 0.73	45.8 ± 1.45	21.3 ± 0.51	31.2 ± 0.73	n = 6	[IA1] i-STAT
11.0 ± 0.00	21.1 ± 0.56	34.1 ± 0.81	16.7 ± 0.55	23.3 ± 0.59	n = 50	[JJ1] Ortho Clinical Diagnostics
12.0 ± 0.00	21.3 ± 0.51	38.3 ± 0.51	18.3 ± 0.51	26.7 ± 0.51	n = 3	[RO8] Roche cobas c111
12.2 ± 0.54	21.8 ± 0.71	38.9 ± 0.88	18.4 ± 0.60	26.6 ± 0.76	n = 26	[RO4] Roche cobas c311/c501/c502/c701/c702
12.6 ± 0.63	22.2 ± 0.76	39.5 ± 0.88	18.9 ± 0.52	27.1 ± 0.70	n = 32	[RO2] Roche Hitachi and Modular D/P
12.4 ± 0.56	21.6 ± 0.56	39.2 ± 0.92	18.5 ± 0.74	26.8 ± 0.66	n = 7	[RO1] Roche Integra and MIRA
12.8 ± 0.61	22.7 ± 0.73	39.7 ± 0.90	18.9 ± 0.64	27.8 ± 0.82	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
12.4 ± 0.66	22.3 ± 0.87	39.5 ± 1.25	18.7 ± 0.67	27.3 ± 0.99	n = 98	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Creatinine (mg/dL)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
1.84 ± 0.11	1.02 ± 0.08	4.03 ± 0.16	1.26 ± 0.18	2.79 ± 0.18	n = 374	[---] All Methods & Instruments
1.83 ± 0.10	1.00 ± 0.08	4.01 ± 0.17	1.25 ± 0.19	2.76 ± 0.17	n = 201	[---] All IDMS Traceable Methods
1.86 ± 0.11	1.04 ± 0.09	4.05 ± 0.15	1.28 ± 0.17	2.84 ± 0.18	n = 169	[---] All Non-IDMS Traceable Methods
1.87 ± 0.10	1.04 ± 0.08	4.04 ± 0.14	1.31 ± 0.15	2.85 ± 0.17	n = 137	[-G-] Alkaline picrate/Jaffe
1.82 ± 0.08	1.00 ± 0.08	3.97 ± 0.13	1.31 ± 0.19	2.75 ± 0.14	n = 140	[-H-] Alkaline picrate/Jaffe-IDMS calibration
1.84 ± 0.16	1.01 ± 0.10	4.13 ± 0.25	1.13 ± 0.16	2.78 ± 0.26	n = 32	[-I-] Enzymatic
1.85 ± 0.15	1.00 ± 0.06	4.12 ± 0.22	1.13 ± 0.12	2.81 ± 0.23	n = 61	[-J-] Enzymatic-IDMS-traceable calibration
						<Instruments>
1.75 ± 0.06	1.05 ± 0.12	3.97 ± 0.09	1.01 ± 0.31	2.47 ± 0.23	n = 4	[AXA] Abaxis Piccolo
1.91 ± 0.04	1.13 ± 0.04	4.19 ± 0.07	1.62 ± 0.06	2.93 ± 0.06	n = 21	[ABJ] Abbott Architect c System
1.81 ± 0.04	1.00 ± 0.02	3.94 ± 0.09	1.29 ± 0.03	2.72 ± 0.07	n = 63	[OLC] Beckman Coulter AU Chemistry System
1.77 ± 0.05	0.97 ± 0.05	4.00 ± 0.00	1.17 ± 0.05	2.70 ± 0.00	n = 3	[BCX] Beckman Coulter LX-20
1.74 ± 0.06	0.93 ± 0.06	3.89 ± 0.07	1.09 ± 0.06	2.59 ± 0.09	n = 18	[BCG] Beckman Coulter UniCel DxC 600
1.77 ± 0.05	0.99 ± 0.03	3.99 ± 0.04	1.20 ± 0.05	2.70 ± 0.05	n = 8	[BCH] Beckman Coulter UniCel DxC 800
1.98 ± 0.08	1.08 ± 0.08	4.68 ± 0.16	1.20 ± 0.00	2.92 ± 0.11	n = 5	[IAA] i-STAT
1.81 ± 0.20	0.90 ± 0.16	4.23 ± 0.21	1.05 ± 0.17	2.87 ± 0.22	n = 10	[JJE] Ortho Vitros 250/350/950
1.93 ± 0.05	1.02 ± 0.05	4.25 ± 0.11	1.20 ± 0.00	2.96 ± 0.07	n = 19	[JJF] Ortho Vitros 5,1FS
1.98 ± 0.04	1.04 ± 0.05	4.27 ± 0.10	1.20 ± 0.01	3.00 ± 0.04	n = 19	[JJG] Ortho Vitros 5600
1.74 ± 0.10	0.94 ± 0.04	3.86 ± 0.07	1.08 ± 0.09	2.56 ± 0.05	n = 3	[ROK] Roche cobas c111
1.74 ± 0.06	0.95 ± 0.07	3.94 ± 0.10	1.00 ± 0.01	2.59 ± 0.04	n = 21	[ROC] Roche cobas c501
1.82 ± 0.03	1.00 ± 0.00	3.95 ± 0.12	1.26 ± 0.18	2.76 ± 0.12	n = 5	[ROH] Roche cobas c701
1.80 ± 0.09	1.00 ± 0.09	3.93 ± 0.14	1.18 ± 0.11	2.58 ± 0.12	n = 5	[ROS] Roche Cobas INTEGRA 400
1.76 ± 0.05	0.95 ± 0.05	3.89 ± 0.09	1.08 ± 0.06	2.59 ± 0.08	n = 3	[ROT] Roche Cobas INTEGRA 800
1.86 ± 0.11	1.03 ± 0.06	4.05 ± 0.11	1.37 ± 0.23	2.81 ± 0.17	n = 31	[ROD] Roche MODULAR D/P
1.84 ± 0.07	1.03 ± 0.06	3.98 ± 0.08	1.43 ± 0.10	2.75 ± 0.09	n = 22	[BYE] Siemens ADVIA 1800
1.85 ± 0.06	1.06 ± 0.05	4.02 ± 0.12	1.49 ± 0.07	2.76 ± 0.14	n = 3	[BYB] Siemens ADVIA 2400
1.90 ± 0.09	1.05 ± 0.09	4.07 ± 0.12	1.31 ± 0.10	2.92 ± 0.12	n = 21	[DUE] Siemens Dimension EXL
1.89 ± 0.11	1.06 ± 0.11	4.05 ± 0.15	1.33 ± 0.16	2.94 ± 0.15	n = 21	[DUR] Siemens Dimension RxL
1.83 ± 0.11	0.99 ± 0.11	4.03 ± 0.13	1.24 ± 0.15	2.86 ± 0.17	n = 41	[DUT] Siemens Dimension Vista
1.86 ± 0.12	1.05 ± 0.10	4.01 ± 0.17	1.27 ± 0.16	2.85 ± 0.18	n = 15	[DUX] Siemens Dimension Xpand
						<Reagents>
1.75 ± 0.06	1.05 ± 0.12	3.97 ± 0.09	1.01 ± 0.31	2.47 ± 0.23	n = 4	[AX1] Abaxis
1.91 ± 0.04	1.13 ± 0.04	4.19 ± 0.07	1.62 ± 0.06	2.93 ± 0.06	n = 22	[AB1] Abbott
1.76 ± 0.06	0.96 ± 0.07	3.93 ± 0.08	1.14 ± 0.08	2.64 ± 0.09	n = 31	[BC1] Beckman Coulter
1.81 ± 0.04	1.00 ± 0.02	3.94 ± 0.08	1.29 ± 0.03	2.72 ± 0.07	n = 58	[OL1] Beckman Coulter AU Series
2.00 ± 0.08	1.07 ± 0.09	4.66 ± 0.18	1.22 ± 0.04	2.89 ± 0.11	n = 4	[IA1] i-STAT
1.95 ± 0.06	1.02 ± 0.05	4.26 ± 0.13	1.20 ± 0.01	2.99 ± 0.07	n = 50	[JJ1] Ortho Clinical Diagnostics
1.74 ± 0.10	0.94 ± 0.04	3.86 ± 0.07	1.08 ± 0.09	2.56 ± 0.05	n = 3	[RO8] Roche cobas c111
1.76 ± 0.07	0.96 ± 0.06	3.94 ± 0.11	1.07 ± 0.15	2.62 ± 0.10	n = 28	[RO4] Roche cobas c311/c501/c502/c701/c702
1.74 ± 0.04	0.97 ± 0.02	3.95 ± 0.08	1.02 ± 0.03	2.58 ± 0.06	n = 10	[RO2] Roche Hitachi and Modular D/P-Enzymatic
1.91 ± 0.07	1.06 ± 0.04	4.10 ± 0.04	1.49 ± 0.06	2.90 ± 0.06	n = 21	[RO2] Roche Hitachi and Modular D/P-Jaffe
1.78 ± 0.06	0.97 ± 0.06	3.91 ± 0.12	1.13 ± 0.09	2.59 ± 0.10	n = 8	[RO1] Roche Integra and MIRA
1.84 ± 0.07	1.04 ± 0.06	3.99 ± 0.09	1.45 ± 0.10	2.75 ± 0.11	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
1.69 ± 0.09	0.91 ± 0.06	3.90 ± 0.10	0.97 ± 0.03	2.50 ± 0.08	n = 9	[DA5] Siemens Dimension-Enzymatic
1.88 ± 0.10	1.04 ± 0.10	4.06 ± 0.13	1.30 ± 0.13	2.91 ± 0.14	n = 89	[DA5] Siemens Dimension-Jaffe

Summary of Participant Performance (Mean and Standard Deviation)

Estimated Glomerular Filtration Rate (mL/min/1.73 m<sup>2</sup>)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
49.3 ± 3.61	100.6 ± 12.01	20.0 ± 1.31	74.2 ± 15.58	30.5 ± 2.55	n = 286	[---] All Methods & Instruments
48.9 ± 3.36	98.4 ± 11.15	19.7 ± 1.24	71.1 ± 14.79	30.3 ± 2.45	n = 156	[-A-] IDMS-traceable MDRD Study Equation
49.5 ± 3.37	96.5 ± 11.93	20.4 ± 1.20	72.7 ± 14.02	30.4 ± 2.41	n = 91	[-B-] Original MDRD Study Equation (4-var)
55.2 ± 11.56	126.0 ± 6.84	22.9 ± 6.65	102.0 ± 0.00	35.1 ± 8.67	n = 6	[-D-] Cockcroft-Gault Equation
51.9 ± 3.52	108.0 ± 7.37	20.1 ± 1.33	83.5 ± 13.34	31.8 ± 2.62	n = 28	[-F-] CKD-EPI Equation
47.1 ± 4.88	108.0 ± 4.56	18.5 ± 1.94	69.5 ± 21.02	28.6 ± 3.79	n = 4	[-Z-] Other

Target values and allowable ranges for Estimated Glomerular Filtration Rate (eGFR):

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Method
49 (36-62)	99 (74-124)	20 (14-25)	76 (57-96)	31 (22-39)	IDMS-traceable MDRD Study Equation
51 (38-65)	100 (75-126)	21 (15-27)	79 (59-99)	31 (23-40)	Original MDRD Study Equation
51 (38-64)	106 (79-133)	20 (14-25)	81 (60-102)	31 (23-39)	CKD-EPI Equation
72 (53-90)	129 (97-162)	33 (24-41)	105 (78-131)	47 (35-60)	Cockcroft-Gault Equation

Laboratories were asked to report Estimated Glomerular Filtration Rate (eGFR) for samples C86-C90 for a 43-year-old African American man weighing 98 kg.

Target values for eGFR calculated by the MDRD Study Equations were derived from participant mean values for serum creatinine for both conventional and isotope dilution mass spectroscopy (IDMS) calibration methods and application of the appropriate 4-variable MDRD Study equation.

Target values for eGFR calculated by the CKD-EPI equation were derived from participant mean values for IDMS-traceable serum creatinine methods and application of the CKD-EPI equation.

Target values for eGFR calculated by the Cockcroft-Gault equation were derived from all-method mean values for serum creatinine and application of the Cockcroft-Gault formula.

Allowable ranges are ± 25% of the target eGFR for all samples/equations.

Note: the NKDEP recommends reporting estimated GFR values greater than or equal to 60 mL/min/1.73 m<sup>2</sup> as "> 60 mL/min/1.73 m<sup>2</sup>" and not as an exact number. However, ranges exceeding 60 mL/min are provided as a relative indicator of acceptability for laboratories that report numeric results above that threshold. Note that some laboratories reported results as > 60 mL/min/1.73 m<sup>2</sup>. These data were removed from the calculations of mean and standard deviation since their inclusion would have skewed results. Participant results for specimens C87 and C89 reported as > 60 mL/min/1.73 m<sup>2</sup> were considered acceptable performance.

## Summary of Participant Performance (Mean and Standard Deviation)

## Uric Acid (mg/dL)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
3.89 ± 0.20	5.30 ± 0.26	2.69 ± 0.19	9.07 ± 0.45	6.41 ± 0.29	n = 333	[---] All Methods & Instruments
<Instruments>						
3.87 ± 0.06	5.40 ± 0.08	2.61 ± 0.06	9.36 ± 0.09	6.53 ± 0.08	n = 21	[ABJ] Abbott Architect c System
4.29 ± 0.10	5.71 ± 0.13	2.89 ± 0.07	9.74 ± 0.19	6.87 ± 0.15	n = 59	[OLC] Beckman Coulter AU Chemistry System
3.83 ± 0.05	5.16 ± 0.10	2.63 ± 0.05	8.48 ± 0.15	6.13 ± 0.14	n = 3	[BCX] Beckman Coulter LX-20
3.73 ± 0.05	5.12 ± 0.07	2.50 ± 0.00	8.55 ± 0.08	6.08 ± 0.08	n = 16	[BCG] Beckman Coulter UniCel DxC 600
3.73 ± 0.09	5.12 ± 0.05	2.49 ± 0.06	8.52 ± 0.14	6.10 ± 0.12	n = 8	[BCH] Beckman Coulter UniCel DxC 800
3.77 ± 0.07	5.32 ± 0.09	2.52 ± 0.07	8.82 ± 0.21	6.23 ± 0.15	n = 6	[JJE] Ortho Vitros 250/350/950
3.78 ± 0.08	5.26 ± 0.08	2.55 ± 0.07	8.93 ± 0.15	6.30 ± 0.13	n = 18	[JJF] Ortho Vitros 5,1FS
3.71 ± 0.07	5.19 ± 0.11	2.48 ± 0.07	8.78 ± 0.15	6.21 ± 0.08	n = 19	[JJG] Ortho Vitros 5600
3.90 ± 0.10	5.40 ± 0.12	2.62 ± 0.07	9.27 ± 0.22	6.52 ± 0.19	n = 19	[ROC] Roche cobas c501
3.85 ± 0.12	5.27 ± 0.16	2.55 ± 0.06	9.02 ± 0.21	6.41 ± 0.11	n = 4	[ROH] Roche cobas c701
3.83 ± 0.05	5.20 ± 0.00	2.57 ± 0.05	9.03 ± 0.05	6.40 ± 0.00	n = 3	[ROT] Roche Cobas INTEGRA 800
3.82 ± 0.06	5.29 ± 0.08	2.56 ± 0.06	9.10 ± 0.17	6.40 ± 0.11	n = 30	[ROD] Roche MODULAR D/P
3.85 ± 0.09	5.33 ± 0.10	2.63 ± 0.07	9.13 ± 0.19	6.43 ± 0.14	n = 22	[BYE] Siemens ADVIA 1800
4.00 ± 0.09	5.57 ± 0.14	2.70 ± 0.09	9.52 ± 0.24	6.67 ± 0.14	n = 3	[BYB] Siemens ADVIA 2400
4.00 ± 0.11	5.30 ± 0.11	2.90 ± 0.08	9.15 ± 0.21	6.51 ± 0.18	n = 19	[DUE] Siemens Dimension EXL
3.92 ± 0.11	5.21 ± 0.11	2.87 ± 0.09	9.04 ± 0.21	6.45 ± 0.11	n = 18	[DUR] Siemens Dimension RxL
3.82 ± 0.09	4.91 ± 0.09	2.82 ± 0.07	8.54 ± 0.18	6.14 ± 0.11	n = 40	[DUT] Siemens Dimension Vista
3.94 ± 0.10	5.20 ± 0.10	2.91 ± 0.11	9.11 ± 0.15	6.44 ± 0.14	n = 10	[DUX] Siemens Dimension Xpand
<Reagents>						
3.87 ± 0.06	5.40 ± 0.08	2.61 ± 0.06	9.36 ± 0.09	6.53 ± 0.08	n = 21	[AB1] Abbott
3.74 ± 0.07	5.13 ± 0.07	2.50 ± 0.04	8.53 ± 0.11	6.08 ± 0.09	n = 29	[BC1] Beckman Coulter
4.29 ± 0.10	5.71 ± 0.12	2.89 ± 0.06	9.75 ± 0.18	6.88 ± 0.14	n = 57	[OL1] Beckman Coulter AU Series
3.74 ± 0.08	5.23 ± 0.11	2.51 ± 0.08	8.84 ± 0.18	6.24 ± 0.12	n = 45	[JJ1] Ortho Clinical Diagnostics
3.89 ± 0.11	5.39 ± 0.14	2.61 ± 0.08	9.23 ± 0.25	6.50 ± 0.19	n = 25	[RO4] Roche cobas c311/c501/c502/c701/c702
3.82 ± 0.06	5.28 ± 0.08	2.56 ± 0.05	9.10 ± 0.17	6.40 ± 0.11	n = 31	[RO2] Roche Hitachi and Modular D/P
3.80 ± 0.00	5.20 ± 0.00	2.52 ± 0.08	8.98 ± 0.08	6.35 ± 0.08	n = 5	[RO1] Roche Integra and MIRA
3.86 ± 0.11	5.34 ± 0.13	2.63 ± 0.08	9.15 ± 0.24	6.44 ± 0.17	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
3.90 ± 0.13	5.09 ± 0.21	2.86 ± 0.09	8.84 ± 0.36	6.31 ± 0.22	n = 87	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Bilirubin (mg/dL)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
1.84 ± 0.18	0.61 ± 0.12	1.03 ± 0.14	3.88 ± 0.31	2.47 ± 0.21	n = 357	[---] All Methods & Instruments
<Instruments>						
1.95 ± 0.06	0.75 ± 0.06	1.15 ± 0.06	4.00 ± 0.08	2.57 ± 0.09	n = 4	[AXA] Abaxis Piccolo
2.13 ± 0.16	0.70 ± 0.09	1.18 ± 0.10	4.23 ± 0.26	2.73 ± 0.20	n = 21	[ABJ] Abbott Architect c System
1.82 ± 0.07	0.70 ± 0.00	1.09 ± 0.04	3.58 ± 0.12	2.36 ± 0.09	n = 61	[OLC] Beckman Coulter AU Chemistry System
1.90 ± 0.18	0.70 ± 0.09	1.16 ± 0.10	3.83 ± 0.32	3.59 ± 1.80	n = 3	[BCX] Beckman Coulter LX-20
2.00 ± 0.09	0.75 ± 0.11	1.19 ± 0.07	4.18 ± 0.11	2.72 ± 0.13	n = 17	[BCG] Beckman Coulter UniCel DxC 600
1.94 ± 0.09	0.73 ± 0.07	1.17 ± 0.11	4.10 ± 0.17	2.65 ± 0.13	n = 7	[BCH] Beckman Coulter UniCel DxC 800
1.96 ± 0.15	0.65 ± 0.13	1.13 ± 0.11	3.82 ± 0.29	2.51 ± 0.20	n = 9	[JJE] Ortho Vitros 250/350/950
1.94 ± 0.09	0.66 ± 0.08	1.11 ± 0.07	4.11 ± 0.13	2.59 ± 0.11	n = 19	[JJF] Ortho Vitros 5,1FS
1.92 ± 0.16	0.62 ± 0.10	1.10 ± 0.10	4.10 ± 0.17	2.58 ± 0.16	n = 19	[JJG] Ortho Vitros 5600
1.56 ± 0.09	0.45 ± 0.06	0.83 ± 0.06	3.56 ± 0.19	2.20 ± 0.11	n = 18	[ROC] Roche cobas c501
1.46 ± 0.11	0.40 ± 0.00	0.80 ± 0.00	3.50 ± 0.23	2.08 ± 0.16	n = 5	[ROH] Roche cobas c701
1.54 ± 0.06	0.46 ± 0.06	0.80 ± 0.00	3.57 ± 0.14	2.16 ± 0.11	n = 5	[ROS] Roche Cobas INTEGRA 400
1.63 ± 0.05	0.47 ± 0.05	0.87 ± 0.05	3.57 ± 0.14	2.27 ± 0.05	n = 3	[ROT] Roche Cobas INTEGRA 800
1.65 ± 0.08	0.50 ± 0.00	0.90 ± 0.00	3.72 ± 0.13	2.32 ± 0.09	n = 31	[ROD] Roche MODULAR D/P
1.98 ± 0.06	0.63 ± 0.05	1.10 ± 0.00	4.33 ± 0.08	2.72 ± 0.08	n = 22	[BYE] Siemens ADVIA 1800
1.97 ± 0.05	0.60 ± 0.00	1.07 ± 0.05	4.25 ± 0.19	2.67 ± 0.05	n = 3	[BYB] Siemens ADVIA 2400
1.82 ± 0.10	0.55 ± 0.06	0.96 ± 0.07	3.94 ± 0.16	2.48 ± 0.11	n = 21	[DUE] Siemens Dimension EXL
1.78 ± 0.10	0.54 ± 0.07	0.94 ± 0.10	3.91 ± 0.19	2.43 ± 0.14	n = 21	[DUR] Siemens Dimension RxL
1.82 ± 0.06	0.60 ± 0.00	1.00 ± 0.00	3.92 ± 0.10	2.47 ± 0.07	n = 41	[DUT] Siemens Dimension Vista
1.78 ± 0.09	0.53 ± 0.06	0.95 ± 0.06	3.88 ± 0.19	2.44 ± 0.12	n = 15	[DUX] Siemens Dimension Xpand
<Reagents>						
1.95 ± 0.06	0.75 ± 0.06	1.15 ± 0.06	4.00 ± 0.08	2.57 ± 0.09	n = 4	[AX1] Abaxis
2.13 ± 0.16	0.70 ± 0.09	1.18 ± 0.10	4.23 ± 0.26	2.73 ± 0.20	n = 21	[AB1] Abbott
1.97 ± 0.12	0.74 ± 0.10	1.18 ± 0.08	4.14 ± 0.17	2.68 ± 0.16	n = 29	[BC1] Beckman Coulter
1.82 ± 0.07	0.70 ± 0.00	1.08 ± 0.04	3.58 ± 0.12	2.36 ± 0.09	n = 58	[OL1] Beckman Coulter AU Series
1.95 ± 0.13	0.65 ± 0.10	1.12 ± 0.09	4.07 ± 0.18	2.59 ± 0.15	n = 50	[JJ1] Ortho Clinical Diagnostics
1.54 ± 0.10	0.44 ± 0.06	0.82 ± 0.06	3.57 ± 0.20	2.19 ± 0.11	n = 25	[RO4] Roche cobas c311/c501/c502/c701/c702
1.65 ± 0.08	0.50 ± 0.00	0.90 ± 0.00	3.72 ± 0.13	2.32 ± 0.09	n = 31	[RO2] Roche Hitachi and Modular D/P
1.57 ± 0.07	0.47 ± 0.05	0.83 ± 0.05	3.57 ± 0.14	2.21 ± 0.10	n = 8	[RO1] Roche Integra and MIRA
1.98 ± 0.06	0.62 ± 0.05	1.11 ± 0.04	4.33 ± 0.09	2.72 ± 0.08	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
1.81 ± 0.08	0.56 ± 0.06	0.97 ± 0.07	3.92 ± 0.14	2.46 ± 0.10	n = 98	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Phosphorus (mg/dL)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
3.81 ± 0.20	2.78 ± 0.18	5.27 ± 0.27	4.23 ± 0.26	3.27 ± 0.21	n = 330	[---] All Methods & Instruments
<Instruments>						
3.83 ± 0.05	2.77 ± 0.07	5.27 ± 0.06	4.22 ± 0.06	3.27 ± 0.06	n = 20	[ABJ] Abbott Architect c System
3.69 ± 0.11	2.67 ± 0.09	5.09 ± 0.15	4.06 ± 0.11	3.15 ± 0.11	n = 58	[OLC] Beckman Coulter AU Chemistry System
3.83 ± 0.05	2.90 ± 0.09	5.57 ± 0.14	4.40 ± 0.09	3.44 ± 0.10	n = 3	[BCX] Beckman Coulter LX-20
3.93 ± 0.08	2.90 ± 0.00	5.38 ± 0.12	4.45 ± 0.11	3.37 ± 0.09	n = 17	[BCG] Beckman Coulter UniCel DxC 600
3.93 ± 0.08	2.88 ± 0.05	5.59 ± 0.10	4.39 ± 0.07	3.40 ± 0.00	n = 8	[BCH] Beckman Coulter UniCel DxC 800
4.27 ± 0.12	3.26 ± 0.07	5.92 ± 0.13	4.82 ± 0.10	3.82 ± 0.10	n = 8	[JJE] Ortho Vitros 250/350/950
4.18 ± 0.10	3.18 ± 0.09	5.83 ± 0.13	4.72 ± 0.11	3.72 ± 0.09	n = 18	[JJF] Ortho Vitros 5,1FS
4.16 ± 0.16	3.18 ± 0.13	5.85 ± 0.23	4.72 ± 0.16	3.74 ± 0.13	n = 19	[JJG] Ortho Vitros 5600
3.87 ± 0.09	2.83 ± 0.06	5.35 ± 0.10	4.32 ± 0.10	3.33 ± 0.10	n = 20	[ROC] Roche cobas c501
3.75 ± 0.08	2.74 ± 0.11	5.20 ± 0.09	4.25 ± 0.08	3.28 ± 0.08	n = 5	[ROH] Roche cobas c701
3.87 ± 0.05	2.80 ± 0.00	5.23 ± 0.05	4.27 ± 0.05	3.20 ± 0.09	n = 3	[ROS] Roche Cobas INTEGRA 400
3.87 ± 0.05	2.77 ± 0.05	5.26 ± 0.10	4.20 ± 0.09	3.27 ± 0.05	n = 3	[ROT] Roche Cobas INTEGRA 800
3.83 ± 0.14	2.78 ± 0.11	5.27 ± 0.17	4.25 ± 0.15	3.27 ± 0.11	n = 29	[ROD] Roche MODULAR D/P
3.83 ± 0.08	2.80 ± 0.05	5.29 ± 0.13	4.26 ± 0.10	3.28 ± 0.09	n = 22	[BYE] Siemens ADVIA 1800
3.80 ± 0.09	2.77 ± 0.05	5.20 ± 0.09	4.20 ± 0.09	3.27 ± 0.05	n = 3	[BYB] Siemens ADVIA 2400
3.73 ± 0.09	2.75 ± 0.06	5.19 ± 0.11	4.12 ± 0.08	3.16 ± 0.07	n = 18	[DUE] Siemens Dimension EXL
3.73 ± 0.12	2.77 ± 0.12	5.20 ± 0.18	4.17 ± 0.17	3.20 ± 0.15	n = 19	[DUR] Siemens Dimension RxL
3.59 ± 0.10	2.62 ± 0.10	5.02 ± 0.14	3.97 ± 0.11	3.05 ± 0.11	n = 40	[DUT] Siemens Dimension Vista
3.77 ± 0.06	2.76 ± 0.08	5.15 ± 0.12	4.11 ± 0.11	3.21 ± 0.05	n = 11	[DUX] Siemens Dimension Xpand
<Reagents>						
3.83 ± 0.05	2.77 ± 0.07	5.27 ± 0.06	4.22 ± 0.06	3.27 ± 0.06	n = 20	[AB1] Abbott
3.92 ± 0.09	2.90 ± 0.00	5.45 ± 0.17	4.42 ± 0.11	3.39 ± 0.08	n = 29	[BC1] Beckman Coulter
3.70 ± 0.11	2.67 ± 0.08	5.09 ± 0.14	4.06 ± 0.10	3.15 ± 0.10	n = 55	[OL1] Beckman Coulter AU Series
4.19 ± 0.13	3.19 ± 0.11	5.85 ± 0.17	4.74 ± 0.13	3.75 ± 0.11	n = 47	[JJ1] Ortho Clinical Diagnostics
3.85 ± 0.10	2.82 ± 0.07	5.33 ± 0.12	4.30 ± 0.10	3.32 ± 0.09	n = 26	[RO4] Roche cobas c311/c501/c502/c701/c702
3.83 ± 0.14	2.78 ± 0.11	5.27 ± 0.17	4.25 ± 0.15	3.27 ± 0.11	n = 29	[RO2] Roche Hitachi and Modular D/P
3.87 ± 0.05	2.80 ± 0.00	5.23 ± 0.07	4.24 ± 0.08	3.24 ± 0.08	n = 6	[RO1] Roche Integra and MIRA
3.82 ± 0.09	2.79 ± 0.06	5.27 ± 0.13	4.24 ± 0.10	3.27 ± 0.09	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
3.67 ± 0.13	2.70 ± 0.12	5.11 ± 0.16	4.06 ± 0.15	3.12 ± 0.13	n = 88	[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Calcium (mg/dL)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
6.97 ± 0.17	9.77 ± 0.26	8.39 ± 0.21	13.10 ± 0.36	9.18 ± 0.27	n = 365	[---] All Methods & Instruments
<Instruments>						
6.95 ± 0.22	9.75 ± 0.12	8.52 ± 0.13	12.97 ± 0.16	9.13 ± 0.23	n = 4	[AXA] Abaxis Piccolo
7.15 ± 0.08	10.11 ± 0.15	8.54 ± 0.08	13.61 ± 0.20	9.47 ± 0.11	n = 21	[ABJ] Abbott Architect c System
7.03 ± 0.10	9.83 ± 0.17	8.45 ± 0.14	13.02 ± 0.22	9.25 ± 0.14	n = 62	[OLC] Beckman Coulter AU Chemistry System
6.94 ± 0.10	9.77 ± 0.14	8.36 ± 0.26	12.73 ± 0.23	9.05 ± 0.19	n = 3	[BCX] Beckman Coulter LX-20
6.88 ± 0.09	9.62 ± 0.11	8.25 ± 0.12	12.59 ± 0.17	8.96 ± 0.14	n = 18	[BCG] Beckman Coulter UniCel DxC 600
6.97 ± 0.12	9.74 ± 0.15	8.34 ± 0.17	12.80 ± 0.15	9.03 ± 0.12	n = 8	[BCH] Beckman Coulter UniCel DxC 800
6.82 ± 0.24	9.90 ± 0.27	8.44 ± 0.22	13.25 ± 0.21	9.33 ± 0.32	n = 8	[JJE] Ortho Vitros 250/350/950
6.89 ± 0.15	9.91 ± 0.17	8.47 ± 0.16	13.26 ± 0.20	9.34 ± 0.15	n = 19	[JJF] Ortho Vitros 5,1FS
6.91 ± 0.19	9.93 ± 0.19	8.53 ± 0.20	13.27 ± 0.23	9.37 ± 0.17	n = 19	[JJG] Ortho Vitros 5600
6.73 ± 0.05	9.57 ± 0.05	8.20 ± 0.00	13.23 ± 0.05	8.93 ± 0.05	n = 3	[ROK] Roche cobas c111
6.98 ± 0.15	10.02 ± 0.29	8.50 ± 0.19	13.64 ± 0.32	9.33 ± 0.27	n = 20	[ROC] Roche cobas c501
7.01 ± 0.18	9.81 ± 0.21	8.35 ± 0.18	13.38 ± 0.31	9.17 ± 0.20	n = 5	[ROH] Roche cobas c701
6.85 ± 0.06	9.90 ± 0.24	8.40 ± 0.00	13.48 ± 0.32	9.13 ± 0.28	n = 5	[ROS] Roche Cobas INTEGRA 400
6.93 ± 0.05	9.83 ± 0.05	8.56 ± 0.10	13.60 ± 0.00	9.18 ± 0.15	n = 3	[ROT] Roche Cobas INTEGRA 800
7.08 ± 0.24	9.87 ± 0.24	8.47 ± 0.22	13.22 ± 0.32	9.30 ± 0.22	n = 31	[ROD] Roche MODULAR D/P
7.07 ± 0.17	9.85 ± 0.14	8.49 ± 0.17	13.16 ± 0.26	9.33 ± 0.22	n = 22	[BYE] Siemens ADVIA 1800
7.13 ± 0.05	9.86 ± 0.10	8.47 ± 0.14	13.10 ± 0.09	9.37 ± 0.14	n = 3	[BYB] Siemens ADVIA 2400
6.92 ± 0.12	9.48 ± 0.21	8.21 ± 0.18	12.91 ± 0.25	8.91 ± 0.21	n = 21	[DUE] Siemens Dimension EXL
6.92 ± 0.14	9.55 ± 0.17	8.25 ± 0.15	12.92 ± 0.26	8.92 ± 0.11	n = 21	[DUR] Siemens Dimension RxL
6.90 ± 0.15	9.52 ± 0.20	8.18 ± 0.13	12.96 ± 0.27	8.96 ± 0.17	n = 41	[DUT] Siemens Dimension Vista
6.91 ± 0.18	9.51 ± 0.21	8.22 ± 0.17	12.87 ± 0.22	8.88 ± 0.18	n = 15	[DUX] Siemens Dimension Xpand
<Reagents>						
6.95 ± 0.22	9.75 ± 0.12	8.52 ± 0.13	12.97 ± 0.16	9.13 ± 0.23	n = 4	[AX1] Abaxis
7.15 ± 0.08	10.11 ± 0.15	8.54 ± 0.08	13.61 ± 0.20	9.47 ± 0.11	n = 21	[AB1] Abbott
6.91 ± 0.11	9.66 ± 0.15	8.29 ± 0.16	12.68 ± 0.20	8.99 ± 0.14	n = 31	[BC1] Beckman Coulter
7.03 ± 0.10	9.84 ± 0.15	8.45 ± 0.13	13.04 ± 0.20	9.26 ± 0.14	n = 59	[OL1] Beckman Coulter AU Series
6.89 ± 0.19	9.91 ± 0.19	8.50 ± 0.20	13.28 ± 0.22	9.36 ± 0.18	n = 49	[JJ1] Ortho Clinical Diagnostics
6.73 ± 0.05	9.57 ± 0.05	8.20 ± 0.00	13.23 ± 0.05	8.93 ± 0.05	n = 3	[RO8] Roche cobas c111
6.99 ± 0.15	9.97 ± 0.22	8.48 ± 0.18	13.56 ± 0.29	9.30 ± 0.24	n = 25	[RO4] Roche cobas c311/c501/c502/c701/c702
7.08 ± 0.24	9.87 ± 0.24	8.48 ± 0.22	13.23 ± 0.30	9.31 ± 0.22	n = 30	[RO2] Roche Hitachi and Modular D/P
6.89 ± 0.07	9.82 ± 0.09	8.49 ± 0.14	13.54 ± 0.24	9.13 ± 0.20	n = 8	[RO1] Roche Integra and MIRA
7.05 ± 0.12	9.60 ± 0.24	8.48 ± 0.21	12.46 ± 0.38	9.13 ± 0.23	n = 4	[GZ1] Sekisui Diagnostics (Genzyme)
7.08 ± 0.15	9.85 ± 0.13	8.50 ± 0.16	13.14 ± 0.24	9.34 ± 0.20	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
6.91 ± 0.14	9.52 ± 0.20	8.21 ± 0.16	12.93 ± 0.26	8.93 ± 0.17	n = 98	[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Magnesium (mg/dL)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
2.05 ± 0.09	1.58 ± 0.09	2.91 ± 0.11	2.57 ± 0.11	1.25 ± 0.09	n = 332	[---] All Methods & Instruments
<Instruments>						
1.89 ± 0.05	1.54 ± 0.06	2.72 ± 0.06	2.40 ± 0.07	1.12 ± 0.05	n = 20	[ABJ] Abbott Architect c System
2.03 ± 0.06	1.58 ± 0.05	2.88 ± 0.07	2.54 ± 0.07	1.27 ± 0.06	n = 58	[OLC] Beckman Coulter AU Chemistry System
2.03 ± 0.05	1.60 ± 0.09	2.90 ± 0.09	2.56 ± 0.10	1.30 ± 0.09	n = 3	[BCX] Beckman Coulter LX-20
2.08 ± 0.07	1.61 ± 0.04	2.94 ± 0.07	2.60 ± 0.05	1.28 ± 0.05	n = 18	[BCG] Beckman Coulter UniCel DxC 600
2.02 ± 0.04	1.57 ± 0.07	2.85 ± 0.06	2.61 ± 0.06	1.28 ± 0.04	n = 8	[BCH] Beckman Coulter UniCel DxC 800
2.05 ± 0.06	1.55 ± 0.06	2.90 ± 0.00	2.55 ± 0.06	1.20 ± 0.00	n = 6	[JJE] Ortho Vitros 250/350/950
2.05 ± 0.08	1.58 ± 0.06	2.87 ± 0.09	2.54 ± 0.08	1.23 ± 0.06	n = 19	[JJF] Ortho Vitros 5,1FS
2.06 ± 0.07	1.58 ± 0.05	2.87 ± 0.09	2.55 ± 0.07	1.24 ± 0.06	n = 18	[JJG] Ortho Vitros 5600
2.08 ± 0.07	1.62 ± 0.06	2.96 ± 0.09	2.63 ± 0.10	1.27 ± 0.06	n = 18	[ROC] Roche cobas c501
2.04 ± 0.06	1.60 ± 0.00	2.90 ± 0.00	2.60 ± 0.00	1.26 ± 0.06	n = 5	[ROH] Roche cobas c701
2.13 ± 0.05	1.63 ± 0.05	2.93 ± 0.05	2.56 ± 0.10	1.30 ± 0.00	n = 3	[ROT] Roche Cobas INTEGRA 800
2.09 ± 0.09	1.60 ± 0.00	2.90 ± 0.00	2.61 ± 0.06	1.30 ± 0.00	n = 28	[ROD] Roche MODULAR D/P
2.17 ± 0.08	1.73 ± 0.07	3.06 ± 0.09	2.83 ± 0.08	1.44 ± 0.07	n = 21	[BYE] Siemens ADVIA 1800
2.27 ± 0.05	1.77 ± 0.05	3.10 ± 0.09	2.94 ± 0.10	1.47 ± 0.05	n = 3	[BYB] Siemens ADVIA 2400
2.01 ± 0.05	1.50 ± 0.04	2.94 ± 0.09	2.58 ± 0.07	1.18 ± 0.07	n = 18	[DUE] Siemens Dimension EXL
2.03 ± 0.07	1.53 ± 0.07	2.97 ± 0.10	2.58 ± 0.09	1.21 ± 0.06	n = 20	[DUR] Siemens Dimension RxL
2.09 ± 0.10	1.51 ± 0.11	2.99 ± 0.12	2.61 ± 0.11	1.23 ± 0.10	n = 41	[DUT] Siemens Dimension Vista
1.98 ± 0.07	1.46 ± 0.09	2.91 ± 0.11	2.52 ± 0.08	1.17 ± 0.08	n = 14	[DUX] Siemens Dimension Xpand
<Reagents>						
1.89 ± 0.05	1.54 ± 0.06	2.72 ± 0.06	2.40 ± 0.07	1.12 ± 0.05	n = 20	[AB1] Abbott
2.05 ± 0.06	1.60 ± 0.06	2.90 ± 0.08	2.60 ± 0.07	1.28 ± 0.05	n = 29	[BC1] Beckman Coulter
2.04 ± 0.06	1.58 ± 0.05	2.89 ± 0.07	2.54 ± 0.06	1.28 ± 0.05	n = 56	[OL1] Beckman Coulter AU Series
2.05 ± 0.07	1.57 ± 0.06	2.88 ± 0.09	2.54 ± 0.07	1.23 ± 0.06	n = 45	[JJ1] Ortho Clinical Diagnostics
2.07 ± 0.06	1.62 ± 0.06	2.94 ± 0.09	2.62 ± 0.09	1.27 ± 0.06	n = 23	[RO4] Roche cobas c311/c501/c502/c701/c702
2.09 ± 0.09	1.60 ± 0.00	2.90 ± 0.00	2.61 ± 0.06	1.30 ± 0.00	n = 28	[RO2] Roche Hitachi and Modular D/P
2.10 ± 0.06	1.64 ± 0.06	2.90 ± 0.00	2.50 ± 0.00	1.30 ± 0.00	n = 5	[RO1] Roche Integra and MIRA
2.19 ± 0.08	1.74 ± 0.07	3.08 ± 0.10	2.84 ± 0.09	1.45 ± 0.08	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
2.04 ± 0.09	1.50 ± 0.09	2.96 ± 0.11	2.58 ± 0.10	1.20 ± 0.08	n = 92	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Iron (µg/dL)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
77.4 ± 3.02	94.5 ± 3.62	92.5 ± 6.58	126.1 ± 6.15	109.1 ± 5.62	n = 269	[---] All Methods & Instruments
<Instruments>						
78.8 ± 2.21	92.7 ± 2.25	97.0 ± 1.16	127.2 ± 2.94	112.9 ± 2.55	n = 15	[ABJ] Abbott Architect c System
78.9 ± 2.55	97.5 ± 3.04	93.6 ± 2.78	128.4 ± 3.44	111.1 ± 3.01	n = 52	[OLC] Beckman Coulter AU Chemistry System
80.3 ± 1.37	93.6 ± 1.02	88.7 ± 1.37	129.0 ± 0.90	106.7 ± 1.37	n = 3	[BCX] Beckman Coulter LX-20
78.6 ± 2.04	96.3 ± 2.32	85.7 ± 4.35	129.9 ± 3.24	106.0 ± 3.89	n = 11	[BCG] Beckman Coulter UniCel DxC 600
78.3 ± 0.52	94.4 ± 2.89	85.6 ± 3.94	127.8 ± 0.73	104.9 ± 3.60	n = 8	[BCH] Beckman Coulter UniCel DxC 800
75.4 ± 5.46	97.8 ± 6.50	109.1 ± 5.00	143.2 ± 6.95	125.2 ± 5.89	n = 17	[JJF] Ortho Vitros 5,1FS
74.6 ± 5.20	97.5 ± 3.74	108.4 ± 5.55	140.6 ± 6.26	125.7 ± 6.43	n = 19	[JJG] Ortho Vitros 5600
79.1 ± 2.93	95.2 ± 2.62	93.9 ± 2.74	125.7 ± 1.64	109.6 ± 1.69	n = 10	[ROC] Roche cobas c501
77.7 ± 1.37	93.7 ± 1.37	93.3 ± 1.37	126.7 ± 0.51	109.3 ± 0.51	n = 3	[ROH] Roche cobas c701
77.1 ± 1.33	94.0 ± 1.66	92.2 ± 1.66	125.1 ± 1.82	108.9 ± 1.79	n = 27	[ROD] Roche MODULAR D/P
78.3 ± 0.77	94.8 ± 1.94	94.1 ± 1.19	124.9 ± 1.66	109.9 ± 1.77	n = 21	[BYE] Siemens ADVIA 1800
79.3 ± 2.26	95.5 ± 2.74	94.3 ± 2.26	126.3 ± 3.16	111.7 ± 3.16	n = 3	[BYB] Siemens ADVIA 2400
75.0 ± 0.95	91.1 ± 1.39	87.3 ± 0.82	119.9 ± 1.11	104.4 ± 0.92	n = 11	[DUE] Siemens Dimension EXL
75.0 ± 0.80	91.7 ± 0.66	87.4 ± 1.03	119.6 ± 0.99	103.8 ± 1.03	n = 12	[DUR] Siemens Dimension RxL
76.1 ± 1.73	92.5 ± 1.60	88.3 ± 2.02	121.6 ± 2.05	105.1 ± 2.24	n = 37	[DUT] Siemens Dimension Vista
74.7 ± 0.51	91.0 ± 0.00	87.7 ± 0.51	119.3 ± 0.51	104.0 ± 0.00	n = 3	[DUX] Siemens Dimension Xpand
<Reagents>						
79.0 ± 2.32	92.9 ± 2.15	96.9 ± 1.27	127.6 ± 2.73	113.2 ± 2.44	n = 13	[AB3] Abbott-Iron/6K95
78.6 ± 1.60	95.3 ± 2.52	86.3 ± 4.18	128.8 ± 2.89	105.8 ± 3.60	n = 24	[BC1] Beckman Coulter
79.3 ± 2.54	98.2 ± 2.63	94.1 ± 2.83	129.3 ± 2.87	111.9 ± 2.64	n = 41	[OL1] Beckman Coulter AU Series
74.5 ± 5.34	97.0 ± 5.39	108.3 ± 5.25	141.1 ± 6.89	124.7 ± 6.55	n = 39	[JJ1] Ortho Clinical Diagnostics
78.3 ± 1.38	95.0 ± 2.16	93.7 ± 2.19	126.1 ± 1.41	109.7 ± 1.55	n = 16	[RO4] Roche cobas c311/c501/c502/c701/c702
77.1 ± 1.33	94.0 ± 1.66	92.2 ± 1.66	125.1 ± 1.82	108.9 ± 1.79	n = 27	[RO2] Roche Hitachi and Modular D/P
80.3 ± 1.58	96.8 ± 2.80	93.8 ± 3.21	126.6 ± 1.80	111.3 ± 2.54	n = 4	[RO1] Roche Integra and MIRA
77.6 ± 1.68	94.7 ± 2.59	92.0 ± 1.25	124.9 ± 2.18	108.0 ± 1.97	n = 8	[GZ1] Sekisui Diagnostics (Genzyme)
78.2 ± 1.25	94.8 ± 1.97	93.9 ± 1.58	124.7 ± 2.27	109.9 ± 1.96	n = 24	[BY1] Siemens ADVIA/ADVIA Centaur
75.4 ± 1.24	92.0 ± 1.54	87.9 ± 1.73	120.6 ± 2.06	104.6 ± 1.75	n = 63	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Sodium (mmol/L)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
129.0 ± 1.87	140.7 ± 1.69	145.8 ± 2.07	161.1 ± 2.65	159.6 ± 2.69	n = 372	[---] All Methods & Instruments
<Instruments>						
124.9 ± 2.04	137.5 ± 1.22	142.5 ± 1.22	160.5 ± 2.83	157.7 ± 3.17	n = 4	[AXA] Abaxis Piccolo
128.5 ± 1.07	140.8 ± 1.22	145.8 ± 1.15	162.5 ± 1.49	161.2 ± 1.28	n = 21	[ABJ] Abbott Architect c System
128.2 ± 1.43	140.0 ± 1.25	144.5 ± 1.44	160.0 ± 1.61	159.0 ± 1.34	n = 62	[OLC] Beckman Coulter AU Chemistry System
129.0 ± 2.70	140.7 ± 1.37	146.7 ± 1.37	162.0 ± 1.80	160.3 ± 3.16	n = 3	[BCX] Beckman Coulter LX-20
128.2 ± 0.88	140.3 ± 0.86	144.9 ± 1.01	160.8 ± 1.40	158.8 ± 1.20	n = 18	[BCG] Beckman Coulter UniCel DxC 600
128.8 ± 1.02	140.4 ± 1.07	145.5 ± 0.90	161.0 ± 0.00	158.7 ± 0.69	n = 8	[BCH] Beckman Coulter UniCel DxC 800
126.2 ± 0.66	138.4 ± 0.56	144.0 ± 0.47	159.2 ± 0.66	155.5 ± 0.74	n = 7	[IAA] i-STAT
130.5 ± 2.40	143.1 ± 2.47	150.7 ± 2.52	170.8 ± 3.11	166.9 ± 2.86	n = 8	[JJE] Ortho Vitros 250/350/950
131.0 ± 1.98	142.8 ± 1.99	150.9 ± 2.17	171.1 ± 2.39	167.4 ± 2.55	n = 19	[JJF] Ortho Vitros 5,1FS
130.1 ± 1.20	142.3 ± 1.65	150.5 ± 1.69	171.2 ± 2.12	167.2 ± 1.95	n = 19	[JJG] Ortho Vitros 5600
126.8 ± 1.54	138.6 ± 1.02	145.0 ± 0.90	159.0 ± 0.90	157.7 ± 1.37	n = 3	[ROK] Roche cobas c111
127.0 ± 1.68	140.3 ± 1.66	144.7 ± 1.37	161.2 ± 1.62	159.8 ± 1.10	n = 19	[ROC] Roche cobas c501
128.8 ± 1.28	141.1 ± 1.27	145.8 ± 0.80	161.7 ± 1.61	159.8 ± 0.80	n = 5	[ROH] Roche cobas c701
127.2 ± 0.41	138.5 ± 1.22	144.5 ± 0.57	159.2 ± 1.27	157.9 ± 1.13	n = 4	[ROS] Roche Cobas INTEGRA 400
127.7 ± 0.51	139.0 ± 0.90	145.0 ± 0.00	159.7 ± 0.51	159.0 ± 0.00	n = 3	[ROT] Roche Cobas INTEGRA 800
129.0 ± 1.53	140.9 ± 1.20	145.6 ± 1.23	162.4 ± 1.63	160.7 ± 1.59	n = 30	[ROD] Roche MODULAR D/P
130.4 ± 1.10	142.3 ± 0.90	147.2 ± 0.88	163.1 ± 0.89	161.2 ± 0.79	n = 22	[BYE] Siemens ADVIA 1800
130.0 ± 0.00	141.3 ± 0.51	146.7 ± 0.51	162.0 ± 0.00	160.7 ± 0.51	n = 3	[BYB] Siemens ADVIA 2400
129.1 ± 1.35	141.0 ± 1.05	145.7 ± 1.14	160.4 ± 1.23	159.5 ± 0.96	n = 22	[DUE] Siemens Dimension EXL
128.6 ± 1.68	139.7 ± 1.77	144.6 ± 1.86	159.2 ± 2.40	157.7 ± 2.35	n = 20	[DUR] Siemens Dimension RxL
130.6 ± 1.45	141.0 ± 1.54	147.2 ± 1.28	159.1 ± 1.53	157.2 ± 1.20	n = 41	[DUT] Siemens Dimension Vista
129.3 ± 0.99	141.2 ± 0.95	145.8 ± 0.80	161.1 ± 0.95	159.8 ± 1.06	n = 15	[DUX] Siemens Dimension Xpand
<Reagents>						
124.9 ± 2.04	137.5 ± 1.22	142.5 ± 1.22	160.5 ± 2.83	157.7 ± 3.17	n = 4	[AX1] Abaxis
128.4 ± 1.10	140.6 ± 1.32	145.7 ± 1.21	162.4 ± 1.57	161.1 ± 1.41	n = 22	[AB1] Abbott
128.3 ± 1.25	140.3 ± 1.00	145.1 ± 1.18	160.9 ± 1.33	158.8 ± 1.14	n = 31	[BC1] Beckman Coulter
128.3 ± 1.39	139.9 ± 1.23	144.5 ± 1.43	160.0 ± 1.64	159.0 ± 1.35	n = 60	[OL1] Beckman Coulter AU Series
126.0 ± 0.55	138.5 ± 0.57	144.0 ± 0.55	159.0 ± 0.55	155.3 ± 0.72	n = 6	[IA1] i-STAT
128.0 ± 0.90	140.3 ± 1.37	145.3 ± 0.51	165.1 ± 2.05	160.0 ± 1.80	n = 3	[IL1] Instrumentation Lab
130.6 ± 1.90	142.7 ± 2.02	150.7 ± 2.13	171.1 ± 2.34	167.3 ± 2.36	n = 49	[JJ1] Ortho Clinical Diagnostics
126.8 ± 1.54	138.6 ± 1.02	145.0 ± 0.90	159.0 ± 0.90	157.7 ± 1.37	n = 3	[RO8] Roche cobas c111
127.6 ± 1.76	140.6 ± 1.52	145.1 ± 1.30	161.4 ± 1.57	159.8 ± 0.99	n = 26	[RO4] Roche cobas c311/c501/c502/c701/c702
129.0 ± 1.53	140.9 ± 1.20	145.6 ± 1.23	162.4 ± 1.63	160.7 ± 1.59	n = 30	[RO2] Roche Hitachi and Modular D/P
127.4 ± 0.56	138.7 ± 1.11	144.8 ± 0.47	159.7 ± 0.59	158.7 ± 0.59	n = 7	[RO1] Roche Integra and MIRA
130.3 ± 0.96	142.2 ± 0.93	147.1 ± 0.83	162.9 ± 0.89	161.1 ± 0.75	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
129.6 ± 1.62	140.8 ± 1.46	146.1 ± 1.62	159.8 ± 1.82	158.1 ± 1.94	n = 97	[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Potassium (mmol/L)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
2.81 ± 0.09	3.97 ± 0.09	4.62 ± 0.10	6.34 ± 0.15	5.44 ± 0.12	n = 372	[---] All Methods & Instruments
<Instruments>						
3.00 ± 0.17	4.29 ± 0.11	5.07 ± 0.20	6.87 ± 0.09	5.90 ± 0.11	n = 4	[AXA] Abaxis Piccolo
2.78 ± 0.07	3.92 ± 0.10	4.57 ± 0.09	6.33 ± 0.10	5.42 ± 0.10	n = 21	[ABJ] Abbott Architect c System
2.87 ± 0.06	4.00 ± 0.00	4.61 ± 0.05	6.30 ± 0.09	5.45 ± 0.07	n = 62	[OLC] Beckman Coulter AU Chemistry System
2.80 ± 0.09	3.93 ± 0.05	4.70 ± 0.09	6.43 ± 0.14	5.53 ± 0.14	n = 3	[BCX] Beckman Coulter LX-20
2.73 ± 0.05	3.93 ± 0.06	4.60 ± 0.00	6.39 ± 0.07	5.46 ± 0.06	n = 18	[BCG] Beckman Coulter UniCel DxC 600
2.73 ± 0.05	3.93 ± 0.05	4.62 ± 0.04	6.43 ± 0.07	5.48 ± 0.04	n = 8	[BCH] Beckman Coulter UniCel DxC 800
2.70 ± 0.00	3.90 ± 0.00	4.56 ± 0.06	6.20 ± 0.00	5.30 ± 0.00	n = 7	[IAA] i-STAT
2.90 ± 0.00	4.11 ± 0.06	4.81 ± 0.06	6.59 ± 0.09	5.66 ± 0.07	n = 8	[JJE] Ortho Vitros 250/350/950
2.91 ± 0.05	4.09 ± 0.05	4.83 ± 0.09	6.61 ± 0.10	5.66 ± 0.09	n = 19	[JJF] Ortho Vitros 5,1FS
2.90 ± 0.07	4.11 ± 0.07	4.83 ± 0.07	6.60 ± 0.09	5.66 ± 0.10	n = 19	[JJG] Ortho Vitros 5600
2.80 ± 0.00	4.00 ± 0.00	4.67 ± 0.05	6.33 ± 0.05	5.47 ± 0.05	n = 3	[ROK] Roche cobas c111
2.70 ± 0.08	3.90 ± 0.07	4.54 ± 0.07	6.30 ± 0.10	5.40 ± 0.07	n = 19	[ROC] Roche cobas c501
2.80 ± 0.06	3.96 ± 0.11	4.62 ± 0.08	6.32 ± 0.08	5.42 ± 0.08	n = 5	[ROH] Roche cobas c701
2.80 ± 0.00	4.00 ± 0.00	4.68 ± 0.04	6.32 ± 0.04	5.45 ± 0.06	n = 4	[ROS] Roche Cobas INTEGRA 400
2.80 ± 0.00	4.00 ± 0.00	4.70 ± 0.00	6.36 ± 0.10	5.50 ± 0.00	n = 3	[ROT] Roche Cobas INTEGRA 800
2.76 ± 0.07	3.90 ± 0.06	4.55 ± 0.07	6.30 ± 0.10	5.37 ± 0.07	n = 30	[ROD] Roche MODULAR D/P
2.90 ± 0.00	4.06 ± 0.06	4.70 ± 0.00	6.43 ± 0.06	5.53 ± 0.06	n = 22	[BYE] Siemens ADVIA 1800
2.90 ± 0.00	4.07 ± 0.05	4.70 ± 0.00	6.43 ± 0.05	5.50 ± 0.00	n = 3	[BYB] Siemens ADVIA 2400
2.75 ± 0.06	3.97 ± 0.05	4.60 ± 0.00	6.32 ± 0.05	5.41 ± 0.04	n = 21	[DUE] Siemens Dimension EXL
2.75 ± 0.06	3.95 ± 0.06	4.61 ± 0.07	6.31 ± 0.08	5.40 ± 0.08	n = 21	[DUR] Siemens Dimension RxL
2.80 ± 0.00	3.90 ± 0.00	4.60 ± 0.00	6.18 ± 0.06	5.30 ± 0.04	n = 41	[DUT] Siemens Dimension Vista
2.76 ± 0.06	3.97 ± 0.05	4.60 ± 0.00	6.40 ± 0.00	5.45 ± 0.06	n = 15	[DUX] Siemens Dimension Xpand
<Reagents>						
3.00 ± 0.17	4.29 ± 0.11	5.07 ± 0.20	6.87 ± 0.09	5.90 ± 0.11	n = 4	[AX1] Abaxis
2.77 ± 0.07	3.92 ± 0.10	4.57 ± 0.08	6.32 ± 0.10	5.41 ± 0.10	n = 22	[AB1] Abbott
2.74 ± 0.06	3.93 ± 0.06	4.60 ± 0.00	6.40 ± 0.08	5.46 ± 0.06	n = 31	[BC1] Beckman Coulter
2.87 ± 0.06	4.00 ± 0.00	4.61 ± 0.05	6.30 ± 0.09	5.45 ± 0.07	n = 60	[OL1] Beckman Coulter AU Series
2.70 ± 0.00	3.90 ± 0.00	4.55 ± 0.06	6.20 ± 0.00	5.30 ± 0.00	n = 6	[IA1] i-STAT
2.63 ± 0.05	3.80 ± 0.00	4.50 ± 0.00	6.23 ± 0.05	5.30 ± 0.00	n = 3	[IL1] Instrumentation Lab
2.90 ± 0.05	4.10 ± 0.06	4.82 ± 0.07	6.60 ± 0.09	5.65 ± 0.09	n = 49	[JJ1] Ortho Clinical Diagnostics
2.80 ± 0.00	4.00 ± 0.00	4.67 ± 0.05	6.33 ± 0.05	5.47 ± 0.05	n = 3	[RO8] Roche cobas c111
2.72 ± 0.08	3.91 ± 0.08	4.56 ± 0.08	6.30 ± 0.09	5.40 ± 0.07	n = 26	[RO4] Roche cobas c311/c501/c502/c701/c702
2.76 ± 0.07	3.90 ± 0.06	4.55 ± 0.07	6.30 ± 0.10	5.37 ± 0.07	n = 30	[RO2] Roche Hitachi and Modular D/P
2.80 ± 0.00	4.00 ± 0.00	4.70 ± 0.00	6.30 ± 0.00	5.48 ± 0.05	n = 7	[RO1] Roche Integra and MIRA
2.90 ± 0.00	4.07 ± 0.06	4.70 ± 0.00	6.43 ± 0.06	5.53 ± 0.06	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
2.78 ± 0.05	3.94 ± 0.06	4.60 ± 0.00	6.27 ± 0.11	5.37 ± 0.08	n = 98	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Chloride (mmol/L)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
89.8 ± 1.90	97.6 ± 1.61	115.3 ± 2.64	118.1 ± 2.21	111.5 ± 2.59	n = 367	[---] All Methods & Instruments
<Instruments>						
88.7 ± 1.58	96.0 ± 0.75	114.5 ± 1.22	115.9 ± 1.13	108.9 ± 1.13	n = 4	[AXA] Abaxis Piccolo
90.6 ± 0.66	97.6 ± 0.81	116.7 ± 0.88	118.5 ± 0.90	111.3 ± 0.85	n = 20	[ABJ] Abbott Architect c System
89.0 ± 1.06	96.7 ± 1.02	114.8 ± 1.26	116.6 ± 1.19	109.9 ± 0.97	n = 61	[OLC] Beckman Coulter AU Chemistry System
90.8 ± 1.54	98.0 ± 0.90	117.1 ± 2.05	120.5 ± 1.86	113.0 ± 1.80	n = 3	[BCX] Beckman Coulter LX-20
90.6 ± 0.92	98.0 ± 0.61	116.7 ± 0.84	119.1 ± 0.93	111.7 ± 0.80	n = 18	[BCG] Beckman Coulter UniCel DxC 600
91.5 ± 0.90	98.4 ± 0.72	118.0 ± 0.91	119.7 ± 1.62	112.3 ± 1.08	n = 8	[BCH] Beckman Coulter UniCel DxC 800
93.5 ± 1.02	99.8 ± 1.00	125.3 ± 0.97	127.4 ± 0.56	120.2 ± 0.73	n = 6	[IAA] i-STAT
90.6 ± 1.52	97.1 ± 1.56	118.1 ± 2.37	120.1 ± 2.32	113.4 ± 2.09	n = 8	[JJE] Ortho Vitros 250/350/950
91.1 ± 1.26	97.7 ± 1.18	118.2 ± 1.26	120.7 ± 1.39	113.6 ± 1.54	n = 19	[JJF] Ortho Vitros 5,1FS
90.8 ± 1.01	97.3 ± 1.04	117.5 ± 1.41	120.1 ± 1.32	113.4 ± 1.42	n = 19	[JJG] Ortho Vitros 5600
89.3 ± 1.37	97.3 ± 0.51	115.0 ± 1.80	118.3 ± 1.37	112.3 ± 1.37	n = 3	[ROK] Roche cobas c111
84.2 ± 1.34	95.2 ± 1.35	111.7 ± 1.53	115.1 ± 1.39	107.4 ± 1.26	n = 19	[ROC] Roche cobas c501
85.4 ± 1.09	96.0 ± 0.00	112.2 ± 0.80	116.0 ± 0.64	107.0 ± 0.64	n = 5	[ROH] Roche cobas c701
90.3 ± 0.90	97.5 ± 0.57	115.9 ± 1.13	118.0 ± 0.75	111.8 ± 0.41	n = 4	[ROS] Roche Cobas INTEGRA 400
90.3 ± 0.51	97.7 ± 0.51	116.7 ± 0.51	119.3 ± 0.51	113.0 ± 0.90	n = 3	[ROT] Roche Cobas INTEGRA 800
87.4 ± 1.42	96.1 ± 1.14	113.7 ± 1.38	116.6 ± 1.42	108.8 ± 1.22	n = 30	[ROD] Roche MODULAR D/P
89.7 ± 0.98	97.4 ± 1.24	116.0 ± 1.56	117.4 ± 1.59	110.0 ± 1.43	n = 22	[BYE] Siemens ADVIA 1800
89.7 ± 0.51	97.7 ± 0.51	115.7 ± 1.37	117.4 ± 1.02	109.7 ± 1.37	n = 3	[BYB] Siemens ADVIA 2400
90.1 ± 0.77	99.4 ± 0.73	112.4 ± 0.68	118.1 ± 1.08	112.9 ± 0.71	n = 21	[DUE] Siemens Dimension EXL
89.2 ± 1.73	98.8 ± 1.63	112.9 ± 1.81	118.9 ± 1.81	112.8 ± 2.06	n = 21	[DUR] Siemens Dimension RxL
90.8 ± 1.28	98.6 ± 1.21	117.2 ± 1.25	120.1 ± 1.28	113.9 ± 1.31	n = 41	[DUT] Siemens Dimension Vista
90.2 ± 0.68	99.2 ± 0.85	112.1 ± 1.17	117.7 ± 0.89	112.6 ± 0.83	n = 15	[DUX] Siemens Dimension Xpand
<Reagents>						
88.7 ± 1.58	96.0 ± 0.75	114.5 ± 1.22	115.9 ± 1.13	108.9 ± 1.13	n = 4	[AX1] Abaxis
90.6 ± 0.67	97.6 ± 0.88	116.6 ± 0.94	118.5 ± 1.00	111.2 ± 0.88	n = 21	[AB1] Abbott
90.8 ± 1.11	98.1 ± 0.77	117.0 ± 1.27	119.3 ± 1.42	111.9 ± 1.12	n = 31	[BC1] Beckman Coulter
89.0 ± 1.01	96.7 ± 0.99	114.8 ± 1.22	116.6 ± 1.15	109.9 ± 0.92	n = 59	[OL1] Beckman Coulter AU Series
93.5 ± 1.02	99.8 ± 1.00	125.3 ± 0.97	127.4 ± 0.56	120.2 ± 0.73	n = 6	[IA1] i-STAT
90.9 ± 1.18	97.4 ± 1.14	117.9 ± 1.57	120.3 ± 1.54	113.5 ± 1.60	n = 49	[JJ1] Ortho Clinical Diagnostics
89.3 ± 1.37	97.3 ± 0.51	115.0 ± 1.80	118.3 ± 1.37	112.3 ± 1.37	n = 3	[RO8] Roche cobas c111
84.4 ± 1.22	95.3 ± 1.13	111.8 ± 1.38	115.2 ± 1.28	107.3 ± 1.17	n = 25	[RO4] Roche cobas c311/c501/c502/c701/c702
87.4 ± 1.42	96.1 ± 1.14	113.7 ± 1.38	116.6 ± 1.42	108.8 ± 1.22	n = 30	[RO2] Roche Hitachi and Modular D/P
90.2 ± 1.03	97.4 ± 0.72	116.0 ± 1.42	118.3 ± 1.24	112.1 ± 1.15	n = 8	[RO1] Roche Integra and MIRA
89.7 ± 0.99	97.5 ± 1.07	115.9 ± 1.59	117.3 ± 1.60	110.0 ± 1.37	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
90.3 ± 1.22	98.9 ± 1.21	114.4 ± 2.91	119.0 ± 1.67	113.3 ± 1.42	n = 98	[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Albumin (g/dL)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
3.10 ± 0.21	4.07 ± 0.18	3.68 ± 0.23	4.95 ± 0.21	4.55 ± 0.18	n = 359	[---] All Methods & Instruments
<Instruments>						
3.35 ± 0.12	4.17 ± 0.08	3.90 ± 0.11	5.29 ± 0.11	4.73 ± 0.09	n = 4	[AXA] Abaxis Piccolo
2.89 ± 0.07	3.84 ± 0.07	3.50 ± 0.12	4.56 ± 0.09	4.23 ± 0.10	n = 21	[ABJ] Abbott Architect c System
3.18 ± 0.07	4.11 ± 0.09	3.78 ± 0.08	4.90 ± 0.10	4.54 ± 0.08	n = 63	[OLC] Beckman Coulter AU Chemistry System
2.83 ± 0.05	3.83 ± 0.05	3.33 ± 0.05	4.43 ± 0.05	4.17 ± 0.05	n = 3	[BCX] Beckman Coulter LX-20
2.84 ± 0.08	3.86 ± 0.06	3.38 ± 0.10	4.52 ± 0.14	4.20 ± 0.11	n = 17	[BCG] Beckman Coulter UniCel DxC 600
2.88 ± 0.04	3.81 ± 0.06	3.40 ± 0.00	4.50 ± 0.05	4.20 ± 0.00	n = 8	[BCH] Beckman Coulter UniCel DxC 800
2.79 ± 0.06	3.74 ± 0.07	3.39 ± 0.09	4.83 ± 0.09	4.44 ± 0.07	n = 8	[JJE] Ortho Vitros 250/350/950
2.82 ± 0.09	3.83 ± 0.11	3.44 ± 0.10	4.95 ± 0.10	4.55 ± 0.13	n = 18	[JJF] Ortho Vitros 5,1FS
2.78 ± 0.06	3.79 ± 0.09	3.40 ± 0.12	4.90 ± 0.15	4.51 ± 0.14	n = 19	[JJG] Ortho Vitros 5600
3.35 ± 0.09	4.25 ± 0.11	3.99 ± 0.11	5.16 ± 0.12	4.77 ± 0.13	n = 18	[ROC] Roche cobas c501
3.25 ± 0.08	4.20 ± 0.10	3.94 ± 0.11	5.12 ± 0.08	4.72 ± 0.08	n = 5	[ROH] Roche cobas c701
3.20 ± 0.08	4.15 ± 0.06	3.85 ± 0.06	5.02 ± 0.15	4.65 ± 0.06	n = 4	[ROS] Roche Cobas INTEGRA 400
3.30 ± 0.00	4.13 ± 0.05	3.87 ± 0.05	4.96 ± 0.10	4.63 ± 0.05	n = 3	[ROT] Roche Cobas INTEGRA 800
3.31 ± 0.13	4.21 ± 0.12	3.98 ± 0.12	5.11 ± 0.15	4.71 ± 0.14	n = 31	[ROD] Roche MODULAR D/P
3.28 ± 0.06	4.14 ± 0.09	3.90 ± 0.07	5.08 ± 0.12	4.62 ± 0.09	n = 22	[BYE] Siemens ADVIA 1800
3.27 ± 0.05	4.17 ± 0.05	3.87 ± 0.05	5.10 ± 0.09	4.67 ± 0.05	n = 3	[BYB] Siemens ADVIA 2400
3.08 ± 0.07	4.12 ± 0.07	3.64 ± 0.09	5.02 ± 0.10	4.57 ± 0.07	n = 21	[DUE] Siemens Dimension EXL
3.08 ± 0.06	4.11 ± 0.08	3.60 ± 0.08	4.99 ± 0.11	4.58 ± 0.08	n = 21	[DUR] Siemens Dimension RxL
3.10 ± 0.07	4.16 ± 0.10	3.67 ± 0.08	5.01 ± 0.12	4.60 ± 0.11	n = 41	[DUT] Siemens Dimension Vista
3.06 ± 0.06	4.09 ± 0.08	3.62 ± 0.07	4.98 ± 0.11	4.55 ± 0.11	n = 15	[DUX] Siemens Dimension Xpand
<Reagents>						
3.35 ± 0.12	4.17 ± 0.08	3.90 ± 0.11	5.29 ± 0.11	4.73 ± 0.09	n = 4	[AX1] Abaxis
2.89 ± 0.07	3.84 ± 0.07	3.50 ± 0.12	4.56 ± 0.09	4.23 ± 0.10	n = 21	[AB1] Abbott
2.85 ± 0.07	3.84 ± 0.09	3.37 ± 0.09	4.49 ± 0.13	4.18 ± 0.10	n = 29	[BC1] Beckman Coulter
3.19 ± 0.07	4.11 ± 0.08	3.78 ± 0.08	4.90 ± 0.09	4.54 ± 0.08	n = 59	[OL1] Beckman Coulter AU Series
2.80 ± 0.07	3.80 ± 0.10	3.42 ± 0.11	4.92 ± 0.12	4.52 ± 0.12	n = 48	[JJ1] Ortho Clinical Diagnostics
3.33 ± 0.09	4.24 ± 0.11	3.98 ± 0.11	5.15 ± 0.11	4.76 ± 0.12	n = 25	[RO4] Roche cobas c311/c501/c502/c701/c702
3.31 ± 0.12	4.21 ± 0.12	3.99 ± 0.11	5.12 ± 0.13	4.72 ± 0.12	n = 31	[RO2] Roche Hitachi and Modular D/P
3.25 ± 0.07	4.14 ± 0.06	3.86 ± 0.06	4.99 ± 0.13	4.64 ± 0.06	n = 7	[RO1] Roche Integra and MIRA
3.28 ± 0.06	4.15 ± 0.08	3.89 ± 0.06	5.08 ± 0.11	4.63 ± 0.09	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
3.08 ± 0.07	4.13 ± 0.09	3.64 ± 0.09	5.00 ± 0.11	4.58 ± 0.10	n = 98	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Total Protein (g/dL)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
5.41 ± 0.14	6.64 ± 0.17	6.62 ± 0.18	8.56 ± 0.24	7.53 ± 0.21	n = 358	[---] All Methods & Instruments
<Instruments>						
5.57 ± 0.09	6.75 ± 0.06	6.77 ± 0.09	8.78 ± 0.13	7.65 ± 0.06	n = 4	[AXA] Abaxis Piccolo
5.40 ± 0.00	6.61 ± 0.06	6.65 ± 0.06	8.64 ± 0.06	7.55 ± 0.07	n = 21	[ABJ] Abbott Architect c System
5.28 ± 0.08	6.50 ± 0.11	6.47 ± 0.10	8.34 ± 0.14	7.33 ± 0.11	n = 62	[OLC] Beckman Coulter AU Chemistry System
5.25 ± 0.12	6.41 ± 0.20	6.38 ± 0.20	8.43 ± 0.16	7.36 ± 0.16	n = 4	[BCX] Beckman Coulter LX-20
5.40 ± 0.09	6.56 ± 0.12	6.55 ± 0.13	8.48 ± 0.12	7.49 ± 0.13	n = 16	[BCG] Beckman Coulter UniCel DxC 600
5.28 ± 0.10	6.50 ± 0.13	6.33 ± 0.09	8.53 ± 0.15	7.32 ± 0.12	n = 8	[BCH] Beckman Coulter UniCel DxC 800
5.37 ± 0.09	6.60 ± 0.09	6.63 ± 0.07	8.52 ± 0.17	7.56 ± 0.13	n = 8	[JJE] Ortho Vitros 250/350/950
5.39 ± 0.14	6.63 ± 0.10	6.61 ± 0.15	8.49 ± 0.13	7.51 ± 0.13	n = 19	[JJF] Ortho Vitros 5,1FS
5.42 ± 0.14	6.67 ± 0.13	6.63 ± 0.13	8.54 ± 0.17	7.56 ± 0.18	n = 19	[JJG] Ortho Vitros 5600
5.38 ± 0.11	6.59 ± 0.14	6.53 ± 0.14	8.42 ± 0.17	7.45 ± 0.12	n = 18	[ROC] Roche cobas c501
5.34 ± 0.06	6.50 ± 0.06	6.46 ± 0.06	8.20 ± 0.00	7.30 ± 0.06	n = 5	[ROH] Roche cobas c701
5.25 ± 0.12	6.50 ± 0.11	6.42 ± 0.04	8.28 ± 0.15	7.33 ± 0.09	n = 4	[ROS] Roche Cobas INTEGRA 400
5.27 ± 0.05	6.50 ± 0.09	6.37 ± 0.14	8.27 ± 0.05	7.32 ± 0.15	n = 3	[ROT] Roche Cobas INTEGRA 800
5.41 ± 0.08	6.59 ± 0.09	6.58 ± 0.10	8.44 ± 0.12	7.46 ± 0.10	n = 31	[ROD] Roche MODULAR D/P
5.43 ± 0.12	6.63 ± 0.13	6.63 ± 0.14	8.65 ± 0.15	7.57 ± 0.15	n = 22	[BYE] Siemens ADVIA 1800
5.50 ± 0.09	6.77 ± 0.05	6.74 ± 0.10	8.77 ± 0.14	7.72 ± 0.15	n = 3	[BYB] Siemens ADVIA 2400
5.55 ± 0.11	6.82 ± 0.11	6.84 ± 0.14	8.86 ± 0.16	7.76 ± 0.16	n = 21	[DUE] Siemens Dimension EXL
5.54 ± 0.09	6.81 ± 0.13	6.81 ± 0.11	8.85 ± 0.14	7.77 ± 0.14	n = 21	[DUR] Siemens Dimension RxL
5.55 ± 0.07	6.80 ± 0.07	6.79 ± 0.09	8.79 ± 0.09	7.73 ± 0.12	n = 41	[DUT] Siemens Dimension Vista
5.52 ± 0.07	6.80 ± 0.09	6.81 ± 0.09	8.82 ± 0.09	7.76 ± 0.10	n = 15	[DUX] Siemens Dimension Xpand
<Reagents>						
5.57 ± 0.09	6.75 ± 0.06	6.77 ± 0.09	8.78 ± 0.13	7.65 ± 0.06	n = 4	[AX1] Abaxis
5.40 ± 0.00	6.61 ± 0.06	6.65 ± 0.06	8.64 ± 0.06	7.55 ± 0.07	n = 21	[AB1] Abbott
5.34 ± 0.12	6.52 ± 0.15	6.46 ± 0.17	8.48 ± 0.15	7.41 ± 0.16	n = 30	[BC1] Beckman Coulter
5.28 ± 0.08	6.50 ± 0.11	6.47 ± 0.10	8.34 ± 0.14	7.33 ± 0.11	n = 59	[OL1] Beckman Coulter AU Series
5.40 ± 0.13	6.64 ± 0.12	6.63 ± 0.13	8.53 ± 0.16	7.54 ± 0.15	n = 48	[JJ1] Ortho Clinical Diagnostics
5.38 ± 0.10	6.56 ± 0.13	6.52 ± 0.13	8.38 ± 0.18	7.41 ± 0.13	n = 24	[RO4] Roche cobas c311/c501/c502/c701/c702
5.41 ± 0.08	6.60 ± 0.09	6.59 ± 0.10	8.44 ± 0.11	7.47 ± 0.10	n = 32	[RO2] Roche Hitachi and Modular D/P
5.26 ± 0.09	6.50 ± 0.10	6.42 ± 0.08	8.27 ± 0.11	7.33 ± 0.11	n = 7	[RO1] Roche Integra and MIRA
5.44 ± 0.11	6.65 ± 0.12	6.65 ± 0.14	8.66 ± 0.15	7.59 ± 0.16	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
5.54 ± 0.08	6.80 ± 0.10	6.80 ± 0.10	8.81 ± 0.12	7.74 ± 0.13	n = 97	[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Cholesterol (mg/dL)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
118.3 ± 5.22	153.3 ± 4.57	232.4 ± 6.57	149.2 ± 8.43	180.6 ± 6.78	n = 327	[---] All Methods & Instruments
<Instruments>						
122.6 ± 1.02	156.7 ± 0.51	234.7 ± 1.37	157.7 ± 1.37	187.3 ± 1.37	n = 3	[AXA] Abaxis Piccolo
122.4 ± 1.67	157.9 ± 1.77	237.9 ± 2.15	152.6 ± 1.72	185.1 ± 2.84	n = 17	[ABJ] Abbott Architect c System
116.7 ± 2.43	151.8 ± 3.19	230.4 ± 4.32	145.3 ± 3.10	177.7 ± 3.34	n = 67	[OLC] Beckman Coulter AU Chemistry System
116.0 ± 2.70	149.3 ± 3.37	229.7 ± 3.37	145.7 ± 3.16	176.5 ± 4.53	n = 3	[BCX] Beckman Coulter LX-20
117.6 ± 2.39	152.7 ± 1.88	231.4 ± 3.22	148.8 ± 3.91	179.5 ± 2.97	n = 14	[BCG] Beckman Coulter UniCel DxC 600
117.9 ± 2.67	151.7 ± 3.46	230.0 ± 4.58	149.0 ± 1.89	179.9 ± 1.98	n = 8	[BCH] Beckman Coulter UniCel DxC 800
122.5 ± 5.40	153.6 ± 6.45	236.6 ± 8.43	164.3 ± 11.27	193.1 ± 9.28	n = 3	[JJE] Ortho Vitros 250/350/950
122.7 ± 4.14	154.7 ± 5.07	236.7 ± 6.57	166.7 ± 5.65	191.7 ± 5.70	n = 17	[JJF] Ortho Vitros 5,1FS
124.5 ± 4.27	156.6 ± 5.55	239.3 ± 9.45	169.1 ± 5.98	195.0 ± 8.10	n = 19	[JJG] Ortho Vitros 5600
122.2 ± 3.19	155.9 ± 3.79	236.8 ± 4.69	153.7 ± 3.47	184.3 ± 3.45	n = 16	[ROC] Roche cobas c501
119.0 ± 1.00	153.2 ± 0.41	232.4 ± 5.04	150.2 ± 2.14	180.5 ± 3.91	n = 5	[ROH] Roche cobas c701
121.5 ± 2.33	157.5 ± 2.25	236.5 ± 4.22	150.0 ± 0.00	183.3 ± 2.15	n = 5	[ROS] Roche Cobas INTEGRA 400
119.7 ± 1.37	157.3 ± 0.51	237.7 ± 3.37	152.2 ± 3.23	182.5 ± 1.86	n = 3	[ROT] Roche Cobas INTEGRA 800
121.8 ± 3.14	155.7 ± 3.43	236.8 ± 5.40	152.3 ± 3.71	184.1 ± 3.87	n = 32	[ROD] Roche MODULAR D/P
120.3 ± 2.44	150.7 ± 3.50	228.6 ± 4.93	154.0 ± 3.80	181.2 ± 4.05	n = 22	[BYE] Siemens ADVIA 1800
122.4 ± 2.56	153.7 ± 1.37	231.0 ± 1.80	155.7 ± 1.37	184.5 ± 1.86	n = 3	[BYB] Siemens ADVIA 2400
110.5 ± 2.73	150.1 ± 2.93	229.6 ± 3.82	139.4 ± 3.49	172.7 ± 3.32	n = 18	[DUE] Siemens Dimension EXL
110.1 ± 3.80	149.4 ± 4.51	228.5 ± 5.27	138.8 ± 5.23	173.1 ± 4.81	n = 16	[DUR] Siemens Dimension RxL
114.7 ± 3.28	152.6 ± 4.90	229.4 ± 8.17	145.8 ± 5.04	178.5 ± 5.58	n = 36	[DUT] Siemens Dimension Vista
111.4 ± 3.96	151.5 ± 4.30	230.4 ± 5.57	140.0 ± 4.76	174.1 ± 3.30	n = 10	[DUX] Siemens Dimension Xpand
<Reagents>						
122.6 ± 1.02	156.7 ± 0.51	234.7 ± 1.37	157.7 ± 1.37	187.3 ± 1.37	n = 3	[AX1] Abaxis
122.4 ± 1.67	157.9 ± 1.77	237.9 ± 2.15	152.6 ± 1.72	185.1 ± 2.84	n = 17	[AB1] Abbott
117.5 ± 2.49	152.1 ± 2.80	230.6 ± 3.56	148.5 ± 3.85	179.4 ± 3.29	n = 28	[BC1] Beckman Coulter
116.8 ± 2.39	152.0 ± 3.08	230.6 ± 4.31	145.3 ± 3.03	177.8 ± 3.31	n = 61	[OL1] Beckman Coulter AU Series
123.6 ± 4.51	155.6 ± 5.57	237.8 ± 8.29	167.9 ± 6.42	193.4 ± 7.52	n = 39	[JJ1] Ortho Clinical Diagnostics
121.2 ± 3.22	155.5 ± 3.57	235.7 ± 5.38	152.7 ± 3.65	183.3 ± 4.21	n = 21	[RO4] Roche cobas c311/c501/c502/c701/c702
121.9 ± 3.32	156.0 ± 3.74	236.9 ± 5.86	152.6 ± 4.34	184.2 ± 4.54	n = 33	[RO2] Roche Hitachi and Modular D/P
120.7 ± 2.09	157.7 ± 1.02	236.9 ± 3.98	150.8 ± 2.09	182.8 ± 1.79	n = 8	[RO1] Roche Integra and MIRA
120.4 ± 2.57	151.0 ± 3.47	228.9 ± 4.56	154.2 ± 3.51	181.6 ± 3.92	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
112.4 ± 4.02	151.2 ± 4.27	229.4 ± 6.35	142.0 ± 5.81	175.4 ± 5.40	n = 80	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## HDL-Cholesterol (mg/dL)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
28.0 ± 2.72	40.3 ± 4.39	74.6 ± 7.47	31.2 ± 3.20	47.1 ± 4.34	n = 313	[---] All Methods & Instruments
27.1 ± 1.83	39.4 ± 3.19	79.9 ± 11.14	33.0 ± 3.68	49.0 ± 3.69	n = 20	[---] All Precipitation Methods
28.1 ± 2.73	40.3 ± 4.41	74.4 ± 6.99	31.1 ± 3.12	46.9 ± 4.34	n = 293	[---] All Homogeneous (Direct) Methods
16.3 ± 1.37	25.6 ± 1.02	55.0 ± 0.90	18.5 ± 1.86	31.7 ± 0.51	n = 3	[AX1] Abaxis
26.8 ± 0.73	45.6 ± 0.82	73.6 ± 1.48	31.9 ± 0.88	45.1 ± 1.03	n = 16	[AB1] Abbott
27.5 ± 1.05	44.9 ± 1.50	80.6 ± 2.51	32.3 ± 1.11	48.2 ± 1.61	n = 24	[BC1] Beckman Coulter
26.4 ± 1.09	45.7 ± 1.60	73.4 ± 2.71	31.3 ± 1.10	44.7 ± 1.40	n = 43	[OL1] Beckman Coulter AU Series
28.0 ± 1.29	41.1 ± 1.67	91.8 ± 3.62	35.5 ± 1.57	53.3 ± 2.06	n = 30	[JJ1] Ortho Clinical Diagnostics
29.4 ± 1.66	37.4 ± 1.85	71.9 ± 2.53	30.2 ± 1.52	46.6 ± 2.12	n = 18	[RO4] Roche cobas c311/c501/c502/c701/c702
31.5 ± 2.15	38.7 ± 1.04	77.0 ± 3.46	32.5 ± 2.14	49.8 ± 2.66	n = 28	[RO2] Roche Hitachi and Modular D/P
31.8 ± 1.96	39.9 ± 1.01	76.7 ± 2.60	32.4 ± 1.65	50.1 ± 2.83	n = 8	[RO1] Roche Integra and MIRA
26.9 ± 1.27	47.4 ± 1.09	75.6 ± 2.47	31.7 ± 1.38	45.8 ± 2.14	n = 5	[GZ1] Sekisui Diagnostics (Genzyme)
18.7 ± 0.81	35.1 ± 1.68	60.1 ± 2.57	22.8 ± 1.13	34.1 ± 1.48	n = 27	[BY1] Siemens ADVIA/ADVIA Centaur
28.5 ± 1.60	37.6 ± 1.26	72.8 ± 2.44	29.4 ± 1.76	46.7 ± 2.16	n = 72	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## LDL-Cholesterol (mg/dL)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
72.8 ± 9.71	92.0 ± 14.66	126.4 ± 15.03	97.6 ± 9.88	108.9 ± 12.54	n = 300	[---] All Methods
77.4 ± 6.56	99.8 ± 6.52	132.5 ± 10.07	100.3 ± 8.40	114.3 ± 7.81	n = 151	[-A-] All Calculated Results Friedewald formula [LDL=TC-HDL-(Trigs÷5)]
66.8 ± 9.42	81.2 ± 13.65	118.5 ± 16.15	94.0 ± 10.42	101.2 ± 13.41	n = 147	[---] All Homogeneous (Direct) Methods
67.3 ± 1.61	77.6 ± 1.52	115.1 ± 2.16	95.9 ± 1.27	100.0 ± 1.66	n = 5	[AB1] Abbott
63.8 ± 4.43	73.4 ± 5.88	111.6 ± 7.93	89.6 ± 6.66	95.8 ± 6.82	n = 13	[BC1] Beckman Coulter
57.9 ± 2.92	69.5 ± 3.82	101.7 ± 5.78	84.8 ± 4.20	87.7 ± 5.39	n = 24	[OL1] Beckman Coulter AU Series
67.5 ± 3.01	80.8 ± 3.53	135.2 ± 4.72	100.5 ± 3.69	109.2 ± 3.91	n = 15	[JJ1] Ortho Clinical Diagnostics
82.1 ± 2.18	105.4 ± 2.06	143.3 ± 2.85	110.6 ± 2.30	122.6 ± 2.77	n = 8	[RO4] Roche cobas c311/c501/c502/c701/c702
79.0 ± 2.92	107.6 ± 2.71	139.7 ± 5.58	106.1 ± 2.99	118.1 ± 3.91	n = 14	[RO2] Roche Hitachi and Modular D/P
46.4 ± 6.45	90.1 ± 4.38	100.8 ± 7.68	63.0 ± 7.21	74.8 ± 4.89	n = 3	[RO1] Roche Integra and MIRA
53.4 ± 0.91	64.8 ± 2.98	96.0 ± 3.38	79.5 ± 2.69	82.2 ± 4.68	n = 10	[GZ1] Sekisui Diagnostics (Genzyme)
61.1 ± 2.67	75.6 ± 2.62	118.0 ± 4.31	89.6 ± 3.47	95.9 ± 4.37	n = 13	[BY1] Siemens ADVIA/ADVIA Centaur
71.4 ± 3.44	85.8 ± 4.55	121.5 ± 6.52	96.5 ± 3.86	106.2 ± 5.42	n = 34	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Triglycerides (mg/dL)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
67.1 ± 4.27	68.0 ± 4.44	126.7 ± 6.50	96.6 ± 6.52	103.8 ± 6.05	n = 319	[---] All Methods & Instruments
<Instruments>						
76.8 ± 2.36	74.3 ± 0.51	145.7 ± 1.37	109.6 ± 1.02	120.8 ± 2.36	n = 3	[AXA] Abaxis Piccolo
67.3 ± 1.45	66.2 ± 2.17	120.6 ± 2.99	98.0 ± 1.39	101.6 ± 2.25	n = 18	[ABJ] Abbott Architect c System
64.5 ± 1.88	66.4 ± 2.09	124.0 ± 3.82	91.5 ± 3.03	100.5 ± 3.09	n = 62	[OLC] Beckman Coulter AU Chemistry System
70.7 ± 3.16	74.7 ± 3.07	136.7 ± 6.73	104.5 ± 5.40	109.7 ± 6.73	n = 3	[BCX] Beckman Coulter LX-20
69.9 ± 2.43	75.0 ± 3.89	136.3 ± 4.32	103.6 ± 3.50	111.3 ± 3.34	n = 13	[BCG] Beckman Coulter UniCel DxC 600
69.6 ± 1.96	74.2 ± 2.08	133.5 ± 2.33	102.3 ± 2.96	109.2 ± 2.05	n = 8	[BCH] Beckman Coulter UniCel DxC 800
73.7 ± 1.37	75.0 ± 0.90	143.7 ± 2.26	114.5 ± 1.86	120.8 ± 1.54	n = 3	[JJE] Ortho Vitros 250/350/950
67.3 ± 3.24	68.1 ± 3.39	130.8 ± 3.51	103.3 ± 3.66	108.7 ± 3.63	n = 17	[JJF] Ortho Vitros 5,1FS
67.7 ± 2.92	68.7 ± 2.47	132.1 ± 3.65	104.1 ± 3.69	110.2 ± 3.45	n = 19	[JJG] Ortho Vitros 5600
69.5 ± 2.27	70.7 ± 2.59	127.8 ± 3.33	96.1 ± 2.91	103.2 ± 3.12	n = 15	[ROC] Roche cobas c501
68.1 ± 1.83	69.6 ± 1.09	126.3 ± 2.61	95.8 ± 1.55	102.0 ± 0.93	n = 5	[ROH] Roche cobas c701
66.5 ± 2.25	68.1 ± 2.23	124.0 ± 3.24	91.8 ± 0.41	99.0 ± 3.15	n = 5	[ROS] Roche Cobas INTEGRA 400
66.0 ± 0.90	68.7 ± 1.37	124.5 ± 3.63	90.3 ± 2.26	97.5 ± 1.86	n = 3	[ROT] Roche Cobas INTEGRA 800
67.9 ± 2.22	67.9 ± 1.88	125.3 ± 3.01	95.8 ± 2.25	101.9 ± 2.79	n = 32	[ROD] Roche MODULAR D/P
67.1 ± 2.19	67.8 ± 2.03	125.5 ± 3.15	97.4 ± 2.53	103.8 ± 2.86	n = 22	[BYE] Siemens ADVIA 1800
69.0 ± 2.70	70.2 ± 2.36	127.5 ± 2.74	99.5 ± 2.74	106.3 ± 3.07	n = 3	[BYB] Siemens ADVIA 2400
60.3 ± 1.74	60.9 ± 1.12	119.9 ± 2.15	90.6 ± 1.63	96.7 ± 1.87	n = 18	[DUE] Siemens Dimension EXL
61.9 ± 2.26	61.5 ± 3.04	120.4 ± 3.91	92.1 ± 3.59	99.2 ± 3.80	n = 16	[DUR] Siemens Dimension RxL
71.9 ± 2.46	72.3 ± 2.77	133.1 ± 3.41	102.3 ± 3.00	110.3 ± 3.65	n = 38	[DUT] Siemens Dimension Vista
59.7 ± 2.60	60.7 ± 3.81	118.6 ± 5.46	89.9 ± 3.57	98.9 ± 4.28	n = 7	[DUX] Siemens Dimension Xpand
<Reagents>						
76.8 ± 2.36	74.3 ± 0.51	145.7 ± 1.37	109.6 ± 1.02	120.8 ± 2.36	n = 3	[AX1] Abaxis
67.3 ± 1.45	66.2 ± 2.17	120.6 ± 2.99	98.0 ± 1.39	101.6 ± 2.25	n = 18	[AB1] Abbott
69.6 ± 2.59	74.2 ± 3.54	134.7 ± 4.34	102.7 ± 3.80	109.8 ± 4.03	n = 26	[BC1] Beckman Coulter
64.6 ± 1.90	66.4 ± 1.89	124.2 ± 3.64	91.5 ± 2.80	100.7 ± 2.95	n = 56	[OL1] Beckman Coulter AU Series
59.4 ± 7.94	65.5 ± 9.94	122.2 ± 11.16	80.2 ± 15.82	93.1 ± 13.80	n = 3	[CR1] Carolina
68.0 ± 3.46	68.9 ± 3.35	131.8 ± 4.32	104.2 ± 4.36	110.0 ± 4.35	n = 39	[JJ1] Ortho Clinical Diagnostics
69.1 ± 2.27	70.4 ± 2.22	127.4 ± 3.32	96.0 ± 2.51	102.9 ± 2.66	n = 20	[RO4] Roche cobas c311/c501/c502/c701/c702
68.1 ± 2.35	67.9 ± 1.84	125.4 ± 3.14	95.9 ± 2.31	102.0 ± 2.90	n = 33	[RO2] Roche Hitachi and Modular D/P
66.4 ± 1.66	67.7 ± 1.86	124.2 ± 3.43	90.8 ± 2.22	98.4 ± 2.81	n = 8	[RO1] Roche Integra and MIRA
67.3 ± 2.29	67.9 ± 2.25	125.7 ± 3.10	97.6 ± 2.56	104.0 ± 2.96	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
65.9 ± 6.66	66.3 ± 6.77	126.0 ± 8.29	96.4 ± 7.16	103.9 ± 7.70	n = 79	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

Homocysteine ( $\mu\text{mol/L}$ )

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
7.56 ± 1.13	10.93 ± 0.86	9.20 ± 1.48	23.10 ± 1.93	10.90 ± 1.54	n = 114	[---] All Methods & Instruments
<Instruments>						
7.57 ± 0.36	9.69 ± 0.46	8.95 ± 0.46	22.20 ± 1.44	10.61 ± 0.74	n = 13	[ABH] Abbott Architect i System
7.91 ± 0.44	10.93 ± 0.45	9.91 ± 0.60	23.29 ± 0.91	11.60 ± 0.63	n = 21	[OLC] Beckman Coulter AU Chemistry System
8.86 ± 0.60	11.77 ± 0.48	11.10 ± 0.53	24.81 ± 0.79	12.97 ± 0.44	n = 5	[JJG] Ortho Vitros 5600
8.11 ± 0.46	10.73 ± 0.82	9.52 ± 0.55	23.38 ± 1.12	10.94 ± 0.99	n = 4	[ROC] Roche cobas c501
8.54 ± 1.45	12.14 ± 1.65	10.75 ± 1.87	24.57 ± 2.60	12.85 ± 2.04	n = 5	[ROD] Roche MODULAR D/P
7.87 ± 0.32	10.84 ± 0.56	9.91 ± 0.44	23.58 ± 0.24	11.87 ± 0.41	n = 3	[BYE] Siemens ADVIA 1800
5.83 ± 0.41	11.25 ± 0.66	7.01 ± 0.54	21.79 ± 1.19	9.00 ± 0.52	n = 23	[COB] Siemens ADVIA Centaur
7.15 ± 0.38	10.13 ± 0.64	8.08 ± 0.45	21.00 ± 0.60	9.31 ± 0.37	n = 7	[DUT] Siemens Dimension Vista
7.76 ± 0.76	10.84 ± 0.68	9.58 ± 0.62	25.41 ± 1.88	11.39 ± 1.09	n = 15	[DPD] Siemens Immulite 2000
<Reagents>						
7.57 ± 0.36	9.69 ± 0.46	8.95 ± 0.46	22.20 ± 1.44	10.61 ± 0.74	n = 13	[AB1] Abbott
8.91 ± 1.47	11.81 ± 1.01	11.16 ± 1.65	24.92 ± 1.58	12.72 ± 0.96	n = 3	[AS1] Axis-Shield
8.25 ± 0.64	10.91 ± 0.58	10.79 ± 0.63	24.68 ± 0.97	12.41 ± 0.71	n = 6	[CR1] Carolina
7.91 ± 0.35	10.98 ± 0.47	9.72 ± 0.49	23.06 ± 0.91	11.41 ± 0.66	n = 30	[DZ1] Diazyme
8.76 ± 0.54	11.74 ± 0.44	11.07 ± 0.47	24.94 ± 0.77	12.92 ± 0.41	n = 6	[JJ1] Ortho Clinical Diagnostics
8.21 ± 0.88	11.21 ± 0.97	9.87 ± 0.79	23.95 ± 1.95	11.84 ± 0.91	n = 4	[GZ1] Sekisui Diagnostics (Genzyme)
5.83 ± 0.41	11.25 ± 0.66	7.01 ± 0.54	21.79 ± 1.19	9.00 ± 0.52	n = 23	[BY1] Siemens ADVIA/ADVIA Centaur
7.06 ± 0.43	9.91 ± 0.54	7.94 ± 0.57	20.79 ± 0.65	9.30 ± 0.35	n = 5	[DA5] Siemens Dimension
7.76 ± 0.76	10.84 ± 0.68	9.58 ± 0.62	25.41 ± 1.88	11.39 ± 1.09	n = 15	[DP5] Siemens Immulite

Summary of Participant Performance (Mean and Standard Deviation)

Troponin I (µg/L)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
0.021 ± 0.019	0.017 ± 0.016	0.564 ± 0.453	4.793 ± 4.099	1.476 ± 0.386	n = 227	[---] All Methods & Instruments
<Instruments>						
0.011 ± 0.004	0.011 ± 0.006	1.877 ± 0.092	16.152 ± 0.609	6.916 ± 0.139	n = 23	[ABH] Abbott Architect i System
0.009 ± 0.008	0.008 ± 0.006	0.344 ± 0.034	2.517 ± 0.208	1.187 ± 0.101	n = 24	[SAA] Beckman Coulter ACCESS
< 0.050	< 0.050	0.050 ± 0.000	2.251 ± 0.692	0.410 ± 0.080	n = 6	[BSA] BioSite Triage
0.023 ± 0.034	0.022 ± 0.032	0.930 ± 0.091	6.099 ± 0.376	2.910 ± 0.109	n = 6	[IAA] i-STAT
0.010 ± 0.000	0.010 ± 0.000	1.170 ± 0.054	11.054 ± 0.296	4.214 ± 0.149	n = 17	[JJG] Ortho Vitros 5600
0.010 ± 0.000	0.010 ± 0.000	1.197 ± 0.060	10.698 ± 0.369	4.235 ± 0.144	n = 11	[JJC] Ortho Vitros ECi/ECiQ
0.009 ± 0.007	0.009 ± 0.007	0.387 ± 0.034	4.345 ± 0.305	1.741 ± 0.104	n = 42	[COB] Siemens ADVIA Centaur
0.018 ± 0.020	0.018 ± 0.020	0.377 ± 0.054	4.202 ± 0.341	1.672 ± 0.155	n = 4	[BYP] Siemens ADVIA Centaur CP
0.062 ± 0.036	0.059 ± 0.039	0.485 ± 0.030	3.115 ± 0.164	1.555 ± 0.082	n = 15	[DUE] Siemens Dimension EXL
0.040 ± 0.000	0.040 ± 0.000	0.285 ± 0.042	2.480 ± 0.270	1.121 ± 0.123	n = 11	[DUR] Siemens Dimension RxL
0.033 ± 0.013	0.020 ± 0.000	0.476 ± 0.037	3.145 ± 0.299	1.575 ± 0.144	n = 39	[DUT] Siemens Dimension Vista
0.029 ± 0.020	0.040 ± 0.000	0.276 ± 0.025	2.397 ± 0.092	1.116 ± 0.043	n = 9	[DUX] Siemens Dimension Xpand
< 0.200	< 0.200	0.956 ± 0.080	6.795 ± 1.545	3.089 ± 0.332	n = 5	[DPD] Siemens Immulite 2000
< 0.060	< 0.060	1.990 ± 0.035	17.675 ± 0.606	7.273 ± 0.290	n = 6	[TOM] Tosoh Bioscience
<Reagents>						
0.011 ± 0.005	0.010 ± 0.006	1.870 ± 0.098	16.122 ± 0.607	6.907 ± 0.147	n = 27	[AB1] Abbott
0.009 ± 0.008	0.008 ± 0.007	0.343 ± 0.033	2.505 ± 0.215	1.182 ± 0.101	n = 25	[BC1] Beckman Coulter
0.050 ± 0.000	0.050 ± 0.000	0.050 ± 0.000	2.251 ± 0.692	0.410 ± 0.080	n = 6	[BS1] Biosite Diagnostics
0.023 ± 0.023	0.018 ± 0.024	0.990 ± 0.073	6.146 ± 0.047	2.894 ± 0.074	n = 3	[IA1] i-STAT
0.010 ± 0.000	0.010 ± 0.000	1.180 ± 0.057	10.915 ± 0.388	4.223 ± 0.147	n = 28	[JJ1] Ortho Clinical Diagnostics
0.303 ± 0.005	0.303 ± 0.005	0.303 ± 0.005	1.535 ± 0.144	0.615 ± 0.034	n = 6	[RO3] Roche Elecsys/Modular E/e411/e601/e602
0.009 ± 0.007	0.009 ± 0.007	0.387 ± 0.035	4.333 ± 0.314	1.736 ± 0.116	n = 46	[BY1] Siemens ADVIA/ADVIA Centaur
0.040 ± 0.000	0.040 ± 0.000	0.282 ± 0.041	2.467 ± 0.245	1.126 ± 0.121	n = 23	[DA5] Siemens Dimension
0.036 ± 0.019	0.020 ± 0.000	0.480 ± 0.034	3.137 ± 0.255	1.570 ± 0.123	n = 51	[DA6] Siemens Dimension LOCI
0.200 ± 0.000	0.200 ± 0.000	0.956 ± 0.080	6.795 ± 1.545	3.089 ± 0.332	n = 5	[DP5] Siemens Immulite
0.060 ± 0.000	0.060 ± 0.000	2.036 ± 0.150	17.617 ± 0.751	7.173 ± 0.326	n = 4	[TO2] Tosoh ST AIA

## Summary of Participant Performance (Mean and Standard Deviation)

## Troponin T (µg/L)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
0.010 ± 0.000	0.010 ± 0.000	0.262 ± 0.019	1.360 ± 0.081	0.695 ± 0.044	n = 34	[---] All Methods & Instruments
						<Instruments>
0.010 ± 0.000	0.010 ± 0.000	0.260 ± 0.018	1.275 ± 0.135	0.690 ± 0.055	n = 3	[ROF] Roche cobas e411
0.010 ± 0.000	0.010 ± 0.000	0.254 ± 0.020	1.341 ± 0.096	0.682 ± 0.047	n = 12	[ROA] Roche cobas e601
0.010 ± 0.000	0.010 ± 0.000	0.269 ± 0.015	1.375 ± 0.074	0.705 ± 0.034	n = 8	[BME] Roche Elecsys
0.010 ± 0.000	0.010 ± 0.000	0.264 ± 0.017	1.380 ± 0.049	0.699 ± 0.042	n = 9	[ROE] Roche MODULAR E
						<Reagents>
0.010 ± 0.000	0.010 ± 0.000	0.263 ± 0.018	1.356 ± 0.079	0.696 ± 0.041	n = 31	[RO3] Roche Elecsys/Modular E/e411/e601/e602

## Summary of Participant Performance (Mean and Standard Deviation)

## Alanine Aminotransferase (U/L 37°C)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
84.7 ± 7.73	138.0 ± 11.03	251.1 ± 17.67	45.5 ± 4.87	295.3 ± 16.77	n = 360	[---] All Methods & Instruments
<Instruments>						
78.7 ± 2.69	121.8 ± 3.42	223.8 ± 3.82	44.5 ± 2.17	261.7 ± 3.89	n = 4	[AXA] Abaxis Piccolo
84.2 ± 2.77	137.9 ± 2.72	256.1 ± 5.81	44.1 ± 1.89	305.3 ± 7.01	n = 21	[ABJ] Abbott Architect c System
75.9 ± 2.27	119.8 ± 3.40	223.8 ± 5.56	41.3 ± 1.38	266.2 ± 7.04	n = 63	[OLC] Beckman Coulter AU Chemistry System
84.0 ± 0.90	133.8 ± 3.23	240.7 ± 3.37	44.4 ± 1.02	282.0 ± 4.60	n = 3	[BCX] Beckman Coulter LX-20
83.8 ± 2.38	138.2 ± 3.32	244.5 ± 5.74	45.5 ± 1.25	288.9 ± 5.60	n = 17	[BCG] Beckman Coulter UniCel DxC 600
84.3 ± 0.72	138.8 ± 3.44	245.8 ± 3.01	45.2 ± 1.31	291.2 ± 1.95	n = 7	[BCH] Beckman Coulter UniCel DxC 800
101.2 ± 4.06	151.3 ± 5.04	280.2 ± 2.95	59.1 ± 4.68	310.6 ± 3.30	n = 8	[JJE] Ortho Vitros 250/350/950
99.8 ± 5.89	149.4 ± 4.71	277.3 ± 5.95	54.4 ± 5.69	307.1 ± 6.84	n = 19	[JJF] Ortho Vitros 5,1FS
100.8 ± 4.18	148.7 ± 3.82	275.1 ± 5.21	57.2 ± 4.38	303.8 ± 7.63	n = 19	[JJG] Ortho Vitros 5600
82.9 ± 1.24	132.6 ± 2.17	248.5 ± 5.45	44.3 ± 0.72	293.0 ± 5.12	n = 18	[ROC] Roche cobas c501
81.2 ± 1.07	131.0 ± 2.53	244.4 ± 5.07	42.1 ± 2.69	289.0 ± 6.29	n = 5	[ROH] Roche cobas c701
83.3 ± 3.67	135.2 ± 6.03	247.6 ± 2.31	42.8 ± 2.61	292.1 ± 3.45	n = 5	[ROS] Roche Cobas INTEGRA 400
81.8 ± 2.36	133.5 ± 1.86	245.5 ± 5.40	42.6 ± 1.02	292.8 ± 3.23	n = 3	[ROT] Roche Cobas INTEGRA 800
81.4 ± 2.78	132.6 ± 4.01	245.4 ± 6.34	43.7 ± 1.92	291.9 ± 8.11	n = 31	[ROD] Roche MODULAR D/P
87.4 ± 3.24	143.5 ± 4.02	262.2 ± 6.51	46.6 ± 2.29	310.5 ± 7.58	n = 22	[BYE] Siemens ADVIA 1800
87.7 ± 1.37	140.7 ± 2.26	256.8 ± 2.36	47.5 ± 1.86	306.0 ± 1.80	n = 3	[BYB] Siemens ADVIA 2400
87.9 ± 1.84	146.6 ± 3.07	261.0 ± 4.57	47.4 ± 1.56	307.6 ± 5.97	n = 21	[DUE] Siemens Dimension EXL
91.3 ± 4.22	148.2 ± 2.87	259.5 ± 4.62	52.8 ± 5.04	305.4 ± 5.19	n = 21	[DUR] Siemens Dimension RxL
82.8 ± 1.90	138.6 ± 2.29	249.3 ± 3.39	44.7 ± 1.24	294.9 ± 3.82	n = 41	[DUT] Siemens Dimension Vista
91.4 ± 5.12	148.1 ± 4.39	261.1 ± 4.84	52.7 ± 5.53	306.3 ± 4.83	n = 15	[DUX] Siemens Dimension Xpand
<Reagents>						
78.7 ± 2.69	121.8 ± 3.42	223.8 ± 3.82	44.5 ± 2.17	261.7 ± 3.89	n = 4	[AX1] Abaxis
84.2 ± 2.77	137.9 ± 2.72	256.1 ± 5.81	44.1 ± 1.89	305.3 ± 7.01	n = 21	[AB1] Abbott
83.9 ± 2.07	137.3 ± 3.88	243.9 ± 4.99	45.1 ± 1.39	288.6 ± 5.32	n = 31	[BC1] Beckman Coulter
75.8 ± 2.20	119.8 ± 3.41	223.7 ± 5.52	41.3 ± 1.42	266.1 ± 6.85	n = 59	[OL1] Beckman Coulter AU Series
100.6 ± 5.11	149.2 ± 4.57	276.9 ± 5.74	56.4 ± 5.34	306.5 ± 7.53	n = 49	[JJ1] Ortho Clinical Diagnostics
82.6 ± 1.51	132.6 ± 2.69	247.9 ± 5.76	44.4 ± 1.56	292.7 ± 5.93	n = 25	[RO4] Roche cobas c311/c501/c502/c701/c702
81.5 ± 2.84	132.7 ± 4.20	245.4 ± 6.39	43.8 ± 2.09	291.9 ± 8.25	n = 32	[RO2] Roche Hitachi and Modular D/P
82.4 ± 2.81	133.5 ± 2.34	246.9 ± 4.07	42.4 ± 1.67	292.4 ± 3.38	n = 8	[RO1] Roche Integra and MIRA
87.3 ± 3.18	142.9 ± 4.23	261.0 ± 6.92	46.6 ± 2.29	309.4 ± 7.81	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
86.8 ± 5.07	143.9 ± 5.80	255.7 ± 7.38	47.6 ± 4.90	301.5 ± 8.00	n = 97	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Aspartate Aminotransferase (U/L 37°C)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
95.5 ± 5.21	143.8 ± 10.49	532.5 ± 43.82	71.4 ± 4.82	194.8 ± 10.39	n = 359	[---] All Methods & Instruments
<Instruments>						
98.0 ± 1.50	145.3 ± 5.26	529.6 ± 8.77	76.2 ± 1.96	199.5 ± 3.45	n = 4	[AXA] Abaxis Piccolo
95.8 ± 1.47	144.4 ± 1.75	534.3 ± 6.77	71.1 ± 1.03	196.0 ± 3.11	n = 21	[ABJ] Abbott Architect c System
85.5 ± 3.04	127.6 ± 4.52	470.8 ± 13.96	64.5 ± 2.16	173.9 ± 5.61	n = 62	[OLC] Beckman Coulter AU Chemistry System
93.7 ± 1.37	144.9 ± 2.86	506.4 ± 14.60	72.0 ± 0.90	191.7 ± 2.26	n = 3	[BCX] Beckman Coulter LX-20
94.8 ± 2.29	145.0 ± 4.77	509.7 ± 11.83	71.5 ± 1.67	193.5 ± 5.23	n = 17	[BCG] Beckman Coulter UniCel DxC 600
94.2 ± 1.65	143.4 ± 1.48	507.4 ± 14.65	70.6 ± 0.72	192.3 ± 2.49	n = 8	[BCH] Beckman Coulter UniCel DxC 800
96.8 ± 2.28	159.6 ± 1.75	621.8 ± 12.54	79.1 ± 1.48	202.7 ± 4.37	n = 8	[JJE] Ortho Vitros 250/350/950
96.6 ± 2.25	160.9 ± 4.22	612.8 ± 18.34	79.4 ± 2.12	203.5 ± 3.92	n = 19	[JJF] Ortho Vitros 5,1FS
96.9 ± 3.26	160.6 ± 4.16	619.0 ± 14.00	78.9 ± 1.91	200.7 ± 5.34	n = 19	[JJG] Ortho Vitros 5600
96.7 ± 2.74	145.0 ± 3.23	540.4 ± 12.88	71.0 ± 1.74	195.7 ± 4.69	n = 18	[ROC] Roche cobas c501
95.4 ± 1.89	143.7 ± 2.81	532.5 ± 8.85	71.3 ± 1.38	196.2 ± 5.84	n = 5	[ROH] Roche cobas c701
98.3 ± 1.81	148.3 ± 2.61	553.9 ± 2.04	70.8 ± 1.28	201.2 ± 3.60	n = 5	[ROS] Roche Cobas INTEGRA 400
100.3 ± 0.51	149.3 ± 3.16	557.6 ± 11.05	73.0 ± 0.90	201.3 ± 5.09	n = 3	[ROT] Roche Cobas INTEGRA 800
94.2 ± 2.41	141.7 ± 2.94	520.1 ± 10.93	69.9 ± 1.98	191.4 ± 4.31	n = 31	[ROD] Roche MODULAR D/P
102.0 ± 3.51	154.4 ± 3.83	566.6 ± 11.42	75.1 ± 2.40	207.2 ± 4.78	n = 22	[BYE] Siemens ADVIA 1800
101.0 ± 1.80	151.7 ± 2.26	559.2 ± 5.12	75.3 ± 1.37	205.2 ± 1.54	n = 3	[BYB] Siemens ADVIA 2400
97.6 ± 3.56	141.4 ± 3.89	532.6 ± 12.29	71.4 ± 2.18	195.3 ± 4.63	n = 21	[DUE] Siemens Dimension EXL
97.9 ± 3.43	142.8 ± 5.16	533.4 ± 17.97	72.6 ± 3.23	197.0 ± 6.57	n = 21	[DUR] Siemens Dimension RxL
97.0 ± 2.35	142.9 ± 3.21	539.8 ± 9.79	71.1 ± 2.03	197.7 ± 4.73	n = 41	[DUT] Siemens Dimension Vista
99.0 ± 2.66	143.5 ± 4.28	539.6 ± 11.68	73.2 ± 1.70	197.7 ± 4.77	n = 15	[DUX] Siemens Dimension Xpand
<Reagents>						
98.0 ± 1.50	145.3 ± 5.26	529.6 ± 8.77	76.2 ± 1.96	199.5 ± 3.45	n = 4	[AX1] Abaxis
95.8 ± 1.47	144.4 ± 1.75	534.3 ± 6.77	71.1 ± 1.03	196.0 ± 3.11	n = 21	[AB1] Abbott
94.3 ± 2.29	144.1 ± 3.48	506.1 ± 18.43	71.0 ± 1.81	192.4 ± 3.92	n = 31	[BC1] Beckman Coulter
85.5 ± 3.06	127.6 ± 4.33	470.9 ± 13.56	64.5 ± 2.21	173.9 ± 5.58	n = 58	[OL1] Beckman Coulter AU Series
96.7 ± 2.62	160.7 ± 3.95	617.6 ± 15.10	79.1 ± 1.97	201.7 ± 5.19	n = 49	[JJ1] Ortho Clinical Diagnostics
96.4 ± 2.89	144.6 ± 3.75	538.5 ± 14.39	71.1 ± 1.83	196.0 ± 5.56	n = 25	[RO4] Roche cobas c311/c501/c502/c701/c702
94.3 ± 2.39	141.8 ± 3.07	520.6 ± 11.06	70.0 ± 1.96	191.6 ± 4.34	n = 32	[RO2] Roche Hitachi and Modular D/P
99.2 ± 1.86	148.7 ± 2.94	556.3 ± 8.74	71.8 ± 1.69	201.2 ± 4.23	n = 8	[RO1] Roche Integra and MIRA
101.7 ± 3.38	153.7 ± 4.11	564.8 ± 12.08	75.1 ± 2.27	206.5 ± 4.93	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
97.6 ± 2.94	142.6 ± 3.91	537.2 ± 12.63	71.7 ± 2.37	197.0 ± 5.17	n = 98	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

 $\alpha$ -Amylase (U/L 37°C)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
281.8 ± 50.99	120.2 ± 16.05	285.6 ± 47.36	189.6 ± 23.17	53.7 ± 9.55	n = 314	[---] All Methods & Instruments
<Instruments>						
316.7 ± 6.14	131.4 ± 2.48	318.1 ± 5.37	209.3 ± 3.68	54.5 ± 1.38	n = 19	[ABJ] Abbott Architect c System
246.5 ± 13.15	100.1 ± 4.97	248.5 ± 13.21	159.5 ± 7.80	42.2 ± 2.27	n = 53	[OLC] Beckman Coulter AU Chemistry System
302.9 ± 3.72	128.8 ± 2.36	302.0 ± 6.37	205.7 ± 4.06	61.0 ± 0.90	n = 3	[BCX] Beckman Coulter LX-20
310.8 ± 6.59	133.9 ± 2.10	312.6 ± 5.99	210.8 ± 3.79	61.6 ± 1.03	n = 14	[BCG] Beckman Coulter UniCel DxC 600
310.8 ± 3.15	134.3 ± 1.25	315.4 ± 5.86	211.6 ± 1.87	62.7 ± 0.74	n = 8	[BCH] Beckman Coulter UniCel DxC 800
190.0 ± 6.11	96.8 ± 1.76	205.3 ± 3.61	160.0 ± 1.01	71.6 ± 1.38	n = 6	[JJE] Ortho Vitros 250/350/950
189.9 ± 5.45	94.1 ± 3.80	211.5 ± 7.24	161.1 ± 6.84	69.4 ± 5.31	n = 18	[JJF] Ortho Vitros 5,1FS
189.8 ± 9.31	95.2 ± 7.43	209.5 ± 10.26	160.4 ± 7.57	67.5 ± 6.46	n = 19	[JJG] Ortho Vitros 5600
278.2 ± 5.45	124.1 ± 2.20	282.1 ± 5.69	192.6 ± 3.94	58.7 ± 1.46	n = 17	[ROC] Roche cobas c501
278.3 ± 6.93	124.1 ± 3.72	277.6 ± 9.30	190.6 ± 3.87	57.5 ± 1.86	n = 3	[ROS] Roche Cobas INTEGRA 400
276.3 ± 5.09	122.3 ± 0.51	275.5 ± 1.86	188.7 ± 0.51	58.3 ± 0.51	n = 3	[ROT] Roche Cobas INTEGRA 800
274.8 ± 4.04	122.8 ± 1.71	278.7 ± 4.23	190.7 ± 2.70	57.9 ± 1.11	n = 27	[ROD] Roche MODULAR D/P
272.5 ± 6.75	122.2 ± 2.64	277.7 ± 6.58	189.6 ± 4.52	57.2 ± 1.15	n = 21	[BYE] Siemens ADVIA 1800
267.2 ± 8.77	118.3 ± 3.07	272.5 ± 8.19	185.4 ± 6.45	56.2 ± 2.36	n = 3	[BYB] Siemens ADVIA 2400
338.8 ± 6.12	132.8 ± 2.24	339.7 ± 5.49	210.9 ± 3.96	48.1 ± 1.46	n = 16	[DUE] Siemens Dimension EXL
336.7 ± 6.00	132.5 ± 2.45	338.3 ± 6.42	210.0 ± 3.43	47.8 ± 1.34	n = 18	[DUR] Siemens Dimension RxL
325.7 ± 7.74	129.2 ± 3.55	327.9 ± 8.41	202.9 ± 5.17	46.0 ± 1.26	n = 40	[DUT] Siemens Dimension Vista
340.2 ± 6.54	133.7 ± 2.05	343.0 ± 5.96	212.7 ± 3.68	48.4 ± 0.96	n = 12	[DUX] Siemens Dimension Xpand
<Reagents>						
316.7 ± 6.14	131.4 ± 2.48	318.1 ± 5.37	209.3 ± 3.68	54.5 ± 1.38	n = 19	[AB1] Abbott
310.9 ± 4.08	135.1 ± 1.23	314.6 ± 4.10	211.4 ± 2.23	62.0 ± 0.82	n = 7	[BC1] Beckman Coulter
246.5 ± 12.60	100.3 ± 4.87	248.5 ± 12.68	159.7 ± 7.67	42.3 ± 2.25	n = 49	[OL1] Beckman Coulter AU Series
309.4 ± 6.80	132.6 ± 2.68	311.7 ± 7.64	210.4 ± 3.88	61.9 ± 1.27	n = 19	[BC2] Beckman Coulter IFCC Standardized
189.9 ± 6.94	95.3 ± 5.63	209.7 ± 8.23	159.9 ± 6.63	68.8 ± 5.46	n = 47	[JJ1] Ortho Clinical Diagnostics
277.6 ± 4.06	124.4 ± 1.92	282.1 ± 4.65	192.4 ± 3.24	58.8 ± 1.28	n = 22	[RO4] Roche cobas c311/c501/c502/c701/c702
275.2 ± 4.41	123.0 ± 1.88	279.0 ± 4.56	191.0 ± 2.98	58.0 ± 1.21	n = 29	[RO2] Roche Hitachi and Modular D/P
277.1 ± 5.93	122.2 ± 0.80	274.5 ± 2.72	188.6 ± 0.55	58.0 ± 1.37	n = 6	[RO1] Roche Integra and MIRA
271.6 ± 7.64	121.6 ± 3.19	276.9 ± 7.48	188.9 ± 5.28	57.1 ± 1.36	n = 25	[BY1] Siemens ADVIA/ADVIA Centaur
332.9 ± 9.53	131.5 ± 3.44	335.0 ± 9.66	207.7 ± 6.19	47.1 ± 1.68	n = 86	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## Alkaline Phosphatase (U/L 37°C)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
261.3 ± 27.94	51.0 ± 9.61	172.4 ± 19.43	82.1 ± 15.03	128.4 ± 15.47	n = 359	[---] All Methods & Instruments
<Instruments>						
232.5 ± 3.67	45.7 ± 2.02	151.0 ± 4.64	71.7 ± 3.77	112.5 ± 5.20	n = 4	[AXA] Abaxis Piccolo
267.7 ± 9.55	47.9 ± 2.01	170.4 ± 6.16	78.0 ± 2.71	125.8 ± 4.17	n = 21	[ABJ] Abbott Architect c System
240.4 ± 14.84	42.5 ± 2.81	152.2 ± 9.32	69.1 ± 4.29	113.0 ± 7.12	n = 62	[OLC] Beckman Coulter AU Chemistry System
238.3 ± 15.90	43.5 ± 1.86	150.0 ± 5.41	68.0 ± 4.51	110.6 ± 6.45	n = 3	[BCX] Beckman Coulter LX-20
236.6 ± 8.72	42.1 ± 1.41	149.3 ± 4.98	68.1 ± 2.24	110.6 ± 4.63	n = 17	[BCG] Beckman Coulter UniCel DxC 600
232.8 ± 10.51	42.2 ± 1.30	152.4 ± 5.34	69.7 ± 1.90	112.4 ± 3.33	n = 8	[BCH] Beckman Coulter UniCel DxC 800
252.3 ± 11.12	63.8 ± 3.77	198.1 ± 9.78	108.1 ± 4.90	144.9 ± 6.51	n = 8	[JJE] Ortho Vitros 250/350/950
236.6 ± 14.55	61.7 ± 3.75	189.1 ± 11.80	106.1 ± 6.30	140.7 ± 7.15	n = 19	[JFF] Ortho Vitros 5,1FS
236.4 ± 8.49	61.6 ± 2.74	186.5 ± 5.35	105.4 ± 2.87	138.8 ± 3.65	n = 19	[JJG] Ortho Vitros 5600
259.0 ± 7.87	47.4 ± 1.35	165.8 ± 3.24	76.0 ± 1.76	123.5 ± 3.29	n = 18	[ROC] Roche cobas c501
255.7 ± 2.35	47.0 ± 0.64	162.5 ± 0.57	74.4 ± 1.09	121.0 ± 1.28	n = 5	[ROH] Roche cobas c701
262.0 ± 4.18	46.2 ± 0.41	163.3 ± 2.47	74.5 ± 1.23	120.5 ± 1.22	n = 4	[ROS] Roche Cobas INTEGRA 400
265.5 ± 1.86	46.6 ± 1.02	166.8 ± 2.36	75.0 ± 0.90	123.5 ± 1.86	n = 3	[ROT] Roche Cobas INTEGRA 800
255.1 ± 6.44	47.3 ± 1.17	164.3 ± 3.06	75.4 ± 1.81	121.3 ± 2.71	n = 30	[ROD] Roche MODULAR D/P
282.7 ± 6.88	50.0 ± 1.82	179.7 ± 4.69	81.6 ± 2.56	132.6 ± 3.34	n = 22	[BYE] Siemens ADVIA 1800
273.0 ± 4.60	46.9 ± 5.72	177.2 ± 5.90	76.7 ± 4.06	127.2 ± 5.00	n = 3	[BYB] Siemens ADVIA 2400
310.6 ± 10.67	69.2 ± 4.97	201.9 ± 7.35	102.5 ± 5.34	153.6 ± 7.41	n = 21	[DUE] Siemens Dimension EXL
301.6 ± 20.68	65.7 ± 6.80	194.9 ± 12.35	99.1 ± 7.56	151.0 ± 11.77	n = 21	[DUR] Siemens Dimension RxL
286.0 ± 13.05	53.9 ± 4.99	181.3 ± 9.02	86.6 ± 5.61	136.9 ± 8.34	n = 41	[DUT] Siemens Dimension Vista
290.6 ± 19.86	61.7 ± 5.65	187.7 ± 14.43	93.0 ± 6.99	141.7 ± 12.49	n = 15	[DUX] Siemens Dimension Xpand
<Reagents>						
232.5 ± 3.67	45.7 ± 2.02	151.0 ± 4.64	71.7 ± 3.77	112.5 ± 5.20	n = 4	[AX1] Abaxis
267.7 ± 9.55	47.9 ± 2.01	170.4 ± 6.16	78.0 ± 2.71	125.8 ± 4.17	n = 21	[AB1] Abbott
234.1 ± 10.93	42.0 ± 1.68	149.3 ± 6.12	68.1 ± 2.93	110.2 ± 5.14	n = 31	[BC1] Beckman Coulter
241.2 ± 14.83	42.6 ± 2.84	152.7 ± 9.35	69.3 ± 4.28	113.4 ± 7.05	n = 59	[OL1] Beckman Coulter AU Series
238.9 ± 13.47	62.0 ± 3.45	189.1 ± 9.81	106.0 ± 5.00	140.4 ± 5.93	n = 49	[JJ1] Ortho Clinical Diagnostics
259.1 ± 7.75	47.5 ± 1.28	165.6 ± 3.78	75.9 ± 2.08	123.2 ± 3.27	n = 25	[RO4] Roche cobas c311/c501/c502/c701/c702
255.0 ± 6.33	47.2 ± 1.19	164.2 ± 3.12	75.3 ± 1.87	121.2 ± 2.70	n = 31	[RO2] Roche Hitachi and Modular D/P
263.8 ± 3.01	46.4 ± 0.68	165.1 ± 2.75	75.0 ± 0.48	121.6 ± 1.94	n = 8	[RO1] Roche Integra and MIRA
281.2 ± 7.95	49.8 ± 1.97	179.2 ± 5.29	81.0 ± 3.39	132.0 ± 4.04	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
295.3 ± 18.81	60.8 ± 8.83	189.7 ± 13.85	93.6 ± 9.57	144.1 ± 12.24	n = 98	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

 $\gamma$ -Glutamyltransferase (U/L 37°C)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
35.1 ± 8.58	33.1 ± 7.17	173.3 ± 40.93	104.8 ± 24.75	76.6 ± 18.61	n = 299	[---] All Methods & Instruments
<Instruments>						
34.4 ± 2.23	32.8 ± 1.65	174.2 ± 9.91	105.1 ± 5.89	75.9 ± 4.21	n = 17	[ABJ] Abbott Architect c System
27.5 ± 1.59	26.4 ± 1.58	133.2 ± 6.57	81.4 ± 4.16	59.2 ± 3.19	n = 54	[OLC] Beckman Coulter AU Chemistry System
32.2 ± 1.54	27.3 ± 1.37	170.9 ± 5.22	98.1 ± 2.86	72.5 ± 1.86	n = 3	[BCX] Beckman Coulter LX-20
32.5 ± 0.83	28.4 ± 1.66	174.6 ± 4.42	102.9 ± 3.30	73.7 ± 2.39	n = 15	[BCG] Beckman Coulter UniCel DxC 600
32.5 ± 1.23	28.9 ± 1.13	174.7 ± 5.33	103.1 ± 3.15	74.6 ± 1.65	n = 8	[BCH] Beckman Coulter UniCel DxC 800
52.4 ± 1.33	44.0 ± 0.93	282.3 ± 7.36	165.2 ± 2.46	120.1 ± 2.88	n = 5	[JJE] Ortho Vitros 250/350/950
52.1 ± 0.84	43.8 ± 0.99	277.6 ± 6.38	163.1 ± 4.03	119.3 ± 2.41	n = 17	[JJF] Ortho Vitros 5,1FS
52.4 ± 1.39	43.8 ± 1.11	278.2 ± 4.87	163.5 ± 4.52	119.0 ± 3.25	n = 19	[JJG] Ortho Vitros 5600
29.3 ± 0.61	29.0 ± 0.00	148.3 ± 2.23	89.8 ± 1.40	65.1 ± 1.49	n = 14	[ROC] Roche cobas c501
28.0 ± 0.90	27.7 ± 1.37	142.3 ± 4.22	87.3 ± 2.26	62.3 ± 1.37	n = 3	[ROH] Roche cobas c701
28.3 ± 0.51	27.3 ± 0.51	146.3 ± 3.37	88.8 ± 1.54	63.7 ± 0.51	n = 3	[ROS] Roche Cobas INTEGRA 400
27.4 ± 1.02	28.0 ± 0.00	151.5 ± 1.86	89.7 ± 1.37	65.3 ± 0.51	n = 3	[ROT] Roche Cobas INTEGRA 800
29.8 ± 0.98	28.8 ± 1.23	151.8 ± 2.49	91.1 ± 1.84	65.5 ± 1.36	n = 28	[ROD] Roche MODULAR D/P
32.5 ± 1.81	30.9 ± 1.72	163.7 ± 3.61	99.3 ± 2.59	71.4 ± 2.01	n = 22	[BYE] Siemens ADVIA 1800
31.8 ± 3.23	30.3 ± 2.26	164.0 ± 5.48	99.1 ± 3.72	72.1 ± 3.72	n = 3	[BYB] Siemens ADVIA 2400
42.7 ± 1.21	41.2 ± 1.81	195.3 ± 3.58	119.6 ± 2.39	89.7 ± 1.73	n = 16	[DUE] Siemens Dimension EXL
41.5 ± 2.35	40.1 ± 2.57	193.3 ± 6.89	117.9 ± 4.27	88.3 ± 3.77	n = 13	[DUR] Siemens Dimension RxL
39.9 ± 2.35	37.9 ± 1.99	200.3 ± 4.23	120.2 ± 2.57	88.4 ± 2.62	n = 38	[DUT] Siemens Dimension Vista
43.3 ± 1.88	41.9 ± 1.98	195.7 ± 2.43	120.7 ± 1.62	90.2 ± 1.31	n = 7	[DUX] Siemens Dimension Xpand
<Reagents>						
34.2 ± 2.01	32.6 ± 1.35	173.3 ± 8.79	104.6 ± 5.21	75.5 ± 3.72	n = 16	[AB1] Abbott
32.4 ± 1.04	28.2 ± 1.79	174.1 ± 4.82	102.5 ± 3.44	74.0 ± 2.21	n = 28	[BC1] Beckman Coulter
27.6 ± 1.55	26.4 ± 1.55	133.6 ± 6.21	81.7 ± 3.93	59.3 ± 3.08	n = 52	[OL1] Beckman Coulter AU Series
52.3 ± 1.18	43.8 ± 1.04	278.3 ± 6.03	163.5 ± 4.08	119.2 ± 2.80	n = 43	[JJ1] Ortho Clinical Diagnostics
29.3 ± 0.91	28.8 ± 0.44	148.3 ± 3.24	89.8 ± 1.85	64.9 ± 1.72	n = 20	[RO4] Roche cobas c311/c501/c502/c701/c702
29.8 ± 1.05	28.9 ± 1.33	151.8 ± 2.49	91.1 ± 1.85	65.5 ± 1.36	n = 29	[RO2] Roche Hitachi and Modular D/P
28.1 ± 0.59	27.7 ± 0.51	149.3 ± 3.83	89.3 ± 1.55	64.5 ± 1.02	n = 6	[RO1] Roche Integra and MIRA
32.5 ± 2.13	31.0 ± 1.91	163.5 ± 4.11	99.1 ± 2.73	71.4 ± 2.11	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
41.2 ± 2.57	39.4 ± 2.70	197.7 ± 5.20	119.9 ± 2.76	88.9 ± 2.59	n = 74	[DA5] Siemens Dimension

Summary of Participant Performance (Mean and Standard Deviation)

Creatine Kinase (U/L 37°C)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
43.0 ± 4.49	91.2 ± 6.01	182.1 ± 13.81	182.4 ± 14.03	268.9 ± 16.94	n = 327	[---] All Methods & Instruments
<Instruments>						
46.6 ± 1.93	95.4 ± 3.10	190.0 ± 6.76	194.9 ± 9.13	282.7 ± 10.05	n = 21	[ABJ] Abbott Architect c System
36.5 ± 1.92	79.8 ± 3.38	160.0 ± 7.02	164.0 ± 8.59	241.5 ± 10.73	n = 57	[OLC] Beckman Coulter AU Chemistry System
45.6 ± 1.02	90.5 ± 1.86	189.5 ± 3.63	192.9 ± 6.58	282.9 ± 8.31	n = 3	[BCX] Beckman Coulter LX-20
45.9 ± 2.54	92.9 ± 3.28	192.8 ± 6.10	196.5 ± 6.56	286.5 ± 7.79	n = 14	[BCG] Beckman Coulter UniCel DxC 600
46.5 ± 2.40	92.5 ± 3.15	194.3 ± 4.38	196.6 ± 1.83	288.2 ± 6.15	n = 8	[BCH] Beckman Coulter UniCel DxC 800
45.3 ± 6.59	90.0 ± 11.81	196.2 ± 25.88	165.0 ± 23.47	272.2 ± 35.15	n = 5	[JJE] Ortho Vitros 250/350/950
46.9 ± 3.18	92.3 ± 4.12	200.1 ± 8.34	168.8 ± 5.95	278.0 ± 10.20	n = 17	[JJF] Ortho Vitros 5,1FS
47.0 ± 2.51	92.2 ± 4.62	201.6 ± 10.28	169.7 ± 8.23	276.8 ± 11.65	n = 19	[JJG] Ortho Vitros 5600
41.4 ± 1.42	94.6 ± 2.77	192.1 ± 3.99	194.9 ± 6.45	287.4 ± 6.56	n = 18	[ROC] Roche cobas c501
46.2 ± 1.96	96.2 ± 1.27	185.9 ± 3.00	192.7 ± 1.51	274.8 ± 5.34	n = 4	[ROH] Roche cobas c701
42.7 ± 5.86	93.8 ± 4.11	189.9 ± 5.22	192.4 ± 5.58	288.7 ± 3.07	n = 3	[ROT] Roche Cobas INTEGRA 800
45.3 ± 1.56	95.6 ± 2.40	183.3 ± 4.54	188.9 ± 4.29	271.4 ± 6.43	n = 30	[ROD] Roche MODULAR D/P
38.5 ± 1.35	89.2 ± 2.04	174.0 ± 4.09	179.6 ± 4.61	261.0 ± 5.02	n = 21	[BYE] Siemens ADVIA 1800
38.0 ± 1.80	88.0 ± 3.61	172.2 ± 5.00	176.3 ± 5.86	256.2 ± 5.00	n = 3	[BYB] Siemens ADVIA 2400
42.5 ± 2.25	91.5 ± 2.42	179.5 ± 2.89	186.0 ± 4.54	265.9 ± 4.43	n = 19	[DUE] Siemens Dimension EXL
43.4 ± 2.38	91.4 ± 3.31	181.0 ± 5.20	188.6 ± 5.92	269.8 ± 6.87	n = 20	[DUR] Siemens Dimension RxL
44.4 ± 1.06	92.7 ± 2.05	180.3 ± 3.49	186.4 ± 3.75	266.3 ± 4.77	n = 40	[DUT] Siemens Dimension Vista
42.2 ± 2.50	91.4 ± 3.03	180.3 ± 3.59	184.6 ± 5.40	269.5 ± 4.77	n = 11	[DUX] Siemens Dimension Xpand
<Reagents>						
46.6 ± 1.93	95.4 ± 3.10	190.0 ± 6.76	194.9 ± 9.13	282.7 ± 10.05	n = 21	[AB1] Abbott
45.8 ± 2.36	92.3 ± 3.51	192.7 ± 5.71	196.4 ± 6.28	286.5 ± 7.33	n = 30	[BC1] Beckman Coulter
36.6 ± 1.87	80.0 ± 3.35	160.7 ± 6.79	164.8 ± 8.36	242.5 ± 10.38	n = 52	[OL1] Beckman Coulter AU Series
46.8 ± 3.07	91.9 ± 4.88	200.0 ± 10.38	168.6 ± 8.09	276.1 ± 12.56	n = 43	[JJ1] Ortho Clinical Diagnostics
41.9 ± 2.01	95.0 ± 2.51	191.2 ± 4.72	194.9 ± 5.60	285.7 ± 8.16	n = 24	[RO4] Roche cobas c311/c501/c502/c701/c702
45.3 ± 1.70	95.6 ± 2.54	183.2 ± 4.76	189.0 ± 4.62	271.4 ± 6.65	n = 30	[RO2] Roche Hitachi and Modular D/P
44.1 ± 4.35	95.4 ± 4.09	191.4 ± 4.57	195.1 ± 5.58	287.6 ± 5.24	n = 5	[RO1] Roche Integra and MIRA
38.4 ± 1.51	88.9 ± 2.46	173.5 ± 4.47	178.8 ± 5.20	260.0 ± 5.81	n = 25	[BY1] Siemens ADVIA/ADVIA Centaur
43.5 ± 1.99	91.9 ± 2.68	180.1 ± 3.89	186.5 ± 4.75	267.3 ± 5.84	n = 86	[DA5] Siemens Dimension
52.6 ± 19.50	107.7 ± 35.90	182.5 ± 2.67	164.5 ± 58.68	267.3 ± 2.54	n = 4	[DA6] Siemens Dimension LOCI

Summary of Participant Performance (Mean and Standard Deviation)

Creatine Kinase-MB (ng/mL, U/L 37°C, %)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
0.50 ± 0.31	0.99 ± 0.38	32.17 ± 4.91	15.60 ± 2.58	0.50 ± 0.25	n = 197	[-A-] All Methods - Results reported in ng/mL
0.35 ± 0.06	0.91 ± 0.08	32.02 ± 3.00	15.38 ± 1.41	0.41 ± 0.04	n = 22	[AB1] Abbott
0.50 ± 0.00	1.22 ± 0.08	38.27 ± 1.27	18.53 ± 0.98	0.59 ± 0.04	n = 15	[SAA] Beckman Coulter ACCESS
0.47 ± 0.06	1.25 ± 0.09	39.60 ± 2.31	19.03 ± 1.18	0.60 ± 0.00	n = 13	[BC1] Beckman Coulter UniCel
< 1.00	< 1.00	19.06 ± 3.21	13.21 ± 1.62	< 1.00	n = 3	[BS1] Biosite Diagnostics
0.23 ± 0.06	0.86 ± 0.10	23.35 ± 1.11	11.57 ± 0.70	0.26 ± 0.06	n = 26	[JJ1] Ortho Clinical Diagnostics
0.93 ± 0.08	1.68 ± 0.07	34.72 ± 1.27	18.11 ± 0.67	0.76 ± 0.09	n = 27	[RO3] Roche Elecsys/Modular E/e411/e601/e602
0.20 ± 0.00	0.72 ± 0.20	30.27 ± 1.82	14.60 ± 0.78	0.20 ± 0.00	n = 32	[BY1] Siemens ADVIA/ADVIA Centaur
0.48 ± 0.28	0.65 ± 0.30	31.89 ± 2.14	14.50 ± 1.51	0.46 ± 0.24	n = 28	[DA5] Siemens Dimension
0.73 ± 0.19	1.00 ± 0.13	32.31 ± 1.02	15.87 ± 0.69	0.52 ± 0.04	n = 26	[DA6] Siemens Dimension LOCI
0.50 ± 0.00	1.25 ± 0.06	38.55 ± 0.97	18.30 ± 0.11	0.60 ± 0.11	n = 2	[TO1] Tosoh
-----						
2.70 ± 0.82	6.35 ± 1.58	26.95 ± 3.30	15.14 ± 3.38	3.30 ± 0.82	n = 4	[-B-] All Methods - Results reported in U/L
-----						
0.00 ± 0.00	0.00 ± 0.00	17.00 ± 4.20	6.12 ± 3.81	0.00 ± 0.00	n = 4	[-P-] All Methods - Results reported as %
0.00 ± 0.00	0.00 ± 0.00	16.04 ± 4.60	5.24 ± 4.10	0.00 ± 0.00	n = 3	[HL1] Helena Laboratories

## Summary of Participant Performance (Mean and Standard Deviation)

## Lactate Dehydrogenase (U/L 37°C)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
86.8 ± 7.68	109.0 ± 9.86	256.7 ± 23.06	186.9 ± 16.65	218.9 ± 18.38	n = 261	[-A-] All Methods - Lactate to Pyruvate
227.2 ± 11.03	272.2 ± 10.65	672.1 ± 20.55	492.1 ± 13.90	572.8 ± 19.89	n = 46	[-B-] All Methods - Pyruvate to Lactate
<Instruments>						
89.6 ± 4.42	112.4 ± 4.23	265.7 ± 6.46	192.8 ± 6.10	229.1 ± 6.25	n = 20	[ABJ] Abbott Architect c System
78.6 ± 3.85	97.8 ± 4.08	230.0 ± 9.63	166.2 ± 6.72	197.5 ± 8.19	n = 54	[OLC] Beckman Coulter AU Chemistry System
73.5 ± 1.86	91.1 ± 2.05	217.4 ± 8.43	155.3 ± 7.74	186.3 ± 7.74	n = 3	[BCX] Beckman Coulter LX-20
74.1 ± 2.04	94.7 ± 1.85	215.7 ± 6.24	157.3 ± 4.80	187.6 ± 5.45	n = 15	[BCG] Beckman Coulter UniCel DxC 600
76.0 ± 1.46	93.2 ± 1.03	217.6 ± 3.36	160.6 ± 1.72	190.0 ± 3.25	n = 8	[BCH] Beckman Coulter UniCel DxC 800
228.7 ± 8.00	275.5 ± 7.27	671.5 ± 16.49	496.6 ± 11.60	575.0 ± 15.07	n = 6	[JJE] Ortho Vitros 250/350/950
226.1 ± 9.79	270.8 ± 11.87	672.6 ± 19.85	490.1 ± 15.06	573.3 ± 22.00	n = 19	[JJF] Ortho Vitros 5,1FS
225.5 ± 11.82	272.0 ± 10.22	668.3 ± 21.13	489.9 ± 13.68	569.1 ± 19.50	n = 19	[JJG] Ortho Vitros 5600
89.5 ± 2.28	113.3 ± 2.87	264.4 ± 4.20	192.7 ± 3.27	222.5 ± 5.09	n = 17	[ROC] Roche cobas c501
89.6 ± 1.80	113.3 ± 0.90	262.8 ± 2.11	192.5 ± 1.22	222.0 ± 2.25	n = 4	[ROH] Roche cobas c701
94.3 ± 2.26	117.7 ± 3.16	268.6 ± 1.02	195.0 ± 1.80	226.5 ± 3.63	n = 3	[ROT] Roche Cobas INTEGRA 800
88.7 ± 1.95	111.6 ± 2.79	263.6 ± 4.39	191.7 ± 4.21	223.1 ± 4.88	n = 29	[ROD] Roche MODULAR D/P
91.8 ± 2.51	114.4 ± 2.92	267.4 ± 6.43	196.2 ± 5.15	228.2 ± 6.17	n = 22	[BYE] Siemens ADVIA 1800
88.0 ± 0.00	109.3 ± 0.51	256.0 ± 1.80	188.5 ± 1.86	219.6 ± 1.02	n = 3	[BYB] Siemens ADVIA 2400
88.6 ± 5.35	114.9 ± 7.25	271.5 ± 8.57	198.0 ± 3.91	231.9 ± 8.83	n = 16	[DUE] Siemens Dimension EXL
93.0 ± 5.36	115.2 ± 4.15	273.9 ± 7.49	197.5 ± 5.47	231.1 ± 7.13	n = 12	[DUR] Siemens Dimension RxL
92.4 ± 3.70	115.7 ± 3.42	274.7 ± 6.47	199.6 ± 4.57	232.8 ± 6.63	n = 41	[DUT] Siemens Dimension Vista
90.4 ± 0.56	114.4 ± 3.49	270.1 ± 3.88	195.1 ± 3.38	231.1 ± 4.61	n = 8	[DUX] Siemens Dimension Xpand
<Reagents>						
89.6 ± 4.42	112.4 ± 4.23	265.7 ± 6.46	192.8 ± 6.10	229.1 ± 6.25	n = 20	[AB1] Abbott
74.7 ± 2.01	93.7 ± 2.03	216.7 ± 5.77	158.5 ± 4.57	188.5 ± 5.15	n = 28	[BC1] Beckman Coulter
78.9 ± 3.75	98.1 ± 4.02	230.6 ± 9.56	166.6 ± 6.58	197.9 ± 8.14	n = 51	[OL1] Beckman Coulter AU Series
226.6 ± 10.70	271.8 ± 10.86	671.4 ± 19.74	491.1 ± 14.38	571.8 ± 19.59	n = 46	[JJ1] Ortho Clinical Diagnostics
89.8 ± 2.28	113.4 ± 2.34	264.5 ± 4.20	192.9 ± 2.99	222.9 ± 4.67	n = 23	[RO4] Roche cobas c311/c501/c502/c701/c702
88.7 ± 1.96	111.5 ± 3.05	263.6 ± 4.49	191.5 ± 4.52	223.1 ± 5.01	n = 30	[RO2] Roche Hitachi and Modular D/P
93.5 ± 1.62	117.6 ± 2.31	270.9 ± 3.62	195.9 ± 1.83	228.7 ± 3.42	n = 5	[RO1] Roche Integra and MIRA
91.1 ± 2.82	113.5 ± 3.45	265.5 ± 7.69	194.9 ± 5.68	226.7 ± 6.65	n = 26	[BY1] Siemens ADVIA/ADVIA Centaur
91.4 ± 4.54	115.3 ± 4.36	273.3 ± 6.93	198.3 ± 4.79	232.1 ± 6.81	n = 75	[DA5] Siemens Dimension

## Summary of Participant Performance (Mean and Standard Deviation)

## LDH Isoenzyme 1 (%)

Specimen: C86	Specimen: C87	Specimen: C88	Specimen: C89	Specimen: C90	Number	[Code] Instrument or Reagent System
30.7 ± 2.21	29.8 ± 2.64	41.8 ± 1.89	43.8 ± 3.03	20.7 ± 3.23	n = 8	[-P-] All Methods & Instruments (%)
						<Instruments>
32.1 ± 1.88	31.2 ± 1.46	43.0 ± 1.50	44.1 ± 2.33	22.8 ± 2.58	n = 4	[HLS] Helena SPIFE
29.3 ± 1.58	27.8 ± 2.80	40.7 ± 1.51	43.8 ± 3.61	18.7 ± 2.47	n = 4	[SEE] Sebia Electrophoresis
						<Reagents>
32.1 ± 1.88	31.2 ± 1.46	43.0 ± 1.50	44.1 ± 2.33	22.8 ± 2.58	n = 4	[HL1] Helena Laboratories
29.3 ± 1.58	27.8 ± 2.80	40.7 ± 1.51	43.8 ± 3.61	18.7 ± 2.47	n = 4	[SE1] Sebia