

Bacteriology – Restricted
Group A *Streptococcus* Antigen Detection
September 2016

Number of participating laboratories: 61

Grade distribution

100%	60
80-89%	1

Correct responses:

A – Positive
B – Negative
C – Negative
D – Positive
E - Positive

Method Used	Number correct (incorrect)	
	A,B,C,D	E
Acceava Strep A	16	16
Sekisui (Genzyme) OSOM Ultra Strep A	9	8(1)
Sekisui (Genzyme) OSOM Strep A Test	7	7
Quidel QuickVue + Strep A	6	6
Cardinal Health SP Brand Strep A Dipstick	4	4
Meridian Bioscience ImmunoCard STAT Strep A	3	3
BD Chek Group A Strep	3	3
Abbott Signify Strep A Dipstick	2	2
Beckman Coulter Icon DS Strep A	2	2
Quidel QuickVue Dipstick Strep A	2	2
Fisher Sure-Vue Signature Strep A Test	2	2
Henry Schein One Step + Strep A Dipstick	2	2
Clearview Exact Strep A Dipstick	1	1
CONSULT Strep A Test	1	1
ProAdvantage by NDC	1	1